



NATIONAL DEVELOPMENT COMPLEX

Deputy Directorate General of Quality Control

Physical Laboratory: Site-3A

TEST REPORT

Test Report No: TR (014)/PL/2022

Job Card No: JC (028)/PL/2022

Date: 04-03-2022

(a) - w - 4/03/22

DATE: 04/03/2022

INDENTER'S DATA			
Name:	Shamim Khan Asif	Designation:	FM-I
Group & Site:	PP / 3C	Directorate/ Department:	ADP
Organization:	NDC	Contact No:	3656

GENERAL DATA			
Sample Name:	Silica Phenolic Ablative	ID No:	PP-Si-10A/1120-D-01~05
Lot No :	Si(820)-10A/1120-221~333	Lot Size:	7922 RM
Shape:	Cubic	Chemical Composition:	Silica Phenol
Project:	All Systems	Quantity:	05
Sample Receiving Date:	03-03-2022	Test Date:	04-03-2022

TEST DATA			
Test Method Used	PL/EOD/78	Qualification Criteria	MDE-MP-SQC, Rev:01
Instrument Type	PL/SAB225i-01	Testing Temperature(°C)	RT
Temperature(°C)	20.4 ±1	Relative Humidity (%)	56 ±2
Test Type	Sample ID	Test Results	
Density	PP-Si-10A/1120-D-01	1.63 g/cm ³	
	PP-Si-10A/1120-D-02	1.62 g/cm ³	
	PP-Si-10A/1120-D-03	1.63 g/cm ³	
	PP-Si-10A/1120-D-04	1.62 g/cm ³	
	PP-Si-10A/1120-D-05	1.62 g/cm ³	
Test Status	Qualified for the above mentioned test i.e. 1600 ± 30 Kg/m ³ (1.6 ± 0.03 g/cm ³)		
Remarks	i) These results apply to the items tested as received and mentioned above. ii) This test report shall not be reproduced except in full without the written approval of the laboratory. iii) The Geometrical measurements (Length, Width and Thickness/Height) are taken from the Test Report No: TR (52)/ADP/2022, dated: 23-02-2022 issued by ADP Directorate.		

Test Performed By:	Imran Shahzad ARO	Contact No. 3301
Verified By:	Haseeb Ahmed Khan AM(T)	Contact No. 3301
Reviewed By:	Amir Khan HoD (CA)	Contact No. 3301
Approved By:	Eng. Dr. Nabi Bakhsh PD (ME)	Contact No. 3141

PL/FM/02, Issue 19 (01-06 -17)
Page 01 of 01

Prepared By: I/C/GL/SRO	Reviewed & Recommended By: GM/HoD	Approved By: PD/GMR
-------------------------	-----------------------------------	---------------------

End of Report