

NB-353: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VIII (reproduction NB-316)

Date: 2025-11-19

Tags: O₂ Test Calibration NB Firing
O₂ sensor H₂ SrTiO₃ Unisense
RhCrO₃:Al:SrTiO₃ H₂ Sensor
temperature In situ Trace range robust
oxygen sensor photocatalysis Unisense
normal range

Category: SrTiO₃


Status: Done

Created by: Nadzeya Brezhneva

Objectives

Reproduction of NB-316: simultaneous detection of H₂ and O₂ evolution in liquid phase for irradiated suspension of Rh,CrO_x:Al:SrTiO₃ suspension (EA-358 sample, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C, 15 min (reference conditions).

Reaction scheme



ChemDraw file linked: [NB-316-SrTiO3-photocatalytic H2O splitting.cdxml](#)

Literature/reference experiments

Literature	/
Reproduction	<p>SrTiO₃ - NB-316: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C</p> <p>SrTiO₃ - NB-320: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) I</p> <p>SrTiO₃ - NB-329: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) II</p> <p>SrTiO₃ - NB-331: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) III</p> <p>SrTiO₃ - NB-336: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C V (reproduction NB-316)</p> <p>SrTiO₃ - NB-339: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VI (reproduction NB-316)</p> <p>SrTiO₃ - NB-348: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VII (reproduction NB-316)</p>
Similar experiments	<p>SrTiO₃ - NB-315: Liquid phase H₂ and O₂ of RhCrO_x,Al:SrTiO₃ (NB-301, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C</p>

Reagents

Name	CAS Number / Experiment Number	Inventor y number	Amount [mmol]	Equivalen ts	Mass _{theo} [mg]	Mass _{exp} [mg]	Molar mass [g/mol]	Density (g/ml)	Volume [ml]	Pressure [bar]	Concentra tion [mM]
milli-Q H ₂ O	/	/	/	/	/	/	/	0.998	25 + 25 (for calibration)	/	/
Al:SrTiO ₃ RhCrOx (EA-358)	SrTiO ₃ - EA-358: Modification of Al:SrTiO ₃ (EA-354) via deposition of Rh, Cr oxide co-catalyst, 350°C, 1h, Upscaling (3.33x)	/	/	/	12.5	12.54	/	/	/	/	/
Hydrogen	1333-74-0	/	/	/	/	/	/	/	1 balloon (ca. 2 L)	ca. 1	/

Excel sheet for reagent calculation

/

Irradiation Parameters

Power measurement was performed using [Power Meter - 843-R-USB + 919P-020-12](#) in [Equipment - Advanced power measurment setup V1.0 I](#)

Power measurement was performed in experiment [Prep work - NB-314: Measuring power output of UHP-365 nm #4 with 18A-4 in advanced irradiation setup](#)

	Name
Used Set-up	Equipment - Advanced irradiation setup V1.0 I
Irradiation setup number	Equipment - Irradiation setup 4 (CEEC II, E002)

	Light Source Name	Power Source Name	Wavelength [nm]	Power Setting [mW]	Analog Setting [0.00 - 10.00]
First light source	Light Source - UHP LED 365 nm-4	Power Sources - BLS-18000-14	365	56	0.19

Used beam combiner [Name or None]	/
Irradiation distance [cm]	6.5
Thermostat temperature [°C]	20
Stirring speed [rpm]	500
Irradiation start: 1. Firesting [relative to start log] 2. Unisense	1. 604 s 2. 16:10:04
Irradiation stop: 1. Firesting [relative to start log] 2. Unisense	1.1517 s 2. 16:25:18

O₂/H₂ sensor equipment

	Equipment	Used protocol
Used Firesting	Equipment - Firesting Fiber-Optic Oxygen Meter 2 Channel (Firesting 2)	Protocol - Operation of Firesting Fiber-Optic Oxygen Meter 2 Channel Software
Used O ₂ sensor	Equipment - Robust probe for liquid O ₂ measurment	Protocol - In-situ hydrogen and oxygen measurment in H ₂ /O ₂ reactor
Used H ₂ sensor	Equipment - H ₂ UniAmp Sensor - Normal range - 2.1 x 80 mm needle	Protocol - In-situ hydrogen and oxygen measurment in H ₂ /O ₂ reactor

Procedure/observations

Date	Time	Step	Observations	Pictures/Files
19.11.2025		The experiment was done according to Protocol - In-situ hydrogen and oxygen measurment in H₂/O₂ reactor Important steps and deviations are listed below	/	/
	9:58-11:05	Conditioning of H ₂ sensor	NB-353-Logger1 1.76 mV at the end of polarization procedure	NB-353.ulog NB-353-Logger1-pre-polarization.bmp NB-353-Logger1-pre-polarization.csv

	11:15-30	Assembling the setup for calibration (25 mL of water was added using graduated cylinder), (LAUDA set to 0 °C) done according to Protocol - Liquid phase calibration of H2 UniAmp sensor with H2 bubbling.	/	/
	11:43	Start of O2 logging.	NB-353-Ch2-1	2025-11-19_114341_NB-353-Ch2-1.txt 2025-11-19_114341_NB-353-Ch2-1.png
	11:43	Start of H2 logging.	NB-353-Logger2 offset -2 mV	NB-353.ulog NB-353-Logger2-calibration.csv NB-353-Logger2-calibration step.bmp NB-353-Logger2-2point calibration.bmp
	11:48	Degassing was started.	/	20251119_114802-degassing of water.jpg
	12:19	Moving cannula to gas phase above liquid.	/	/
	12:22	Introducing H2 sensor into the reactor under Ar flow.	/	/
	12:24	0 ppm was taken.	/	/
	12:25	H2 bubbling of the reactor was started	/	20251119_122724-H2 bubbling.jpg
	12:42	1.000.000 ppm point was taken	837 mV	/
	12:46	The previous point was deleted, new point was taken and calibration was saved	839 mV, slope: 1.058, 793 uM	20251119_124232-H2 table.jpg
	12:47	Stop of H2 logging.	/	/
	12:47	Stop of O2 logging.	/	/
	13:00	Deassembling the setup, drying the reactor with acetone and compressed air .	/	/
		Sample preparation		
	14:43	Weighing EA-358 photocatalyst in a 50 mL vial.	Creamy solid.	/
	14:45	Addition of 25 mL H2O to the vial via graduated cylinder.	/	/
	14:52-55	The suspension was vortexed for 3 min (Equipment - VWR® VV3, Vortex Mixer , stage 4/6), covered with Al foil before further use.	/	20251119_145537-suspension after vortex.jpg

		Continue in Protocol - In-situ hydrogen and oxygen measurment in H2/O2 reactor from step 6		
	15:00-05	The suspension was transferred to the reactor using glass pipette (preliminary the vial was manually shaken ca. 15 s) .	/	/
	15:10	Assembling the setup.	Currently, stopper instead of H ₂ sensor, PT100, PT1000 and O ₂ robust probe are inside the reactor immersed in the liquid phase	/
	15:15	Start of O2 logging.	NB-353-Ch2-2	2025-11-19_151536_NB-353-Ch2-2.txt 2025-11-19_151536_NB-353-Ch2-2.png
	15:17	The degassing was started	/	20251119_151820-degassing of the suspension.jpg
	15:49	Cannula was transferred to gas phase, above the suspension.	/	/
	15:52	H ₂ sensor was added in Ar counterflow.	/	/
	15:57	The degassing was stopped by removing the cannula and closing the valve.	/	/
	15:59	Stop of O2 logging.	/	/
	16:00	Start of O2 logging.	NB-353-Ch2-3	2025-11-19_160001_NB-353-Ch2-3.txt 2025-11-19_160001_NB-353-Ch2-3.png
	16:00	Start of H2 logging.	NB-353-Logger3	NB-353.ulog NB-353-Logger3-during irradiation.csv NB-353-Logger3-during irradiation.bmp
	16:00-10	Equilibration time.	/	/
	16:10	The irradiation was started	/	20251119_161145-start of irradiation.jpg
	16:25	The irradiation was stopped.	/	/
	16:25-35	Equilibration time.	/	/
	16:35	Stop of O2 and H2 logging.	/	/

	ca. 16:50	Deassembling the setup, cleaning the reactor.	Tips of the sensors and reactor were covered with attached photocatalyst particles. Tip: After preliminary cleaning with sticks, wipes, the residual particles attached to the walls of the reactor could be removed by sonication - fill the reactor with water and place it in ultrasonic bath for ca. 20 s (Eco mode).	20251119_164458-after irradiation.jpg 20251119_164405-after irradiation-1.jpg
--	-----------	---	--	--

Analysis

Used calibration for Firing: [20250611-BOLA-fitting-liquid-phase-trace-oxygen-sensor-H2-O2 reactor.ini](#)

Used calibration for UniSense: NB-353-Logger2

Date	Time	Sample name	Analysis method	Analytical device	Solvent	Raw Data	Python script	Processed Data	Comparative Data	Interpretation
19.11.2025	9:58	NB-353-Logger1	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-353.ulong NB-353-Logger1-pre-polarization.csv	/	NB-353-Logger1-pre-polarization.bmp	/	Pre-polarization of H2 sensor.
	11:43	NB-353-Logger2	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-353.ulong NB-353-Logger2-calibration.csv	/	NB-353-Logger2-calibration step.bmp NB-353-Logger2-2point calibration.bmp	/	2 point calibration, signal at 10 ⁶ ppm - 839 mV
	16:00	NB-353-Logger3	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-353.ulong NB-353-Logger3-during irradiation.csv	NB-353-O2 and H2 curve.py	NB-353-Logger3-during irradiation.bmp NB-353-O2 and H2 curves.png	SrTiO3 - NB-316: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C	H2 evolution during irradiation.
	11:43	NB-353-Ch2-1	Optical O2 detection	Equipment - Firing Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-19_114341_NB-353-Ch2-1.txt	/	2025-11-19_114341_NB-353-Ch2-1.png	/	Degassing, followed by calibration of H2 sensor.
	15:15	NB-353-Ch2-2	Optical O2 detection	Equipment - Firing Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-19_151536_NB-353-Ch2-2.txt	/	2025-11-19_151536_NB-353-Ch2-2.png	/	Degassing of the photocatalyst suspension.
	16:00	NB-353-Ch2-3	Optical O2 detection	Equipment - Firing Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-19_160001_NB-353-Ch2-3.txt	NB-353-O2 and H2 curve.py	2025-11-19_160001_NB-353-Ch2-3.png NB-353-O2 and H2 curves.png	SrTiO3 - NB-316: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C	O2 evolution during irradiation of the suspension.

Results

Reproduction of NB-316: simultaneous H₂ and O₂ measurements of irradiated suspension of EA-358 (0.5 mg/mL) in O₂/H₂ photoreactor under 365 nm irradiation (50 mW/cm², 20 °C, 15 min, reference conditions) were performed.

Linked experiments

SrTiO₃ - NB-316: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C

SrTiO₃ - NB-320: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) I

SrTiO₃ - NB-329: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) I I

SrTiO₃ - NB-331: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C (reproduction NB-316) I I I

SrTiO₃ - NB-336: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C V (reproduction NB-316)

SrTiO₃ - NB-339: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VI (reproduction NB-316)

SrTiO₃ - NB-345: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 20 mW/cm², 20 °C (reproduction NB-325) I I

SrTiO₃ - NB-348: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VII (reproduction NB-316)

Linked resources

Equipment - [Firesting Fiber-Optic Oxygen Meter 2 Channel \(Firesting 2\)](#)

Equipment - [Irradiation setup 4 \(CEEC II, E002\)](#)

Protocol - [Getting hydrogen from hydrogen bottle in CEEC II E014](#)

Protocol - [Liquid phase calibration of H₂ UniAmp sensor](#)

Protocol - [In-situ hydrogen and oxygen measurment in H₂/O₂ reactor](#)


Attached files

NB-353-O2 and H2 curve.py

sha256: 81bec22d78fb150a20db64b6b030cc3cf0ebb020d0010d8174ca51c03d184240

NB-353-O2 and H2 curves.png

sha256: 40b25d3244dec4654ecc09fc9951b8c1a356f874456bf9fdf1499a76447b659e



NB-353.ulog

sha256: 815e131d426d2cd950a7cf68e1f1d6eef94079cd7e87d2839c859bb9ba27fbd2

20251119_164405-after irradiation-1.jpg

sha256: 3f996503b378b00a5ed311f64f01f0c17c2d0a8287a3d7916fcccb338f6d11da



20251119_164458-after irradiation.jpg

sha256: 0323ecb4bc6d3ab7e6baf91428656b2d1ead6211b891b052591dbc60091d8542



20251119_161145-start of irradiation.jpg

sha256: 9e6544c6223f4afd66e8ba804c83a4ce36b818d31f6a6772c6920c02e5f7fbe2



20251119_151820-degassing of the suspension.jpg

sha256: 707e622b12cc8fbeeabac9fa92806754c84de0c5340b6dd63e7bd234066fbe50




20251119_145537-suspension after vortex.jpg

sha256: 90ddf1cdf8bd2cc6181a641f5477516440fcb696bae69b5fd876c850fa310e45



20251119_124232-H2 table.jpg

sha256: 505db541d7f56dd03a885d3f8ded50a0d29ce7c35d76ca8dad7524e2f28158a8



20251119_114802-degassing of water.jpg

sha256: 669fc1e6179b4f088eb05fc86ce03ffe5b757315ae1a1bee3a839a1faa4855a6




20251119_122724-H2 bubbling.jpg

sha256: 454db3973bdd0072f9f80cd6a7c6c9be5c27603627ff1a1c4f495dc4ea986573



NB-353-Logger3-during irradiation.bmp

sha256: cbdacfaa12dc20ce6a32efd0d6c5b326bb235ec5c2ea58ce89dcd9e9be109b56




NB-353-Logger3-during irradiation.csv

sha256: ed71798b6ddaf8df1b600cf29e4a11fa0394edf9a928be07f25bbffd338f3f80


NB-353-Logger2-2point calibration.bmp

sha256: e2219a96f2e1760589ab1067fa409e79f700cd2002fdae7dac5fdd25d15134a4



NB-353-Logger2-calibration step.bmp

sha256: d02b5831e1876e75b4cf2bc6e7dccf4313620978a50f207a1b47984893b42984




NB-353-Logger2-calibration.csv

sha256: d25f9245ce18c0dcf041aa7153c20d0d2b57a23af4ce878e545e9826f50019c4

NB-353-Logger1-pre-polarization.bmp

sha256: 2758b9ad0f06e17236c8a9623a785e244a6ecb74b400c571948205dd6efe3754



NB-353-Logger1-pre-polarization.csv


sha256: 9001207611ccc714d1b97bd396766230bc474392ff648f5691bbff4439b2db05

2025-11-19_160001_NB-353-Ch2-3.txt

sha256: bae0fbf2d6aaeb9407bec903165d7ee096b503bb465806588b2b0c86ff163cc5

2025-11-19_160001_NB-353-Ch2-3.png

sha256: 32db611c10afba7945a29158ebe06687cabea0ea582e3f38160a8100046344fd



2025-11-19_151536_NB-353-Ch2-2.txt

sha256: 5b871ce4f7f569cf365c7ba2b3c4cf40dcaae3f547651a365d394d8ca55d2e46

2025-11-19_151536_NB-353-Ch2-2.png

sha256: a76369c7036269270e26a3f84e47d74947c170fb0d71c13f51e660251aba9192



2025-11-19_114341_NB-353-Ch2-1.png

sha256: 93dcafcbb3fc8aa9105d0cafc23e3d0f27f247d5862d74da08e21e631c72a9bc



2025-11-19_114341_NB-353-Ch2-1.txt

sha256: 4717fd3a341c9627edf5a302dcf054a7fb48b2312f2d15ac9256b37d84b4ec9c



Unique eLabID: 20251119-6ce705b0e03c7fd1464e937fb3f69708bf42720a

Link: <https://elab.water-splitting.org/experiments.php?mode=view&id=3486>