

NB-348: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C VII (reproduction NB-316)

Date: 2025-11-13

Tags: O₂ Test Calibration NB Firing
O₂ sensor H₂ SrTiO₃ Unisense
RhCrO₃:Al:SrTiO₃ H₂ Sensor
temperature In situ Trace range robust
oxygen sensor photocatalysis Unisense
normal range

Category: SrTiO₃


Status: Done

Created by: Nadzeya Brezhneva

Objectives

Reproduction of NB-316: simultaneous detection of H₂ and O₂ evolution in liquid phase for irradiated suspension of Rh,CrOx:Al:SrTiO₃ suspension (EA-358 sample, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C, 15 min (reference conditions).

Reaction scheme



ChemDraw file linked: [NB-316-SrTiO3-photocatalytic H2O splitting.cdxml](#)

Literature/reference experiments

Literature	/
Reproduction	SrTiO₃ - NB-316: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C
Similar experiments	SrTiO₃ - NB-313: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (NB-301, 0.5 mg/mL), 365 nm, 50 mW SrTiO₃ - NB-315: Liquid phase H₂ and O₂ of RhCrOx,Al:SrTiO₃ (NB-301, 0.5 mg/mL), 365 nm, 50 mW/cm², 20 °C

Reagents

Name	CAS Number / Experiment Number	Inventor y number	Amount [mmol]	Equivalents	Mass _{theo} [mg]	Mass _{exp} [mg]	Molar mass [g/mol]	Density (g/ml)	Volume [ml]	Pressure [bar]	Concentration [mM]
milli-Q H ₂ O	/	/	/	/	/	/	/	0.998	25 + 25 (for calibration)	/	/

Al:SrTiO ₃ RhCrO _x (EA-358)	SrTiO ₃ - EA-358: Modification of Al:SrTiO ₃ (EA-354) via deposition of Rh, Cr oxide co- catalyst, 350°C, 1h, Upscaling (3.33x)	/	/	/	12.5	12.45	/	/	/	/	/
Hydrogen	1333-74-0	/	/	/	/	/	/	/	1 balloon (ca. 2 L)	ca. 1	/

Excel sheet for reagent calculation

/

Irradiation Parameters

Power measurement was performed using [Power Meter - 843-R-USB + 919P-020-12](#) in [Equipment - Advanced power measurment setup V1.0 I](#)

Power measurement was performed in experiment [Prep work - NB-314: Measuring power output of UHP-365 nm #4 with 18A-4 in advanced irradiation setup](#)

	Name
Used Set-up	Equipment - Advanced irradiation setup V1.0 I
Irradiation setup number	Equipment - Irradiation setup 4 (CEEC II, E002)

	Light Source Name	Power Source Name	Wavelength [nm]	Power Setting [mW]	Analog Setting [0.00 - 10.00]
First light source	Light Source - UHP LED 365 nm-4	Power Sources - BLS-18000-1 4	365	56	0.19

Used beam combiner [Name or None]	/
Irradiation distance [cm]	6.5
Thermostat temperature [°C]	20

Stirring speed [rpm]	500
Irradiation start: 1. Firesting [relative to start log] 2. Unisense	1. 603 s 2. 17:37:25
Irradiation stop: 1. Firesting [relative to start log] 2. Unisense	1. 1520 s 2. 17:52:42

O₂/H₂ sensor equipment

	Equipment	Used protocol
Used Firesting	Equipment - Firesting Fiber-Optic Oxygen Meter 2 Channel (Firesting 2)	Protocol - Operation of Firesting Fiber-Optic Oxygen Meter 2 Channel Software
Used O ₂ sensor	Equipment - Robust probe for liquid O ₂ measurment	Protocol - In-situ hydrogen and oxygen measurment in H ₂ /O ₂ reactor
Used H ₂ sensor	Equipment - H ₂ UniAmp Sensor - Normal range - 2.1 x 80 mm needle	Protocol - In-situ hydrogen and oxygen measurment in H ₂ /O ₂ reactor

Procedure/observations

Date	Time	Step	Observations	Pictures/Files
13.11.2025		The experiment was done according to Protocol - In-situ hydrogen and oxygen measurment in H₂/O₂ reactor Important steps and deviations are listed below	/	/
	12:14-14:00	Conditioning of H ₂ sensor	NB-348-Logger1 ca. 1.8 mV at the end of polarization procedure	NB-348.ulog NB-348-Logger1-pre-polarization.csv NB-348-Logger1-pre-polarization.bmp
	14:10	Assembling the setup for calibration (25 mL of water was added using graduated cylinder), done according to Protocol - Liquid phase calibration of H₂ UniAmp sensor with H₂ bubbling .	/	/

	14:33	Start of O2 logging.	NB-348-Ch2-1	2025-11-13_143340_NB-348-Ch2-1.txt 2025-11-13_143340_NB-348-Ch2-1.png
	14:33	Start of H2 logging.	NB-348-Logger2 offset - 2 mV	NB-348.ulog NB-348-Logger2-calibration.csv NB-348-Logger2-calibration step.bmp NB-348-Logger2-2point calibration curve.bmp
	14:36	Degassing was started.	/	20251113_143712-degassing of water.jpg
	15:02	Moving cannula to gas phase above the liquid.	/	/
	15:05	Introducing H2 sensor into the reactor under Ar flow.	/	/
	15:10	0 ppm was taken.	/	/
	15:17	Removing cannula, closing the valve.	/	/
	15:18	H2 bubbling of the reactor was started	/	20251113_151843-H2 bubbling.jpg
	15:38	1.000.000 ppm point was taken	846 mV	/
	15:42	Deleting the previous point, taking the new one, calibration was saved.	847 mV, slope: 1.074, 788 uM	20251113_153640-H2 table.jpg
	15:43	Stop of H2 logging.	/	/
	15:43	Stop of O2 logging.	/	/
	ca. 16:00	Deassembling the setup, drying the reactor with acetone and compressed air .	/	/
		Sample preparation		
	16:25	Weighing EA-358 photocatalyst in a 50 mL vial.	Creamy solid.	/
	16:27	Addition of 25 mL H2O to the vial via graduated cylinder.	/	/
	16:27-30	The suspension was vortexed for 3 min (Equipment - VWR® VV3, Vortex Mixer, stage 4/6), covered with Al foil before further use.	/	20251113_163014-suspension after vortex.jpg
		Continue in Protocol - In-situ hydrogen and oxygen measurment in H2/O2 reactor from step 6		

	ca. 16:40	The suspension was transferred to the reactor using glass pipette (preliminary the vial was manually shaken ca. 15 s) .	/	
	16:45	Assembling the setup.	Currently, stopper instead of H ₂ sensor, PT100, PT1000 and O ₂ robust probe are inside the reactor immersed in the liquid phase	
	16:47	Start of O2 logging.	NB-348-Ch2-2	2025-11-13_164739_NB-348-Ch2-2.txt 2025-11-13_164739_NB-348-Ch2-2.png
	16:49	The degassing was started	/	20251113_165040-degassing of the suspension.jpg
	17:17	Cannula was transferred to gas phase, above the suspension.	/	/
	17:19	H ₂ sensor was added in Ar counterflow.	/	/
	17:24	The degassing was stopped by removing the cannula and closing the valve.	/	/
	17:26	Stop of O2 logging.	/	/
	17:27	Start of O2 logging.	NB-348-Ch2-3	2025-11-13_172722_NB-348-Ch2-3.txt 2025-11-13_172722_NB-348-Ch2-3.png
	17:27	Start of H2 logging.	NB-348-Logger3	NB-348.ulog NB-348-Logger3-during irradiation.csv NB-348-Logger3-during irradiation.bmp
	17:27-37	Equilibration time.	/	/
	17:37	The irradiation was started	/	20251113_174011-after start of irradiation.jpg
	17:52	The irradiation was stopped.	/	/
	17:52-18:02	Equilibration time.	/	/

	18:02	Stop of O2 and H2 logging.	/	/
	ca. 18:20	Deassembling the setup, cleaning the reactor.	Tips of the sensors and reactor were covered with attached photocatalyst particles. Tip: After preliminary cleaning with sticks, wipes, the residual particles attached to the walls of the reactor could be removed by sonication - fill the reactor with water and place it in ultrasonic bath for ca. 20 s (Eco mode).	20251113_181137-after irradiation.jpg

Analysis

Used calibration for Firesting: [20250611-BOLA-fitting-liquid-phase-trace-oxygen-sensor-H2-O2 reactor.ini](#)

Used calibration for UniSense: NB-348-Logger2

Date	Time	Sample name	Analysis method	Analytical device	Solvent	Raw Data	Python script	Processed Data	Comparative Data	Interpretation
13.11.2025	12:14	NB-348-Logger1	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-348.ulong NB-348-Logger1-pre-polarization.csv	/	NB-348-Logger1-pre-polarization.bmp	/	Pre-polarization of the H2 sensor.
	14:33	NB-348-Logger2	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-348.ulong NB-348-Logger2-calibration.csv	/	NB-348-Logger2-calibration step.bmp NB-348-Logger2-2point calibration curve.bmp	/	Calibration, 10 ⁶ ppm corresponds to 847 mV, slope 1.074
	17:27	NB-348-Logger3	electrochemical H2 detection	Equipment - H2 UniAmp Sensor - Normal range - 2.1 x 80 mm needle	water	NB-348.ulong NB-348-Logger3-during irradiation.csv	NB-348-O2 and H2 curve.py	NB-348-Logger3-during irradiation.bmp NB-348-O2 and H2 curve.png	SrTiO3 - NB-316: Liquid phase H2 and O2 of RhCrOx, Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C	H2 evolution during irradiation of the suspension, small step in signal during irradiation.
	14:33	NB-348-Ch2-1	Optical O2 detection	Equipment - Firesting Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-13_143340_NB-348-Ch2-1.txt	/	2025-11-13_143340_NB-348-Ch2-1.png	/	Degassing of water followed by calibration of H2 sensor.

16:47	NB-348-Ch2-2	Optical O2 detection	Equipment - Firesting Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-13_164739_NB-348-Ch2-2.txt	/	2025-11-13_164739_NB-348-Ch2-2.png	/	Degassing of the suspension.
17:27	NB-348-Ch2-3	Optical O2 detection	Equipment - Firesting Fiber-Optic Oxygen Meter 2 Channel	water	2025-11-13_172722_NB-348-Ch2-3.txt	NB-348-O2 and H2 curve.py	2025-11-13_172722_NB-348-Ch2-3.png	SrTiO3 - NB-316: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C	O2 evolution during irradiation.

Results

Reproduction of NB-316 - simultaneous H₂ and O₂ measurements of irradiated suspension of EA-358 (0.5 mg/mL) in O₂/H₂ photoreactor under 365 nm irradiation (50 mW/cm², 20 °C, 15 min, reference conditions) - was performed.

Linked experiments

SrTiO3 - NB-313: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (NB-301, 0.5 mg/mL), 365 nm, 50 mW

SrTiO3 - NB-315: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (NB-301, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C

SrTiO3 - NB-316: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C

SrTiO3 - NB-336: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C V (reproduction NB-316)

SrTiO3 - NB-339: Liquid phase H2 and O2 of RhCrOx,Al:SrTiO3 (EA-358, 0.5 mg/mL), 365 nm, 50 mW/cm2, 20 °C VI (reproduction NB-316)

Linked resources

Equipment - [Firesting Fiber-Optic Oxygen Meter 2 Channel \(Firesting 2\)](#)

Equipment - [Irradiation setup 4 \(CEEC II, E002\)](#)

Protocol - [Getting hydrogen from hydrogen bottle in CEEC II E014](#)

Protocol - [Liquid phase calibration of H2 UniAmp sensor](#)

Protocol - [In-situ hydrogen and oxygen measurment in H2/O2 reactor](#)


Attached files

NB-348-O2 and H2 curve.py

sha256: 99350f377dde9afd20b4f9d68ef7a91969e57064a7236eb0729a6cc8265058db

NB-348-O2 and H2 curve.png

sha256: 5af10f9755c7cc045adcb34d69521eeff713c4ced8b989615f0e30429bfb404b



20251113_143712-degassing of water.jpg

sha256: 51df69c7c0ae7d52cee267a87593af6eea04b05625dfe63496dc62b18a1ae8b7



20251113_151843-H2 bubbling.jpg

sha256: 60892c6399afaba07c95780f1e51f3dc25ca4c6b4d06770bbc280e9a489e88c1



20251113_153640-H2 table.jpg

sha256: 0cee996ec0523c2a5db0d86b4f6e602d98654f1df2a4fab10e2452f88e2fd0a1



20251113_163014-suspension after vortex.jpg

sha256: 327d15076242609b6ec9bfd9e86ad3442f097907f97fb2404db1dda352a703c6



20251113_165040-degassing of the suspension.jpg

sha256: 82de8295756cf1f35e204f9ff3f316b7f7578be2e57ea63007e2789bc652f38e



20251113_181137-after irradiation.jpg

sha256: 7a9f9f0ca779df3391c5c57d2c2ea8fbb2856794be1310ddd79abea2a491f837



20251113_174011-after start of irradiation.jpg

sha256: d0823dd5502e9b7fbc05fce3b90f2d2ecfad179a87dcfd0fe81b7fe09e58c215




2025-11-13_172722_NB-348-Ch2-3.txt

sha256: cd132d54681ee4210b4b8b9884cb9400c0e1d3ace8fc9920f727fe62f6f4bebe

2025-11-13_172722_NB-348-Ch2-3.png

sha256: c94899e66c6345deb8a3fa80c69f39a9df5466db582979df6d8f7b95427580b4



2025-11-13_164739_NB-348-Ch2-2.txt

sha256: 329b2176ba6f5c71772c2d5a71e7bc9f873ade9519a57b523000d4617d9366ff

2025-11-13_164739_NB-348-Ch2-2.png

sha256: eb48c3cbd05a45fb2ec9c62f8842c57942a70bace6bf43f4150a97dc12401cb2



2025-11-13_143340_NB-348-Ch2-1.txt

sha256: 7a086b9a99e1b0d804617c1fdf3a41d390aa584c2a6c991ae451d2f3b2434e67

2025-11-13_143340_NB-348-Ch2-1.png

sha256: b9d1a3946b87c1d2414501ceef02650f266fdbb166f9d2fe8852d6605e945c74




NB-348-Logger3-during irradiation.csv

sha256: aa3191355e2d47c59794710d7cd29ab1c6dbf36a5d78aab9821bcafab7c8b897

NB-348-Logger3-during irradiation.bmp

sha256: 2bb3ef49e43aa61c8257db42789f20ccc523f86bb4b2d45bd72871b5c867812a




NB-348-Logger2-calibration.csv

sha256: 32126ce8dd27c5e1a1d41426aee887f74508dfb6a16251318c9c4aa4a51ebb48


NB-348-Logger2-calibration step.bmp

sha256: 5b9167fbaaec0a86de2430987b12692d995663259855ade0ca0da20932bcbd0



NB-348-Logger2-2point calibration curve.bmp

sha256: 3c402381ba11d7856ba11d013839fd01f1fe349065edc2e9b87d1e0d7b244460




NB-348-Logger1-pre-polarization.csv

sha256: 8865cf86aa477bd068dfd70ccd8bc119832e9c5ba0a4d3981ae731519690feb0

NB-348-Logger1-pre-polarization.bmp

sha256: 92a58536b0e91bb288de7f942dc48206d00ad5f0a69a0a07b80ec8b255ea2d59



NB-348.ulog

sha256: dba5e90e51d937bc0765088ed92524211b507061dc7d57e8c4c7688de7efd33d



Unique eLabID: 20251113-8ffe0425ea5502329530add130e7c9fa57b959a2

Link: <https://elab.water-splitting.org/experiments.php?mode=view&id=3466>