

Parameters - Overview		(1st interrog.)
Mode		VVI-CLS
Basic rate/Night rate [bpm]		60/-----
CLS		120
Sensor/Rate fading [bpm]		120/-----
Upper rate response [bpm]		
Home Monitoring		OFF
Pulse amplitude [V]	⚡	1.6
Pulse width [ms]		0.4
Capture control		ON
Sensitivity [mV]		AUTO
Refract. period [ms]		250
Sensing polarity	⚡	BIPL
Pacing polarity	⚡	BIPL
Calculated ERI		10 Y. 9 Mo.



Parameters - Bradycardia		(1st interrog.)
Mode		VVI-CLS
Basic rate/Night rate [bpm]		60/-----
Basic rate [bpm]		60
Night rate [bpm]		
Night begins		
Night ends		
Hysteresis [bpm]		
Repetitive/Scan cycles		
Magnet response		ASync
CLS		120
CLS response		Medium
CLS resting rate control [bpm]		+20
Vp required		No
Sensor/Rate fading [bpm]		120/-----
Max. activity rate [bpm]		120
Sensor gain	⚡	AUTO
Sensor threshold		Medium
Rate fading		
Rate increase [bpm/cycle]		2
Rate decrease [bpm/cycle]		0.5
Upper rate response [bpm]		

Parameters - Bradycardia		(1st interrog.)
Pulse amplitude [V]	⌘	1.6
Pulse width [ms]		0.4
Capture control		ON
Threshold test start [V]		2.4
Safety margin [V]		1.0
Search type		Time of day
Interval [h]		
Time of day		00:30
Status		OK
Note		
Last measured threshold [V]		0.6
Date		03/14/2025
Time		08:46
Sensitivity [mV]		AUTO
Refract. period [ms]		250
Sensing polarity	⌘	BIPL
Pacing polarity	⌘	BIPL
Calculated ERI		10 Y. 9 Mo.

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 4/28

Parameters - Home Monitoring	(1st interrog.)
------------------------------	-----------------

General settings	
------------------	--

PID	89
-----	----

Home Monitoring	OFF
-----------------	-----

Time of transmission [hh:mm]	
------------------------------	--

High rate	
-----------	--

Event-based IEGM	
------------------	--

Last message	
--------------	--

Message type	---
--------------	-----

Message created on	
--------------------	--

Parameters - Diagnostics		(1st interrog.)
Recording triggers		
High rate		ON
Patient trigger		OFF
Pre-trigger recording [%]		75
IEGM signal		Filtered
Statistics		
Start resting period [hh:mm]		02:00
Resting period duration [h]		4.0
Thoracic impedance		ON
High rate		
High rate limit [bpm]		150
High rate counter [Events]		8
Enable lead check		ON

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 6/28

Parameters - Patient	(1st interrog.)
----------------------	-----------------

ID

Last name	295131-MRI
First name	Chen-Yan
Date of birth	04/25/1938
Gender	Female
Date of implant	03/24/2023
Hospital, City	CSMUH
Physician	Dr.Su
Phone	
Symptom	B2 Dizzy spells B3 Bradycardia

ECG indication	E6 Atrial flutter/fib. + Brady
----------------	--------------------------------

Etiology	XXX
----------	-----

LVEF [%]	XXX
----------	-----

NYHA	XXX
------	-----

Remark	
--------	--

Leads

Lead model	Manufacturer	Serial number	Type	Implantation	Channels
Solia S 60	Biotronik	8000717314	BIPL	03/24/2023	RV
				XX/XX/XXXX	
				XX/XX/XXXX	
				XX/XX/XXXX	

Diagnostics - Timing


(1st interrog.)

Rate histogram [% time]


■ Vs

□ Vp

181 days

**Event episodes and events****Pacing trend**

—V




Start 09/13/2024


Diagnostics - HF monitor

(1st interrog.)


Mean rate



Patient activity



Thoracic impedance



Start 09/13/2024




Diagnostics - 24 hours

(1st interrog.)


Rate

—V



% paced

—V




Start 09/13/2024

Diagnostics - More diag. - Event counters

(1st interrog.)

Event counters




Diagnostics - More diag. - Pulse ampl. and threshold**(1st interrog.)**

Ventricle

✕ Threshold

◇ Amplitude



Start 09/13/2024

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023



BIOTRONIK

excellence for life


PSW 2302.A/1

Page 12/28

Diagnostics - More diag. - Sensor rate

(1st interrog.)

Sensor rate [% time]



Start 09/13/2024

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 13/28



BIOTRONIK

excellence for life

Diagnostics - More diag. - Arrhythmia

(1st interrog.)

Ven. arrhythmia

High ventricular rate episodes [Count]

0

Start 09/13/2024

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023



BIOTRONIK

excellence for life

PSW 2302.A/1


Page 14/28

Diagnostics - More diag. - P/R wave trend

(1st interrog.)

P/R wave trend

—V




Start 09/13/2024

Diagnostics - More diag. - Lead impedance trend**(1st interrog.)**

Lead impedance trend

—V




Start 09/13/2024

Follow-up - Details


(1st interrog.)

Rate histogram [% time]




Long-term trends

Rate



Patient activity




Start 09/13/2024

TrendView

(1st interrog.)


Mean rate

—V




Pacing threshold

—V




P/R wave amplitude

—V



Lead impedance

—V



Start 09/13/2024

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 18/28

TrendView

(1st interrog.)

Pacing in V [%]

78

Start 09/13/2024

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 19/28



BIOTRONIK
excellence for life

Recordings - Episodes

No.	Date	Duration	Trigger	MVR	PC
-----	------	----------	---------	-----	----

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

**BIOTRONIK**

excellence for life

PSW 2302.A/1

Page 20/28

Follow-up	(1st interrog.)
-----------	-----------------

Patient

Name	295131-MRI, Chen-Yan
Last follow-up	09/13/2024
Date of implant	03/24/2023

Device status

Program count	4
Home Monitoring	OFF
MRI program	OFF

Mode	VVI-CLS
Basic rate [bpm]	60
Pulse amplitude [V]	1.6
Pulse width [ms]	0.4

Expected ERI	10 Y. 9 Mo.
Battery status	OK
Remaining battery capacity [%]	85

Test results

Sensing amplitude [mV]	5.5
Sensing polarity	BIPL

Threshold [V]	0.6
Pacing polarity	BIPL
Pulse width [ms]	0.4

Lead impedance [Ω]	721
Pacing polarity	BIPL

Diagnostics

Pacing [%]	78
------------	----

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:33

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 21/28



BIOTRONIK
excellence for life

Follow-up	(1st interrog.)
-----------	-----------------

Alerts and episodes

New episodes



0 recording(s)

Test - Sensing**Test program**

Mode	R measurement VVI
Basic rate [bpm]	40
Sensitivity [mV]	AUTO
Sensing polarity	BIPL

Measured values

Mean rate [bpm]	51
Min. amplitude [mV]	5.2
Mean amplitude [mV]	5.8
Max. amplitude [mV]	6.5

Freeze**(R measurement VVI)**

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:34

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023



BIOTRONIK


excellence for life

PSW 2302.A/1

Page 24/28

Freeze

(R measurement VVI)




Test - Pacing threshold



Test program

Mode	Ven. VVI
Basic rate [bpm]	90
Start amplitude [V]	1.6
Pulse width [ms]	0.4
Pacing polarity	BIPL
Number of pulses	2

Selected value

Threshold [V]	0.6
---------------	-----



Freeze

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:38

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

**BIOTRONIK**

excellence for life

PSW 2302.A/1

Page 27/28

Follow-up

Patient

Name	295131-MRI, Chen-Yan
Last follow-up	09/13/2024
Date of implant	03/24/2023

Device status

Program count	4
Home Monitoring	OFF
MRI program	OFF

Mode	VVI-CLS
Basic rate [bpm]	60
Pulse amplitude [V]	⌀ 1.6
Pulse width [ms]	0.4

Expected ERI	10 Y. 9 Mo.
Battery status	OK
Remaining battery capacity [%]	85

Test results

Sensing amplitude [mV]	⌀ 5.8
Sensing polarity	BIPL

Threshold [V]	0.6
Pacing polarity	BIPL
Pulse width [ms]	0.4

Lead impedance [Ω]	⌀ 721
Pacing polarity	BIPL

Diagnostics

Pacing [%]	78
------------	----

295131-MRI, Chen-Yan

Date of birth: 04/25/1938

Treating physician: Dr.Su

Recent follow-up: 03/14/2025 10:38

Last follow-up: 09/13/2024

Enitra 8 SR-T (S/N: 70321996 PID: 89)

Implantation: 03/24/2023

PSW 2302.A/1

Page 28/28

Follow-up

Alerts and episodes

New episodes

0 recording(s)