

BATCH RECORD · LOT-2026-300

Product: mAb-X Date: 2026-03-15

Step 1: Buffer Preparation

pH: 7.2

Temperature: 25.0 °C

Volume: 500 mL

Start Time: 08:30 Operator: JKL

Notes: Completed successfully.

Step 2: Centrifugation

Speed: 3000 g

Duration: ·

Temperature: ·

Start Time: 09:15

*** RUN ABORTED · Equipment failure at centrifugation step ***

Centrifuge vibration detected at 3000g. Run stopped per SOP.