






# Correcting Presbyopia With Autofocusing Liquid-Lens Eyeglasses

Mohit U. Karkhanis , Chayanjit Ghosh , Aishwaryadev Banerjee , Nazmul Hasan, Rugved Likhite ,  
Tridib Ghosh, Hanseup Kim, and Carlos H. Mastrangelo 

## I. INTRODUCTION

**Abstract—Objective:** Presbyopia, an age-related ocular disorder, is characterized by the loss in the accommodative abilities of the human eye. Conventional methods of correcting presbyopia divide the field of view, thereby resulting in significant vision impairment. We demonstrate the design, assembly and evaluation of autofocusing eyeglasses for restoration of accommodation without dividing the field of view. **Methods:** The adaptive optics eyeglasses comprise of two variable-focus liquid lenses, a time-of-flight range sensor and low-power, dual microprocessor control electronics, housed within an ergonomic frame. **Subject-specific accommodation deficiency models** were utilized to demonstrate high-fidelity accommodative correction. The abilities of this system to reduce accommodation deficiency, its power consumption, response time, optical performance and MTF were evaluated. **Results:** Average age corrected accommodation deficiencies for 5 subjects ranged from -0.021 D to 0.016 D. Each accommodation correction calculation was performed in  $\sim 67$  ms which consumed 4.86 mJ of energy. The optical resolution of the system was 10.5 cycles/degree, and featured a restorative accommodative range of 4.3 D. This system was capable of running for up to 19 hours between charge cycles and weighed  $\sim 132$  g. **Conclusion:** The design, assembly and performance of an autofocusing eyeglasses system to restore accommodation in presbyopes has been demonstrated. **Significance:** The new autofocusing eyeglasses system presented in this article has the potential to restore pre-presbyopic levels of accommodation in subjects diagnosed with presbyopia.

**Index Terms—**Autofocusing eyeglasses, presbyopia, accommodation, adaptive optics, smart eyeglasses, variable focus lens.

PRESBYOPIA is an age-related condition, which stems from the gradual loss in the eye's ability to change its optical power [1]. It manifests as an inability to focus on objects, often leading to symptoms like visual discomfort, eye strain and headache, after around 45 years of age [2]. The amount of accommodation loss as a function of age is characterized by Duane's curve [3], shown in Fig. 1(a).

This curve shows the accommodative range (or accommodative amplitude AA) of the human visual system versus age. A young person's visual system has about 12 D of accommodative range, but that range is reduced to about 1 D after age 45 (physically, this translates into being able to focus objects located between 0 m and 1/AA m). A young individual can thus see objects clearly between infinity up to 8 cm in front of them. An older, 60-year-old presbyope can only see objects between infinity to 1 m or worse as loss of accommodation often is accompanied by a fixed power offset. It should be noted that the Duane's curve describes the subjective amplitude of accommodation in the visual system with progressing age. This amplitude is also influenced by the size of the pupil which determines the depth of focus of the eye.

The precise mechanism underpinning the development and progression of presbyopia is still debated, but the principal reasons behind this condition are the loss of power in the ciliary muscles and the morphological changes in the crystalline lens of the eye [3]–[22]. In 2018, globally, more than 1.8 billion people were afflicted with this condition [23]. With increasing life expectancies, the number of people affected by this debilitating condition is expected to keep on rising [24]. Currently presbyopia can be partially corrected using several optical devices and methods listed below, none of which can restore normal vision.

### A. Multifocal Corrective Eyeglasses

Traditional methods used to correct vision loss due to presbyopia include the use of bifocal, trifocal, multifocal and progressive eyeglasses [25]. These devices feature lenses which divide the field of view into zones of different optical powers. The lens zoning enables a user to focus on near or far objects by adjusting their viewing angle. However, these mature technologies have serious drawbacks. Bifocal and trifocal eyeglasses provide a very fragmented field of view per zone which causes "image jumps" when the user adjusts their viewing angle [26]. Design of highly customized progressive lenses is now possible using free-form

Manuscript received January 21, 2021; revised April 23, 2021 and June 16, 2021; accepted June 18, 2021. Date of publication July 7, 2021; date of current version December 23, 2021. This work was supported by the U.S. National Institutes of Health NIBIB 1U01EB023048-01 cooperative Agreement. (Correspondence: Carlos H. Mastrangelo.)

Mohit U. Karkhanis, Chayanjit Ghosh, Aishwaryadev Banerjee, and Hanseup Kim are with the Electrical and Computer Engineering Department, University of Utah, USA.

Nazmul Hasan and Rugved Likhite are with Intel Corporation, USA.

Tridib Ghosh is with NewEyes, Inc, USA and also with the Electrical and Computer Engineering Department, University of Utah, USA.

Carlos H. Mastrangelo is with the Electrical and Computer Engineering Department, University of Utah, Salt Lake City, UT 84112 USA (e-mail: carlos.mastrangelo@utah.edu).

Digital Object Identifier 10.1109/TBME.2021.3094964




Fig. 1. (a) Representation of Duane's curve of accommodation loss versus age [3]. The amplitude of accommodation progressively reduces to  $<1.0$  D as a person enters their fifth decade. Note that the curve flattens out but does not become zero indicating that some degree of remnant accommodation is present. (b) Schematic of a smart eyeglasses system. The adaptive lenses continuously adjust the lens power to bring the object in focus for the observer along the line of sight. The lenses must provide the accommodation deficiency, a function of the object distance.

surfacing and sophisticated computer-aided modelling [27]–[29]. However, such lenses are expensive, have a steep learning curve and come with a penalty of unwanted astigmatism which distorts the user's peripheral vision [30]–[32]. Presbyopes who use multifocal eyeglasses are also more likely to experience accidents and mobility issues due to dioptric blurs in their field of view [33]–[39].

### B. Monovision and Multifocal Contact Lenses

The idea of correcting presbyopia using contact lenses has been around since the last five decades [40] but it has enjoyed limited success [26], [41], [42]. Presbyopic contact lens technologies have been extensively reviewed in literature [43]–[48] and can be divided into two main categories—monovision and multifocal contact lens correction. In monovision correction, the dominant eye is usually corrected for distance vision, while the other is corrected for near vision [49]–[54]. Many clinical trials have shown that after an adaptation period to monovision correction, presbyopes experienced an improved range of clear focus, at the expense of reduced contrast sensitivity and stereopsis [55]–[64]. There are also hard limitations on the optimum optical power of the near addition lens in monovision correction, as higher optical powers negatively impact stereopsis [49]–[51], [65]–[67]. Studies report that presbyopes successfully adapt to

multifocal contact lenses despite the presence of optical aberrations [68]–[74]. Despite such successes, contact lens correction of presbyopia remains unpopular due to discomfort and inconvenience, poor visual experience, and cost [75], [76].

### C. Intra-Ocular Lens Replacement

Many surgical approaches have been suggested for the correction of presbyopia, which involve replacing the natural crystalline lens of the eye with an artificial intra-ocular lens (IOL). These IOLs generally follow the same principles of refraction as the monovision and multifocal contact lenses mentioned above [47], [77]. However, this correction strategy requires detailed knowledge of the subject's ocular system, involves very precise lens calibrations before surgery, and a very delicate surgical procedure to implant the lens in the eye. Successes of this correction strategy are debatable as subjects have reported haloes and glare in their vision and low contrast sensitivity [78].

### D. Accommodating Lenses

A major problem with all the above methods is that they only provide a partial and unnatural tiled or warped correction of vision. To restore normal vision, the accommodation must be corrected actively, which means it must be dependent on the observed object distance, without any division in the field of view or viewing angle adjustments. Restoring active accommodation in such subjects, requires a variable optical power change of at least  $3.0$  D [26]. The earliest practical attempts at developing variable-focus eyeglasses were spearheaded by B. F. Edwards with his “polyfocal spectacles” in the 1950s [79]. These eyeglasses featured variable-focus, fluid-filled lenses [80] made from deformable membranes, a reservoir to store the optical fluid and a mechanism to “actuate” or control the amount of fluid in the variable-focus lens, thereby changing its optical power. Improvements in the actuation mechanism of such variable-focus lenses have been suggested by multiple inventors and researchers [81]–[84]. J. D. Silver's pioneering work to provide low-cost, variable-focus eyeglasses [85] was taken further by Adlens Ltd. with their adaptive eyewear products [86]. In addition to liquid-filled lenses, such variable-focus lenses can be implemented utilizing other technologies such as the sliding Alvarez lenses [87], electro-wetting lenses [88] and liquid crystal (LC) lenses [89]–[92], each holding its own set of advantages and pitfalls [93], [94]. The fabrication of slim, lightweight variable-focus lenses with 30–50 mm apertures necessary for eyeglasses, however, still remains a challenge.

### E. Smart Autofocusing Wearables

Recent advances in fluid-filled lenses [95]–[98] have enabled successful integration of large aperture variable-focus lenses into eyeglasses [98]–[101] integrated with object range sensors which measure the distance between the eyeglasses and the object a subject is trying to focus on. The range measured is utilized to automatically configure the power of the variable-focus lenses that corrects the accommodative deficiency [98], [99], as shown in the schematic of Fig. 1(b). Previously described unique

designs [100], [101] feature binocular eye tracking systems to analyze the gaze of the wearer and vary the optical power of the variable-focus lens. Such systems require investments in expensive computational devices and smartphones. However, authors in these articles show through clinical trials, that variable-focus eyeglasses outperform the traditional methods of correcting presbyopia.

A successful implementation of variable-focus eyeglasses requires the autofocusing system to reliably **tune correct optical powers with minimal delay**, while at the same time, crucial factors such as the system's electrical power consumption, affordability and aesthetics cannot be ignored. This paper attempts to provide a solution to the drawbacks of the existing methods of alleviating presbyopia. We utilize the adaptive optics technology developed by Hasan et.al. [95] **to implement autofocusing eyeglasses**. We demonstrate the performance of this system in terms of its ability to provide automatic, adequate optical correction to a presbyopic eye at 5 different distances ranging from 1 m to 30 cm, as presbyopes usually find it difficult to focus on objects placed within 1m. We discuss the electrical performance of this system and its ability to operate without an external power source. We also demonstrate the optical performance of this system using a 1951 USAF optical resolution test chart and we compare the modulation transfer function (MTF) of our variable-focus lens to that of an average human eye.

## II. MATERIALS AND METHODS

### A. System Assembly

The housing for the eyeglasses was designed in Blender 2.79b, an open source, 3D creation platform. This design was 3D printed at a 3D printing service provider (Xometry) using a nylon-based polymer and selective laser sintering (SLS) process. Electronic PCBs were designed in the electronic design automation (EDA) software, Eagle 9.6.2. These designs were manufactured on low profile, two-layered PCBs. Surface mount technology (SMT) components were used while designing the PCBs to keep their height as small as possible. Flat pack connectors (FPC) and cables were used to facilitate communications between different PCBs. Software necessary for the proper functioning of the eyeglasses, including subject-specific algorithms were programmed into the microcontrollers with Arduino integrated development environment (IDE). Finally, the lenses and the PCBs were fixed in their respective housing with the means of micro screws.

### B. Optical Measurements

The performance of the optical components and subsystems of the set were measured using the following methods:

**Accommodation vs Object Distance:** The performance of the variable-focus lenses, being the most critical subsystem within the autofocusing eyeglasses, were evaluated for their optical response with respect to different actuator voltages. The optical power of each lens was measured using a Shack-Hartman wavefront sensor (Thorlabs WFS150-7AR) while varying the




Fig. 2. Schematic of the experimental setup used to measure the power –distance relation of the variable-focus lens.

bimorph actuator voltage from -200V to +200V. This process was repeated 5 times.

To evaluate the system response (i.e., lens optical response) in terms of the accommodation stimulus (reciprocal of the object distance), an object was placed at 5 different distances (1 m, 70 cm, 50 cm, 40 cm, 30 cm) in front of the autofocusing system, as shown in Fig. 3. The ToF distance sensor and control electronics were attached to the same lens-holder which held the variable-focus lens. This allowed the distance sensor to gather real time distance data of an object placed directly in its line of sight and communicate this distance to the control electronics which set the corrective optical powers on the variable-focus lens. The relationship between the object distance and the lens corrective power was determined using subject-specific empirical accommodation deficiency models [102] programmed into the control electronics. Subject data necessary to test these eyeglasses were obtained from a recent, unpublished clinical trial study undertaken by authors of this article [102], [103].

We used a 633 nm light source and the Shack-Hartman wavefront sensor to measure the various optical powers generated by this system at various object distances, as shown in the schematic of Fig. 2. The distance sensor and the control electronics were attached to the variable-focus lens holder. The object was moved towards or away from the distance sensor while the autofocusing system automatically changed the optical power of the lens in response to the object distance. The minimum and maximum object distances were 30 cm and 1 m, respectively.

**Optical Resolution:** To evaluate the approximate optical resolution of the variable-focus lens undergoing automatic focus correction, we repeated the experiment described above while utilizing a 1951 USAF optical resolution test chart. Fig. 3 shows a schematic of the setup used for this experiment.

This test chart was placed at the 5 different distances from the eyeglasses, mentioned above, and a DSLR camera (Canon EOS 1200D) was used to obtain raw images of the resolution test chart through the tuned variable-focus lenses. Test chart illumination was kept constant using a diffused, LED studio lighting system. The correlated color temperature of the lighting system was kept constant at 5000 K. The DSLR camera's ISO, aperture and shutter speed were fixed at 1600, f/22 and 0.5 s, respectively.






Fig. 3. Schematic of the experimental setup to measure the optical resolution of the variable-focus lens. The 1951 USAF resolution test chart was mounted on a board.

The distance sensor and control electronics were attached to the variable-focus lens holder. The board was moved towards and away from the distance sensor, with minimum and maximum board-distance sensor distances being 30 cm and 1 m, respectively. The autofocusing system automatically tuned the variable-focus lens in response to the board distance.

**Modulation Transfer Function (MTF):** The MTF of an optical system is its intensity response as a function of spatial frequency [104] and a measure of the system optical quality. We estimated the 1-D MTF along the X-axis and Y-axis of the variable-focus lens used in the autofocusing eyeglasses directly from the aberration Zernike polynomial coefficients up to vertical quadrafoil (OSA/ANSI index Z14) [105], [106], which were measured with the help of the Shack-Hartman sensor. These estimated MTFs were then compared to those of an average human eye with different pupil diameters.

### III. RESULTS AND DISCUSSION

#### A. System Design

It is important to note that as a person ages, their accommodative ability is reduced, but not completely lost [107], [108]. Hence, remnant accommodation plays a crucial role when a presbyope is trying to achieve clear focus. In order to restore normal vision, the variable-focus lens must reproduce the accommodation deficiency (AD) curve [102] as shown in Fig. 4, which is the difference between the expected ideal response of the accommodative system (clear focus) and the degraded accommodation response (AR) of the presbyopic eye.

The AD must be provided by the accommodative optics in order to restore normal vision. Therefore, in practical terms, the subject's accommodation deficiency data should be programmed into an individual's autofocusing eyeglasses. Such data are a function of the accommodation stimulus (AS), which is the reciprocal of the distance between the eyes and the object plane where clear focus is expected.

In order to compensate for the AD, the autofocusing eyeglasses thus require a distance or a depth sensor that measures the range between the observer and the object being imaged. The autofocusing eyeglasses discussed in this paper thus consists of




Fig. 4. Example presbyope accommodation deficiency versus accommodative stimulus curve. Accommodation stimulus is the reciprocal of the object distance from the eyes, while the accommodation deficiency is the uncorrected defocus error present in the human ocular system. The solid line represents a typical AD model, developed in a recent study [102].

four key parts: (1) variable-focus lenses, (2) distance/ depth sensor, (3) control electronics, and (4) subject-specific AD models. Each of these subsystems is discussed below:

**1) Subject-Specific AD Models:** The AD curves vary quite a bit between presbyopes and, while fundamentally one can perform a full measurement of the AD and program it into the autofocusing eyeglasses, the cost of such a test can be prohibitive and a barrier to the use of the autofocusing system. Therefore, we chose to use model functions for the AD that require just a few parameters (and measurements), to describe the AD more economically. Multiple simplified accommodation response models, which express the accommodation response of an ocular system in terms of the object distance, have been suggested [102], [109]. In practice, utilization of these simple models result in significant correction power errors and ultimately, degradation of visual acuity [110]; therefore subject-specific accommodation deficiency models are necessary to provide optimal correction using the smart autofocusing eyeglasses.

We use a newly developed model described in [102], which relates the accommodation response to the object distance using a sigmoid function:

$$AR(AS, a, b, k) = \frac{a}{1 + e^{(-k(AS-b))}} \quad (1)$$

where  $AR$  is the accommodation response,  $AS$  is the accommodation stimulus (reciprocal of the object distance),  $a$  determines the accommodative amplitude of the individual i.e., the useful linear range of the experimental accommodation response curve,  $k$  determines range of stimuli for which the  $AR$  curve exhibits a linear response and parameter  $b$  corresponds to the midpoint of the non-flat accommodation stimulus range. It should be noted that the parameters  $a$ ,  $b$  and  $k$  are subject-specific and vary from eye to eye.

This model allows us to predict how a subject's accommodation system behaves as an object is moved within their field of view. The solid curve of Fig. 4 shows a fit of the model to an experimental AD data. Accommodation can then be restored in subjects by simply adding the deficiency in their accommodation




Fig. 5. (a) Working principle of liquid-filled variable-focus lens. (b) Simplified schematic of the variable-focus lens without the piezoelectric bimorph actuators [95].

system's behavior with the help of the distance data from the distance sensor and properly tuning lenses thereafter.

**2) Variable-Focus Lenses:** We utilize two squeezable-type liquid-filled lenses which do not require external fluid reservoirs. This significantly reduces the weight and profile of the variable-focus lens subsystem. Each lens consists of a cylindrical liquid-filled cavity, bound by two flexible polydimethylsiloxane (PDMS) membranes. Additional details of this lens, including its structure and electro-optical performance, have been described in [95].

Fig. 5(a) and (b) show the working principle and a simplified cross-section of this variable-focus lens without the actuators, respectively. Three curved piezoelectric bimorphs, placed along the periphery of the lens, form the actuators. One end of these bimorphs is fixed to the lens rim, while the other is connected to a hollow piston with three extended arms. The piston is attached to one of the deformable membranes. When voltage is applied to the actuators, they move the piston into or out of the lens, depending on the applied voltage amplitude and polarity. This makes a concave or a convex profile on the other membrane, changing the optical power of the lens.

Fig. 6(a) shows a photograph of a 30 mm aperture diameter tunable liquid lens. The lens weighs less than 15 g, making them suitable for eyeglasses applications, much lighter than other piezoelectric liquid lenses (eg. Optotune EL-16-40-TC [111]). Fig. 6(b) shows the optical response of the lens as a function of the applied voltage across the bimorph actuator. This lens features a repetitive linear optical response to the applied voltage, which greatly simplifies the electronics system required to control the lens. The response time of this lens is  $\sim 40$  ms with a variable-focus range of  $\sim 4.3$  D.

**3) Object Distance Sensor and Range Cutoff:** The role of a distance sensor in this system is crucial and requires precise measurement of the distance between the eyeglasses and the object to be focused. Distance sensors are characterized by their resolution, refresh rates (time taken per measurement) and the maximum range of reliable operation. IR (infrared)




Fig. 6. (a) Photograph of the variable-focus lens [95]. (b) Optical response of the variable-focus lens as a function of bimorph actuator voltage. Bars represent the standard deviation. The optical power of the lens is a linear function of the applied voltage.

time-of-flight (ToF) distance sensors (eg. RF Digital Simblee RFD77402) provide very fine resolution measurements, with rapid refresh rates and a reasonable working range of up to 2 m, corresponding to an accommodation stimulus of 0.5 D. The autofocus system regards objects located at distances  $> 2$  m as being reasonably well focused at the 0.5 D cut off. This is a reasonable first order approximation as the average RMS aberration of the human eye [112]–[116] translates to an equivalent defocus error of 0.05 D–0.49 D; thus lens adjustments smaller than about 0.5 D are marginally noticed. The RF Digital's Simblee RFD77402 distance sensor is an 880 nm (Class I laser) rangefinder with a  $29^\circ$  field of illumination and a  $55^\circ$  field of view. It should be noted that the distance sensor used in this system is only capable of detecting large, flat objects situated in front of the sensor, due to its  $55^\circ$  field of view. This sensor essentially computes its distance to the 2D projection of the 3D object placed in front of the sensor and is therefore not capable of perfectly replicating the mechanism through which the human visual system perceives multiple focal planes.

**4) High-Voltage Circuits and Feedback Control Electronics:** Fig. 7 shows a block diagram of the feedback control electronics. As evident from Fig. 6(b), the variable-focus lenses used within this system require a variable voltage between +200 V and -200 V. Efficient high-voltage DC-DC converters are necessary to convert the on-board battery voltage of 3.3 V to +200 V. We use a high voltage, ultra-miniature DC-DC piezo transformer for this operation. The bimorph actuators of each lens are connected to the high voltage converter in a H-bridge configuration with the help of four high-voltage optoelectronic switches (Ixys LAA100P). These optoelectronic switches help in setting the proper polarity of the voltage across the lenses.




Fig. 7. Block diagram of the entire autofocus system.

The high-voltage converter is operated within a feedback loop, such that it does not need to be switched on all the time, which increases the electrical power efficiency of this system.

We use two microcontrollers in a master-slave configuration to control the autofocus eyeglasses. The master microcontroller continuously fetches distance data from the ToF distance sensor, through an I<sup>2</sup>C bus and calculates the optical powers for the two lenses, using the distance data and the subject-specific correction algorithm. These two optical powers are then mapped to the corresponding voltage values, using the optical power-actuation voltage relation of the variable-focus lens, shown in Fig. 6(b). The master microcontroller sends these two voltage values to the slave microcontroller through a standard UART bus, operating at 19200 baud. The slave microcontroller is responsible for continuously converting the voltage data to pulse-width-modulated (PWM) control signals, which digitally control the analog output of the ultra-miniature DC-DC piezo transformer. A resistor-divider feedback network provides real-time measurements of the lens voltage, allowing our system to be precisely and power-efficiently controlled.

Prescriptions of subjects are bound to change as time and presbyopia progress within their ocular systems. It has also been shown that visual acuity and accommodation response are dependent on illumination [102], [117]–[119]. We incorporated a Bluetooth Low Energy (BLE) module within our system, which works in conjunction with a smartphone application. This allows the subjects to easily enter their updated prescriptions and user settings into the eyeglasses' control electronics. This BLE subsystem is connected to the master microcontroller through its principal hardware SPI bus, enabling over-the-air (OTA) upgrades for the entire system-controlling software, when required.

### B. Power Consumption

In practice, bulky power packs or batteries to reliably power sophisticated presbyopia correction systems for long hours, are unfeasible, inconvenient, and unacceptable to users. Presbyopes

require *continuous* accommodation correction and such technologies must stay operational, preferably without charging, to allow the users unimpeded functionality. We specifically utilize low power microcontrollers (Microchip ATmega32U4) and BLE SoC (Nordic nRF51822) in our design to facilitate better power efficiency. Key systems in the master microcontroller which are not used, like the analog-digital converter (ADC) and a few timers have been disabled to conserve power. The microcontrollers run with a clock frequency of 8 MHz to save power, without affecting the computation speed. The entire system is operated with a 3.7 V, 400 mAh, rechargeable lithium-ion polymer battery. We have also incorporated a battery charger (Microchip MCP73831) and a fuel gauge (Texas Instruments bq27441) which measures the remaining battery capacity and communicates battery status with the master microcontroller through a high speed I<sup>2</sup>C bus. Power is routed to the entire electronics system through a 3.3 V, low-noise, low dropout voltage regulator (MaxLinear SPX3819). A low power on/off switch (Maxim Integrated MAX16054) in conjunction with a push button constitute the power switching subsystem. We measured the current consumption of our system at 3.3 V with a 6 ½ digit multimeter (Agilent 34401A), to be ~21 mA. This value increased to 22 mA, during BLE communications with a smartphone application. However, software and prescription upgrades, which require BLE functionality, are rare and hence, the average steady-state current consumption of 21 mA can be safely approximated for this system. With the 400 mAh capacity lithium-ion polymer battery we use in our design, this ideally yields an average operational time of ~19 hours between every charge cycle, when the system is configured to perform 1 correction per second. However, battery age and frequency of use will determine the practical limits of this operational time.

### C. Accommodation Correction and Image Quality

The autofocus system's ability to provide adequate optical correction in presbyopes was evaluated by comparing their AD before and after optical correction with our system. Fig. 8(a)






Fig. 8. (a) Uncorrected accommodation deficiencies of 5 subjects before optical and active correction, plotted using models developed in [102]. (b) Average corrected accommodation deficiencies of the same 5 subjects after our autofocus system correction. Bars represent the standard deviation. Corrected AD does not exceed 0.2 D.

shows the deficiency models (solid lines) and deficiency data (markers) for 5 different eyes of uncorrected presbyopes, presented in [102]. The increase in their AD corresponding to increasing stimulus (i.e., decreasing object distance) can be clearly seen from the deficiency curves in Fig. 8(a). The AD increases from  $\sim 0$  D to  $>3.0$  D as the stimulus is progressively increased from 1.0 D (object distance = 1.0 m) to 3.3 D (object distance = 30 cm). Ideally, optical corrections using autofocus eyeglasses should reduce this stimulus-dependent deficiency in presbyopes to 0 D. Fig. 8(b) shows the deficiency in the same 5 presbyopia subjects after correction using our autofocus system. The curves represent difference in the optical correction determined by the subject-specific AD models and our system's response measured at 5 object distances (1 m, 70 cm, 50 cm, 40 cm and 30 cm). Our results indicate that the average corrected AD for the 5 subjects ranged from  $-0.021$  D to  $0.016$  D over the 5 object distances, using our autofocus system. The maximum corrected deficiency in presbyopes did not exceed 0.4 D. The system was able to repeatedly generate the optical powers as determined by the presbyopia correction algorithms for different subjects.

Optical resolution plays an important part in determining the effectiveness of autofocus systems to be used for presbyopia correction. To that extent, we also measured the approximate, subjective optical resolution of this system with the help of a




Fig. 9. Photos of the USAF optical resolution test chart through the variable-focus lens, placed at (a) 30 cm, (b) 40 cm and (c) 50 cm. The accommodation deficiency model for subject no. 3 was used in conjunction with the autofocus system for these images [102]. The photos highlight the smallest, clearly visible element on the optical resolution test chart.

1951 USAF optical resolution test chart. Fig. 9(a)-(c) show the images of the resolution test chart placed at different distances away from variable-focus lens. The autofocus system was running a subject-specific algorithm, developed in [102], and automatically set the appropriate optical powers (necessary to correct that subject's AD) on the variable-focus lens. We observed the smallest, clearly visible elements on the test chart. Our observations show that the resolution of this system was between 0.75 – 1.00 lp/mm.

An autofocus system intended to provide optical correction to a presbyope should also have optical quality comparable to that of an average human eye. We measured the 1-D vertical and horizontal MTFs for the variable-focus lens and compared it to that of a best-corrected human eye. An analytical formula




Fig. 10. Comparison of the 1-D MTF of variable-focus lens used in the autofocus system and the radial MTF of the human eye. The solid red and blue lines represent the 1-D MTFs of the variable focus lens along the X (horizontal) and Y (vertical) axes, respectively.

constructed by Watson [120] was used to model the pupil-diameter dependent MTFs for a human eye. Fig. 10 compares the 1-D MTFs, measured for our lens to those modelled for a best-corrected human eye. We report an average MTF-50 (i.e., spatial frequency at 50% MTF) value of 10.5 cycles/degree for our variable-focus lens.

The green shaded area represents the range of the MTFs for a best corrected human eye. Comparison of the X and Y profiles of the 1-D MTFs of the variable-focus lenses shows the presence of astigmatism. However, it should be noted that such liquid-filled lenses are subject to such optical aberrations due to gravity-induced coma [121] and unavoidable lens fabrication errors.

#### D. Response Time

The electronics control system and the lenses introduce inherent time delays in the autofocus system. As reported before, the lenses have a response time of  $\sim 40$  ms, i.e., it takes  $\sim 40$  ms from the moment a particular voltage is applied to the actuators of the lenses till the point where the lenses produce the expected optical power.

Our measurements show that the average time taken by the electronics subsystem to produce the required voltages for the lenses is  $\sim 67$  ms. This response time takes into account the time required by the distance sensor to complete its ranging operation, the time required by the master microcontroller to extrapolate the correct optical powers and their related voltages, and the time required by the slave microcontroller to set these voltages across the lens actuators. Hence, the total time taken by this system from the moment distance ranging begins, till the point where the lenses produce the appropriate optical power, is 107 ms. This enables us to have a maximum refresh rate of  $\sim 10$  Hz, i.e., the system can change the optical powers of the lenses up to 10 times a second.

#### E. Ergonomics, Aesthetics and Portability

Autofocusing systems are meant to be worn all the time, throughout the daily routine of a presbyope, in order to provide




Fig. 11. Dimensions of the lens and electronics housing in the autofocus eyeglasses system. (a) and (b) show the top and front views of the lens and ToF distance sensor housing, respectively, while (c) and (d) show the side and front views of the electronics housing, respectively. Dimensions are in mm.

seamless accommodation. In practical terms, this requires the system to be lightweight, comfortable and fashionable enough to be worn during daily engagements. Engineering limitations impose strict restrictions on the sizes of various components within our system, for example, the lenses cannot be thinned down any further, without sacrificing their electro optical performance. Similarly, the housing for electronics cannot be any thinner without sacrificing battery capacity. Despite such limitations, our system attempts to provide a frame design which is much *sleeker* compared to recent similar technologies [98], [100], [101]. Figs. 11(a)-(d) show the mechanical dimensions of the autofocus eyeglasses.

Fig. 12(a) shows the arrangement of the internal components in these eyeglasses. The maximum *thickness* of the frame is 14.46 mm, while the maximum *height* of the electronics housing is 28.00 mm. The diameter of the lens housing is 51 mm to house lenses with diameter 50 mm. The 0.5 mm annular gap between the lens and its housing helps in snugly fitting FPC communications cables and the HV wires for the lenses. The total weight of these autofocus eyeglasses is 132.06 g. Assuming a vertex distance of 15 mm, each variable-focus lens utilized within this design provides an angular field-of-view of  $93.69^\circ$ . Our system has been designed to automatically focus on the object which is directly in the line-of-sight of the ToF distance sensor.

Spherical and cylindrical solid lenses, which are used to correct myopia/hyperopia and astigmatism, respectively, can be incorporated into our system by designing slots to house such lenses within the back cover of the autofocus eyeglasses system. The slots which house the solid lenses are 3D printed using a flexible resin which enables easy replacement of additional spherical or cylindrical corrections necessary for the subject. It should be noted that adding spherical corrections will linearly shift the maximum and minimum optical powers of the varifocal lens, but not change the range of the entire optical system, thereby maintaining the ability to provide additional accommodation.






Fig. 12. (a) Autofocusing eyeglasses from the inside. This photo shows the various subsystems and their placements within the eyeglasses. (b) Ergonomic fit of autofocused eyeglasses resembles oversized sunglasses. The reflective covers hide the variable-focus lenses.

#### IV. CONCLUSION

We demonstrated the design, assembly, and performance of an autofocusing eyeglasses system which can potentially restore accommodation in presbyopes. These autofocusing eyeglasses consist of two, variable-focus, liquid-filled lenses, a ToF distance sensor and a rechargeable battery-powered electronics control system, which utilizes subject-specific accommodation deficiency models to restore pre-presbyopic levels of accommodation in them. This system was evaluated on its ability to perform high fidelity and accurate optical corrections based on presbyopic subject data from a clinical trial. The optical resolution (MTF-50) of our autofocusing eyeglasses is 10.5 cycles/degree. Our system can provide up to 4.3 D in accommodation correction.

#### REFERENCES

- [1] W. J. Benjamin and I. M. Borish, *Borish's Clinical Refraction*. 2nd ed., St. Louis, MO, USA: Butterworth Heinemann/Elsevier, 2006. Accessed: Sep. 02, 2020. [Online]. Available: <https://www.worldcat.org/title/borishs-clinical-refraction/oclc/785829290>
- [2] I. Patel and S. K. West, "Presbyopia: Prevalence, impact, and interventions," *Community Eye Health J.*, vol. 20, no. 63, pp. 40–41, Sep. 2007. Accessed: Sep. 02, 2020. [Online]. Available: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2040246/>
- [3] A. Duane, "Studies in monocular and binocular accommodation with their clinical applications," *Amer. J. Ophthalmol.*, vol. 5, no. 11, pp. 865–877, 1922, doi: [10.1016/S0002-9394\(22\)90793-7](https://doi.org/10.1016/S0002-9394(22)90793-7).

- [4] F. C. Donders, *On the Anomalies of Accommodation and Refraction of the Eye*. London, U.K.: New Sydenham Society, 1864, doi: [10.1001/archophth.1986.01050210035014](https://doi.org/10.1001/archophth.1986.01050210035014).
- [5] C. Hess, "Arbeiten aus dem Gebiete der Accommodationslehre," *Albrecht von Graefes Arch. für Ophthalmol.*, vol. 46, pp. 243–276, 1896.
- [6] R. Weale, "Presbyopia toward the end of the 20th century," *Surv. Ophthalmol.*, vol. 34, no. 1, pp. 15–30, Jul. 1989, doi: [10.1016/0039-6257\(89\)90126-4](https://doi.org/10.1016/0039-6257(89)90126-4).
- [7] J. L. Semmlow, et al., "The relationship between ciliary muscle contraction and accommodative response in the presbyopic eye," in *Presbyopia Research. Perspectives in Vision Research*. Boston, MA, USA: Springer, L. W. Stark, Ed., 1991, pp. 245–253, doi: [10.1007/978-1-4757-2131-7\\_24](https://doi.org/10.1007/978-1-4757-2131-7_24).
- [8] B. K. Pierscionek, "What we know and understand about presbyopia," *Clin. Exp. Optometry*, vol. 76, no. 3, pp. 83–90, May 1993, doi: [10.1111/j.1444-0938.1993.tb05095.x](https://doi.org/10.1111/j.1444-0938.1993.tb05095.x).
- [9] D. A. Atchison, "Accommodation and presbyopia," *Ophthalmic Physiol. Opt.*, vol. 15, no. 4, pp. 255–272, Jul. 1995, doi: [10.1046/j.1475-1313.1995.9500020e.x](https://doi.org/10.1046/j.1475-1313.1995.9500020e.x).
- [10] B. Gilmartin, "The aetiology of presbyopia: A summary of the role of lenticular and extralenticular structures," *Ophthalmic Physiol. Opt.*, vol. 15, no. 5, pp. 431–437, 1995.
- [11] J. A. Mordi and K. J. Ciuffreda, "Static aspects of accommodation: Age and presbyopia," *Vis. Res.*, vol. 38, no. 11, pp. 1643–1653, Jun. 1998, doi: [10.1016/S0042-6989\(97\)00336-2](https://doi.org/10.1016/S0042-6989(97)00336-2).
- [12] M. A. Croft, A. Glasser, and P. L. Kaufman, "Accommodation and presbyopia," *Ophthalmic Physiol. Opt.*, vol. 41, no. 2, pp. 33–46, 2001, doi: [10.1097/00004397-200104000-00005](https://doi.org/10.1097/00004397-200104000-00005).
- [13] W. N. Charman, "The eye in focus: Accommodation and presbyopia," *Clin. Exp. Optometry*, vol. 91, no. 3, pp. 207–225, May 2008, doi: [10.1111/j.1444-0938.2008.00256.x](https://doi.org/10.1111/j.1444-0938.2008.00256.x).
- [14] A. Gullstrand, *Die Optische Abbildung in Heterogen Medien Die Dioptrik Der Kristalllinse Des Menschen*. Uppsala, Sweden: Almqvist & Wiksell, 1908.
- [15] A. Glasser, "Presbyopia and aging in the crystalline lens," *J. Vis.*, vol. 3, no. 12, pp. 22–22, Mar. 2010, doi: [10.1167/3.12.22](https://doi.org/10.1167/3.12.22).
- [16] S. A. Strenk, L. M. Strenk, and J. F. Koretz, "The mechanism of presbyopia," *Prog. Retinal Eye Res.*, vol. 24, no. 3, pp. 379–393, May 2005, doi: [10.1016/j.preteyeres.2004.11.001](https://doi.org/10.1016/j.preteyeres.2004.11.001).
- [17] R. F. Fisher, "The ciliary body in accommodation," *Trans. Ophthalmol. Soc. U.K.*, vol. 105, no. Pt 2, pp. 208–219, 1986, Accessed: Apr. 13, 2020. [Online]. Available: <http://www.ncbi.nlm.nih.gov/pubmed/3467496>
- [18] H. von Helmholtz, *Helmholtz's Treatise on Physiological Optics*. Washington, DC, USA: Optical Society of America, 1924, doi: [10.1037/13536-000](https://doi.org/10.1037/13536-000).
- [19] E. F. Fincham, *The Mechanism of Accommodation*. London, U.K.: G. Pulman & Sons, Limited, 1937, pp. 5–80.
- [20] D. Adler-Grinberg, "Questioning our classical understanding of accommodation and presbyopia," *Optometry Vis. Sci.*, vol. 63, no. 7, pp. 571–580, 1986, doi: [10.1097/00006324-198607000-00012](https://doi.org/10.1097/00006324-198607000-00012).
- [21] M. W. Morgan, "The ciliary body in accommodation and accommodative-convergence," *Optometry Vis. Sci.*, vol. 31, no. 5, pp. 219–229, 1954, doi: [10.1097/00006324-195405000-00001](https://doi.org/10.1097/00006324-195405000-00001).
- [22] E. F. Fincham, "The proportion of ciliary muscular force required for accommodation," *J. Physiol.*, vol. 128, no. 1, pp. 99–112, Apr. 1955, doi: [10.1113/jphysiol.1955.sp005293](https://doi.org/10.1113/jphysiol.1955.sp005293).
- [23] T. R. Fricke, et al., "Global prevalence of presbyopia and vision impairment from uncorrected presbyopia systematic review, meta-analysis, and modelling," *Ophthalmology*, vol. 125, pp. 1492–1499, 2018, doi: [10.1016/j.ophtha.2018.04.013](https://doi.org/10.1016/j.ophtha.2018.04.013).
- [24] W. He, D. Goodkind, and P. Kowal, "An aging world: 2015," pp. 3–11, Mar. 2016. [Online]. Available: <https://www.census.gov/content/dam/Census/library/publications/2016/demo/p95-16-1.pdf>
- [25] M. Jalie, *Ophthalmic Lenses and Dispensing*, 3rd ed. London, U.K.: Butterworth-Heinemann, 2008, Accessed: Sep. 10, 2020. [Online]. Available: <https://pdfs.semanticscholar.org/051f/a0e8f17f0d4a3b96cf1f8d4a6058d3f410b9.pdf>
- [26] W. N. Charman, "Developments in the correction of presbyopia I: Spectacle and contact lenses," *Ophthalmic Physiol. Opt.*, vol. 34, no. 1, pp. 8–29, Jan. 2014, doi: [10.1111/opo.12091](https://doi.org/10.1111/opo.12091).
- [27] D. J. Meister and S. W. Fisher, "Progress in the spectacle correction of presbyopia. Part 1: Design and development of progressive lenses," *Clin. Exp. Optometry*, vol. 91, no. 3, pp. 240–250, May 2008, doi: [10.1111/j.1444-0938.2007.00245.x](https://doi.org/10.1111/j.1444-0938.2007.00245.x).

- [28] D. J. Meister and S. W. Fisher, "Progress in the spectacle correction of presbyopia. Part 2: Modern progressive lens technologies," *Clin. Exp. Optometry*, vol. 91, no. 3, pp. 251–264, May 2008, doi: [10.1111/j.1444-0938.2008.00246.x](https://doi.org/10.1111/j.1444-0938.2008.00246.x).
- [29] G. Savio, G. Concheri, and R. Meneghello, "Progressive lens design by discrete shape modelling techniques," *Int. J. Interact. Des. Manuf.*, vol. 7, no. 3, pp. 135–146, Aug. 2013, doi: [10.1007/s12008-012-0170-z](https://doi.org/10.1007/s12008-012-0170-z).
- [30] G. Minkwitz, "Über den Flächenastigmatismus Bei Gewissen Symmetrischen Asphären," *Opt. Acta Int. J. Opt.*, vol. 10, no. 3, pp. 223–227, Jul. 1963, doi: [10.1080/713817794](https://doi.org/10.1080/713817794).
- [31] J. E. Sheedy, *et al.*, "Progressive powered lenses: The Minkwitz theorem," *Optometry Vis. Sci.*, vol. 82, no. 10, pp. 916–922, Oct. 2005, doi: [10.1097/01.opx.0000181266.60785.c9](https://doi.org/10.1097/01.opx.0000181266.60785.c9).
- [32] T. L. Alvarez, E. H. Kim, and B. Granger-Donetti, "Adaptation to progressive additive lenses: Potential factors to consider," *Sci. Rep.*, vol. 7, no. 2529, pp. 1–14, May 2017, doi: [10.1038/s41598-017-02851-5](https://doi.org/10.1038/s41598-017-02851-5).
- [33] S. R. Lord, J. Dayhew, and A. Howland, "Multifocal glasses impair edge-contrast sensitivity and depth perception and increase the risk of falls in older people," *J. Amer. Geriatr. Soc.*, vol. 50, no. 11, pp. 1760–1766, Nov. 2002, doi: [10.1046/j.1532-5415.2002.50502.x](https://doi.org/10.1046/j.1532-5415.2002.50502.x).
- [34] J. C. Davies, *et al.*, "Bifocal/varifocal spectacles, lighting and missed-step accidents," *Saf. Sci.*, vol. 38, no. 3, pp. 211–226, Aug. 2001, doi: [10.1016/S0925-7535\(01\)00002-9](https://doi.org/10.1016/S0925-7535(01)00002-9).
- [35] D. B. Elliott and G. J. Chapman, "Adaptive gait changes due to spectacle magnification and dioptric blur in older people," *Investig. Ophthalmol. Vis. Sci.*, vol. 51, no. 2, pp. 718–722, Feb. 2010, doi: [10.1167/iov.09-4250](https://doi.org/10.1167/iov.09-4250).
- [36] G. J. Chapman, A. J. Scally, and D. B. Elliott, "Adaptive gait changes in older people due to lens magnification," *Ophthalmic Physiol. Opt.*, vol. 31, no. 3, pp. 311–317, May 2011, doi: [10.1111/j.1475-1313.2010.00807.x](https://doi.org/10.1111/j.1475-1313.2010.00807.x).
- [37] A. A. Black, J. A. Kimlin, and J. M. Wood, "Stepping accuracy and visuomotor control among older adults: Effect of target contrast and refractive blur," *Ophthalmic Physiol. Opt.*, vol. 34, no. 4, pp. 470–478, 2014, doi: [10.1111/opo.12141](https://doi.org/10.1111/opo.12141).
- [38] E. Supuk, *et al.*, "Dizziness, but not falls rate, improves after routine cataract surgery: The role of refractive and spectacle changes," *Ophthalmic Physiol. Opt.*, vol. 36, no. 2, pp. 183–190, Mar. 2016, doi: [10.1111/opo.12243](https://doi.org/10.1111/opo.12243).
- [39] D. B. Elliott, *et al.*, "Intermediate addition multifocals provide safe stair ambulation with adequate 'short-term' reading," *Ophthalmic Physiol. Opt.*, vol. 36, no. 1, pp. 60–68, Jan. 2016, doi: [10.1111/opo.12236](https://doi.org/10.1111/opo.12236).
- [40] S. Duke-Elder and D. Abrams, *System of Ophthalmology Vol V: Ophthalmic Optics and Refraction*. St. Louis, MO, USA: The CV Mosby Company, 1970, p. 783.
- [41] P. B. Morgan and N. Efron, "Contact lens correction of presbyopia," *Contact Lens Anterior Eye*, vol. 32, no. 4, pp. 191–192, Aug. 2009, doi: [10.1016/j.clae.2009.05.003](https://doi.org/10.1016/j.clae.2009.05.003).
- [42] P. B. Morgan, *et al.*, "An international survey of contact lens prescribing for presbyopia," *Clin. Exp. Optometry*, vol. 94, no. 1, pp. 87–92, Jan. 2011, doi: [10.1111/j.1444-0938.2010.00524.x](https://doi.org/10.1111/j.1444-0938.2010.00524.x).
- [43] W. J. Benjamin and I. M. Borish, "Presbyopic correction with contact lenses," in *Borish's Clinical Refraction*. W. J. Benjamin, Ed., Philadelphia, PA, USA: W.B. Saunders, 1998, pp. 1022–1050.
- [44] E. S. Bennett, "Contact lens correction of presbyopia," *Clin. Exp. Optometry*, vol. 91, no. 3, pp. 265–278, May 2008, doi: [10.1111/j.1444-0938.2007.00242.x](https://doi.org/10.1111/j.1444-0938.2007.00242.x).
- [45] J. Meyler, "Presbyopia," in *Contact Lens Practice*. 2nd ed., N. Efron, Ed. London, U.K.: Butterworth-Heinemann, 2010, pp. 252–265. Accessed: Sep. 11, 2020. [Online]. Available: <https://www.elsevier.com/books/contact-lens-practice/efron/978-0-7506-8869-7>
- [46] P. Kallinikos, J. Santodomingo-Rubido, and S. Plainis, "Correction of presbyopia with contact lenses," in *Presbyopia: Origins, Effects, and Treatment*. I. G. Pallikaris, S. Plainis, and W. N. Charman, Eds., Thorofare, NJ, USA: Slack, 2012, pp. 127–137. Accessed: Sep. 11, 2020. [Online]. Available: <https://www.ofthalmologos.org.ar/catalogo/items/show/5482>
- [47] J. S. Wolffsohn and L. N. Davies, "Presbyopia: Effectiveness of correction strategies," *Prog. Retinal Eye Res.*, vol. 68, pp. 124–143, Jan. 2019, doi: [10.1016/j.preteyeres.2018.09.004](https://doi.org/10.1016/j.preteyeres.2018.09.004).
- [48] P. S. Kollbaum and A. Bradley, "Correction of presbyopia: old problems with old (and new) solutions," *Clin. Exp. Optometry*, vol. 103, no. 1, pp. 21–30, Jan. 2020, doi: [10.1111/cxo.12987](https://doi.org/10.1111/cxo.12987).
- [49] B. J. W. Evans, "Monovision: A review," *Ophthalmic Physiol. Opt.*, vol. 27, no. 5, pp. 417–439, Sep. 2007, doi: [10.1111/j.1475-1313.2007.00488.x](https://doi.org/10.1111/j.1475-1313.2007.00488.x).
- [50] S. Jain, I. Arora, and D. T. Azar, "Success of monovision in presbyopes: Review of the literature and potential applications to refractive surgery," *Surv. Ophthalmol.*, vol. 40, no. 6, pp. 491–499, May 1996, doi: [10.1016/S0039-6257\(96\)82015-7](https://doi.org/10.1016/S0039-6257(96)82015-7).
- [51] K. R. Johannsdottir and L. B. Stelmach, "Monovision: A review of the scientific literature," *Optometry Vis. Sci.*, vol. 78, no. 9, pp. 646–651, Sep. 2001, Accessed: Sep. 11, 2020. [Online]. Available: [https://journals-lww-com.ezproxy.lib.utah.edu/optvissci/Fulltext/2001/09000/Monovision\\_a\\_Review\\_of\\_the\\_Scientific\\_Literature.9.aspx](https://journals-lww-com.ezproxy.lib.utah.edu/optvissci/Fulltext/2001/09000/Monovision_a_Review_of_the_Scientific_Literature.9.aspx)
- [52] R. Legras, *et al.*, "Effect of induced anisometropia on binocular through-focus contrast sensitivity," *Optometry Vis. Sci.*, vol. 78, no. 7, pp. 503–509, 2001, doi: [10.1097/00006324-200107000-00013](https://doi.org/10.1097/00006324-200107000-00013).
- [53] G. Vandermeer, *et al.*, "Subjective through-focus quality of vision with various versions of modified monovision," *Brit. J. Ophthalmol.*, vol. 99, no. 7, pp. 997–1003, Jul. 2015, doi: [10.1136/bjophthalmol-2014-305437](https://doi.org/10.1136/bjophthalmol-2014-305437).
- [54] S. Ravikumar, *et al.*, "Expanding binocular depth of focus by combining monovision with diffractive bifocal intraocular lenses," *J. Cataract Refract. Surg.*, vol. 42, no. 9, pp. 1288–1296, Sep. 2016, doi: [10.1016/j.jcrs.2016.04.036](https://doi.org/10.1016/j.jcrs.2016.04.036).
- [55] N. Gupta, J. S. W. Wolffsohn, and S. A. Naroo, "Comparison of near visual acuity and reading metrics in presbyopia correction," *J. Cataract Refract. Surg.*, vol. 35, no. 8, pp. 1401–1409, Aug. 2009, doi: [10.1016/j.jcrs.2009.03.026](https://doi.org/10.1016/j.jcrs.2009.03.026).
- [56] J. Woods, C. A. Woods, and D. Fonn, "Early symptomatic presbyopes—What correction modality works best?," *Eye Contact Lens*, vol. 35, no. 5, pp. 221–226, Sep. 2009, doi: [10.1097/ICL.0b013e3181b5003b](https://doi.org/10.1097/ICL.0b013e3181b5003b).
- [57] J. Woods, C. Woods, and D. Fonn, "Visual performance of a multifocal contact lens versus monovision in established presbyopes," *Optometry Vis. Sci.*, vol. 92, no. 2, pp. 175–182, 2015, doi: [10.1097/OPX.0000000000000476](https://doi.org/10.1097/OPX.0000000000000476).
- [58] A. Sivardeen, D. Laughton, and J. S. Wolffsohn, "Randomized crossover trial of silicone hydrogel presbyopic contact lenses," *Optometry Vis. Sci.*, vol. 93, no. 2, pp. 141–149, Feb. 2016, doi: [10.1097/OPX.0000000000000780](https://doi.org/10.1097/OPX.0000000000000780).
- [59] L. Imbeau, *et al.*, "Presbyopia compensation: Looking for cortical predictors," *Brit. J. Ophthalmol.*, vol. 101, no. 2, pp. 223–226, Feb. 2017, doi: [10.1136/bjophthalmol-2015-307581](https://doi.org/10.1136/bjophthalmol-2015-307581).
- [60] F. Zeri, *et al.*, "Immediate cortical adaptation in visual and non-visual areas functions induced by monovision," *J. Physiol.*, vol. 596, no. 2, pp. 253–266, Jan. 2018, doi: [10.1113/JP274896](https://doi.org/10.1113/JP274896).
- [61] J. J. Castro, *et al.*, "Binocular summation and visual function with induced anisocoria and monovision," *Biomed. Opt. Exp.*, vol. 7, no. 10, pp. 4250–4262, Oct. 2016, doi: [10.1364/boe.7.004250](https://doi.org/10.1364/boe.7.004250).
- [62] R. Nabie, *et al.*, "Effect of artificial anisometropia in dominant and nondominant eyes on stereoacuity," *Can. J. Ophthalmol.*, vol. 52, no. 3, pp. 240–242, Jun. 2017, doi: [10.1016/j.jcjo.2016.11.008](https://doi.org/10.1016/j.jcjo.2016.11.008).
- [63] E. J. Fernández, *et al.*, "Impact on stereo-acuity of two presbyopia correction approaches: Monovision and small aperture inlay," *Biomed. Opt. Exp.*, vol. 4, no. 6, Jun. 2013, Art. no. 830, doi: [10.1364/boe.4.000822](https://doi.org/10.1364/boe.4.000822).
- [64] M. H. Freeman and W. N. Charman, "An exploration of modified monovision with diffractive bifocal contact lenses," *Contact Lens Anterior Eye*, vol. 30, no. 3, pp. 189–196, Jul. 2007, doi: [10.1016/j.clae.2006.12.006](https://doi.org/10.1016/j.clae.2006.12.006).
- [65] K. Hayashi, *et al.*, "Optimal amount of anisometropia for pseudophakic monovision," *J. Refract. Surg.*, vol. 27, no. 5, pp. 332–338, May 2011, doi: [10.3928/1081597X-20100817-01](https://doi.org/10.3928/1081597X-20100817-01).
- [66] Z. F. Pollard, *et al.*, "Strabismus precipitated by monovision," *Amer. J. Ophthalmol.*, vol. 152, no. 3, Sep. 2011, Art. no. 482, doi: [10.1016/j.ajo.2011.02.008](https://doi.org/10.1016/j.ajo.2011.02.008).
- [67] C. E. Smith, *et al.*, "Monovision: Consequences for depth perception from large disparities," *Exp. Eye Res.*, vol. 183, pp. 62–67, Jun. 2019, doi: [10.1016/j.exer.2018.09.005](https://doi.org/10.1016/j.exer.2018.09.005).
- [68] P. R. B. Fernandes, *et al.*, "Adaptation to multifocal and monovision contact lens correction," *Optometry Vis. Sci.*, vol. 90, no. 3, pp. 228–235, Mar. 2013, doi: [10.1097/OPX.0b013e318282951b](https://doi.org/10.1097/OPX.0b013e318282951b).
- [69] D. Rio, K. Woog, and R. Legras, "Effect of age, decentration, aberrations and pupil size on subjective image quality with concentric bifocal optics," *Ophthalmic Physiol. Opt.*, vol. 36, no. 4, pp. 411–420, Jul. 2016, doi: [10.1111/opo.12300](https://doi.org/10.1111/opo.12300).
- [70] A. Bradley, *et al.*, "Impact of contact lens zone geometry and ocular optics on bifocal retinal image quality," *Ophthalmic Physiol. Opt.*, vol. 34, no. 3, pp. 331–345, 2014, doi: [10.1111/opo.12110](https://doi.org/10.1111/opo.12110).
- [71] R. Xu, *et al.*, "Modelling the effects of secondary spherical aberration on refractive error, image quality and depth of focus," *Ophthalmic Physiol. Opt.*, vol. 35, no. 1, pp. 28–38, Jan. 2015, doi: [10.1111/opo.12185](https://doi.org/10.1111/opo.12185).



- [72] C. Fedtke, *et al.*, "Impact of spherical aberration terms on multifocal contact lens performance," *Optomatory Vis. Sci.*, vol. 94, no. 2, pp. 197–207, Feb. 2017, doi: [10.1097/OPX.0000000000001017](https://doi.org/10.1097/OPX.0000000000001017).
- [73] E. Kim, R. C. Bakaraju, and K. Ehrmann, "Power profiles of commercial multifocal soft contact lenses," *Optomatory Vis. Sci.*, vol. 94, no. 2, pp. 183–196, Feb. 2017, doi: [10.1097/OPX.0000000000000998](https://doi.org/10.1097/OPX.0000000000000998).
- [74] P. Fernandes, *et al.*, "Light disturbance with multifocal contact lens and monovision for presbyopia," *Contact Lens Anterior Eye*, vol. 41, no. 4, pp. 393–399, Aug. 2018, doi: [10.1016/j.clae.2018.03.006](https://doi.org/10.1016/j.clae.2018.03.006).
- [75] A. D. Pucker and A. A. Tichenor, "A review of contact lens dropout," *Clin. Optomatory*, vol. 2, pp. 85–94, Jun. 2020, doi: [10.2147/OPTO.S198637](https://doi.org/10.2147/OPTO.S198637).
- [76] B. Hutchins and B. Huntjens, "Patients' attitudes and beliefs to presbyopia and its correction," *J. Optomatory*, vol. 14, no. 2, pp. 127–132, Jun. 2021, doi: [10.1016/j.optom.2020.02.001](https://doi.org/10.1016/j.optom.2020.02.001).
- [77] W. N. Charman, "Developments in the correction of presbyopia II: Surgical approaches," *Ophthalmic Physiol. Opt.*, vol. 34, no. 4, pp. 397–426, 2014, doi: [10.1111/opo.12129](https://doi.org/10.1111/opo.12129).
- [78] N. E. de Vries and R. M. M. A. Nuijts, "Multifocal intraocular lenses in cataract surgery: Literature review of benefits and side effects," *J. Cataract Refract. Surg.*, vol. 39, no. 2, pp. 268–278, Feb. 2013, doi: [10.1016/j.jcrs.2012.12.002](https://doi.org/10.1016/j.jcrs.2012.12.002).
- [79] B. F. Edwards, "Polyfocal spectacles," US Patent 2576581A, Nov. 27, 1951.
- [80] C.-M. Gariel, "La lentille à foyer variable du Dr Cusco," *J. Phys. Théorique Appliquée*, vol. 10, no. 1, pp. 76–79, 1881, doi: [10.1051/jphystap:018810010007601](https://doi.org/10.1051/jphystap:018810010007601).
- [81] M. R. Okun, "Variable focus eyeglasses," US Patent 4403840A, Sep. 13, 1983.
- [82] H. Krueger-Beuster, "Spectacles of variable focus," DE Patent 3335468A1, May 02, 1985.
- [83] R. B. Stoner, "Eyeglasses with controllable refracting power," US Patent 5182585A, Jan. 26, 1993.
- [84] J. E. Floyd, "Fluid filled and pressurized lens with flexible optical boundary having variable focal length," US Patent 5684637A, Nov. 04, 1997.
- [85] J. D. Silver, A. Roberson, and M. Newberry, "Variable focus lens and spectacles," US Patent 20100045930A1, Feb. 25, 2010.
- [86] J. Buch, *et al.*, "Variable focus lens and spectacles," US Patent 8441737B2, May 14, 2013.
- [87] L. W. Alvarez, "Two-element variable-power spherical lens," US Patent 3305294A, Feb. 21, 1967.
- [88] S. Kuiper and W. Hendricks, and H. Bernardus, "Variable focus spectacles," WO/2005/003842, Jan. 13, 2005.
- [89] S. Sato, "Liquid-crystal lens-cells with variable focal length," *Jpn. J. Appl. Phys.*, vol. 18, no. 9, pp. 1679–1684, 1979, doi: [10.1143/JJAP.18.1679](https://doi.org/10.1143/JJAP.18.1679).
- [90] A. F. Naumov, *et al.*, "Liquid-crystal adaptive lenses with modal control," *Opt. Lett.*, vol. 23, no. 13, Jul. 1998, Art. no. 994, doi: [10.1364/ol.23.000992](https://doi.org/10.1364/ol.23.000992).
- [91] Y.-H. Lin and H.-S. Chen, "Electrically tunable-focusing and polarizer-free liquid crystal lenses for ophthalmic applications," *Opt. Expr.*, vol. 21, no. 8, Apr. 2013, Art. no. 9436, doi: [10.1364/oe.21.009428](https://doi.org/10.1364/oe.21.009428).
- [92] H. S. Chen, M. S. Chen, and Y. H. Lin, "Electrically tunable ophthalmic lenses for myopia and presbyopia using liquid crystals," *Mol. Cryst. Liq. Cryst.*, vol. 596, no. 1, pp. 88–96, Jun. 2014, doi: [10.1080/15421406.2014.918321](https://doi.org/10.1080/15421406.2014.918321).
- [93] H. Zappe and C. Duppe, Eds., *Tunable Micro-Optics*. Cambridge, U.K.: Cambridge Univ. Press, 2015, doi: [10.1017/cbo9781139506052](https://doi.org/10.1017/cbo9781139506052).
- [94] H. Jiang and X. Zeng, *Microlenses*, 1st ed.. Boca Raton, FL, USA: CRC Press, 2013, doi: [10.1201/b14496](https://doi.org/10.1201/b14496).
- [95] N. Hasan, *et al.*, "Tunable-focus lens for adaptive eyeglasses," *Opt. Expr.*, vol. 25, no. 2, pp. 1221–1233, Jan. 2017, doi: [10.1364/oe.25.001221](https://doi.org/10.1364/oe.25.001221).
- [96] N. Hasan, H. Kim, and C. H. Mastrangelo, "Large aperture tunable-focus liquid lens using shape memory alloy spring," *Opt. Expr.*, vol. 24, no. 12, Jun. 2016, Art. no. 13342, doi: [10.1364/oe.24.013344](https://doi.org/10.1364/oe.24.013344).
- [97] H. Ren and S. Wu, *Introduction to Adaptive Lenses*. Hoboken, NJ, USA: Wiley, 2012, doi: [10.1002/9781118270080](https://doi.org/10.1002/9781118270080).
- [98] J. Jarosz, *et al.*, "Adaptive eyeglasses for presbyopia correction: An original variable-focus technology," *Opt. Expr.*, vol. 27, no. 8, Apr. 2019, Art. no. 10533, doi: [10.1364/oe.27.010533](https://doi.org/10.1364/oe.27.010533).
- [99] N. Hasan, *et al.*, "Lightweight smart autofocusing eyeglasses," in *Proc. 17th MOEMS Miniaturized Syst.*, Feb. 2018, vol. 10545, Art. no. 6, doi: [10.1117/12.2300737](https://doi.org/10.1117/12.2300737).
- [100] N. Padmanaban, R. Konrad, and G. Wetzstein, "Autofocals: Evaluating gaze-contingent eyeglasses for presbyopes," *Sci. Adv.*, vol. 5, no. 6, Art. no. eaav6187, Jun. 2019, doi: [10.1126/sciadv.aav6187](https://doi.org/10.1126/sciadv.aav6187).
- [101] J. Mompeán, J. L. Aragón, and P. Artal, "Portable device for presbyopia correction with optoelectronic lenses driven by pupil response," *Sci. Rep.*, vol. 10, no. 1, pp. 1–9, Dec. 2020, doi: [10.1038/s41598-020-77465-5](https://doi.org/10.1038/s41598-020-77465-5).
- [102] M. U. Karkhanis *et al.*, *Models of Accommodation Deficiency in Presbyopia Patients*. New York, NY, USA: Cold Spring Harbor Lab. Press, Aug. 2020, doi: [10.1101/2020.08.12.20173724](https://doi.org/10.1101/2020.08.12.20173724).
- [103] ClinicalTrials.gov [Internet]. "Bethesda (MD): National library of medicine (US), 2000 Feb 29 -. Identifier NCT03911596, smart autofocusing eyeglasses," Apr. 11, 2019. [Online]. Available: <https://clinicaltrials.gov/ct2/show/NCT03911596>
- [104] G. D. Boreman, *Modulation Transfer Function in Optical and Electro-Optical Systems*, vol. TT52. Bellingham, WA, USA: SPIE, 2001, Accessed: Jan. 10, 2021. [Online]. Available: <https://spie.org/Publications/Book/419857>
- [105] J. C. Wyant, "PSF and MTF," Accessed: Jan. 10, 2021. [Online]. Available: <http://wyant.optics.arizona.edu/psfMtf/psfMtf.htm>
- [106] D. G. Voelz, *Computational Fourier Optics: A MATLAB Tutorial*. Bellingham, WA, USA: SPIE, 2011, doi: [10.1117/3.858456](https://doi.org/10.1117/3.858456).
- [107] M. Millodot and S. Millodot, "Presbyopia correction and the accommodation in reserve," *Ophthalmic Physiol. Opt.*, vol. 9, no. 2, pp. 126–132, 1989, doi: [10.1111/j.1475-1313.1989.tb00831.x](https://doi.org/10.1111/j.1475-1313.1989.tb00831.x).
- [108] J. S. Wolffsohn, *et al.*, "Accommodative amplitude required for sustained near work," *Ophthalmic Physiol. Opt.*, vol. 31, no. 5, pp. 480–486, Sep. 2011, doi: [10.1111/j.1475-1313.2011.00847.x](https://doi.org/10.1111/j.1475-1313.2011.00847.x).
- [109] K. J. Ciuffreda, "Accommodation, the pupil and presbyopia," in *Borish's Clinical Refraction*, 2nd ed., W. J. Benjamin, Ed., Philadelphia, PA, USA: Butterworth-Heinemann, 2006, pp. 77–120.
- [110] R. Blendowske, "Unaided visual acuity and blur," *Optomatory Vis. Sci.*, vol. 92, no. 6, pp. 121–125, May 2015, doi: [10.1097/OPX.0000000000000592](https://doi.org/10.1097/OPX.0000000000000592).
- [111] "EL-16-40-TC lens — Optotune," 2019, Accessed: Jan. 17, 2021. [Online]. Available: <https://www.optotune.com/el-16-40-tc-lens>
- [112] J. Porter, *et al.*, "Monochromatic aberrations of the human eye in a large population," *J. Opt. Soc. America A*, vol. 18, no. 8, Aug. 2001, Art. no. 1793, doi: [10.1364/josaa.18.001793](https://doi.org/10.1364/josaa.18.001793).
- [113] L. N. Thibos, *et al.*, "Statistical variation of aberration structure and image quality in a normal population of healthy eyes," *J. Opt. Soc. America A*, vol. 19, no. 12, Dec. 2002, Art. no. 2329, doi: [10.1364/josaa.19.002329](https://doi.org/10.1364/josaa.19.002329).
- [114] T. O. Salmon and C. Van De Pol, "Normal-eye Zernike coefficients and root-mean-square wavefront errors," *J. Cataract Refract. Surg.*, vol. 32, no. 12, pp. 2064–2074, 2006, doi: [10.1016/j.jcrs.2006.07.022](https://doi.org/10.1016/j.jcrs.2006.07.022).
- [115] R. A. Applegate, *et al.*, "Three-dimensional relationship between high-order root-mean-square wavefront error, pupil diameter, and aging," *J. Opt. Soc. America A*, vol. 24, no. 3, pp. 578–587, Mar. 2007, doi: [10.1364/josaa.24.000578](https://doi.org/10.1364/josaa.24.000578).
- [116] A. S. Bruce and L. J. Catania, "Clinical applications of wavefront refraction," *Optomatory Vis. Sci.*, vol. 91, no. 10, pp. 1278–1286, Oct. 2014, doi: [10.1097/OPX.0000000000000377](https://doi.org/10.1097/OPX.0000000000000377).
- [117] C. A. Johnson, "Effects of luminance and stimulus distance on accommodation and visual resolution," *J. Opt. Soc. America*, vol. 66, no. 2, pp. 138–142, Feb. 1976, doi: [10.1364/JOSA.66.000138](https://doi.org/10.1364/JOSA.66.000138).
- [118] C. A. Johnson and E. J. Casson, "Effects of luminance, contrast, and blur on visual acuity," *Optomatory Vis. Sci.*, vol. 72, no. 12, pp. 864–869, 1995, doi: [10.1097/00006324-199512000-00004](https://doi.org/10.1097/00006324-199512000-00004).
- [119] L. P. Tidbury, G. Czanner, and D. Newsham, "Fiat Lux: the effect of illuminance on acuity testing," *Graefes's Archive Clin. Exp. Ophthalmol.*, vol. 254, no. 6, pp. 1091–1097, Jun. 2016, doi: [10.1007/s00417-016-3329-7](https://doi.org/10.1007/s00417-016-3329-7).
- [120] A. B. Watson, "A formula for the mean human optical modulation transfer function as a function of pupil size," *J. Vis.*, vol. 13, no. 6, May 2013, Art. no. 18, doi: [10.1167/13.6.18](https://doi.org/10.1167/13.6.18).
- [121] P. Zhao, Ç. Ataman, and H. Zappe, "Gravity-immune liquid-filled tunable lens with reduced spherical aberration," *Appl. Opt.*, vol. 55, no. 28, Oct. 2016, Art. no. 7816, doi: [10.1364/ao.55.007816](https://doi.org/10.1364/ao.55.007816).