

Received 30 May 2022, accepted 21 June 2022, date of publication 29 June 2022, date of current version 5 July 2022.

Digital Object Identifier 10.1109/ACCESS.2022.3187036

RESEARCH ARTICLE

Compact Models of Presbyopia Accommodative Errors for Wearable Adaptive-Optics Vision Correction Devices

MOHIT U. KARKHANIS¹, AISHWARYADEV BANERJEE¹, CHAYANJIT GHOSH¹,
RUGVED LIKHITE², (Member, IEEE), ERFAN POURSHABAN¹, HANSEUP KIM¹,
DAVID A. MEYER³, AND CARLOS H. MASTRANGELO¹, (Senior Member, IEEE)

¹Department of Electrical and Computer Engineering, The University of Utah, Salt Lake City, UT 84112, USA

²Intel Corporation, Hillsboro, OR 97124, USA

³John A. Moran Eye Center, The University of Utah, Salt Lake City, UT 84112, USA

Corresponding author: Mohit U. Karkhanis (mohit.karkhanis@utah.edu)

This work involved human subjects or animals in its research. Approval of all ethical and experimental procedures and protocols was granted by the University of Utah Institutional Review Board under Application No. IRB 00114415.

ABSTRACT Subjective accommodation refractive error models are essential for implementing adaptive vision correction devices that utilize varifocal optics. This article describes compact empirical models of subjective accommodative refractive errors in subjects with advanced presbyopia. The models are based on measurements of subjective refractive errors from fifteen presbyopes over the age of 45 using commercially available focus-tunable eyeglasses under three different illumination conditions over a 3.08D accommodation stimulus range. The resulting average residual root-mean-squared (RMS) error values for the best fitting 8-parameter model was 0.25D compared to an average RMS error of 0.4D for the conventional DDF and HHG models. The RMS error for the best-fitting model is below the average refractive error of the human eye.

INDEX TERMS Presbyopia, adaptive optics, vision correction, smart eyewear, autofocus eyeglasses.

I. INTRODUCTION

Presbyopia is an age-related refractive disorder caused by the loss of focal tunability of the eye crystalline lens. The tunability loss manifests as an inability to focus on nearby objects and an overall object distance-dependent defocusing error. According to recent surveys [1], [2], more than a quarter of the world's population is affected by presbyopia and this number is slated to increase with an improvement in the life expectancy of the global population.

Current presbyopia mitigation devices such as multifocal and progressive lenses do not restore normal vision. Instead, these devices fragment the field of view into zones of different but fixed object distance. Therefore, these mitigation devices reduce the in-focus field-of-view resulting in major

reductions in the visual acuity when averaged over the entire view field.

Fundamentally the restoration of normal vision in presbyopes requires not fixed but variable/ tunable focus lenses that can adjust their optical power and cancel out the presbyope refractive error at all object distances. Recent advances in adaptive optical systems [3]–[6] suggest that adaptive smart eyewear systems can be implemented in lightweight form and contact lens configurations [7], [8]. Autofocusing eyewear like smart-eyeglasses for example can provide automatic accommodation correction and potentially restore near-normal vision in presbyopes [5], [9]–[13]. Such devices hold great promise as they can potentially restore normal vision and improve the lives of billions of people affected by presbyopia.

The restoration of normal vision with these devices essentially requires detailed knowledge of a particular presbyope accommodative response (AR) curve. The accommodative

The associate editor coordinating the review of this manuscript and approving it for publication was Yogendra Kumar Prajapati¹.

response curve is defined as the change in the optical power of the crystalline lens a function of the object distance. The object distance is specified by the accommodation stimulus (AS) parameter, which is the reciprocal of the object distance [14]. The difference between the AS and the AR is often termed as accommodative lead (positive difference, over-accommodation) and accommodative lag (negative difference, under-accommodation) or, more generally, accommodative error ($AE = AS - AR$). This curve must be measured for each individual and stored in one's autofocus-ing device. For all practical purposes, the empirical model must be compact and developed using minimum number of measurements and fitting parameters, thereby ensuring low test costs.

A. OBJECTIVE AND SUBJECTIVE AR

The accommodative response can be measured in several ways. Objective AR is measured optically using, for example, an autorefractor instrument. However, the objective AR [15]–[28] does not necessarily represent the presbyope's defocus perception as it does not take into account the effects of the eye depth-of-focus under different lighting and pupil aperture conditions, the static refractive errors like myopia, hyperopia and astigmatism or any other higher order aberrations. In this scenario, measurement of the *subjective* accommodation response (AR), i.e., the refractive correction required by an individual to *perceive* the clearest possible focus at an object plane, provides better information about real-world performance of the presbyopic eye. The *subjective* accommodation error (AE) thus represents the optical correction which the autofocusing device would need to provide the clearest vision in presbyopes.

In practical terms, to realize such correction, a compact software mathematical model of the subjective AE would need to be programmed into the autofocusing device to produce the appropriate optical power as a function of the object distance and illumination I such that $AE = f(AS, I)$ as shown in Fig. 1. At this juncture it is important to clarify that this function is a purely empirical model extracted from patient data, and it does not directly represent the underlying phenomena responsible for presbyopia.

While this correction concept can, in principle, be performed for any presbyope, the procedure requires detailed measurement of the individual's subjective AE as a function of the object distance and lighting conditions.

B. CONVENTIONAL MODELS OF OBJECTIVE AR

Two theoretical models of *objective* AR viz., Donders-Duane-Fincham (DDF) model and the Helmholtz-Hess-Gullstrand (HHG) model have been discussed extensively in literature so far, which originate from the two major variants of the lenticular theory [29], [30]. In the HHG theory, the loss in accommodation is completely attributed only to the morphological changes in the human eye's crystalline lens capsule. This model shows a near-ideal response to stimulus over the manifest zone and then a sharp transition into a hard saturation




FIGURE 1. Schematic of an accommodation correction system. The accommodation error model is programmed into the correction device. The variable-focus lenses vary their optical power in real-time based on the programmed accommodation error (AE) model for the individual wearing the correction device. Variables d and I represent the object distance and illumination, respectively.

region, where the lens can no longer accommodate. The AR curve corresponding to HHG theory is shown in Fig. 2(a). In contrast, DDF theory attributes loss in accommodation with age to changes in only the bio-mechanical properties of ciliary muscle which controls the crystalline lens. The DDF model exhibits a linear response between the objective AR and the applied stimulus, with the slope of the line reducing progressively with age, until it is nearly flat and stagnated for advanced presbyopes. Fig. 2(b) shows the corresponding accommodation characteristics of DDF model.

To our knowledge, no studies have been conducted to date to model the subjective AE of presbyopes at various object distances and under various illumination levels.

The objective of this paper is the development of individual-specific compact empirical models of subjective AE as functions of object distance and illumination, for use in autofocusing eyewear to provide adequate vision correction for presbyopes. The models are constructed from subjective AE data collected from a group of advanced presbyopes in a registered clinical study. For each of these presbyopes the model is fitted to the presbyope data using as few parameters as possible. We also compare the RMS fitting errors of these models in relation to accommodation restoration using smart autofocusing eyewear technologies.

C. IMPACT OF DEFOCUS AE ON VISUAL ACUITY

Visual Acuity (VA) is a unit less number that quantifies the clarity of vision and rates the ability of the human visual system to recognize/ resolve small details with precision. It is measured in terms of the logarithm of the minimum angle of resolution (MAR) or logMARs [31]–[35]. Standard vision is defined as logMAR of 0.0 (Snellen 20/20), and it helps to define normal vision. Positive logMAR values indicate vision loss, while negative values denote normal or better visual acuity. It is well known that the visual acuity deteriorates in presence of defocus errors [36]. A simple empirical model,




FIGURE 2. Comparison of accommodation response characteristics from the (a) Helmholtz-Hess-Gullstrand and the (b) Duane-Donders-Fincham models [29]. The black line is the ideal accommodation response. Neither model fits actual individual AD characteristics adequately for performing a good AD correction with adaptive smart corrective eyewear.

suggested by Blendowske [37], relates the defocus error and deterioration in the visual acuity as:

$$\frac{V}{V_{bc}} = \frac{1}{1 + D^2} \quad (1)$$

where V is the deteriorated decimal visual acuity. V_{bc} is the best corrected decimal visual acuity and D is the defocus error in diopters. For example, a defocus error of 1.0D reduces the visual acuity by 50%, thereby degrading the visual acuity by 0.3 logMAR. Any model used for restoration of the subject-specific AE curve must ideally be sufficiently accurate to minimize the reduction of visual acuity at all object distances and under all illumination conditions.

In practical terms, patient-specific AE models with low RMS errors will have to be programmed into autofocusing eyewear devices to avoid over- or under-compensation and avoid related deterioration in the vision of the presbyope. We have shown the implementation of one of the models in a recently published article which demonstrates the design and implementation of a pair of smart autofocusing, liquid-lens eyeglasses to mitigate presbyopia [13].

II. COMPACT MODEL DEVELOPMENT

Every individual's eye behavior is characterized by their AR(AS). Ideally, the AR curve should be a line of unity slope, but as presbyopia progresses, the curve flattens to lower optical powers [14], [29], [38]–[42] as shown in Fig. 3.

In this article, we seek to develop compact empirical equations that fit the subjective AR curve for a particular eye of an individual with the smallest number of parameters. Several types of models were fitted to patient data based on the underlying physics and pure empirical observations of the patient data. These models are discussed in the sections below.

A. SIGMOID MODEL

The typical subjective AR curve displays a saturation type of response with flat regions at low and high accommodation stimuli and an approximately uniform slope in between (Fig. 3). Empirical data [14], [29], [39]–[41], [43]–[48] of AR in the literature exhibits a typical “S”-shaped or sigmoid curve that can be modelled using the sigmoid function




FIGURE 3. Example of typical subjective accommodation response curve for the human eye. The curve progressively flattens for higher powers with age [29], [38].

family [49]. The simplest continuous static sigmoid function [50] is the shifted logistics equation

$$AR_{sig} = \frac{a}{1 + (e^{-k \cdot [AS - b]})} - d \quad (2)$$

where a , b , d and k are fitting parameters. a is the maximum subjective accommodative amplitude, k represents the range of AS for which the AR exhibits a linear response and b represents the shift along the horizontal axis. The parameter d represents any vertical shift in the subjective AR due to the presence of other refractive errors.

B. PIECEWISE LINEAR MODEL

The subjective AR in Fig. 3 can also be modeled with simple piecewise linear models with one or two breakpoints, using the Heaviside step function $H(x)$. We can define a ramp function R

$$R(x, S, \alpha) \equiv (x - \alpha) \cdot S \cdot H(x - \alpha) \quad (3)$$

where S is the slope of the ramp function. Using these definitions, we can approximate the curve with one breakpoint piecewise linear model

$$AR_{PWL1} = R(AS, k_1, 0) + R(AS, k_2, x_0) + f \quad (4)$$

where, k_1 , k_2 , are slopes of the two segments and are fitting parameters and x_0 is the location of the breakpoint. Similarly, a two- breakpoint piecewise linear model can be constructed as:

$$AR_{PWL2} = R(AS, k_1, 0) + R(AS, k_2, x_0) + R(AS, k_3, x_1) + f \quad (5)$$

where, k_1 , k_2 , k_3 are the slopes the three segments, respectively and x_0 , x_1 are the locations of the two breakpoints and f is an initial offset. We choose to limit the number of breakpoints used in the piecewise linear models to a maximum of two since any more breakpoints would increase the number of fitting parameters beyond the number of observations.

C. ILLUMINATION DEPENDENCE

The eye pupil changes diameter with different illumination levels to control the light intensity at the retina. At high

illumination the pupil constricts to a smaller diameter. This increases the depth-of-focus of the visual system and improves vision. Thus, the subjective *AR* will correspondingly vary under different illumination levels. Recent studies have shown that the subjective *AR* of presbyopes bears a linear relation with the logarithm of luminance levels [51] and the slope of the *AR* curve decreases with reduction in object luminance [52]. Since luminance is proportional to illumination (in lux) we modify the sigmoid model of Eq. (2) as

$$AR_{sig}(AS, I) = \left(\frac{a}{1 + (e^{-k \cdot [AS-b]})} - d \right) \cdot \log(I) \quad (6)$$

where I is the illumination in lux. Similarly, the illumination dependence is included in the piecewise models as a multiplicative factor.

$$\begin{aligned} AR_{PWL1,I} &= AR_{PWL1} \cdot \log(I), \\ AR_{PWL2,I} &= AR_{PWL2} \cdot \log(I) \end{aligned} \quad (7)$$

Since the range of the illumination parameter in our recorded observations was narrow and approximately within the same orders of magnitude (75–800 lux), we also utilized a simple exponential function to model the illuminance dependence on the subjective *AR*. Accordingly, using Eq. (2), (4) and (5) the sigmoid and piecewise linear models can be expressed as:

$$AR_{sig}(AS, I, \beta) = \left(\frac{a}{1 + (e^{-k \cdot [AS-b]})} - d \right) \cdot I^\beta \quad (8)$$

and the piecewise approximations

$$AR_{PWL1,I} = AR_{PWL1} \cdot I^\beta, AR_{PWL2,I} = AR_{PWL2} \cdot I^\beta \quad (9)$$

where β is an empirical exponent.

In addition, we also utilized a piecewise linear model with one breakpoint with slopes and breakpoints that are linear functions of the illumination

$$f_I = f_{00} + f_{01} \cdot I, \quad (10)$$

$$x_I = x_{00} + x_{01} \cdot I, \quad (11)$$

$$k_{I1} = k_{10} + k_{11} \cdot I, \quad (12)$$

and,

$$k_{I2} = k_{20} + k_{21} \cdot I \quad (13)$$

where $f_{00}, f_{01}, x_{00}, x_{01}, k_{10}, k_{11}, k_{20}$ and k_{21} are fitting parameters. The linear dependence of parameters appears to fit the characteristics of the patient data and it is purely empirical.

III. METHODS

A. STUDY DESIGN

Many researchers have explicitly studied objective *AR* or *AE* in order to analyze the effects of age on the morphological changes observed within the eye and how the *AR* changes with age [15], [16], [24]–[28], [40], [43]–[48], [53], [54]. Such objective *AR* and *AE* are measured using sophisticated, computer controlled optometers, retinoscopes and aberrometers. In these studies, the pupil diameter is tightly controlled



FIGURE 4. A presbyope undertaking a visual task at 1m chart distance.

with the help of phenylephrine [55] in order to decouple the effects of the depth of field on the *AR/AE* measurements. Therefore, such measurements cannot be directly or easily converted to subjective *AR/AE* [15], [16], [19]–[23], [27], [56], [57] for practical use in autofocusing eyewear.

The focus of this study is to develop *AE* models which can be directly used in smart autofocusing eyewear algorithms for personalized corrections. In order to simulate the real-world conditions in which such algorithms and systems will be used, we measured the subjective accommodative errors which the presbyopic subjects exhibited during their visual task at different distances as described in the subsections below. The methodology used in this study resembles the “push-up” method of analyzing accommodation amplitude, where the target or an eye chart is moved progressively closer to the subjects’ eyes till a blurry image is reported [24], [26]. In our method, the distance is fixed at a few locations away from the observer and the image at each location is made subjectively as clear as possible utilizing a set of commercially available tunable power eyeglasses which cancel the distance-dependent subjective refractive error. Additionally, we utilize calibrated ETDRS charts for various chart distances in which the angular detail of the optotypes is maintained [31]–[35]. The details of the experimental method used is described in the sections below.

B. HUMAN STUDY PARTICIPANTS AND EXCLUSION CRITERIA

Human study approval was acquired from the University of Utah Institutional Review Board (IRB 00114415), and experiments were performed according to the ethical standards laid down in the Declaration of Helsinki, 1964. A total of 15 advanced presbyopia subjects ages 45–68 years, with a mean age of 54.6 (S.D. = 6.8) years, were recruited from a population of patients from the University of Utah Moran Eye Center and associated clinics. Since emmetropes and early presbyopes generally exhibit a non-zero objective accommodation amplitude, they were not considered for this study. All subjects provided informed, signed consent before

entry into the study. A record on clinical trials performed in this study has been registered with ClinicalTrials.gov [58] (NCT03911596). Individuals with astigmatism $> 1.0\text{D}$, artificial intraocular lenses, or those having any ocular pathology that would inhibit accommodation of their natural lenses were excluded. The recruited subjects had prescriptions between -1.5 D and $+2.5\text{ D}$ and were correctable to logMAR 0.0 (Snellen 20/20).

C. STUDY CONDITIONS

Testing was carried out in a light proofed optometry exam room at the University of Utah Moran Eye Center. Chart retro-illumination was not used. In order to explore the effects of illumination on the accommodative insufficiency, we conducted the study under 3 chart illumination levels- 75 lux to simulate dark conditions, 500 lux to simulate normal indoor lighting conditions and 800 lux to simulate outdoor conditions [59]. Chart illumination was controlled using an LED studio lighting system which also provided diffused lighting for the exam room. The correlated color temperature of the lighting system was fixed at 5000 K. Chart illumination was kept constant while measuring the accommodation insufficiency. Fig. 4 shows a subject undertaking a visual task.

D. STUDY PROCEDURE

The visual task consisted of reading optotypes on 7 different Early Treatment Diabetic Retinopathy Study (ETDRS) charts calibrated for 7 distances (4 m, 2 m, 1 m, 70 cm, 50 cm, 40 cm and 30 cm) under 3 chart illumination conditions (75 lux, 500 lux and 800 lux). We utilized a commercially available manually-tunable variable-focus Adlens Hemisphere eyeglasses to assess the subjective *AE*. The standard deviation of the eyeglasses optical power was 0.03 D. Subjects were assisted by the study staff in manually tuning the eyeglasses until they could correctly identify, to the best of their ability, the optotypes corresponding to logMAR 0.0 (Snellen 20/20) line on the ETDRS charts. Each lens in the eyeglasses was monocularly tuned before patients undertook the visual task, binocularly, under every test distance and illumination condition. Once optimum subjective refraction was reached, the visual acuity of the subjects was measured in logMAR. This was done to verify if the variable-focus eyeglasses provided the optimum correction necessary for subjects during all the visual tasks. The eyeglasses were then taken from the subjects and their optical power was measured using a Thorlabs WFS150C-AR Shack-Hartmann wavefront sensor. Subjects who were unable to completely identify the optotypes corresponding to logMAR 0.0 line, were instead asked to identify the optotypes on the previous line (logMAR 0.1) and their visual acuity was recorded accordingly. Those who could correctly identify the optotypes corresponding to the logMAR 0.0 line were then asked to identify optotypes on the logMAR -0.1 line and their visual acuity was recorded accordingly. Calibrated ETDRS charts, which were placed at distances of 1 m, 70 cm, 50 cm and 30 cm, were carefully designed and printed on high quality optical white

paper. Subjects were tested under chart illumination levels of 500 lx, 800 lx and 75 lx, sequentially. The study was paused after changes in the illumination levels till the subjects were comfortable with their visual experience. The study sessions lasted between 1.5–3 hours in duration for each subject.

E. MODEL FITTING

The subjective *AE* models were surface-fitted to the recorded subjective *AE* data for each eye of every subject with the Python 3.8 SciPy 1 optimization library [60]. We utilized the Levenberg-Marquardt nonlinear least squares method for surface-fitting Eq. (6)–(9) to the *AE* data [61]–[63] using the k 's, x 's and f as fitting parameters. The independent variables-*AS* and the illumination were not bounded during the surface-fitting process as the simple Levenberg-Marquardt method does not handle bounds on the independent variables [63], [64]. The surface-fitting process produced a separate characteristic equation and a surface for every eye of every subject. The surface-fits were analyzed and compared with each other using the R^2 metric (goodness of fit/ coefficient of determination) and the residual RMS errors exhibited by the models. It is important to note that these models are going to be used for personalized corrections using autofocusing eyewear and hence the fitting parameters extracted from the curve fitting process would vary from eye to eye.

IV. RESULTS AND DISCUSSION

A. ACCOMMODATION ERRORS

The observed subjective *AE*s of the subjects were plotted against the *AS*. The recorded data has been made available in Data File 1. Figs. 5 – 7 show the observed *AE* for 15 subjects (30 eyes) under chart illumination levels 75, 500 and 800 lux, respectively. The solid lines with circular markers represent the right eye and the dashed lines with square markers represent the left eye. The size of the markers is proportional to the standard deviation of the measurement instrument. The markers are 5x the standard deviation for better visibility.

It can be clearly seen from the plots that the subjective *AE* kept increasing with increasing stimulus, under all illumination conditions. All subjects required higher positive additions to perform near vision tasks at *AS* of 2.5 D and 3.33 D, corresponding to chart distances of 40 cm and 30 cm, respectively, when compared to far vision tasks at *AS* of 0.25 D and 0.5 D corresponding to chart distances of 4m and 2m, respectively. This trend is consistent with the existing theories of presbyopia [29] and subjective/ objective *AR* data reported throughout literature [14], [29], [39]–[42], [44]–[48], [54] where the *AR* reduces or stagnates irrespective of the increase in the stimulus and causes *AE* to increase with increasing stimulus. It can also be seen from Figs. 5 – 7 that the recorded *AE* in the left and the right eyes of all subjects are very similar under all illumination conditions, except for subject #5 whose right *AE* is significantly higher compared to their left *AE*. This could be attributed to the presence of a high-power static refractive error in one of their eyes.




FIGURE 5. Measured accommodation error for 15 subjects (30 eyes) under 75 lux chart illumination. The markers show the recorded data from the subjects. The solid lines with circular markers and the dashed lines with square markers represent subjective AE in the right and left eye, respectively. The size of the markers is 5 times the standard deviation of the manually tunable eyeglasses, for better visibility.




FIGURE 6. Measured accommodation errors for 15 subjects (30 eyes) under 500 lux chart illumination. The markers show the recorded data from the subjects. The solid lines with circular markers and the dashed lines with square markers represent subjective AE in the right and left eye, respectively. The size of the markers is 5 times the standard deviation of the manually tunable eyeglasses, for better visibility.

B. AE AND ILLUMINATION

We also explored the dependence of subjective AE on the illumination levels. Fig. 8. shows the average of the absolute subjective AE for 30 eyes under 3 different illumination conditions of 75, 500 and 800 lux. The blue bars represent the left eyes while the purple bars represent the right eyes. A trend can be seen from the plots where the average subjective AE progressively decreases with illumination for most subjects. However, this trend cannot be quantified in this study for two main reasons-(a) the AE was studied under only 3 illumination levels and (b) only 15 presbyopes (30 eyes) were studied, which are insufficient to conduct any meaningful statistical analysis.

The depth of field of the human eye is dependent on the pupil diameter [65] and the pupil diameter is a function of the illumination levels. With increasing illumination levels, the pupil size decreases [66], thereby increasing the depth-of-focus of the visual system and reducing the required

amplitude of accommodation to focus on an object. One of the principal reasons why presbyopes tend to squint is because doing so improves their visual acuity at the expense of much lower brightness. At lower illumination levels, the pupil size increases, reducing the depth of field and requiring more accommodation amplitude [53]. It has also been shown that accommodation depends on cone activity [67], which ceases completely under low illumination conditions. These factors can help explain why a majority of the subjects exhibited decreasing average AE with increasing illumination. Fig. 8 also shows that for few participant's eyes, the AE at 800 lux was greater than that at 500 lux. This seems counter-intuitive as accommodation error decreases under higher object illumination levels due to improved depth-of-focus [68]. However, there are some clues to explain this discrepancy. As light enters the eye, part of its energy is scattered in the crystalline lens [69] and gives rise to a phenomenon called disability glare [70], [71]. Disability




FIGURE 7. Measured accommodation errors for 15 patients (30 eyes) under 800 lux chart illumination. The markers show the recorded data from the subjects. The solid lines with circular markers and the dashed lines with square markers represent AE in the right and left eye, respectively. The size of the markers is 5 times the standard deviation of the manually tunable eyeglasses, for better visibility.

glare is a physiological glare which generally impairs vision, decreases visual acuity without causing any discomfort [72] and decreases contrast sensitivity [73]. Due to anatomical changes in the visual system with age [74]–[77], senior people were found to be more sensitive to glare [78]–[81]. This could possibly explain why some eyes in this study exhibit higher accommodation error amplitudes at 800 lux chart illumination compared to those at 500 lux chart illumination. Detailed analyses of subjective AE with respect to the illumination levels is beyond the scope of this work.

C. MODEL FITTING RESULTS AND DISCUSSION

At any given fixed power, the crystalline lens in the human eye has an average RMS wavefront error of $0.35 \mu\text{m}$ [82]. This is roughly equivalent to a natural refractive defocus of 0.3D [83]–[86] at a pupil size of 6 mm diameter, for healthy, young individuals and steadily increases with age [86], [87] which cannot be corrected by conventional eyeglasses.




FIGURE 8. Average accommodation errors for 15 subjects (30 eyes). A majority of subjects show decreasing AE with increasing illumination levels. Subplots (a)–(o) shows the average accommodation errors for Subjects 1–15, respectively.




FIGURE 9. RMS fitting errors for 15 subjects corresponding to the older (a) DDF and (b) HHG models in literature and the (c) proposed best fitting one-break PWL model with illumination dependent parameters (PWL-IDP). The blue bars represent the left eye and the purple bars represent the right eye. RMS fitting errors for the one-break PWL-IDP models are significantly smaller compared to the older model.

Therefore, in order to restore pre-presbyopic accommodation we seek subjective AE models with average errors $< 0.3\text{D}$ and any model that meets this criterion is sufficient for describing the AE. To find such model we have fitted each of our proposed AE models for each of the 30 eyes considered in this study, and from the RMS errors for each fit we report the averaged RMS error for the eye set. The model fitting plots and the fitting parameters for 15 subjects (30 eyes) and all proposed models are available in the supplemental document.




FIGURE 10. Average corrected visual acuities (VA) of 15 subjects recorded under various chart illumination conditions. The bars represent one standard deviation. Average corrected VA of all 15 subjects is less than +0.1 logMAR. The blue and green dashed lines correspond to spherical refractive errors of 0.5 D and 0.25 D, respectively. Subplots (a)–(o) shows the average corrected VA for Subjects 1–15, respectively.

TABLE 1. Average RMS fitting errors for 15 subjects (30 eyes).

Fitting Model	# of parameters	Avg. RMS error (D)
DDF	2	0.409
HHG	3	0.408
Sigmoid- Log (S-L)	4	0.331
One-break PWL – Log (1-PWL-L)	4	0.376
Two-break PWL – Log (2-PWL-L)	6	0.364
Sigmoid- Exponential (S-E)	5	0.291
One-break PWL- Exponential (1-PWL-E)	5	0.294
Two-break PWL- Exponential (2-PWL-E)	7	0.283
One-break PWL- Illumination dependent parameters (PWL-IDP)	8	0.25
One-break PWL with individual illumination fits	12	0.206

Table 1 shows the averaged RMS error over the 30 eyes for each of our models. The last line shown in the table corresponds to individual fits for each illumination level.

Out of the 7 proposed models investigated, 3 models featuring an exponential dependence on illumination parameter (S-E, 1-PWL-E and 2-PWL-E) and the illumination-dependent parameters (PWL-IDP) model had RMS fitting errors smaller than the 0.3D threshold. The fits with the most fitting parameters have the lowest average RMS errors. The best fitting PWL-IDP model displayed nearly half the average RMS fitting error (0.25D) than the DDF (0.4D) and the HHG models (0.4D). Neither the DDF and HHG models feature illumination dependence.

Fig. 9 shows a comparison of the residual RMS fitting errors obtained for 15 subjects (30 eyes) after fitting the older DDF, HHG models and the one-break PWL with illumination dependent parameters (PWL-IDP). The blue bars represent

TABLE 2. Average corrected visual acuities of 15 subjects under 3 chart illumination levels.

Subject	Corrected VA (75 lux)	Corrected VA (500 lux)	Corrected VA (800 lux)
1	-0.04	-0.05	-0.01
2	0.01	-0.07	0.01
3	0.01	-0.04	-0.07
4	0.01	0.0	-0.02
5	0.01	0.0	-0.02
6	0.02	0.0	-0.04
7	0.02	0.0	-0.01
8	0.01	0.0	0.0
9	0.04	0.0	0.0
10	0.02	0.0	-0.04
11	0.04	0.0	-0.02
12	0.0	0.0	0.0
13	0.0	0.02	0.02
14	0.01	0.0	-0.01
15	0.01	-0.01	0.01

the left eyes while the purple bars represent the right eyes. The red dashed line represents a 0.5D reference. Our best illumination-dependent fitting model (PWL-IDP) exhibits fitting errors less 0.3D, for 23 out of the 30 eyes (77%). The PWL-IDP model outperforms the older DDF and HHG models and, on the average, meet the acceptability threshold <0.3D for our experimental data set.

D. VISUAL ACUITY MEASUREMENTS

The results of our study show that the corrected visual acuities of all subjects were mostly equal to or better than +0.1 logMAR (Snellen 20/25) under all stimuli and illumination conditions when we use the best fitting model, i.e., the piecewise linear model with one breakpoint and illumination dependent parameters (PWL-IDP). Visual acuities recorded at different distances and chart illumination conditions for each participant have been included (Data File 2). Fig. 10 shows the average corrected visual acuities of the subjects under varying accommodation stimuli and chart illumination levels. None of the 15 subjects exhibited average corrected visual acuity worse than +0.04 logMAR.

In order to better relate an individual's visual acuity to lens prescription, visual acuity scores corresponding to defocus errors of 0.25D and 0.5D were calculated using the simple relation of Eq. (1). More accurate VA formulas corresponding to spherical refractive errors are available in literature [88] if the pupil size is known. The blue and green lines in Fig. 10 show these calculated visual acuity scores of 0.026 logMAR and 0.096 logMAR corresponding to spherical refractive errors of 0.25D and 0.5D, respectively. Corrected visual acuities falling within the blue region correspond to an acceptable spherical refractive error between 0.25D and 0.5D, while those falling within the green region correspond to a minimum desirable spherical refractive error of < 0.25 D in the autofocusing eyewear. In optometry, visual impairment is defined as visual acuity of worse than +0.3 logMAR (Snellen 20/40) [89]. Since the corrected visual acuities of all subjects while wearing tunable eyeglasses were better than or equal to +0.1 logMAR and the subjects responded favorably after each optical adjustment, it can be concluded

the subjects in this study were adequately corrected. It can also be seen from Fig. 10 that a majority of the subjects were properly corrected for their AE at 500 lux chart illumination ($VA \approx 0.0$ logMAR).

Table 2 shows the recorded visual acuity of the subjects, averaged over all 7 distances, under the chart illumination levels of 75, 500 and 800 lux. Overall, the corrections provided by the focus-tunable Adlens eyeglasses were able to adequately compensate the subjective accommodation errors present in the subjects' eyes, resulting in an average corrected VA of -0.01 logMAR for all eyes at 500 lux standard illumination condition [59].

V. CONCLUSION

We present new empirical models for presbyopes, which relate their subjective accommodative errors to the object distance and the illumination levels. Measured accommodation errors of all subjects under all illumination conditions increased with a reduction in the test-chart distance, indicating a loss in the accommodative abilities of their visual system. The proposed subjective accommodation error models were surface-fitted to the measured data of 15 subjects and the best fitting subjective accommodation error model exhibited an RMS fitting error of less than 0.3D, the average equivalent refractive error of a human eye, for a majority (77%) of the eyes under examination. These empirical AE models are sufficiently accurate for use in future personalized adaptive optics based autofocusing devices.

REFERENCES

- [1] T. R. Fricke, N. Tahhan, S. Resnikoff, E. Papas, A. Burnett, S. M. Ho, T. Naduvilath, and K. S. Naidoo, "Global prevalence of presbyopia and vision impairment from uncorrected presbyopia: Systematic review, meta-analysis, and modelling," *Ophthalmology*, vol. 125, pp. 1492–1499, Oct. 2018.
- [2] R. R. A. Bourne, J. D. Steinmetz, S. Flaxman, P. S. Briant, H. R. Taylor, S. Resnikoff, R. J. Casson, A. Abdoli, E. Abu-Gharbieh, A. Afshin, and H. Ahmadi, "Trends in prevalence of blindness and distance and near vision impairment over 30 years: An analysis for the global burden of disease study," *Lancet Glob. Heal.*, vol. 9, no. 2, pp. e130–e143, Feb. 2021.
- [3] N. Hasan, A. Banerjee, H. Kim, and C. H. Mastrangelo, "Tunable-focus lens for adaptive eyeglasses," *Opt. Exp.*, vol. 25, no. 2, pp. 1221–1233, Jan. 2017.
- [4] N. Hasan, H. Kim, and C. H. Mastrangelo, "Large aperture tunable-focus liquid lens using shape memory alloy spring," *Opt. Exp.*, vol. 24, no. 12, p. 13342, Jun. 2016.
- [5] J. Jarosz, N. Molliex, G. Chenon, and B. Berge, "Adaptive eyeglasses for presbyopia correction: An original variable-focus technology," *Opt. Exp.*, vol. 27, no. 8, p. 10533, Apr. 2019.
- [6] A. Jamali, D. Bryant, A. K. Bhowmick, and P. J. Bos, "Large area liquid crystal lenses for correction of presbyopia," *Opt. Exp.*, vol. 28, no. 23, pp. 33982–33993, Nov. 2020.
- [7] C. Ghosh, A. Mastrangelo, M. Karkhanis, A. Deshpande, A. Banerjee, H. Kim, and C. H. Mastrangelo, "Low-profile induced-voltage distance ranger for smart contact lenses," *IEEE Trans. Biomed. Eng.*, vol. 68, no. 7, pp. 2203–2210, Jul. 2021.
- [8] A. Banerjee, C. Ghosh, M. Karkhanis, E. Pourshaban, H. Kim, and C. H. Mastrangelo, "Low-power, thin and flexible, stacked digital LC lens for adaptive contact lens system with enhanced tunability," in *Proc. Conf. Lasers Electro-Optics*, 2021, p. AF2Q.8.
- [9] N. Hasan, M. Karkhanis, F. Khan, T. Ghosh, H. Kim, and C. H. Mastrangelo, "Adaptive optics for autofocusing eyeglasses," in *Proc. Imag. Appl. Opt. (3D, AIO, COSI, IS, MATH, pCAOP)*, 2017, p. AM3A.1.
- [10] C. Mastrangelo, F. Khan, N. Hasan, C. Ghosh, T. Ghosh, H. Kim, and M. Karkhanis, "Lightweight smart autofocusing eyeglasses," in *Proc. 17th MOEMS Miniaturized Syst.*, Feb. 2018, p. 6.
- [11] N. Padmanaban, R. Konrad, and G. Wetzstein, "Autofocals: Evaluating gaze-contingent eyeglasses for presbyopes," *Sci. Adv.*, vol. 5, no. 6, Jun. 2019, Art. no. eaav6187.
- [12] J. Mompeán, J. L. Aragón, and P. Artal, "Portable device for presbyopia correction with optoelectronic lenses driven by pupil response," *Sci. Rep.*, vol. 10, no. 1, pp. 1–9, Dec. 2020.
- [13] M. U. Karkhanis, C. Ghosh, A. Banerjee, N. Hasan, R. Likhite, T. Ghosh, H. Kim, and C. H. Mastrangelo, "Correcting presbyopia with autofocusing liquid-lens eyeglasses," *IEEE Trans. Biomed. Eng.*, vol. 69, no. 1, pp. 390–400, Jan. 2022.
- [14] W. N. Charman, "The eye in focus: Accommodation and presbyopia," *Clin. Experim. Optometry*, vol. 91, no. 3, pp. 207–225, May 2008.
- [15] L. N. Thibos, X. Hong, A. Bradley, and R. A. Applegate, "Accuracy and precision of objective refraction from wavefront aberrations," *J. Vis.*, vol. 4, no. 4, p. 9, Apr. 2004.
- [16] J. R. Bennett, G. M. Stalboerger, D. O. Hodge, and M. M. Schornack, "Comparison of refractive assessment by wavefront aberrometry, autorefraction, and subjective refraction," *J. Optometry*, vol. 8, no. 2, pp. 109–115, Apr. 2015.
- [17] D. H. Burns, P. M. Allen, D. F. Edgar, and B. J. W. Evans, "Sources of error in clinical measurement of the amplitude of accommodation," *J. Optometry*, vol. 13, no. 1, pp. 3–14, Jan. 2020.
- [18] D. Burns, P. Allen, D. Edgar, and B. Evans, "A review of depth of focus in measurement of the amplitude of accommodation," *Vision*, vol. 2, no. 3, p. 37, Sep. 2018.
- [19] R. Aboumourad and H. A. Anderson, "Comparison of dynamic retinoscopy and autorefraction for measurement of accommodative amplitude," *Optometry Vis. Sci.*, vol. 96, no. 9, pp. 670–677, Sep. 2019.
- [20] D. López-Alcón, I. Marín-Franch, V. Fernández-Sánchez, and N. López-Gil, "Optical factors influencing the amplitude of accommodation," *Vis. Res.*, vol. 141, pp. 16–22, Dec. 2017.
- [21] R. Rampat, G. Debellemannière, J. Malet, and D. Gatineau, "Using artificial intelligence and novel polynomials to predict subjective refraction," *Sci. Rep.*, vol. 10, no. 1, pp. 1–9, May 2020.
- [22] P. Kanclerz, K. Pluta, H. Momeni-Moghaddam, and R. Khoramnia, "Comparison of the amplitude of accommodation measured using a new-generation closed-field autorefractor with conventional subjective methods," *Diagnostics*, vol. 12, no. 3, p. 568, Feb. 2022.
- [23] S. K. Dhallu, A. L. Sheppard, T. Drew, T. Mhashi, J. F. Zapata-Díaz, H. Radhakrishnan, D. R. Iskander, and J. S. Wolffsohn, "Factors influencing pseudo-accommodation—The difference between subjectively reported range of clear focus and objectively measured accommodation range," *Vision*, vol. 3, no. 3, p. 34, Jun. 2019.
- [24] D. Burns, B. Evans, and P. Allen, "Clinical measurement of amplitude of accommodation: A review," *Optom. Pract.*, vol. 15, no. 3, pp. 75–86, 2014.
- [25] A. León, M. Rosenfield, J. M. Estrada, S. M. Medrano, and M. M. Marquez, "Lag of accommodation between 5 and 60 years of age," *Optom. Vis. Perform.*, vol. 5, no. 3, pp. 103–108, Jun. 2017.
- [26] L. A. Ostrin and A. Glasser, "Accommodation measurements in a prepresbyopic and presbyopic population," *J. Cataract Refractive Surgery*, vol. 30, no. 7, pp. 1435–1444, Jul. 2004.
- [27] H. A. Anderson and K. K. Stuebing, "Subjective versus objective accommodative amplitude: Preschool to presbyopia," *Optometry Vis. Sci.*, vol. 91, no. 11, pp. 1290–1301, Nov. 2014.
- [28] A. León, J. M. Estrada, and M. Rosenfield, "Age and the amplitude of accommodation measured using dynamic retinoscopy," *Ophthalmic Physiol. Opt.*, vol. 36, no. 1, pp. 5–12, Jan. 2016.
- [29] K. J. Ciuffreda, "Accommodation, the pupil and presbyopia," in *Borish's Clinical Refraction*, 2nd ed., W. J. Benjamin, Ed. Philadelphia, PA, USA: Butterworth-Heinemann, 2006, pp. 77–120.
- [30] M. B. Hamill, Ed., "Accommodative and nonaccommodative treatment of presbyopia," in *Proc. BCSC (Basic Clin. Sci. Course), Refractive Surgery*, Amer. Acad. Ophthalmology, 2019, p. 159.
- [31] I. L. Bailey and J. E. Lovie, "New design principles for visual acuity letter charts," *Optometry Vis. Sci.*, vol. 53, no. 11, pp. 740–745, Nov. 1976.
- [32] D. A. Rosser, S. N. Cousens, I. E. Murdoch, F. W. Fitzke, and D. A. H. Laidlaw, "How sensitive to clinical change are ETDRS logMAR visual acuity measurements?" *Invest. Ophthalmol. Vis. Sci.*, vol. 44, no. 8, pp. 3278–3281, Aug. 2003.

- [33] L.-A. Lim, N. A. Frost, R. J. Powell, and P. Hewson, "Comparison of the ETDRS logMAR, 'compact reduced logMar' and snellen charts in routine clinical practice," *Eye*, vol. 24, no. 4, pp. 673–677, Jun. 2009.
- [34] I. L. Bailey and J. E. Lovie-Kitchin, "Visual acuity testing. From the laboratory to the clinic," *Vis. Res.*, vol. 90, pp. 2–9, Sep. 2013.
- [35] D. B. Elliott, "The good (logMAR), the bad (Snellen) and the ugly (BCVA, number of letters read) of visual acuity measurement," *Ophthalmic Physiol. Opt.*, vol. 36, no. 4, pp. 355–358, Jun. 2016.
- [36] L. P. Tidbury, G. Czanner, and D. Newsham, "Fiat lux: The effect of illuminance on acuity testing," *Graefes Arch. Clin. Experim. Ophthalmol.*, vol. 254, no. 6, pp. 1091–1097, Jun. 2016.
- [37] R. Blendowske, "Unaided visual acuity and blur," *Optom. Vis. Sci.*, vol. 92, no. 6, pp. 121–125, May 2015.
- [38] K. J. Ciuffreda, "Accommodation and its anomalies," in *Vision and Visual Dysfunction*, vol. 1, W. N. Charman, Ed. London, U.K.: Macmillan, 1991, pp. 231–279.
- [39] C. Ramsdale and W. N. Charman, "A longitudinal study of the changes in the static accommodation response," *Ophthalmic Physiol. Opt.*, vol. 9, no. 3, pp. 255–263, Jul. 1989.
- [40] H. Radhakrishnan and W. N. Charman, "Age-related changes in static accommodation and accommodative miosis," *Ophthalmic Physiol. Opt.*, vol. 27, no. 4, pp. 342–352, Jul. 2007.
- [41] W. N. Charman, "The path to presbyopia: Straight or crooked?" *Ophthalmic Physiol. Opt.*, vol. 9, no. 4, pp. 424–430, Oct. 1989.
- [42] W. N. Charman, "Developments in the correction of presbyopia I: Spectacle and contact lenses," *Ophthalmic Physiol. Opt.*, vol. 34, no. 1, pp. 8–29, Jan. 2014.
- [43] M. S. Almutairi, B. H. Altoaimi, and A. Bradley, "Accommodation and pupil behaviour of binocularly viewing early presbyopes," *Ophthalmic Physiol. Opt.*, vol. 37, no. 2, pp. 128–140, Mar. 2017.
- [44] M. J. Turner, "Observations on the normal subjective amplitude of accommodation," *Brit. J. Physiol. Opt.*, vol. 15, no. 2, pp. 70–100, 1958.
- [45] S. R. Cobb, "An investigation into accommodation by Ayrshire study circle," *British J. Ophthalmol.*, vol. 21, pp. 31–35, 1964.
- [46] A. Duane, "Studies in monocular and binocular accommodation with their clinical applications," *Amer. J. Ophthalmol.*, vol. 5, no. 11, pp. 865–877, Nov. 1922.
- [47] D. Hamasaki, J. Ong, and E. Marg, "The amplitude of accommodation in presbyopia," *Optometry Vis. Sci.*, vol. 33, no. 1, pp. 3–14, Jan. 1956.
- [48] M. Kalsi, G. Heron, and W. N. Charman, "Changes in the static accommodation response with age," *Ophthalmic Physiol. Opt.*, vol. 21, no. 1, pp. 77–84, Jan. 2001.
- [49] N. Kyurkchiev and S. Markov, *Sigmoid Functions: Some Approximation and Modelling Aspects*. New York, NY, USA: Academic, 2015.
- [50] A. Menon, K. Mehrotra, C. K. Mohan, and S. Ranka, "Characterization of a class of sigmoid functions with applications to neural networks," *Neural Netw.*, vol. 9, no. 5, pp. 819–835, Jul. 1996.
- [51] F. Lara, A. J. Del Águila-Carrasco, I. Marín-Franch, R. Riquelme-Nicolás, and N. López-Gil, "The effect of retinal illuminance on the subjective amplitude of accommodation," *Optometry Vis. Sci.*, vol. 97, no. 8, pp. 641–647, Aug. 2020.
- [52] C. A. Johnson, "Effects of luminance and stimulus distance on accommodation and visual resolution," *J. Opt. Soc. Amer. A, Opt. Image Sci.*, vol. 66, no. 2, pp. 138–142, Feb. 1976.
- [53] F. Lara, P. Bernal-Molina, V. Fernández-Sánchez, and N. López-Gil, "Changes in the objective amplitude of accommodation with pupil size," *Optometry Vis. Sci.*, vol. 91, no. 10, pp. 1215–1220, Oct. 2014.
- [54] J. A. Mordi and K. J. Ciuffreda, "Static aspects of accommodation: Age and presbyopia," *Vis. Res.*, vol. 38, no. 11, pp. 1643–1653, Jun. 1998.
- [55] A. J. Del Águila-Carrasco, F. Lara, P. Bernal-Molina, R. Riquelme-Nicolás, I. Marín-Franch, J. J. Esteve-Taboada, R. Montés-Micó, P. B. Kruger, and N. López-Gil, "Effect of phenylephrine on static and dynamic accommodation," *J. Optometry*, vol. 12, no. 1, pp. 30–37, Jan. 2019.
- [56] L. C. Bittencourt, M. R. Alves, D. O. Dantas, P. F. Rodrigues, and E. D. Santos-Neto, "An evaluation of estimation methods for determining addition in presbyopes," *Arquivos Brasileiros de Oftalmologia*, vol. 76, no. 4, pp. 218–220, Aug. 2013.
- [57] A. Gil, C. S. Hernández, A. S. Nam, V. Varadaraj, N. J. Durr, D. Lim, S. R. Dave, and E. Lage, "Predicting subjective refraction with dynamic retinal image quality analysis," *Sci. Rep.*, vol. 12, no. 1, pp. 1–10, Mar. 2022.
- [58] (2000). *ClinicalTrials.gov [Internet]. Bethesda (MD): National Library of Medicine (U.S.)*. Apr. 11, 2019. [Online]. Available: <https://clinicaltrials.gov/ct2/show/NCT03911596>
- [59] *BS EN 12464-1:2021 Light and lighting*, British Standards Inst., London, U.K., 2021.
- [60] P. Virtanen, R. Gommers, T. E. Oliphant, M. Haberland, T. Reddy, D. Cournapeau, E. Burovski, P. Peterson, W. Weckesser, J. Bright, and S. J. Van Der Walt, "SciPy 1.0: Fundamental algorithms for scientific computing in Python," *Nature Methods*, vol. 17, pp. 261–272, Feb. 2020.
- [61] K. Levenberg, "A method for the solution of certain non-linear problems in least squares," *Quart. J. Appl. Math.*, vol. 2, no. 2, pp. 164–168, Jul. 1944.
- [62] D. W. Marquardt, "An algorithm for least-squares estimation of nonlinear parameters," *J. Soc. Ind. Appl. Math.*, vol. 11, no. 2, pp. 431–441, 1963.
- [63] J. More, "The Levenberg–Marquardt algorithm: implementation and theory," in *Numerical Analysis (Lecture Notes in Mathematics)*, vol. 630, G. Watson, Ed. Berlin, Germany: Springer 1978, pp. 105–116.
- [64] S. Shan, "A Levenberg-Marquardt method for large-scale bound-constrained nonlinear least-squares," M.S. thesis, Dept. Comput. Sci., Univ. British Columbia, Vancouver, BC, Canada, 2008.
- [65] B. Wang and K. J. Ciuffreda, "Depth-of-focus of the human eye: Theory and clinical implications," *Surv. Ophthalmol.*, vol. 51, no. 1, pp. 75–85, Jan. 2006.
- [66] S. Shlaer, "The relation between visual acuity and illumination," *J. Gen. Physiol.*, vol. 21, no. 2, pp. 165–188, Nov. 1937.
- [67] F. W. Campbell, "The minimum quantity of light required to elicit the accommodation reflex in man," *J. Physiol.*, vol. 123, no. 2, pp. 357–366, Feb. 1954.
- [68] M. Fernandez-Alonso, A. Finch, G. Love, and J. Read, "Changes in the pupillary and accommodative responses of the human eye under different illuminations," *J. Vis.*, vol. 18, no. 10, p. 583, Sep. 2018.
- [69] R. A. Weale, "Real light scatter in the human crystalline lens," *Graefes Arch. Clin. Experim. Ophthalmol.*, vol. 224, no. 5, pp. 463–466, Sep. 1986.
- [70] T. M. Aslam, D. Haider, and I. J. Murray, "Principles of disability glare measurement: An ophthalmological perspective," *Acta Ophthalmol. Scandinavica*, vol. 85, no. 4, pp. 354–360, Feb. 2007.
- [71] M. A. Mainster and P. L. Turner, "Glare's causes, consequences, and clinical challenges after a century of ophthalmic study," *Amer. J. Ophthalmol.*, vol. 153, no. 4, pp. 587–593, Apr. 2012.
- [72] J. J. Vos, "Reflections on glare," *Lighting Res. Technol.*, vol. 35, no. 2, pp. 163–175, Jun. 2003.
- [73] T. J. T. P. van den Berg, L. J. (René) van Rijn, R. Kaper-Bongers, D. J. Vonhoff, H. J. Völker-Dieben, G. Grabner, C. Nischler, M. Emesz, H. Wilhelm, and D. Gamer, "Disability glare in the aging eye. Assessment and impact on driving," *J. Optometry*, vol. 2, no. 3, pp. 112–118, 2009.
- [74] P. D. Spear, "Neural bases of visual deficits during aging," *Vis. Res.*, vol. 33, no. 18, pp. 2589–2609, Dec. 1993.
- [75] A. Glasser and M. C. W. Campbell, "Presbyopia and the optical changes in the human crystalline lens with age," *Vis. Res.*, vol. 38, no. 2, pp. 209–229, Jan. 1998.
- [76] C. A. Curcio, C. Owsley, and G. R. Jackson, "Spare the rods, save the cones in aging and age-related maculopathy," *Investig. Ophthalmol. Vis. Sci.*, vol. 41, no. 8, pp. 2015–2018, Jul. 2000.
- [77] A. Glasser, "Presbyopia and aging in the crystalline lens," *J. Vis.*, vol. 3, no. 12, p. 22, Mar. 2010.
- [78] E. Wolf, "Glare and age," *Arch. Ophthalmol.*, vol. 64, no. 4, pp. 502–514, Oct. 1960.
- [79] M. S. M. Lasa, M. B. Datiles, M. J. Podgor, and B. V. Magno, "Contrast and glare sensitivity: Association with the type and severity of the cataract," *Ophthalmology*, vol. 99, no. 7, pp. 1045–1049, Jul. 1992.
- [80] M. S. M. Lasa, M. J. Podgor, M. B. Datiles, R. C. Caruso, and B. V. Magno, "Glare sensitivity in early cataracts," *British J. Ophthalmol.*, vol. 77, no. 8, pp. 489–491, 1993.
- [81] A. Facchin, R. Daini, and D. Zavagno, "The glare effect test and the impact of age on luminosity thresholds," *Frontiers Psychol.*, vol. 8, p. 1132, Jun. 2017.
- [82] M. B. Hamill, Ed., "The science of refractive surgery," in *Proc. BCSC (Basic Clin. Sci. Course), Refractive Surgery*, Amer. Acad. Ophthalmol., 2019, p. 9.
- [83] J. Porter, A. Guirao, I. G. Cox, and D. R. Williams, "Monochromatic aberrations of the human eye in a large population," *J. Opt. Soc. Amer. A, Opt. Image Sci.*, vol. 18, no. 8, p. 1793, Aug. 2001.

- [84] L. N. Thibos, X. Hong, A. Bradley, and X. Cheng, "Statistical variation of aberration structure and image quality in a normal population of healthy eyes," *J. Opt. Soc. Amer. A, Opt. Image Sci.*, vol. 19, no. 12, p. 2329, Dec. 2002.
- [85] T. O. Salmon and C. van de Pol, "Normal-eye Zernike coefficients and root-mean-square wavefront errors," *J. Cataract Refractive Surgery*, vol. 32, no. 12, pp. 2064–2074, 2006.
- [86] R. A. Applegate, W. J. Donnelly, J. D. Marsack, D. E. Koenig, and K. Pesudovs, "Three-dimensional relationship between high-order root-mean-square wavefront error, pupil diameter, and aging," *J. Opt. Soc. Amer. A, Opt. Image Sci.*, vol. 24, no. 3, pp. 578–587, Mar. 2007.
- [87] E. Berrio, J. Tabernero, and P. Artal, "Optical aberrations and alignment of the eye with age," *J. Vis.*, vol. 10, no. 14, p. 34, 2010.
- [88] G. Smith, "Relation between spherical refractive error and visual acuity," in *Optometry Vis. Sci.*, vol. 1991, vol. 68, no. 8, pp. 591–598.
- [89] E. M. Chia, J. J. Wang, E. Rochtchina, W. Smith, R. R. Cumming, and P. Mitchell, "Impact of bilateral visual impairment on health-related quality of life: The blue mountains eye study," *Investig. Ophthalmol. Vis. Sci.*, vol. 45, no. 1, pp. 71–76, Jan. 2004.

MOHIT U. KARKHANIS received the bachelor's degree in electronics engineering from the University of Pune, India, in 2014. He is currently pursuing the Ph.D. degree with the Department of Electrical and Computer Engineering, The University of Utah, USA. His graduate work is focused around developing adaptive optics based smart, wearable, vision correction devices, and systems for ophthalmic applications, including the world's first truly "autofocusing smart-eyeglasses" system, featured in multiple press articles. His current research interests include the design, development, and integration of low-power electronic systems featuring adaptive optics elements, smart autofocusing eyewear, embedded systems for MEMS devices and energy harvesters, and low-power oculometers.

AISHWARYADEV BANERJEE was born in Kolkata, India, in 1988. He received the bachelor's degree from the West Bengal University of Technology, India, in 2011, and the M.S. and Ph.D. degrees from The University of Utah, USA, in 2019.

His graduate work dealt with nano-scale engineering and advanced bio-sensing technologies. He is currently working as a Research Assistant Professor with The University of Utah. His active areas of research include adaptive optics, energy harvesting, ultra-low powered nano-sensors for detection of plant hormones, and neural-probe arrays. He has coauthored multiple technical articles and is a co-inventor of two provisional patents dealing with microfluidic actuators and low-power sensors. He has won multiple "best-paper" awards in international conferences and is a regular contributor to the research community by participating in the peer-reviewed journal publication process. He was part of the group which has recently created the world's first "smart eye-glasses" and this work was featured in multiple press articles. He is also part of a group which has developed a new bio-sensing mechanism which can be potentially used for smart, portable and low-power bio-hazard detection.

CHAYANJIT GHOSH received the bachelor's degree in electronics and communications engineering from the West Bengal University of Technology, India, in 2011, and the M.S. degree in electrical engineering from the State University of New York, Buffalo, USA, in 2015. He is a Graduate Student with the Department of Electrical and Computer Engineering, The University of Utah, USA. His research interests include adaptive optics and microfabrication technology.

RUGVED LIKHITE (Member, IEEE) received the bachelor's degree in metallurgical engineering and materials science from the Indian Institute of Technology Bombay, in 2014, and the M.S. and Ph.D. degrees in electrical engineering from The University of Utah, in 2017 and 2020, respectively. The focus of his dissertation was the development of low-power gas sensors for IoT applications. He is an expert in MEMS-based gas sensor design and microfabrication process development. He is currently working as a Process Engineer at Intel Corporation on development of copper barrier seed interconnect technology for 4 nm and beyond process nodes.

ERFAN POURSHABAN received the M.S. degree from the University of Tehran, in 2016. He is currently pursuing the Ph.D. degree with The University of Utah, under the supervision of Prof. Carlos Mastrangelo. His current research is focused on design, fabrication, and characterization of energy harvesting, energy storage, and power management circuits to supply power to wearable IoTs.

HANSEUP KIM received the B.S. degree in electrical engineering from Seoul National University, in 1997, and the M.S. and Ph.D. degrees in electrical engineering from the University of Michigan, in 2002 and 2006, respectively. From 2006 to 2009, he held a post-doctoral research fellow position with the Center for Wireless Integrated MicroSystems (WIMS), University of Michigan, working on a micro-scale gas chromatography system for environmental monitoring system. He is currently an USTAR Associate Professor in electrical and computer engineering and the Director of the Utah Nanofabrication Facility, The University of Utah. His present research at The University of Utah focuses on the development of micro sensors, actuators, and integrated IoT microsystems for health care and environmental monitoring applications. He was a recipient of both the prestigious national awards: NSF CAREER Award, in 2012, and the DARPA Young Faculty Award, in 2011. He has been a leading PI for multiple mid-to-large scale projects sponsored by DARPA and ARPAE. He has served as a Technical Program Committee Member of premier conferences in the fields of microelectromechanical systems (MEMS), including IEEE MEMS, Transducers and Hilton Head Workshops.

DAVID A. MEYER received the B.S. degree in exercise physiology from Brigham Young University, Provo, UT, USA, in 2005, and the O.D. degree from the Southern California College of Optometry, Fullerton, CA, USA, in 2009.

He also received a primary care residency from the VA Hospital, Salt Lake City, from 2009 to 2010, and finally a fellowship at the Moran Eye Center, Salt Lake City, from 2014 to 2015. From 2009 to 2014, he worked at various clinics, primarily focusing on specialty contact lenses. In 2014, he joined the John A. Moran Eye Center, The University of Utah, and continued his work seeing patients with keratoconus and other vision-disabling conditions. He has contributed heavily to Moran CORE, an educational resource for providers all over the world. He is currently completing a study on the use of scleral contact lenses.

Dr. Meyer is a member and a fellow of the American Academy of Optometry and the American Optometric Association. According to the Press Ganey survey, he is ranked in the top 10

CARLOS H. MASTRANGELO (Senior Member, IEEE) was born in Buenos Aires, Argentina, in 1960. He received the B.S., M.S., and Ph.D. degrees in electrical engineering and computer science from the University of California at Berkeley, in 1985, 1988, and 1991, respectively. His graduate work concentrated on the applications of microbridges in microsensor technology. From 1991 to 1992, he was with the Scientific Research Laboratory, Ford Motor Company, Dearborn MI, USA, developing microsensors for automotive applications. From 1993 to 2002, he was an Associate Professor in electrical engineering and computer science with the Center for Integrated Microsystems, University of Michigan, Ann Arbor. From 2000 to 2005, he was the Vice President of Engineering at Corning Intellisense, Wilmington, MA, USA, and the Director of the Biochemical Technologies Research Group, Corning, NY, USA. From 2005 to 2008, he was an Associate Professor of electrical engineering and computer science with Case Western Reserve University. He is currently an USTAR Professor of electrical engineering and bioengineering with The University of Utah, Salt Lake City. His research focuses on microelectromechanical system applications and technology, microfluidic systems, and integration, design, and modeling of MEMS fabrication processes. He is on the editorial boards of sensors and actuators was on the IEEE/ASME JOURNAL OF MICROELECTROMECHANICAL SYSTEMS, and he has participated in technical and organizing committees of numerous SPIE and IEEE conferences in the MEMS area.

...