

Metadata: Winston-Lutz Analysis

Função: Físico Number of images: 8

Acelerador Linear: Synergy Full Maximum 2D CAX->BB distance: 1.24mm Modelo MLC: Agility Median 2D CAX->BB distance: 0.89mm

Shift to iso: facing gantry, move BB: RIGHT 0.07mm; IN 0.33mm; DOWN 0.04mm

Gantry 3D isocenter diameter: 1.16mm (4/8 images considered)

Maximum Gantry RMS deviation (mm): 0.44mm Maximum EPID RMS deviation (mm): 2.60mm


Gantry+Collimator 3D isocenter diameter: 1.53mm (6/8 images considered)

Collimator 2D isocenter diameter: 0.40mm (3/8 images considered)

Maximum Collimator RMS deviation (mm): 1.24

Couch 2D isocenter diameter: 0.94mm (3/8 images considered)

Maximum Couch RMS deviation (mm): 0.98



Análise de Winston-Lutz com pylinac

Campo 2 cm X 2 cm Tolerância: 1 mm Limite de ação: 0.8 mm

Local das imagens: ./WL\WL\_2019-10-10\_analisado




Metadata:


Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility






Metadata:


Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility





Metadata:

Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility

