Winston-Lutz Analysis

Metadata:
Autor: William A. P. dos Santos
Função: Físico
Acelerador Linear: Synergy Full
Modelo MLC: Agility

Winston-Lutz Analysis
==========
Number of images: 9
Maximum 2D CAX->BB distance: 1.09mm
Median 2D CAX->BB distance: 0.69mm
Shift to iso: facing gantry, move BB: RIGHT 0.37mm; IN 0.05mm; DOWN 0.67mm
Gantry 3D isocenter diameter: 1.20mm (4/9 images considered)
Maximum Gantry RMS deviation (mm): 0.92mm
Maximum EPID RMS deviation (mm): 2.71mm
Gantry+Collimator 3D isocenter diameter: 1.52mm (7/9 images considered)
Collimator 2D isocenter diameter: 0.57mm (4/9 images considered)
Maximum Collimator RMS deviation (mm): 0.98
Couch 2D isocenter diameter: 1.17mm (3/9 images considered)
Maximum Couch RMS deviation (mm): 1.09









Notes:

Análise de Winston-Lutz com pylinac
Campo 2 cm X 2 cm
Tolerância: 1 mm
Limite de ação: 0.8 mm
Local das imagens: ./WL_IOSP_20201216\WL_2020-10-28

Winston-Lutz Analysis

Metadata:
Autor: William A. P. dos Santos
Função: Físico
Acelerador Linear: Synergy Full
Modelo MLC: Agility



Winston-Lutz Analysis

Metadata:
Autor: William A. P. dos Santos
Função: Físico
Acelerador Linear: Synergy Full
Modelo MLC: Agility

Collimator images



Winston-Lutz Analysis

Metadata:
Autor: William A. P. dos Santos
Função: Físico
Acelerador Linear: Synergy Full
Modelo MLC: Agility