



# Picket Fence Analysis

## Metadata:

Autor: William A. P. dos Santos

Função: Físico

Unidade: ICESP

Acelerador Linear: AL06

Modelo MLC: Beam Modulator

## Picket Fence results:

Magnification factor (SID/SAD): 1.60

Tolerance (mm): 0.5

Leaves passing (%): 100.0

Absolute median error (mm): 0.053



Mean picket spacing (mm): 20.2

Picket offsets (mm): 81.8 61.5 41.2 21.1 0.9 -19.2 -39.4 -59.7 -80.0

Max Error: 0.214 mm on Picket: 8, Leaf: 1

Gantry Angle: 0.00

Collimator Angle: 0.00



## Notes:

Análise da variação de MU e largura da faixa

Campo 16 cm X 20 cm

Sem erros propositais

Tolerância: 0.5 mm

Limite de ação: 0.3 mm

Arquivo de imagem: 10mm-3MU.tif