

Picket Fence Analysis

Metadata:

Autor: William A. P. dos Santos

Função: Físico Unidade: ICESP

Acelerador Linear: AL02 Modelo MLC: MLCi2 Picket Fence results:

Magnification factor (SID/SAD): 1.60


Tolerance (mm): 0.5 Leaves passing (%): 100.0


Absolute median error (mm): 0.043 Mean picket spacing (mm): 20.4

Picket offsets from CAX (mm): 81.0 60.6 40.2 20.0 -0.4 -20.8 -41.3 -61

Max Error: 0.195 mm on Picket: 6, Leaf: 8

Gantry Angle: 0.00 Collimator Angle: 0.00





Notes:

Testando análise variação de MU e largura da faixa Campo 20 cm X 20 cm Sem erros propositais Tolerância: 0.5 mm Limite de ação: 0.3 mm

Arquivo de imagem: 5mm-00Y1x20Y2-5MU.tif