

Metadata: Winston-Lutz Analysis

Função: Físico Number of images: 9

Acelerador Linear: Synergy Full Maximum 2D CAX->BB distance: 1.24mm Modelo MLC: Agility Median 2D CAX->BB distance: 0.71mm

Shift to iso: facing gantry, move BB: RIGHT 0.03mm; IN 0.19mm; DOWN 0.17mm

Gantry 3D isocenter diameter: 1.22mm (4/9 images considered)

Maximum Gantry RMS deviation (mm): 0.61mm Maximum EPID RMS deviation (mm): 4.09mm


Gantry+Collimator 3D isocenter diameter: 1.61mm (7/9 images considered)

Collimator 2D isocenter diameter: 0.40mm (4/9 images considered)

Maximum Collimator RMS deviation (mm): 1.00

Couch 2D isocenter diameter: 1.22mm (3/9 images considered)

Maximum Couch RMS deviation (mm): 1.24



Análise de Winston-Lutz com pylinac

Campo 2 cm X 2 cm Tolerância: 1 mm Limite de ação: 0.8 mm

Local das imagens: ./WL\_IOSP\_20201216\WL\_2020-12-02\_2




Metadata:

Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility






Metadata:


Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility





Metadata:

Autor: William A. P. dos Santos

Função: Físico

Acelerador Linear: Synergy Full

Modelo MLC: Agility

