

Airedale Metrology Services Ltd

Unit 7, Aire Valley Business Park, Ainsdale Road, Keighley, West Yorkshire, BD21 4LZ
Tel 01535 640 218 | info@airedale-metrology.co.uk | Company No. 07421996

CERTIFICATE OF CALIBRATION

Accredited calibration - issued in accordance with the laboratory's accredited scope

Certificate number	AMS-C-26-0418	Date of calibration	17/02/2026
Customer	Ansford Manufacturing Group	Date of issue	19/02/2026
Site	Line 2 Quality Room	Recommended next calibration	17/02/2027

Instrument under calibration

Description	Coordinate measuring machine (CMM)
Manufacturer / Model	Mitutoyo Crysta-Apex S 574
Asset number	AMG-CMM-02
Serial number	MT-574-118207
Controller / software	Mitutoyo MCOSMOS
Location	Ansford Manufacturing Group, Line 2 Quality Room

Calibration method

Acceptance and reverification test of the length measurement error carried out in accordance with BS EN ISO 10360-2:2009. The length measurement error E0 was determined using a calibrated gauge block artefact measured in five positions within the measuring volume, three repeat measurements at each position. Ambient temperature 20.0 deg C +/- 1.0 deg C. Reference standards traceable to national measurement standards.

Results - length measurement error (ISO 10360-2)

Nominal length (mm)	E0 measured (um)	MPE_E (um)	Result
30	0.8	+/- 1.9	Pass
105	1.1	+/- 2.2	Pass
205	1.4	+/- 2.5	Pass
305	1.7	+/- 2.8	Pass
405	2.3	+/- 3.1	Pass

MPE_E is the maximum permissible error stated by the manufacturer for this model, $MPE_E = +/- (1.7 + L/300)$ um where L is in mm. Each E0 value is the largest magnitude error observed across the repeat verification measurements at the stated nominal length.

Verdict: PASS - all measured length measurement errors are within the maximum permissible error for the instrument.

Calibrated and approved by

D. Hartley
Calibration Engineer
Airedale Metrology Services Ltd

Signature: D. Hartley
Date: 19/02/2026