Water Quality Report

Sample Pod · 2026-08-04 to 2026-08-11

Watch

Coordinates	33.7400° N, 118.1000° W (Location Scrubbed)
Water Body Type	Marine (from device registry)
Client / Contract	Not available from device registry
Report Date	2026-10-10
Last Reading	2026-08-11 19:37 UTC (60 days before this report)
Prepared By	Clean Earth Rovers
Omitted	turbidity notes (at the requester's request)

This pod has not reported for 60 days. This report describes the water during the reporting period, not current conditions.

1. Summary

- Overall status: Watch.
- Temperature (°F) moved outside the site baseline for the 2026-08-04 to 2026-08-11 reporting period.
- 1 candidate event(s) were identified.
- Turbidity (Relative): Turbid (relative index mean 805.2, held steady across the period).
- Recommendation: see Recommendations below for the operational, investigative, and stakeholder follow-up.

2. Parameter Data

Parameter	Site Baseline	Min	Max	Mean	Median	Out of range	Flag
Dissolved Oxygen (mg/L)	4-15	7.61	10.77	9.13	9.14	0%	Normal
ORP (mV)	50-400	200.20	221.80	213.72	215.95	0%	Normal
pH	6-10	6.98	7.57	7.37	7.47	0%	Normal
Conductivity (µS/cm)	40000-75000	52825	63040	59005	60225	0%	Normal
Temperature (°F)	50-80	66.13	83.52	72.12	68.58	17%	Elevated
Turbidity (Relative)	—	520.00	1005	805.19	846.00	—	Turbid

Flag values: Normal, Elevated, Low, or Exceedance relative to the site baseline. Out of range is the share of the period's series buckets whose average sat outside that baseline — approximate, at bucket resolution, and shown because a flag alone cannot separate a single stray reading from a sustained offset.
Site Baseline for Dissolved Oxygen (mg/L), ORP (mV), pH, Conductivity (µS/cm), Temperature (°F) is this device's operator-configured threshold from the device registry — an alert limit the operator set, not a scientifically reviewed range. A parameter with no usable registry threshold shows no baseline at all ("Not established") rather than falling back to a fixed table.

3. Parameter Analysis

Each chart plots the period's series bucket means, oldest to newest. The shaded band is the site baseline where one exists; the dots mark the plotted high and low.

Temperature (°F)

Elevated

Showed no clearly classified pattern; against the 50-80 °F operator-set threshold for this device it moved slightly above it to 83.52 °F, recorded as an elevated. Outside baseline in 17% of the period's series buckets. First left baseline at 2026-08-10 20:30. Operator-set threshold for this device (50-80 °F, from the device registry).

Turbidity (Relative)

Turbid

Water clarity read as Turbid for the period (relative index mean 805.2, range 520.0-1005.0), held steady across the period. The period spanned moderate to turbid conditions.

4. Event Detection

Event 1 — Threshold crossing

Low

2026-08-10 20:30 to 2026-08-10 23:30

Parameter movements: Temperature (°F) rose to 83.46 °F on bucket averages (period max 83.52 °F)

Interpretation: Readings crossed this pod's configured thresholds. No approved explanation covers this pattern, so the report does not name a cause. A sensor problem can look like a water-quality change. A single isolated point, a reading stuck at the same value for a long time, or every reading changing at the same instant are more likely signs of a sensor or connection problem. A real event more often shows a combination of readings that fits a physical explanation and lasts across several readings. The change lasted 3.0 hours.

Follow-up: Inspect the pod; Confirm with an independent measurement; Check for nearby activity; Record what you see; Keep watching the readings; Share the report (see Recommendations).

5. Data Quality

Check	Result	Notes
Data completeness	100.0%	47 of 47 readings returned for the period were usable for at least one parameter. 0 readings were left out because no requested parameter was usable. Share of returned readings, not coverage against an expected sampling cadence.
Calibration status	Pass	No readings outside plausibility bounds in the period. Probe error flags were clear for all values used above.
Drift indicators	Not assessed	Not assessed — this pipeline has no drift detector. Requires comparison against a calibration record, which is not available from the device API.
Biofouling indicators	Not assessed	Not assessed — this pipeline has no biofouling detector. Requires service history or a fouling-sensitive baseline, neither of which is available here.
Sensor agreement	Not assessed	Not assessed — cross-sensor agreement needs a second co-located pod or a grab-sample result to compare against.

6. Recommendations

Operational: Inspect the sensors for fouling or damage and check that the pod is sitting properly in the water, so an instrument problem can be ruled out. Clean the sensors only if you can do so safely and within your site's safety rules; if a sewage, bloom or chemical release is suspected, avoid touching the water. If the battery is low or a sensor looks damaged, contact Clean Earth Rovers support. Don't let this delay other steps if people, pets or fish may be at risk. Contact: sales@cleanearthrovers.com, or your usual contact at Clean Earth Rovers if you already work with one

Investigative: Ask a qualified person to confirm the flagged readings with an independent measurement, such as a calibrated handheld meter or a laboratory sample, before drawing conclusions. Avoid contact with water that may be contaminated. Check for activity near or upstream of the pod at the time of the change, such as construction, dredging, pumping, outfall discharges or heavy boat traffic. Note what you can see from a safe distance, such as water color, debris, dead fish or nearby work, with times and photos. Mention strong odors, but don't lean over or approach water you suspect is contaminated. Keep watching the readings over the following days. A change that lasts across several readings or recovers gradually is more likely real. A single isolated point is more often a sensor glitch, but a brief spike can also be real, so note when it happened.

Stakeholder: Share this report with the people responsible for the site, such as your site manager or monitoring team.