

Certificate of Calibration

INSTRUMENT

Model Number: iPA-0850A
Description: Portable PIM Analyzer - iPA-0850A
Serial Number: TX2152400189

CALIBRATION EQUIPMENT USED

ACE-2000	S/N: TX2254000039	S/N: TX2254000021
ACE-3000	S/N: TX2254400068	S/N: TX2254400074

CERTIFICATION

Certificate No : 202607130906TX2152400189
Date of Calibration : 13 July 2026 (UTC)

This certifies that the instrument detailed above has been calibrated in accordance with the comparison method, in compliance with a quality management system registered to ISO9001:2015 and the manufacturer's specification, which can be accessed on the web address of

<http://www.kaelus.com/test-and-measurement-solutions/>.

All activities were conducted in an environment at or around +23° Celsius, using test equipment calibrated against national and/or international standards.

PIM equipment calibration and measurements are conducted in accordance with IEC62037.

Kaelus recommends a calibration interval of 12 months to ensure measurement accuracy.

CALIBRATION PERFORMED BY

Name: Nutan Adhikari
Company Name: Entelso Pty Ltd

AUTHORISATION

Authorised By: Kaelus Service Manager