

## Section Thickness Verification Sheet

### Routine verification (3–5 µm unless specified)

Use a reference block or film gauge. Record instrument, target, observed values, and adjustments. For kidney/PAS, verify 2–3  $\mu\text{m}$ . Investigate deviations and document corrective action.

[illegible]