
## RAPORT Z POMIARU MODULU PV, AGH SOLAR PLANE



Pmmp = 19.4W Vmmp = 7.039V Immp = 2.763A Voc = 11.532V Isc = 3.935A

**Sprawnosc:** 7.55%

Czas trwania pomiaru: 6.81s

## Warunki pomiaru

Natezenie swiatla = 700W/m2 Pole ogniwa = 153cm2 liczba ogniw = 24