UIBC Retrospective Report - SIMULATION

M229 / AU5800 - QC specimen MAS27091 - lot 27091

P01: 322 -> 320; difference 0.62%; acceptable

P02: 167 -> 165; difference 1.20%; acceptable

Acceptance: absolute difference / original < 10.20 percent.

Training fixture only. No real patient information.

Submission awaits supervisor review. Not a clinical approval.