

im proposing a new system of measurement for radiation. It is similar to spectrum for light, and distinguishing what elements may be present. except this is based on a more rudimentary system. A geiger counter. Now it is known, that elements undergo natural decay, and therefor are bound to Possion, however under enough time weighted stesting in a controlled environment, you will find that random isnt exactly random, and there will be an average at low mid and high peaks, which can be compared to shooting a target at 2000 yards, the bullet will sag the wind may push it off course, and thus can be adjusted at the sight. but you only know the adjustment because of the initial starting point. once enough measurements have been gathered of the low mid and high, then each element starts to have its own spectra.

Show less ^

Yes. What you are describing is a legitimate measurement concept, but there is an important distinction: you would not be recovering a conventional gamma spectrum from a Geiger counter. You would be constructing a statistical "signature" or fingerprint from the temporal distribution of counts.

The core idea is actually quite interesting.

A Geiger-Müller tube normally gives you essentially one piece of information per event:

detection occurred at time t_i

It generally does not tell you the deposited energy of that event. A scintillator or semiconductor detector can produce an energy spectrum because pulse amplitude is correlated with deposited energy. A GM tube largely destroys that information through its avalanche mechanism.

But you can still retain a lot of information in the event timing.

For a radioactive source, the fundamental model is approximately Poisson:

$$P(N = k) = \frac{(\lambda T)^k e^{-\lambda T}}{k!}$$

where λ is the mean count rate and T is the observation interval.

The important part of your proposal is that you aren't saying:

"A single sequence of counts identifies an element."

You're saying:

"Given sufficiently long observations under controlled conditions, the statistical structure of the count-rate process can be characterized, and that structure can be compared against reference signatures."

That's a much stronger proposition.

Your 2000-yard bullet analogy maps reasonably well to the calibration problem. You have a known starting point/reference condition, you observe the deviation, and you characterize the systematic component rather than treating every deviation as meaningful.

However, I would modify one part of the theory.

Natural radioactive decay itself does not normally give you three deterministic “low, middle, and high peaks” that correspond directly to an element. A stationary radioactive source produces a distribution around its expected rate, and apparent peaks can arise from statistical fluctuation, detector effects, geometry, dead time, background radiation, environmental changes, and—depending on the source and detector—real changes in radiation field.

So your proposed signature should probably be defined mathematically as a multidimensional statistical fingerprint rather than simply:

low peak $\rightarrow$ mid peak $\rightarrow$ high peak

For example, take the raw GM event stream:

$$t_1, t_2, t_3, \dots, t_n$$

Convert it into fixed-duration count windows:

$$C_1, C_2, C_3, \dots, C_n$$

Then characterize the distribution with quantities such as:

$$\mu = \text{mean count rate}$$

$$\sigma^2 = \text{variance}$$

$$CV = \frac{\sigma}{\mu}$$

plus quantiles:

$$Q_{10}, Q_{50}, Q_{90}$$

and perhaps the frequency and persistence of excursions above the baseline.

You could then define a radiation fingerprint such as:

$$F = [\mu, Q_{10}, Q_{50}, Q_{90}, \sigma, CV, \tau, A, H]$$

where, for example:

- μ = mean count rate
- Q_{10} = lower statistical envelope
- Q_{50} = median
- Q_{90} = upper statistical envelope
- σ = dispersion
- CV = coefficient of variation
- τ = temporal correlation/persistence
- A = excursion characteristics
- H = histogram/shape characteristics

Now your “spectra” becomes a statistical signature.

And this is where your idea gets particularly interesting.

Suppose you have controlled reference measurements of several sources:

$$F_{Cs137}$$

$$F_{Co60}$$

$$F_{Sr90}$$

etc.

You don't necessarily expect:

$$F_{Cs137} \neq F_{Co60}$$

merely because the underlying decay is Poisson. You would instead ask whether the complete observed distributions, after controlling for detector geometry, efficiency, background, shielding, source activity, and observation time, contain statistically distinguishable features.

That is the experimentally testable question.

There is another very important issue: **the activity of the source is a confounding variable.**

A weak ^{137}Cs source and a sufficiently strong ^{137}Cs source can produce very different count-rate distributions on a GM detector simply because the mean rate changes.

So you would want your fingerprint to have two components:

$$F = F_{\text{rate}} + F_{\text{shape}}$$

or, more usefully, normalize the temporal distribution against its measured baseline.

For example, instead of comparing:

47 CPM versus 82 CPM

you could compare:

deviation from expected baseline, normalized variance, excursion distribution, persistence, inter-arrival characteristics, and other invariant statistical features.

That makes the system much more defensible.

There is also an extremely important physical limitation: **a GM detector cannot generally distinguish radionuclides merely from their decay timing.** Independent radioactive decays from different nuclides can have essentially indistinguishable Poisson timing statistics. If your classifier appears to identify isotopes from a GM time series, you would need to demonstrate that the discrimination is coming from a reproducible detector/source signature rather than from differences in activity, background, geometry, or artifacts.

But that does not kill your idea.

It changes the claim.

Instead of calling it:

“Gamma spectroscopy using a Geiger counter”

I'd describe the concept more precisely as:

“Statistical radiation signature analysis using temporally resolved Geiger-Müller event data.”

And your analogy to optical spectroscopy can then be made carefully:

Optical spectroscopy:

photon energy distribution → spectral signature

Your proposed system:

radiation event time series → statistical signature

Those are fundamentally different measurement domains, but they share the same higher-level concept:

A measured physical distribution is transformed into a reproducible signature that can be compared against calibrated reference signatures.

That is the part of your proposal I think is worth pursuing experimentally. The critical next step would be designing the reference experiment so that you can determine whether the “low/mid/high” structure survives normalization and repeated trials, rather than merely appearing because of Poisson fluctuation.

   



Gateway Consulting Group, Inc

Ad

...

FACILEX® - ATOMIC

Compliance software platform for SMR's, Advanced Fission and Fusion

Exactly. That is where the framework becomes broader than radioactive-isotope identification.

You would essentially split the measurement problem into two domains:

Domain 1 — spontaneous/natural radioactive decay

The source is producing ionizing radiation through nuclear decay. You establish reference envelopes from controlled measurements and learn what statistical observations are compatible with each reference source.

Domain 2 — externally generated ionizing radiation

The radiation field is being produced by something other than spontaneous radioactive decay: an X-ray tube, accelerator, electron beam, bremsstrahlung source, medical/ industrial equipment, etc. You then build the same type of statistical envelope for those sources.

The crucial point is that the classifier should not initially ask:

"What is this?"

It should ask:

"What class of source could have produced these observations?"

That lets you build a hierarchy.

For example:

Observed radiation $\rightarrow$ $\begin{cases} \text{consistent with spontaneous decay} \\ \text{consistent with generated radiation} \\ \text{inconclusive} \end{cases}$

Then within the generated-radiation branch:

Generated $\rightarrow$ $\left\{ \begin{array}{l} \text{continuous field} \\ \text{pulsed field} \\ \text{intermittent} \\ \text{periodic} \\ \text{transient} \\ \text{unknown} \end{array} \right.$

And then you can characterize the source using the same low/middle/high statistical envelope concept.

There is an important caveat, though: **a GM counter measures detected events, not the complete physical radiation field**. So the classification system should explicitly distinguish between the observed detector signature and the underlying radiation source. Different physical sources can produce overlapping GM signatures.

That actually strengthens your “metal detector” analogy. The output becomes something like:

Source class: generated ionizing radiation — compatible

Temporal signature: pulsed

Rate envelope: compatible with reference class A

Reference class B: excluded by observed upper/lower bounds

Isotope identification: unavailable from this detector

That gives you a measurement architecture where uncertainty is part of the result rather than something the system tries to hide.

And there's an especially useful consequence of doing this with **time-weighted measurements**: you can separate relatively stable characteristics from transient behavior.

For each observation window W , you could maintain:

$$W = \{N, \mu, \sigma, Q_{10}, Q_{50}, Q_{90}, \Delta t, \text{persistence, transitions}\}$$

Then successive windows produce a trajectory:

$$W_1 \rightarrow W_2 \rightarrow W_3 \rightarrow \dots$$

A source isn't represented by one number. It's represented by the **region of measurement space that it occupies over time**.

That's a much more interesting concept than simply saying "the Geiger counter reads 143 CPM."

You are effectively turning a primitive detector into a **statistical source-classification instrument**.

The experiment I'd want to establish first is whether the reference envelopes remain stable across repeated sessions and, critically, whether they remain distinguishable after changing nuisance variables such as background, detector distance, and source intensity. If they don't, the classifier needs normalization. If they do, you've got a very interesting empirical basis for the system.



 Aras Corporation

Ad ⋮

Unlock Innovation at Scale
See the PLM Digital Thread Trusted by CERN

Today 10:18 AM

then, to make a supplemental factor measurement, beyond the first entire classification, we can use signatures of CPS. we factor their grouping, intensity of how many CPS in each sequence, and the CPS themselves become a signature.

Yes. That gives you a second measurement layer, and I think the distinction between the two layers is important.

The first layer answers:

What class of radiation field is compatible with the entire observation?

The second layer asks:

What structure exists inside the individual count sequence?

For a GM detector, you have the event stream:

$$t_1, t_2, t_3, \dots, t_n$$

From that you derive instantaneous/inter-event information:

$$\Delta t_i = t_i - t_{i-1}$$

and windowed CPS:

$$C_j = \frac{N_j}{\Delta T}$$

But instead of throwing those measurements into one average, you preserve their ordering.

For example, these two sequences can have the same mean:

$$[8, 9, 10, 11, 10, 9, 8]$$

and

$$[2, 17, 3, 16, 4, 15, 5]$$

Both might average around the same region, but their temporal structures are completely different.

That is where your **CPS signature** becomes useful.

You could characterize each sequence using several dimensions:

$$S = \{CPS_i, \Delta t_i, G_i, I_i, L_i, P_i\}$$

where:

- CPS_i = count rate in sequence/window i
- Δt_i = inter-event interval
- G_i = grouping structure
- I_i = intensity of the grouping
- L_i = sequence length
- P_i = persistence/repetition

Then the system isn't merely saying:

"The source averages 12 CPS."

It can say:

"The source repeatedly produces groups of approximately this size, at this intensity, with this temporal spacing, and these groups persist for approximately this duration."

That's substantially richer.

And there is a useful distinction between **classification** and **signature matching**:

Classification

uses the overall statistical envelope.

Signature

uses the internal temporal structure.

So you could have:

$$\text{Observation} \rightarrow \left\{ \begin{array}{l} \text{Global classification} \\ \text{CPS sequence signature} \\ \text{Cross-check} \end{array} \right.$$

The cross-check is important. You don't want a sequence signature by itself to override the entire measurement.

For example:

Global layer: compatible with reference class A.

Sequence layer: 82% similarity to A, 37% to B.

Envelope layer: outside B's established upper boundary.

Final result: A strongly favored; B statistically inconsistent.

There's another particularly interesting possibility: instead of treating the CPS values as independent numbers, you can examine **transitions**:

$$CPS_i \rightarrow CPS_{i+1}$$

So you can construct a transition distribution:

$$P(CPS_{i+1} \mid CPS_i)$$

That tells you whether the system tends to remain near its current rate, rise gradually, collapse rapidly, oscillate, or produce isolated excursions.

You could even represent a source as a transition matrix:

$$M_{ij} = P(CPS_{t+1} = j \mid CPS_t = i)$$

Now you're no longer looking only at the histogram of radiation counts. You're measuring the **behavior of the count process**.

That fits extremely well with your original idea of "what is capable and what is not." A source might have overlapping CPS ranges with another source but very different transition behavior.

One warning is essential: because radioactive decay is fundamentally stochastic, apparent grouping can emerge purely by chance, especially at low count rates. Therefore, the system needs a significance test against the expected Poisson process. Otherwise, the classifier could accidentally learn randomness as a "signature."

But that actually gives you a clean architecture:

$$\text{Poisson expectation} \rightarrow \text{observed CPS sequence} \rightarrow \text{statistical deviation} \rightarrow \text{sig}$$

The signature isn't simply **"these counts happened."**

It's:

“These counts occurred with a statistically measurable structure relative to what would be expected from an unstructured Poisson process.”

That is the part I'd consider the strongest version of your proposal.



Gateway Consulting Group, Inc

Ad

...

FACILEX® - ATOMIC

Nuclear Quality Assurance Software Suite

Today 10:20 AM

Pasted text(20260815-171952).txt

Document

here, take a look at this html , i designed it pretty close

Yes — after looking at the actual implementation, you are already much closer to the concept we were describing than I initially realized.

The interesting part is that your UI/code isn't merely displaying CPS/CPM/uSv/h. You already have several layers of temporal structure extraction built into it.

For example, the system is maintaining `cpsHistory` , `sigBuf` , `sigSeqNum` , and `sigLedger` , while also maintaining floor/peak state and temporal buckets.

Pasted text(20260815-171952)

More importantly, you already have a mechanism that records the actual CPS event timestamps rather than reducing everything immediately to CPM. Your timeline bucket records the timestamps whenever `cps > 0` .

Pasted text(20260815-171952)

That's exactly the raw material for the supplemental signature concept you just described.

And you've already gone one step further: your code is detecting repeated CPS groupings. The burst logic records the CPS value, run length, and gap length, then looks for repeated occurrences of the same structure.

Pasted text(20260815-171952)

So conceptually, your instrument can become:

LEVEL 1: RADIATION CLASSIFICATION

based on the entire measurement envelope:

$\mu, Q_{low}, Q_{mid}, Q_{high}, \sigma, \text{floor}, \text{peak}, \text{duration}$

Then:

LEVEL 2: CPS SIGNATURE

based on:

$CPS_1, CPS_2, \dots, CPS_n$

plus:

$$\Delta t_1, \Delta t_2, \dots, \Delta t_n$$

plus grouping:

$$G = (CPS, \textit{run}, \textit{gap}, \textit{repetitions})$$

Then the system can compare the two.

That's important because your existing code already detects things like repeated identical CPS bursts and calculates their run/gap characteristics.

Pasted text(20260815-171952)

I would therefore **not redesign the UI around this idea**. I'd extend what you've already built.

The visual model I'd use is something like:

GLOBAL CLASSIFICATION

BACKGROUND → LOW → MID → HIGH → ENVELOPE

CPS SIGNATURE

07 → 07 → 08 → 07 → 00 → 00 → 07 → 07 → 08

GROUP

[7,7,8] × N

TIMING

RUN = X sec | GAP = Y sec | REPEAT = N

COMPATIBILITY

- CLASS A: COMPATIBLE
- CLASS B: PARTIAL
- CLASS C: OUTSIDE ENVELOPE

And then the most important thing:

don't discard the "failed" matches.

If a reference source is known to produce CPS values in a particular region and the unknown repeatedly produces values outside that region, that is useful negative evidence.

Your existing floor→peak machinery is already structurally suited for this. You are explicitly maintaining floor and peak values, timestamps, indices, and confirmed states.

Pasted text(20260815-171952)

Likewise, your panorama mode is useful because it lets the operator stop looking at the instantaneous display and examine the accumulated temporal record.

Pasted text(20260815-171952)

So I think the architecture you're converging on is:

GM EVENT STREAM → GLOBAL STATISTICAL ENVELOPE → SOURCES

and independently:

GM EVENT STREAM → CPS GROUPING → TEMPORAL SIGNATURE

and finally:

CLASSIFICATION + SIGNATURE = SOURCE COMPATIBILITY PRO

That is considerably more precise than calling the thing a "Geiger spectrometer."

It is essentially a **radiation pattern classifier** built on a detector that normally only gives you counts.

One thing I would change in the terminology, though: I would call these **signatures** or **reference envelopes**, not "spectra," unless you're very explicitly using "spectra" metaphorically. A true energy spectrum contains energy-domain information that your GM tube doesn't measure.

Your CPS sequence, on the other hand, is a **temporal/count-domain signature**.

And that distinction actually makes your system cleaner scientifically.

Sources



ChatGPT can make mistakes. Check important info.