

Using XyphroLabs UsbGpib adapter with Linux & Visa

(Ubuntu + R&S Visa and Python with PyVisa)

Contents

Introduction.....	2
SW setup	2
Installation of R&S Visa.....	2
Installation of PyVisa	5
HW Setup	6
First interaction with the instrument	8
Confirming that the USB device is detected	8
Testing if the instrument is detected by R&S Visa	9
Why are no USBTMC devices found?	15
Usage with Python & PyVisa.....	16
Scanning for connected devices	18
Some more pyvisa Hints	20
Stripping of read termination.....	20
Automatic write Termination addition.....	20
Timeout handling.....	20

Xyphro - Kai Gossner 19th October 2025
In case of comments, contact xyphro@gmail.com

Introduction

This document provides guidance on utilizing the UsbGpib adapter with Linux, specifically in conjunction with Python.

While the UsbGpib Adapter is no different to other UsbTmc instruments, because it exposes a standard UsbTmc interface, these instructions are mainly intended for first time users of Instrument control to give a kick-start in using it and setting it up.

Various approaches are available, as the USB GPIB Adapter operates via USBTMC. Users may directly employ the USBTMC Linux driver and interface through C or other programming languages, use pyvisa-py or pyvisa-usbtc (which leverage a more direct access to the USB devices), or adopt the standard practice of engaging a Visa provider.

The installation process detailed herein uses R&S Visa, which is controlled from Python using Pyvisa. R&S Visa offers a standardized interface compatible with PyVisa, LabVIEW, and other software packages. Visa providers are also available from other companies like NI or Keysight, but R&S Visa is the Visa provider which requires the least amount of hard disk space (~35MB). Most other visa providers install Gigabytes of files on your computer, which is not required nor helpful.

The primary objective of such a Visa is to deliver a unified API for managing multiple types of interfaces - regardless of their physical connection (USB, different Ethernet protocols, Serial port) - and it is widely used as a standard tool used for automating test and measurement setups.

For demonstration purposes, Ubuntu 24.04.3 LTS serves as the platform in this guide. A fresh installation was made to create this guide, to ensure that all dependencies are captured.






SW setup

Installation of R&S Visa


Please visit https://www.rohde-schwarz.com/us/applications/r-s-visa-application-note_56280-148812.html

In case this links does not work in future, use a google search for "R&S Visa Download" to find the most recent correct link.

Download the RS Visa Setup for Ubuntu:


 RS VISA Setup for Mac By downloading R&S®VISA you are agreeing to be bound by the Terms and Conditions for Royalty Free Software	Application Note File	5.12.11	03-Apr-2024	100 MiB
 RS VISA Setup for Raspbian By downloading R&S®VISA you are agreeing to be bound by the Terms and Conditions for Royalty Free Software	Application Note File	5.12.9	01-Sep-2023	3 MiB
 RS VISA Setup for CentOS By downloading R&S®VISA you are agreeing to be bound by the Terms and Conditions for Royalty Free Software	Application Note File	5.12.9	28-Jun-2023	3 MiB
 RS VISA Setup for openSUSE By downloading R&S®VISA you are agreeing to be bound by the Terms and Conditions for Royalty Free Software	Application Note File	5.12.9	28-Jun-2023	3 MiB
 RS VISA Setup for Ubuntu By downloading R&S®VISA you are agreeing to be bound by the Terms and Conditions for Royalty Free Software	Application Note File	5.12.9	28-Jun-2023	3 MiB

Once downloaded, the file will be in your Downloads folder:




To proceed, double-click the file or install it via the command line.

A warning may appear, indicating that the file originates from a third party (R&S):



Select "Install" and confirm your intention to proceed. When prompted, enter your user password to grant SUDO privileges.

Upon completion of the installation, three new applications will become available:



Although not strictly required, RsVisa Tester is a graphical user interface application that facilitates instrument scanning and enables basic read/write tests, making it particularly valuable during initial system bring-up.

Note, that the release notes of R&S Visa contain also installation instructions – also for Windows, Mac OSX or other Linux flavors; Even Raspberry PI with Raspbian.


We will continue with the hardware setup in the next section.

Installation of PyVisa

PyVisa is necessary for interfacing with the instrument using Python. PyVisa will interface to the visa provider which then interfaces to the instrument.

This guide utilizes Python 3. Please ensure that both Python 3 and pip3 are installed prior to proceeding.

To ensure that execute from terminal: `sudo apt-get install python3 python3-pip python3-pyvisa`

A terminal window with a dark purple background and light green text. The prompt is 'xyphro@xyphro-VirtualBox:~\$'. The command entered is 'sudo apt-get install python3-pyvisa'. The output shows the package lists being read, the dependency tree being built, and state information being read. It then lists additional packages to be installed (python3-pyvisa-py, python3-usb) and suggested packages (libvisa0). Finally, it lists the new packages to be installed (python3-pyvisa, python3-pyvisa-py, python3-usb) and shows the upgrade status: 0 upgraded, 3 newly installed, 0 to remove, and 82 not upgraded.

```
xyphro@xyphro-VirtualBox:~$ sudo apt-get install python3-pyvisa
Reading package lists... Done
Building dependency tree... Done
Reading state information... Done
The following additional packages will be installed:
  python3-pyvisa-py python3-usb
Suggested packages:
  libvisa0
The following NEW packages will be installed:
  python3-pyvisa python3-pyvisa-py python3-usb
0 upgraded, 3 newly installed, 0 to remove and 82 not upgraded.
```

(The actual output will vary depending on which packages are already installed on your machine)

All above steps are one-time actions.

Let's proceed to the HW setup.

HW Setup

In this example I use a HP3457A Multimeter with GPIB:



Initial one-time steps:

- Connect it to power but keep it powered off.
 - Connect the USB GPIB converter to the GPIB port and the USB cable to your PC.
- Note: The PC won't detect a USB device yet as the instrument is powered off. This is a feature to prevent inaccessible devices are showing up on your PC.



- Power on the GPIB device.
- Note: The PC does now detect a new USB Device.

It is not necessary to follow a specific sequence for connecting or disconnecting devices, powering them up or down, and similar actions. However, the following order is provided to present the instructions in a logical, step-by-step manner. As indicated in the procedure, the adapter will only be detected by a PC when the instrument is powered on and GPIB functionality is enabled.

The adapter monitors USB connections to identify whether any devices with GPIB enabled are present.

Regardless of the specific GPIB address, it systematically scans all available addresses on its GPIB port. Upon detecting a responsive GPIB address, the adapter enumerates the device on the USB interface, thereby enabling access from the PC.

First interaction with the instrument

Confirming that the USB device is detected

The following steps are not mandatory but can be beneficial for debugging. They verify that the USBTMC device is recognized by the PC and that the instrument is properly powered and configured for GPIB communication.

Execute from terminal: `lsusb`

```
xyphro@xyphro-VirtualBox:~$ lsusb
Bus 001 Device 001: ID 1d6b:0001 Linux Foundation 1.1 root hub
Bus 001 Device 002: ID 80ee:0021 VirtualBox USB Tablet
Bus 001 Device 003: ID 03eb:2065 Atmel Corp. LUFA Test and Measurement Demo Application
Bus 002 Device 001: ID 1d6b:0002 Linux Foundation 2.0 root hub
```

The output above displays an entry with VID:PID 03eb:2065, identifying the USB GPIB Adapter.

This device will appear when all connections are established according to the HW-Setup section, GPIB functionality is enabled on the instrument, and the instrument is powered on.

Further information can be seen when executing: `lsusb -v`

```
Bus 001 Device 003: ID 03eb:2065 Atmel Corp. LUFA Test and Measurement Demo Application
Couldn't open device, some information will be missing
Device Descriptor:
  bLength                18
  bDescriptorType         1
  bcdUSB                  1.10
  bDeviceClass             0 [unknown]
  bDeviceSubClass          0 [unknown]
  bDeviceProtocol          0
  bMaxPacketSize0          8
  idVendor                0x03eb Atmel Corp.
  idProduct               0x2065 LUFA Test and Measurement Demo Application
  bcdDevice               0.01
  iManufacturer           1 XyphroLabs
  iProduct                2 GpibUsb
  iSerial                 3 GPIB_22_85032303739351D00141
  bNumConfigurations      1
Configuration Descriptor:
  bLength                9
```

Note: This "message" is normal and does appear on other UsbTmc devices too.

The green highlighted area displays the USBGPIB adapter serial number. In this instance, it indicates the detected GPIB address, which is 22. The subsequent numbers represent a unique


serial number, allowing users to distinguish between multiple USB GPIB adapters when identical GPIB addresses are assigned to different instruments.

Please note that the GPIB adapter may also display the instrument name instead of the GPIB address if the AutoID feature is enabled.


Testing if the instrument is detected by R&S Visa

You can close open terminal windows, etc.

Start the RsVisa Tester application:




After starting the below GUI window is shown:



While the interface may appear complex to first-time users, it contains numerous standard features that help in initial instrument communication and debug. To verify detection of the GPIB instrument, select "Find Resource" in the upper left corner.

A new window opens:




Select the green-highlighted button to initiate an instrument scan.

If no USBTMC devices are detected, please refer to the section titled "Why are no USBTMC devices found?".


This situation typically occurs during initial installations, and the section provides guidance on how to enable device availability on a persistent basis.

If the operation is successful, a USB device should be visible as shown. Please note that the previously displayed serial number forms part of the resource string.



Select the found devices and click OK.

Now you are back to the main window. The resource field should show the previously detected device:




Now you can click on the connect button to connect to the instrument.

The log lines should show a successful connection:

48	92 ms	VI_SUCCESS	viOpen(sesn= 1, rsrc= USB0::0x03EB::...
49	6 µs	VI_SUCCESS	viSetAttribute(vi= 14,type= ...

The right side of the main GUI window now gives the ability to interact with the instrument:



If your instrument supports the *IDN? query, you may enter "*IDN?" in the designated field and click Query; the instrument's response will then be displayed.


However, many legacy instruments do not support this command and may also lack EOI termination during read operations, which can lead to unsatisfactory timeout errors giving the impression that communication is not possible

For this reason, I have prepared this manual for the older HP3457A, which requires some additional configuration steps. These requirements are not unique to this USB GPIB Adapter - they apply to any GPIB adapter and are dictated by the instrument itself.

The equipment's GPIB programming manuals offer detailed guidance on configuring the appropriate settings, and additional support is available in the troubleshooting section if necessary. For certain legacy instruments, these settings may be challenging to locate and may require some degree of experimentation.

Point #1: The used HP3457A does default wise not terminate with EOI, but with \n. A query will result in a timeout.

To enable \n termination, go to the Attributes tab, select VI_ATTR_TERMCHAR_EN, enter 1 in New AttributeValue field and click on SetAttribute.




This is a non permanent setting, which is only valid for this currently active session.

The attribute VI_ATTR_TERMCHAR does select the actual termination character which is used to detect a termination of read GPIB messages. Default wise it is 0xa (10), which matches \n and is for most older HP instruments the correct settings.

Point #2: HP3457A does not support the *IDN? SCPI standard query. Older HP Equipment required "ID?\r\n" as query.

After adjusting the termination settings as stated in Point #1 you can try to execute the ID?\r\n query:



The green highlighted section shows the response and the log window shows no timeout, but VI_SUCCESS on the read and write instructions. This confirms that the communication is working.

As mentioned, most equipment does not require an adjustment of the read termination as they generate an EOI on read to terminate.

Note: Above items show the usage using RSVisaTester in an interactive way. If you just want to access your instrument using Python, this is typically not required, but it is a nice tool to debug your instrument. It can also create logs of your visa instructions which you execute using Python & PyVisa which is a powerful tool for debugging during development phase.

Why are no USBTMC devices found?

A "Find Resource scan" performed on RsVisaTester may yield no detected instruments. Furthermore, each time the scan button is pressed, the main GUI screen's log displays entries

indicating this outcome:

14	3 ms	VI_ERROR_RSRC_NFOUND	viFindRsrc(vi= 1, expr= ?*, findList= 0, retcnt= 0, ...
15	4 ms	VI_ERROR_RSRC_NFOUND	viFindRsrc(vi= 1, expr= ?*, findList= 0, retcnt= 0, ...
16	3 ms	VI_ERROR_RSRC_NFOUND	viFindRsrc(vi= 1, expr= ?*, findList= 0, retcnt= 0, ...

This message indicates that no instruments were detected. However, this situation may occur even when an instrument is properly connected, but permissions to usbtmc devices are not properly set up yet on the machine and access would only work with SUDO rights.

To set them correctly, add them to the usbtmc.rules file:

Execute: `sudo nano /etc/udev/rules.d/99-usbtmc.rules`

Add this line to the file:

```
SUBSYSTEM=="usb", ATTR{idVendor}=="03eb", ATTR{idProduct}=="2065",  
MODE="0666"
```

The 2 highlighted parts are the vendor ID and product ID, which were shown earlier in the lsusb output. For this USB GPIB adapter, the numbers will never change and always be the same. But in case you want to support other USB connected measurement instruments, adjust those numbers.

After adding this line and saving with <CTRL>+O and exiting, execute:

```
sudo udevadm control --reload-rules  
sudo udevadm trigger
```

or reboot the machine.

Also disconnect and reconnect the USB cable once after executing those lines.

Note: Instead of using the described line, you can also add this line:

```
KERNEL=="usbtmc[0..9]", SUBSYSTEM=="usbmisc", MODE="0666"
```

It will set usbtmc0 to usbtmc9 devices access rights, without requiring to specify a vendor and product ID, thus is a bit more generic.

Usage with Python & PyVisa

If Python and PyVISA have been installed as outlined in previous sections, you should now be able to communicate with your instrument. This section provides suggestions and example code snippets to assist with this process.

While step-by-step interactive usage within the Python console is demonstrated here, the provided snippets may also be integrated into standalone Python scripts as needed.

Scanning for connected devices

Start python3:

```
xyphro@xyphro-VirtualBox:~$ python3
Python 3.12.3 (main, Aug 14 2025, 17:47:21) [GCC 13.3.0] on linux
Type "help", "copyright", "credits" or "license" for more information.
```

Import pyvisa:

```
>>> import pyvisa
>>> 
```

And create a visa resource manager (which allows listing resources at a later stage):

```
>>> rm = pyvisa.ResourceManager()
>>> 
```

When you enter: `rm.list_resources()` all available visa resources will be shown:

```
>>> rm.list_resources()
('ASRL/dev/ttyS31::INSTR', 'ASRL/dev/ttyS30::INSTR', 'ASRL/dev/ttyS29::INSTR', 'ASRL/dev/ttyS28::INSTR', 'ASRL/dev/ttyS27::INSTR', 'ASRL/dev/ttyS26::INSTR', 'ASRL/dev/ttyS25::INSTR', 'ASRL/dev/ttyS24::INSTR', 'ASRL/dev/ttyS23::INSTR', 'ASRL/dev/ttyS22::INSTR', 'ASRL/dev/ttyS21::INSTR', 'ASRL/dev/ttyS20::INSTR', 'ASRL/dev/ttyS19::INSTR', 'ASRL/dev/ttyS18::INSTR', 'ASRL/dev/ttyS17::INSTR', 'ASRL/dev/ttyS16::INSTR', 'ASRL/dev/ttyS15::INSTR', 'ASRL/dev/ttyS14::INSTR', 'ASRL/dev/ttyS13::INSTR', 'ASRL/dev/ttyS12::INSTR', 'ASRL/dev/ttyS11::INSTR', 'ASRL/dev/ttyS10::INSTR', 'ASRL/dev/ttyS9::INSTR', 'ASRL/dev/ttyS8::INSTR', 'ASRL/dev/ttyS7::INSTR', 'ASRL/dev/ttyS6::INSTR', 'ASRL/dev/ttyS5::INSTR', 'ASRL/dev/ttyS4::INSTR', 'ASRL/dev/ttyS3::INSTR', 'ASRL/dev/ttyS2::INSTR', 'ASRL/dev/ttyS1::INSTR', 'ASRL/dev/ttyS0::INSTR', 'USB0::1003::8293::GPIB_22_85032303739351D00141::0::INSTR')
```

Our instrument is also part of this, as highlighted with the green box.

To open the instrument, type:

```
>>> inst = rm.open_resource('USB0::1003::8293::GPIB_22_85032303739351D00141::0::INSTR')
>>> 
```

Your actual instrument name (called the resource string) will vary of course.

In case a “[Errno 16] Resource busy” error is shown, it is possible, that it is e.g. still open in RsVisaTester – ensure that you are disconnect there. It is also possible, that the device is still open from previous python experiments – in that case execute `inst.close()` before or create a new python session.

If successful, no error is shown, and we can interact with the instrument using the “inst” object.

If your instrument is SCPI compliant, you can now simply execute `instr.query('*IDN')` to read the instrument name or better identification string over GPIB.

As the used HP3457a is not SCPI compliant and requires specific termination settings, we apply those settings now first during the next steps.

To set \n termination for that instrument, execute those 2 lines:

```
>>> inst.set_visa_attribute(pyvisa.constants.VI_ATTR_TERMCHAR, ord('\n') )
<StatusCode.success: 0>
>>> inst.set_visa_attribute(pyvisa.constants.VI_ATTR_TERMCHAR_EN, True)
<StatusCode.success: 0>
>>> 
```

Afterwards the read termination will be set to \n and queries or reads will be successful from this instrument:

```
>>> inst.query('ID?\r\n')
'HP3457A\r\n'
```

You can now write to the instrument using commands like `inst.write()` for ascii strings, `inst.write_binary_values` to send raw binary data, among other variants.

To read data use `inst.read()` for ascii strings, `inst.read_binary_values()`, etc.

The previously used `inst.query` command is just a combination of write followed by read. It will write the provided text strings (here 'ID?') to the instrument, followed by a read and returns the information.

To disconnect from the instrument, execute `inst.close()`.

Some more pyvisa Hints

Stripping of read termination

An instrument which returns read termination, like the one used here will contain in reads always read termination. As visible in previous example, the ID? Query answer is terminated with `\r\n`.

Pyvisa can “strip away” the `\r\n` termination automatically, if you specify a `read_termination` to it:

```
>>> inst.read_termination = '\r\n'
>>> inst.query('ID?\r\n')
'HP3457A'
```

This `read_termination` setting is just a string handling within pyvisa, it does not set any read termination of the GPIB adapter itself. So it is no substitution of the previous `TERMCHAR` and `TERMCHAR_EN` attributes which were used before.

Automatic write Termination addition

As you saw in the examples, the query command string included `\r\n` as termination. PyVisa can add this automatically, if you specify a write termination, making the code a bit easier to read.

```
>>> inst.write_termination = '\r\n'
>>> inst.query('ID?')
'HP3457A'
```

Again, `write_termination` is just a string based handling. Before the string gets sent to the device, the set `write_termination` string is added to the message internally.

Timeout handling

Some instruments can take a longer time to return a response. This can e.g. happen for large data transfers. There are sophisticated ways using service requests that an instrument can signal with interrupts it's readiness, but sometimes it is easier to just increase the timeout.

The property `timeout` will return the current timeout in units of milliseconds.

```
>>> inst.timeout  
2000
```

2000 means 2000ms which is 2s. This is the default timeout value.

You can change it by assigning another number to it. To increase the timeout to 5s, execute:

```
>>> inst.timeout = 5000
```

Don't forget: Timeouts can indicate actual errors like an unsupported command; It can also indicate wrong write or read termination but also be just simply the instrument taking a longer time to process a command.