

1. General Information concerning the Type

This report is aligned to the applicable forms in OIML R 76-2:2007. It is generated by the laboratory application and is not issued or endorsed by OIML.

Report / application reference	NAWI-2026-8HRD1DD6OZU7
Internal report ID	cmtr5a4s60003rmh201i9aj0r
Type / model designation	AVX-500
Manufacturer	Yadnesh
Manufacturer country	India
Manufacturer registration code	101
Serial number	Not recorded
Instrument type / category	SINGLE_RANGE
Accuracy class	III
Minimum capacity Min	0.2 kg
Maximum capacity Max	30 kg
Verification scale interval e	0.01 kg
Actual scale interval d	0.005 kg
Load-receptor support points	4
Additive tare device	Not fitted
Automatic zero / tracking	Not fitted
Initial zero-setting device	Not fitted
Initial zero-setting range	0
Fine display device	Not fitted
Verification context	INITIAL_VERIFICATION
Ruleset version	OIML R76-1:2006
Evaluation period	2026-09-06T12:15:58.990Z to 2026-09-07T10:40:24.578Z
Report date	2026-09-07T11:15:55.447Z
Observer / technician	Development Lab Technician
Session reference	cmtprzimm00088hrd1dd6ozu7
Environment / equipment / applicant	Not recorded

2. Summary of Type Evaluation

This generated report contains the OIML R 76-2 sections supported by the currently implemented test modules.

Test	R76-2	Result	Remarks
Weighing Performance	1	PASS	R76-1 (2006) clause 3.6
Eccentricity using weights	3.1	PASS	R76-1 (2006) clause 3.6.2, A.4.7
Repeatability	5	PASS	R76-1 (2006) clause 3.6.1, A.4.10 (3rd paragraph)

Overall persisted technical result

PASS

Human approval status

APPROVED

Unsupported R76-2 test forms are not represented as completed or passed in this report.

1. Weighing Performance

Persisted test outcome: **PASS**

Weighing performance test PASSED with max absolute error 0.00505.

L kg	I kg	Delta L kg	E kg	Ec kg	MPE kg	Result
12.34567	12.34573	0.00001	0.00505	0.00505	0.01	PASS

Persisted R76 reference	R76-1 (2006) clause 3.6
Maximum absolute error	0.00505 kg
MPE	0.01 kg

3.1. Eccentricity using weights

Persisted test outcome: **PASS**

Eccentric loading test overall outcome: PASS. Max absolute error: 0 kg.

Position	L kg	I kg	Delta L	E kg	Ec kg	Result
position-1	10	10	0.005	0	0	PASS

Persisted R76 reference

Maximum absolute error

MPE

R76-1 (2006) clause 3.6.2, A.4.7

0 kg

0.01 kg

5. Repeatability

Persisted test outcome: **PASS**

Test load: 10 kg, Weighings: 3. I_{max} = 10.005, I_{min} = 9.995. R = 0.01. MPE is +/-0.01. R <= MPE -> PASS.

Run	Load kg	Indication I kg	Per-run error
1	10	10	Not recorded
2	10	10.005	Not recorded
3	10	9.995	Not recorded

E _{max} - E _{min} / repeatability range	0.01
Applicable MPE	0.01 kg
Result	PASS
Persisted R76 reference	R76-1 (2006) clause 3.6.1, A.4.10 (3rd paragraph)
Maximum absolute error	0.01 kg
MPE	0.01 kg

Supplement A. Software-generated Compliance Trace

This supplementary traceability section is generated by the application from persisted engine output. It is not an original OIML R 76-2 form.

Weighing Performance

Outcome: PASS. Weighing performance test PASSED with max absolute error 0.00505.

Reference: Table 6 - 3.5.1 - Maximum permissible errors on verification - OIML R 76-1:2006 (E)

Reference: Annex A.4.4.1 - Weighing test - general procedure - OIML R 76-1:2006 (E)

Reference: Annex A.4.4.3 - Error calculation with rounding correction (digital indication) - OIML R 76-1:2006 (E)

Indication before rounding (P): $P = I + 0.5e - \Delta L$ | $P = 12.34573 + 0.5(0.01) - 0.00001 = 12.35072$ kg

Error before zero correction (E): $E = P - L$ | $E = 12.35072 - 12.34567 = 0.00505$ kg

Corrected indication error (Ec): $E_c = E - E_0$ | $E_c = 0.00505 - 0 = 0.00505$ kg

Decision: $|0.00505 \text{ kg}| \leq 0.01 \text{ kg}$

Repeatability

Outcome: PASS. Test load: 10 kg, Weighings: 3. $I_{\max} = 10.005$, $I_{\min} = 9.995$. $R = 0.01$. MPE is ± 0.01 . $R \leq \text{MPE} \rightarrow \text{PASS}$.

Reference: 3.6.1 - Repeatability - maximum permissible variation of indication - OIML R 76-1:2006 (E)

Reference: Annex A.4.10 - Repeatability test procedure (3rd paragraph: $R \leq \text{MPE}$ criterion) - OIML R 76-1:2006 (E)

Reference: Table 6 - 3.5.1 - Maximum permissible errors on verification (MPE reference for repeatability) - OIML R 76-1:2006 (E)

Maximum indication ($I_{\max}$): $I_{\max} = \max(I_1, I_2, \dots, I_n)$ | $I_{\max} = \max(10, 10.005, 9.995) = 10.005$ kg

Minimum indication ($I_{\min}$): $I_{\min} = \min(I_1, I_2, \dots, I_n)$ | $I_{\min} = \min(10, 10.005, 9.995) = 9.995$ kg

Repeatability range (R): $R = I_{\max} - I_{\min}$ | $R = 10.005 - 9.995 = 0.01$ kg

Decision: $0.01 \text{ kg} \leq 0.01 \text{ kg}$

Eccentricity using weights

Outcome: PASS. Eccentric loading test overall outcome: PASS. Max absolute error: 0 kg.

Reference: 3.6.2 - Eccentric loading - general requirement - OIML R 76-1:2006 (E)

Reference: 3.6.2.1 - Eccentric loading - test load rule: $\frac{1}{3} \times (\text{Max} + \text{maximum additive tare effect})$, applied at each quarter segment - OIML R 76-1:2006 (E)

Reference: Annex A.4.7 - Eccentric loading test procedure; retest condition when zero is not determined before each loading - OIML R 76-1:2006 (E)

Reference: Annex A.4.7.1 - Eccentric loading procedure for load receptors with not more than four support points - OIML R 76-1:2006 (E)

Reference: Table 6 - 3.5.1 - Maximum permissible errors on verification (MPE reference for eccentric loading) - OIML R 76-1:2006 (E)

Indication before rounding (P): $P = I + 0.5e - \Delta L$ | $P = 10 + 0.5(0.01) - 0.005 = 10$ kg

Error before zero correction (E): $E = P - L$ | $E = 10 - 10 = 0$ kg

Corrected indication error (Ec): $E_c = E - E_0$ | $E_c = 0 - 0 = 0$ kg

Decision: $|0 \text{ kg}| \leq 0.01 \text{ kg}$

Supplement B. Review and Approval

Final session status	APPROVED
Technician	Development Lab Technician
Reviewing Officer	Development Reviewing Officer
Approving Officer	Development Approving Officer

Workflow action	Officer	Date	Comment
SUBMIT_FOR_REVIEW	Development Lab Technician	2026-09-07	No comment recorded
REVIEW_APPROVE	Development Reviewing Officer	2026-09-07	Development technical review completed.
FINAL_APPROVE	Development Approving Officer	2026-09-07	Development final approval completed.

Human workflow status and technical R76 compliance outcomes are distinct. Approval confirms the persisted evaluation record; it does not recalculate the regulatory result.

Public report verification



Report reference: NAWI-2026-8HRD1DD6OZU7

<http://localhost:3100/verify/cmtr5a4s60004rmh2sj04cu1j>

Scan to verify report