

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-E8BAB1AC

Verification Request ID	32
Application ID	TW-0B4ADEC0
Inspection ID	30
Instrument ID	13

INSTRUMENT DETAILS

Instrument Type	12
Manufacturer	21
Model	11
Serial Number	22
Capacity	30
Location	pune

INSPECTION RESULTS

Standard Weight	2.000 kg
Machine Reading	2.000 kg
Calculated Error	0.000 kg
Permissible Error	0.050 kg
Result	PASS
Remarks	No remarks

INSPECTOR

Name: Test Inspector

Email: inspector@test.com

INSPECTION LOCATION

Latitude: 20.911690273041295

Longitude: 74.76282897036427

GPS Accuracy: 183.00 m

VERIFICATION RESULT: PASS

BLOCKCHAIN VERIFICATION

Status: NOT_CONFIGURED

Certificate Hash: 43ab951a1ac8bc842787b484511698eef62efdb8f7127eb06d4605ea1eee78d8

SCAN TO VERIFY



<http://192.168.29.48:5173/verify/TW-CERT-E8BAB1AC>

Certificate issued: 2026-09-29 21:12:44.836177+05:30
Generated by TrueWeight Verification Platform