

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-3091FB20

Verification Request ID	25
Application ID	TW-E529158A
Inspection ID	23
Instrument ID	13

INSTRUMENT DETAILS

Instrument Type	12
Manufacturer	21
Model	11
Serial Number	22
Capacity	30
Location	pune

INSPECTION RESULTS

Standard Weight	1.000 kg
Machine Reading	1.000 kg
Calculated Error	0.000 kg
Permissible Error	0.050 kg
Result	PASS
Remarks	No remarks

INSPECTOR

Name: Test Inspector

Email: inspector@test.com

VERIFICATION RESULT: PASS

BLOCKCHAIN VERIFICATION

Status: NOT_CONFIGURED

Certificate Hash: c03606a8f0b0a8485da9ca77e3c8624dd10147605c06438f2e3b59f7c070b40d

SCAN TO VERIFY



<http://192.168.29.48:5173/verify/TW-CERT-3091FB20>

Certificate issued: 2026-09-29 19:47:51.223270+05:30
Generated by TrueWeight Verification Platform