

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-35676C4F

Verification Request ID	26
Application ID	TW-37D4F089
Inspection ID	24
Instrument ID	13

INSTRUMENT DETAILS

Instrument Type	12
Manufacturer	21
Model	11
Serial Number	22
Capacity	30
Location	pune

INSPECTION RESULTS

Standard Weight	2.000 kg
Machine Reading	2.000 kg
Calculated Error	0.000 kg
Permissible Error	0.050 kg
Result	PASS
Remarks	No remarks

INSPECTOR

Name: Test Inspector
Email: inspector@test.com

INSPECTION LOCATION

Latitude: 20.911728658149357
Longitude: 74.76283450589322
GPS Accuracy: 187.00 m

VERIFICATION RESULT: PASS

BLOCKCHAIN VERIFICATION

Status: NOT_CONFIGURED
Certificate Hash: 56a1dbce49e17b9f29b1a17a8d85523a7753157a79df9cf1c859da51a5e51a13

SCAN TO VERIFY



<http://192.168.29.48:5173/verify/TW-CERT-35676C4F>

Certificate issued: 2026-09-29 20:00:29.788411+05:30
Generated by TrueWeight Verification Platform