

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-45EE3C5D

Verification Request ID: 23

Inspection ID: 21

Instrument ID: 13

Instrument Type: 12

Manufacturer: 21

Model: 11

Serial Number: 22

Capacity: 30

Location: pune

Standard Weight: 30.0 kg

Machine Reading: 30.02 kg

Calculated Error: 0.019999999999999574 kg

Permissible Error: 0.05 kg

Result: PASS

Remarks:

No remarks

VERIFICATION RESULT: PASS



SCAN TO VERIFY

<http://10.23.93.59:5173/verify/>