

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-59EED846

Verification Request ID	24
Application ID	TW-3F006CF0
Inspection ID	22
Instrument ID	13

INSTRUMENT DETAILS

Instrument Type	12
Manufacturer	21
Model	11
Serial Number	22
Capacity	30
Location	pune

INSPECTION RESULTS

Standard Weight	1.000 kg
Machine Reading	1.000 kg
Calculated Error	0.000 kg
Permissible Error	0.050 kg
Result	PASS
Remarks	No remarks

INSPECTOR

Name: Test Inspector
Email: inspector@test.com

VERIFICATION RESULT: PASS

BLOCKCHAIN VERIFICATION

Status: NOT_CONFIGURED
Certificate Hash: 58b83c1ed1e3fa5f46205a5f0159584e9d0c60bb0d43e9060dbabdf0e658e9b3

SCAN TO VERIFY



<http://192.168.29.48:5173/verify/TW-CERT-59EED846>

Certificate issued: 2026-09-29 19:29:45.007536+05:30
Generated by TrueWeight Verification Platform