

TRUEWEIGHT

VERIFICATION CERTIFICATE

Certificate Number: TW-CERT-E570787E

Verification Request ID	27
Application ID	TW-A9D60B21
Inspection ID	25
Instrument ID	13

INSTRUMENT DETAILS

Instrument Type	12
Manufacturer	21
Model	11
Serial Number	22
Capacity	30
Location	pune

INSPECTION RESULTS

Standard Weight	2.000 kg
Machine Reading	2.000 kg
Calculated Error	0.000 kg
Permissible Error	0.050 kg
Result	PASS
Remarks	No remarks

INSPECTOR

Name: Test Inspector
Email: inspector@test.com

INSPECTION LOCATION

Latitude: 20.911728658149357
Longitude: 74.76283450589322
GPS Accuracy: 187.00 m

VERIFICATION RESULT: PASS

BLOCKCHAIN VERIFICATION

Status: NOT_CONFIGURED
Certificate Hash: 3e800c928832b2d49c2207e1829f13e706aad8137c7afa4050b6a94ae7ae47f

SCAN TO VERIFY



<http://192.168.29.48:5173/verify/TW-CERT-E570787E>

Certificate issued: 2026-09-29 20:12:31.689038+05:30
Generated by TrueWeight Verification Platform