
Published code (mnl_ccepAge, ccep_detect_n1peak_ECoG.m)



- Noise: SD of the average between −2 and −0.1 s = 18 µV, raised to a FLOOR of 50 µV → SD used = 50 µV. Reference level = MEDIAN.
- Trough: only dips at least 20 µV deep count (× = too shallow; first sample after 9 ms skipped; > 3000 µV refused as saturation).
- Decision: N1 if the trough is deeper than k × SD: 3.4 × 50 = 170 µV (never less than 170 µV).
- Same average, here: trough 132 µV deep < 170 µV → N1 NO.