Lakeside Community Clinic Equipment Maintenance Memo

To: Clinical operations. From: Biomedical engineering. Date: 2025-08-20.

Autoclave unit AC-2 failed its weekly spore test on 2025-08-14. All sterilization has moved to unit AC-1 until AC-2 passes three consecutive tests.

Imaging

The linear ultrasound probe L12 was replaced under a service agreement with Medira Systems. The replacement cost was 7,450 dollars.

Audit

The next biomedical equipment audit is scheduled for 2026-02-09. The responsible technician is Priya Raman.