

# Feature-Conditioned Cascaded Video Diffusion Models for Precise Echocardiogram Synthesis

Hadrien Reynaud<sup>1,2(⊠)</sup>, Mengyun Qiao<sup>2,3</sup>, Mischa Dombrowski<sup>4</sup>, Thomas Day<sup>5,6</sup>, Reza Razavi<sup>5,6</sup>, Alberto Gomez<sup>5,7</sup>, Paul Leeson<sup>7,8</sup>, and Bernhard Kainz<sup>2,4</sup>

- <sup>1</sup> UKRI CDT in AI for Healthcare, Imperial College London, London, UK hadrien.reynaud19@imperial.ac.uk
- Department of Computing, Imperial College London, London, UK
   Department of Brain Sciences and DSI, Imperial College London, London, UK
   Friedrich-Alexander University Erlangen-Nürnberg, Erlangen, Germany
   School of BMEIS, King's College London, London, UK
  - <sup>6</sup> Guy's and St Thomas' NHS Foundation Trust, London, UK
    <sup>7</sup> Ultromics Ltd., Oxford, UK
- <sup>8</sup> John Radcliffe Hospital, Cardiovascular Clinical Research Facility, Oxford, UK

**Abstract.** Image synthesis is expected to provide value for the translation of machine learning methods into clinical practice. Fundamental problems like model robustness, domain transfer, causal modelling, and operator training become approachable through synthetic data. Especially, heavily operator-dependant modalities like Ultrasound imaging require robust frameworks for image and video generation. So far, video generation has only been possible by providing input data that is as rich as the output data, e.g., image sequence plus conditioning in  $\rightarrow$  video out. However, clinical documentation is usually scarce and only single images are reported and stored, thus retrospective patient-specific analysis or the generation of rich training data becomes impossible with current approaches. In this paper, we extend elucidated diffusion models for video modelling to generate plausible video sequences from single images and arbitrary conditioning with clinical parameters. We explore this idea within the context of echocardiograms by looking into the variation of the Left Ventricle Ejection Fraction, the most essential clinical metric gained from these examinations. We use the publicly available EchoNet-Dynamic dataset for all our experiments. Our image to sequence approach achieves an  $\mathbb{R}^2$  score of 93%, which is 38 points higher than recently proposed sequence to sequence generation methods. Code and weights are available at https://github.com/HReynaud/EchoDiffusion.

**Keywords:** Generative · Diffusion · Video · Cardiac · Ultrasound

**Supplementary Information** The online version contains supplementary material available at https://doi.org/10.1007/978-3-031-43999-5 14.

<sup>©</sup> The Author(s), under exclusive license to Springer Nature Switzerland AG 2023 H. Greenspan et al. (Eds.): MICCAI 2023, LNCS 14229, pp. 142–152, 2023. https://doi.org/10.1007/978-3-031-43999-5\_14

#### 1 Introduction

Ultrasound (US) is widely used in clinical practice because of its availability, realtime imaging capabilities, lack of side effects for the patient and flexibility. US is a dynamic modality that heavily relies on operator experience and on-the-fly interpretation, which requires many years of training and/or Machine Learning (ML) support that can handle image sequences. However, clinical reporting is conventionally done via single, selected images that rarely suffice for clinical audit or as training data for ML. Simulating US from anatomical information, e.q. Computed Tomography (CT) [28], Magnetic Resonance Imaging (MRI) [25] or computational phantoms [11,27], has been considered as a possible avenue to provide more US data for both operator and ML training. However, simulations are usually very computationally expensive due to complex scattering, reflection and refraction of sound waves at tissue boundaries during image generation. Therefore, the image quality of US simulations has not yet met the necessary quality to support tasks such as cross-modality registration, multi-modal learning, and robust decision support for image analysis during US examinations. More recently, generative deep learning methods have been proposed to address this issue. While early approaches show promising results, they either focus on generating individual images [16] or require video input data and further conditioning to provide useful results [17,21]. Research in the field of imageconditioned video generation is very scarce [33] and, to the best of our knowledge, we are the first to apply it to medical imaging.

Contribution: In this paper, we propose a new method for video diffusion [7,30] based on the Elucidated Diffusion Model (EDM) [13] that allows to synthesise plausible video data from single frames together with precise conditioning on interpretable clinical parameters, e.g., Left Ventricular Ejection Fraction (LVEF) in echocardiography. This is the first time diffusion models have been extended for US image and video synthesis. Our contributions are three-fold: (1) We show that discarding the conventional text-embeddings [7,20,23,24,30] to control the reverse diffusion process is desirable for medical use cases where very specific elements must be precisely controlled; (2) We quantitatively improve upon existing methods [21] for counterfactual modelling, e.g., when doctors try to answer questions like "how would the scan of this patient look like if we would change a given clinical parameter?"; (3) We show that fine-grained control of the conditioning leads to precise data generation with specific properties and outperforms the state-of-the-art when using such data, for example, for the estimation of LVEF in patients that are not commonly represented in training databases.

Related Work: Video Generation has been a research area within computer vision for many years now. Prior works can be organized in three categories: (1) pixel-level autoregressive models [2,4,12], (2) latent-level autoregressive model coupled with generators or up-samplers [1,14] and (3) latent-variable

transformer-based models with up-samplers [5,37]. Diffusion models have shown reasonable performance on low temporal and spatial resolutions [10] as well as on longer samples with high definition image quality [7,30] conditioned on text inputs. Recently, [38] combined an autoregressive pixel-level model with a diffusion-based pipeline that predicts a correction of the frame, while [3] presents an autoregressive latent diffusion model.

Ultrasound simulation has been attempted with three major approaches: (1) physics-based simulators [11,28], (2) cross-modality registration-based methods [15] and (3) deep-learning based methods, usually conditioned on US, MRI or CT image priors [25,34,35] to condition the anatomy of the generated US images. Cine-ultrasound has also attracted some interest. [17] presents a motion-transfer-based method for pelvic US video generation, while [21] proposes a causal model for generating echocardiograms conditioned on arbitrary LVEF.

**LVEF** is a major metric in the assessment of cardiac function and diagnosis of cardiomyopathy. The EchoNet-dynamic dataset [19] is used as the go-to benchmark for LVEF-regression methods. Various works [18,22] have attempted to improve on [19] but the most reproducible method remains the use of an R2+1D model trained over fixed-length videos. The R2+1D\_18 trained for this work achieves an  $R^2$  score of 0.81 on samples of 64 frames spanning 2 s.

# 2 Method

Diffusion probabilistic models [8,31,32] are the most recent family of generative models. In this work, we follow the definition of the EDM from [13]. Let  $q(\mathbf{x})$  represent the real distribution of our data, with a standard deviation of  $\sigma_q$ . A family of distributions  $p(\mathbf{x}; \sigma)$  can be obtained by adding i.i.d Gaussian noise with a standard deviation of  $\sigma$  to the data. When  $\sigma_{\text{max}} \gg \sigma_q$ , the distribution  $p(\mathbf{x}; \sigma_{\text{max}})$  is essentially the same as pure Gaussian noise. The core idea of diffusion models is to sample a pure noise data point  $\mathbf{x}_0 \sim \mathcal{N}(\mathbf{0}, \sigma_{\text{max}}^2 \mathbf{I})$  and then progressively remove the noise, generating images  $\mathbf{x}_i$  with standard deviation  $\sigma_i$  such that  $\sigma_{\text{max}} = \sigma_0 > \sigma_1 > ... > \sigma_N = 0$ , and  $\mathbf{x}_i \sim p(\mathbf{x}; \sigma_i)$ . The final image  $\mathbf{x}_N$  produced by this process is thus distributed according to  $q(\mathbf{x})$ , the true distribution of the data. To perform the reverse diffusion process, we define a denoising function  $D(\mathbf{x}, \sigma)$  trained to minimize the  $L_2$  denoising error for all samples drawn from q for every  $\sigma$  such that:

$$\mathcal{L} = \mathbb{E}_{\boldsymbol{y} \sim q} \mathbb{E}_{\boldsymbol{n} \sim \mathcal{N}(\boldsymbol{0}, \sigma^2 \mathbf{I})} || D(\boldsymbol{y} + \boldsymbol{n}; \sigma) - \boldsymbol{y} ||_2^2$$
 (1)

where  $\boldsymbol{y}$  is a training data point and  $\boldsymbol{n}$  is noise. By following the definition of ordinary differential equations (ODE) we can continuously increase or decrease the noise level of our data point by moving it forward or backward in the diffusion process, respectively. To define the ODE we need a schedule  $\sigma(t)$  that sets the noise level given the time step t, which we set to  $\sigma(t) = t$ . The probability flow ODE's characteristic property is that moving a sample  $\boldsymbol{x}_a \sim p(\boldsymbol{x}_a; \sigma(t_a))$  from the diffusion step  $t_a$  to  $t_b$  with  $t_a > t_b$  or  $t_a < t_b$  should


result in a sample  $x_b \sim p(x_b; \sigma(t_b))$  and this requirement is satisfied by setting  $dx = -\dot{\sigma}(t)\sigma(t)\nabla_{\tau}\log p(x;\sigma(t))dt$  where  $\dot{\sigma}$  denotes the time derivative and  $\nabla_x \log p(x;\sigma)$  is the score function. From the score function, we can thus write  $\nabla_x \log p(x;\sigma) = (D(x;\sigma) - x)/\sigma^2$  in the case of our denoising function, such that the score function isolates the noise from the signal x and can either amplify it or diminish it depending on the direction we take in the diffusion process. We define  $D(x;\sigma)$  to transform a neural network F, which can be trained inside D by following the loss described in Eq. (1). The EDM also defines a list of four important preconditionings which are defined as  $c_{\text{skip}}(\sigma) = (\sigma_q^2)/(\sigma^2 + \sigma_q^2)$ ,  $c_{\text{out}}(\sigma) = \sigma * \sigma_q * 1/(\sigma_q^2 * \sigma^2)^{0.5}, c_{\text{in}}(\sigma) = 1/(\sigma_q^2 * \sigma^2)^{0.5} \text{ and } c_{\text{noise}}(\sigma) = \log(\sigma_t)/4$ where  $\sigma_q$  is the standard deviation of the real data distribution. In this paper, we focus on generating temporally coherent and realistic-looking echocardiograms. We start by generating a low resolution, low-frame rate video  $v_0$  from noise and condition on arbitrary clinical parameters and an anatomy instead of the commonly used text-prompt embeddings [7,30]. Then, the video is used as conditioning for the following diffusion model, which generates a temporally and/or spatially upsampled video  $v_1$  resembling  $v_0$ , following the Cascaded Diffusion Model (CDM) [9] idea. Compared to image diffusion models, the major change to the Unet-based architecture is to add time-aware layers, through attention, at various levels as well as 3D convolutions (see Fig. 1 and Appendix Fig. 1). For the purpose of this research, we extend [7] to handle our own set of conditioning inputs, which are a single image  $I_c$  and a scalar value  $\lambda_c$ , while following the EDM setup, which we apply to video generation. We formally define the denoising models in the cascade as  $D_{\theta_s}$  where s defines the rank (stage) of the model in the cascade, and where  $D_{\theta_0}$  is the base model. The base model is defined as:

$$D_{\theta_0}(\boldsymbol{x}; \sigma, \boldsymbol{I}_c, \lambda_c) = c_{\text{skip}}(\sigma)\boldsymbol{x} + c_{\text{out}}(\sigma)F_{\theta_0}(c_{\text{in}}(\sigma)\boldsymbol{x}; c_{\text{noise}}(\sigma), \boldsymbol{I}_c, \lambda_c)),$$

where  $F_{\theta_0}$  is the neural network transformed by  $D_{\theta_0}$  and  $D_{\theta_0}$  outputs  $v_0$ . For all subsequent models in the cascade, the conditioning remains similar, but the models also receive the output from the preceding model, such that:

$$D_{\theta_s}(\boldsymbol{x}; \sigma, \boldsymbol{I}_c, \lambda_c, \boldsymbol{v}_{s-1}) = c_{\text{skip}}(\sigma)\boldsymbol{x} + c_{\text{out}}(\sigma)F_{\theta_s}(c_{\text{in}}(\sigma)\boldsymbol{x}; c_{\text{noise}}(\sigma), \boldsymbol{I}_c, \lambda_c, \boldsymbol{v}_{s-1})).$$

This holds  $\forall s > 0$  and inputs  $I_c$ ,  $v_{s-1}$  are rescaled to the spatial and temporal resolutions expected by the neural network  $F_{\theta_s}$  as a pre-processing. We apply the robustness trick from [9], *i.e.*, we add a small amount of noise to real videos  $v_{s-1}$ 



**Fig. 1.** Summarized view of our Model. Inputs (blue): a noised sample  $x_i$ , a diffusion step  $t_i$ , one anatomy image  $I_c$ , and one LVEF  $\lambda_c$ . Output (red): a slightly denoised version of  $x_i$  named  $x_{i+1}$ . See Appendix Fig. 1 for more details. (Color figure online)

during training, when using them as conditioning, in order to mitigate domain gaps with the generated samples  $v_{s-1}$  during inference.

Sampling from the EDM is done through a stochastic sampling method. We start by sampling a noise sample  $x_0 \sim \mathcal{N}(\mathbf{0}, t_0^2 \mathbf{I})$ , where t comes from our previously defined  $\sigma(t_i) = t_i$  and sets the noise level. We follow [13] and set constants  $S_{\text{noise}} = 1.003, S_{t_{\text{min}}} = 0.05, S_{t_{\text{max}}} = 50$  and one constant  $S_{\text{churn}}$  dependent on the model. These are used to compute  $\gamma_i(t_i) = \min(S_{\text{churn}}/N, \sqrt{2} - 1)$  $\forall t_i \in [S_{t_{\min}}, S_{t_{\max}}]$  and 0 otherwise, where N is the number of sampling steps. Then  $\forall i \in \{0,...,N-1\}$ , we sample  $\epsilon_i \sim \mathcal{N}(\mathbf{0}, S_{\text{noise}}\mathbf{I})$  and compute a slightly increased noise level  $\hat{t}_i = (\gamma_i(t_i) + 1)t_i$ , which is added to the previous sample  $\hat{x}_i = x_i + (\hat{t}_i^2 - t_i^2)^{0.5} \epsilon_i$ . We then execute the denoising model  $D_{\theta}$  on that sample and compute the local slope  $d_i = (\hat{x}_i - D_{\theta}(\hat{x}_i; \hat{t}_i))/\hat{t}_i$  which is used to predict the next sample  $x_{i+1} = \hat{x}_i + (\hat{t}_{i+1} - \hat{t}_i)d_i$ . At every step but the last (i.e:  $\forall i \neq N-1$ ), we apply a correction to  $x_{i+1}$  such that:  $d'_{i} = (x_{i+1} - D_{\theta}(x_{i+1}; t_{i+1}))/t_{i+1}$  and  $x_{i+1} = \hat{x}_{i} + (t_{i+1} - \hat{t}_{i})(d_{i} + d'_{i})/2$ . The correction step doubles the number of executions of the model, and thus the sampling time per step, compared to DDPM [8] or DDIM [32]. The whole sampling process is repeated sequentially for all models in the cascaded EDM. Models are conditioned on the previous output video  $v_{s-1}$  inputted at each step of the sampling process, with the frame conditioning  $I_c$  as well as the scalar value  $\lambda_c$ .

Conditioning: Our diffusion models are conditioned on two components. First, an anatomy, which is represented by a randomly sampled frame  $I_c$ . It defines the patient's anatomy, but also all the information regarding the visual style and quality of the target video. These parameters cannot be explicitly disentangled, and we therefore limit ourselves to this approach. Second, we condition the model on clinical parameters  $\lambda_c$ . This is done by discarding the text-encoders that are used in [10,30] and directly inputting normalized clinical parameters into the conditional inputs of the Unets. By doing so, we give the model fine-grained control over the generated videos, which we evaluate using task-specific metrics.

Parameters: As video diffusion models are still in their early stage, there is no consensus on which are the best methods to train them. In our case, we define, depending on our experiment, 1-, 2- or 4-stages CDMs. We also experiment with various schedulers and parametrizations of the model. [26,32] show relatively fast sampling techniques which work fine for image sampling. However, in the case of video, we reach larger sampling times as we sample 64 frames at once. We therefore settled for the EDM [13], which presents a method to sample from the model in much fewer steps, largely reducing sampling times. We do not observe any particular speed-up in training and would argue, from our experience, that the v-parametrization [32] converges faster. We experimentally find our models to behave well with parameters close to those suggested in [13].

## 3 Experiments

Data: We use the EchoNet-Dynamic [19] dataset, a publicly available dataset that consists of 10,030 4-chamber cardiac ultrasound sequences, with a spatial resolution of  $112 \times 112$  pixels. Videos range from 0.7 to 20.0 s long, with frame rates between 18 and 138 frames per second (fps). Each video has 3 channels, although most of them are greyscale. We keep the original data split of EchoNet-Dynamic which has 7465 training, 1288 validation and 1277 testing videos. We only train on the training data, and validate on the validation data. In terms of labels, each video comes with an LVEF score  $\lambda \in [0,100]$ , estimated by a trained clinician. At every step of our training process, we pull a batch of videos, which are resampled to 32 fps. For each video, we retrieve its corresponding ground truth LVEF as well as a random frame. After that, the video is truncated or padded to 64 frames, in order to last 2 s, which is enough to cover any human heartbeat. The randomly sampled frame is sampled from the same original video as the 64-frames sample, but may not be contained in those 64 frames, as it may come from before or after that sub-sample.

Architectural Variants: We define three sets of models, and present them in details in Table 1 of the Appendix. We call the models X-Stage Cascaded Models (XSCM) and present the models' parameters at every stage. Every CDM starts with a Base diffusion model that is conditioned on the LVEF and one conditional frame. The subsequent models perform either temporal super resolution (TSR), spatial super resolution (SSR) or temporal and spatial super resolution (TSSR). TSR, SSR and TSSR models receive the same conditioning inputs as the Base model, along with the output of the previous-stage model. Note that [7] does not mention TSSR models and [30] states that extending an SSR model to perform simultaneous temporal and spatial up-sampling is too challenging.

**Training:** All our models are trained from scratch on individual cluster nodes, each with  $8 \times \text{NVIDIA}$  A100. We use a per-GPU batch size of 4 to 8, resulting in batches of 32 to 64 elements after gradient accumulation. The distributed training is handled by the accelerate library from HuggingFace. We did not see any speed-up or memory usage reduction when enabling mixed precision and thus used full precision. As pointed out by [9] all models in a CDM can be trained in parallel which significantly speeds up experimentation. We empirically find that training with a learning rate up to  $5*10^{-4}$  is stable and reaches good image quality. We use an Exponential Moving Average (EMA) copy of our model to smooth out the training. We train all our models' stages for 48h, *i.e.*, the 2SCM and 4SCM CDMs are proportionally more costly to train than the 1SCM. As noted by [7,30] training on images and videos improves the overall image quality. As our dataset only consists of videos, we simply deactivate the time attention layers in the Unet with a 25% chance during training, for all models.

**Results:** We evaluate our models' video synthesis capabilities on two objectives: LVEF accuracy ( $R^2$ , MAE, RMSE) and image quality (SSIM, LPIPS, FID, FVD). We formulate the task as counterfactual modelling, where we set

(1) a random conditioning frame as confounder, (2) the ground-truth LVEF as a factual conditioning, and (3) a random LVEF in the physiologically plausible range from 15% to 85% as counterfactual conditioning. For each ground truth video, we sample three random starting noise samples and conditioning frames. We use the LVEF regression model to create a feedback loop, following what [21] did, even though their model was run  $100\times$  per sample instead of  $3\times$ . For each ground truth video, we keep the sample with the best LVEF accuracy to compute all our scores over 1288 videos for each model.

The results in Table 1 show that increasing the frame rate improves model fidelity to the given LVEF, while adding more models to the cascade decreases image quality. This is due to a distribution gap between true low-resolution samples and sequentially generated samples during inference. This issue is partially addressed by adding noise to real low-resolution samples during training, but the 1SCM model with only one stage still achieves better image quality metrics. However, the 2SCM and 4SCM models perform equally well on LVEF metrics and outperform the 1SCM model thanks to their higher temporal resolution that precisely captures key frames of the heartbeat. The TSSR model, used in the 2SCM, yields the best compromise between image quality, LVEF accuracy, and sampling times, and is compared to previous literature.


We outperform previous work for LVEF regression: counterfactual video generation improves with our method by a large margin of 38 points for the  $R^2$  score as shown in Table 2. The similarity between our factual and counterfactual results show that our time-agnostic confounding factor (*i.e.* an image instead of a video) prevents entanglement, as opposed to the approach taken in [21]. Our method does not score as high for SSIM as global image similarity metric, which is expected because of the stochasticity of the speckle noise. In [21] this was mitigated by their data-rich confounder. Our results also match other video diffusion models [3,7,10,30] as structure is excellent, while texture tends to be more noisy as shown in Fig. 2.

Table 1. Metrics for all CDMs. The *Gen.* task is the counterfactual generation comparable to [21], the *Rec.* task is the factual reconstruction task. *Frames* is the number of frames generated by the model, always spanning 2 s.  $^{\ddagger}$ Videos are temporally upsampled to 64 frames for metric computation. *S. time* is the sampling time for one video on an RTX A5000.  $R^2$ , MAE and RMSE are computed between the conditional LVEF  $\lambda_c$  and the regressed LVEF using the model described in Sect. 1. SSIM, LPIPS, FID and FVD are used to quantify the image quality. LPIPS is computed with VGG [29], FID [6] and FVD [36] with I3D. FID and FVD are computed over padded frames of  $128 \times 128$  pixels.

Model	Task	Res.	Frames	S. time	$R^2\uparrow$	$MAE \downarrow$	$RMSE\downarrow$	SSIM↑	LPIPS↓	FID↓	$\mathrm{FVD}\!\!\downarrow$
1SCM	Gen.	112	16 <sup>‡</sup>	$62\mathrm{s}^{\ddagger}$	0.64	9.65	12.2	0.53	0.21	12.3	60.5
2SCM	Gen.	112	64	$146\mathrm{s}$	0.89	4.81	6.69	0.53	0.24	31.7	141
4SCM	Gen.	112	64	$279\mathrm{s}$	0.93	3.77	5.26	0.48	0.25	24.6	230
1SCM	Rec.	112	16 <sup>‡</sup>	62 s ‡	0.76	4.51	6.07	0.53	0.21	13.6	89.7
2SCM	Rec.	112	64	$146\mathrm{s}$	0.93	2.22	3.35	0.54	0.24	31.4	147
4SCM	Rec.	112	64	$279\mathrm{s}$	0.90	2.42	3.87	0.48	0.25	24.0	228

**Table 2.** Comparison of our 2SCM model to previous work. We try to reconstruct a ground truth video or to generate a new one. Our model is conditioned on a single frame and an LVEF, while [21] conditions on the entire video and an LVEF. In both cases the LVEF is either the ground truth LVEF (*Rec.*) or a randomly sampled LVEF (*Gen.*).

Method	Conditioning	Task	S. time	$R^2 \uparrow$	$\mathrm{MAE}\downarrow$	$\mathrm{RMSE}\downarrow$	SSIM ↑
Dartagnan [21]	Video+EF	Gen.	$\sim 1 \mathrm{s}$	0.51	15.7	18.4	0.79
2SCM	Image+EF	Gen.	$146\mathrm{s}$	0.89	4.81	6.69	0.53
Dartagnan [21]	Video+EF	Rec.	~1 s	0.87	2.79	4.45	0.82
2SCM	Image+EF	Rec.	$146\mathrm{s}$	0.93	2.22	3.35	0.54



**Fig. 2.** Top: Ground truth frames with 29.3% LVEF. Middle: Generated factual frames, with estimated 27.9% LVEF. Bottom: Generated counterfactual frames, with estimated 64.0% LVEF. (Counter-)Factual frames are generated with the 1SCM, conditioned on the ground-truth anatomy.

Qualitative Study: We asked three trained clinicians (Consultant cardiologist > 10 years experience, Senior trainee in cardiology > 5 years experience, Chief cardiac physiologist > 15 years experience) to classify 100 samples, each, as real or fake. Experts were not given feedback on their performance during the evaluation process and were not shown fake samples beforehand. All samples were generated with the 1SCM model or were true samples from the EchoNet-Dynamic dataset, resampled to 32fps and 2s. The samples were picked by alphabetical order from the validation set. Among the 300 samples evaluated, 130 (43.33%) were real videos, 89 (29.67%) were factual generated videos, and 81 (27.0%) were counterfactual generated videos. The average expert accuracy was 54.33%, with an inter-expert agreement of 50.0%. More precisely, experts detected real samples with an accuracy of 63.85%, 50.56% for factual samples and 43.21% for the counterfactual samples. The average time taken to evaluate each sample was 16.2 s. We believe that these numbers show the video quality that our model reaches, and can put in perspective the SSIM scores from Table 1.

**Downstream Task:** We train our LVEF regression model on rebalanced datasets and resampled datasets. We rebalance the datasets by using our 4SCM

model to generate samples for LVEF values that have insufficient data. The resampled datasets are smaller datasets randomly sampled from the real training set. We show that, in small data regimes, using generated data to rebalance the dataset improves the overall performance. Training on 790 real data samples yields an  $R^2$  score of 56% while the rebalanced datasets with 790 samples,  $\sim 50\%$  of which are real, reaches a better 59% on a balanced validation set. This observation is mitigated when more data is available. See Appendix Table 2 for all our results.

**Discussion:** Generating echocardiograms is a challenging task that differs from traditional computer vision due to the noisy nature of US images and videos. However, restricting the training domain simplifies certain aspects, such as not requiring a long series of CDMs to reach the target resolution of  $112 \times 112$  pixels and limiting samples to 2s, which covers any human heartbeat. The limited pixel-intensity space of the data also allows for models with fewer parameters. In the future, we plan to explore other organs and views within the US domain, with different clinical conditionings and segmentation maps.

### 4 Conclusion

Our application of EDMs to US video generation achieves state-of-the-art performance on a counterfactual generation task, a data augmentation task, and a qualitative study by experts. This significant advancement provides a valuable solution for downstream tasks that could benefit from representative foundation models for medical imaging and precise medical video generation.

Acknowledgements. This work was supported by Ultromics Ltd., the UKRI Centre for Doctoral Training in Artificial Intelligence for Healthcare (EP/S023283/1) and HPC resources provided by the Erlangen National High Performance Computing Center (NHR@FAU) of the Friedrich-Alexander-Universität Erlangen-Nürnberg (FAU) under the NHR project b143dc. NHR funding is provided by federal and Bavarian state authorities. NHR@FAU hardware is partially funded by the German Research Foundation (DFG) - 440719683. We thank Phil Wang (https://github.com/lucidrains) for his open source implementation of [7]. Support was also received from the ERC - project MIA-NORMAL 101083647 and DFG KA 5801/2-1, INST 90/1351-1.

# References

- Babaeizadeh, M., Finn, C., Erhan, D., Campbell, R.H., Levine, S.: Stochastic variational video prediction. arXiv:1710.11252 (2018)
- Babaeizadeh, M., Saffar, M.T., Nair, S., Levine, S., Finn, C., Erhan, D.: FitVid: overfitting in pixel-level video prediction. arXiv:2106.13195 (2021)
- 3. Esser, P., Chiu, J., Atighehchian, P., Granskog, J., Germanidis, A.: Structure and content-guided video synthesis with diffusion models. arXiv:2302.03011 (2023)
- Finn, C., Goodfellow, I., Levine, S.: Unsupervised learning for physical interaction through video prediction. In: Advances in Neural Information Processing Systems, vol. 29 (2016)

- Gupta, A., Tian, S., Zhang, Y., Wu, J., Martín-Martín, R., Fei-Fei, L.: MaskViT: masked visual pre-training for video prediction. arXiv:2206.11894 (2022)
- Heusel, M., Ramsauer, H., Unterthiner, T., Nessler, B., Hochreiter, S.: GANs trained by a two time-scale update rule converge to a local Nash equilibrium. arXiv:1706.08500 (2018)
- Ho, J., et al.: Imagen video: high definition video generation with diffusion models (2022). arXiv:2210.02303
- Ho, J., Jain, A., Abbeel, P.: Denoising diffusion probabilistic models. In: Advances in Neural Information Processing Systems, vol. 33, pp. 6840–6851 (2020)
- Ho, J., Saharia, C., Chan, W., Fleet, D.J., Norouzi, M., Salimans, T.: Cascaded diffusion models for high fidelity image generation. J. Mach. Learn. Res. 23, 1–33 (2022)
- Ho, J., Salimans, T., Gritsenko, A., Chan, W., Norouzi, M., Fleet, D.J.: Video diffusion models (2022). arXiv:2204.03458
- Jensen, J.: Simulation of advanced ultrasound systems using Field II. In: 2004 2nd IEEE International Symposium on Biomedical Imaging: Nano to Macro (IEEE Cat No. 04EX821), pp. 636–639, vol. 1 (2004)
- 12. Kalchbrenner, N., et al.: Video pixel networks. In: ICML, pp. 1771–1779 (2017)
- 13. Karras, T., Aittala, M., Aila, T., Laine, S.: Elucidating the design space of diffusion-based generative models. arXiv:2206.00364 (2022)
- Kumar, M., et al.: VideoFlow: a conditional flow-based model for stochastic video generation. arXiv:1903.01434 (2020)
- Ledesma-Carbayo, M., et al.: Spatio-temporal nonrigid registration for ultrasound cardiac motion estimation. IEEE TMI 24, 1113–1126 (2005)
- Liang, J., et al.: Sketch guided and progressive growing GAN for realistic and editable ultrasound image synthesis. Med. Image Anal. 79, 102461 (2022)
- Liang, J., et al.: Weakly-supervised high-fidelity ultrasound video synthesis with feature decoupling. In: Wang, L., Dou, Q., Fletcher, P.T., Speidel, S., Li, S. (eds.) MICCAI 2022. LNCS, vol. 13434, pp. 310–319. Springer, Cham (2022). https://doi.org/10.1007/978-3-031-16440-8 30
- Mokhtari, M., Tsang, T., Abolmaesumi, P., Liao, R.: EchoGNN: explainable ejection fraction estimation with graph neural networks. In: Wang, L., Dou, Q., Fletcher, P.T., Speidel, S., Li, S. (eds.) MICCAI 2022. LNCS, vol. 13434, pp. 360–369. Springer, Cham (2022). https://doi.org/10.1007/978-3-031-16440-8
- Ouyang, D., et al.: Video-based AI for beat-to-beat assessment of cardiac function. Nature 580, 252–256 (2020)
- 20. Ramesh, A., et al.: Zero-shot text-to-image generation. arXiv:2102.12092 (2021)
- Reynaud, H., et al.: D'ARTAGNAN: counterfactual video generation. In: Wang, L., Dou, Q., Fletcher, P.T., Speidel, S., Li, S. (eds.) MICCAI 2022. LNCS, vol. 13438, pp. 599–609. Springer, Cham (2022). https://doi.org/10.1007/978-3-031-16452-1 57
- Reynaud, H., Vlontzos, A., Hou, B., Beqiri, A., Leeson, P., Kainz, B.: Ultrasound video transformers for cardiac ejection fraction estimation. In: de Bruijne, M., et al. (eds.) MICCAI 2021. LNCS, vol. 12906, pp. 495–505. Springer, Cham (2021). https://doi.org/10.1007/978-3-030-87231-1 48
- Rombach, R., Blattmann, A., Lorenz, D., Esser, P., Ommer, B.: High-resolution image synthesis with latent diffusion models. arXiv:2112.10752 (2022)
- Saharia, C., et al.: Photorealistic text-to-image diffusion models with deep language understanding. arXiv:2205.11487 (2022)

- Salehi, M., Ahmadi, S.-A., Prevost, R., Navab, N., Wein, W.: Patient-specific 3D ultrasound simulation based on convolutional ray-tracing and appearance optimization. In: Navab, N., Hornegger, J., Wells, W.M., Frangi, A.F. (eds.) MICCAI 2015. LNCS, vol. 9350, pp. 510–518. Springer, Cham (2015). https://doi.org/10.1007/978-3-319-24571-3 61
- Salimans, T., Ho, J.: Progressive distillation for fast sampling of diffusion models. arXiv:2202.00512 (2022)
- Segars, W.P., Sturgeon, G., Mendonca, S., Grimes, J., Tsui, B.M.W.: 4D XCAT phantom for multimodality imaging research. Med. Phys. 37, 4902–4915 (2010)
- Shams, R., Hartley, R., Navab, N.: Real-time simulation of medical ultrasound from CT images. In: Metaxas, D., Axel, L., Fichtinger, G., Székely, G. (eds.) MICCAI 2008. LNCS, vol. 5242, pp. 734–741. Springer, Heidelberg (2008). https://doi.org/ 10.1007/978-3-540-85990-1 88
- 29. Simonyan, K., Zisserman, A.: Very deep convolutional networks for large-scale image recognition. arXiv:1409.1556 (2015)
- 30. Singer, U., et al.: Make-a-video: text-to-video generation without text-video data. arXiv:2209.14792 (2022)
- 31. Sohl-Dickstein, J., Weiss, E.A., Maheswaranathan, N., Ganguli, S.: Deep unsupervised learning using nonequilibrium thermodynamics. arXiv:1503.03585 (2015)
- 32. Song, J., Meng, C., Ermon, S.: Denoising diffusion implicit models. arXiv:2010.02502 (2022)
- Song, Y., Zhu, J., Li, D., Wang, X., Qi, H.: Talking face generation by conditional recurrent adversarial network. arXiv:1804.04786 (2019)
- 34. Teng, L., Fu, Z., Yao, Y.: Interactive translation in echocardiography training system with enhanced cycle-GAN. IEEE Access 8, 106147–106156 (2020)
- 35. Tomar, D., Zhang, L., Portenier, T., Goksel, O.: Content-preserving unpaired translation from simulated to realistic ultrasound images. In: de Bruijne, M., et al. (eds.) MICCAI 2021. LNCS, vol. 12908, pp. 659–669. Springer, Cham (2021). https://doi.org/10.1007/978-3-030-87237-3 63
- 36. Unterthiner, T., Steenkiste, S.V., Kurach, K., Marinier, R., Michalski, M., Gelly, S.: FVD: a new metric for video generation. In: ICLR 2022 Workshop: Deep Generative Models for Highly Structured Data (2019)
- 37. Villegas, R., et al.: Phenaki: variable length video generation from open domain textual description. arXiv:2210.02399 (2022)
- 38. Yang, R., Srivastava, P., Mandt, S.: Diffusion probabilistic modeling for video generation. arXiv:2203.09481 (2022)