

GOVERNMENT APPROVED TEST CENTRE

(IND/GATC/KL/26/04)

Certificate of Verification

Interweighing Pvt. Ltd, 49/470 D1,3rd floor, Asian Tower,Vyttila ,Cochin, Kerala -682019, 682019, Ph: 8803333444

Certificate No: IND/GATC/KL/26/04/26/3636

Date of Verification: 22.08.2026

I hereby certify that I have this day verified and stamped/rejected the under mentioned **Non-Automatic Weighing Instrument of Accuracy Class III 150kg** belonging to **M/s-INTERWEIGHING PVT LTD**, Address- 49/470 D1 3RD FLOOR,ASIAN TOWER,VYTILTA,COCHIN, Kerala, Ernakulam, 682019, Ph:- 8590603061

Type of Instrument	Manufacturer / Model / Brand	Serial Number	Year of Manufacture	Accuracy Class (III)	Maximum Capacity (Max upto 150 kg)	Minimum Capacity (Min)	Verification Scale Interval (e)	Unit of Measurement: kg	Actual Scale Interval (d)	No. of Verification Intervals (n = Max / e)	Maximum Permissible Error (MPE)
Platform Scale	YESWEIGH	YZ01408	2026	III	50kg	100g	5g	kg	5g	10000	7.5g

Visual Examination (Pass / Fail)	Zero Setting / Zero Tracking Test (Pass / Fail)	Eccentricity Test (Pass / Fail)	Repeatability Test (Pass / Fail)	Accuracy / Weighing Performance Test (Pass / Fail)	Tare Device Test (Pass / Fail)	Overall Verification Result (Pass / Fail)	Ambient Temperature (°C)	Relative Humidity (%)	Supply Voltage (if electronic)	Verification Seal Affixed	Seal Identification No(s)	Software Identification / Version / Checksum (if applicable)
Pass	Pass	Pass	Pass	Pass	Pass	Pass	27	76	230 V AC	Yes	IND/GATC/KL/26/04/C26	Nil

Instrument conforms to OIML Recommendation / LM (Gen) Rules, 2011 (Yes/No)						Verified and stamped for use in commercial transactions (Yes / No)				Remarks (if any)	
Yes						Yes				Original verification by IWP	

Model Approval No(s) - IND/09/21/780

Verification Fee (Rs. + GST) - 0; Carriage / Conveyance, adjusting Charges, etc. (+ GST) - 0

Total Rs. 0 deposited vide Money Receipt No. 0, Dated 2026-08-22

Repaired by / Used by: Interweighing Pvt. Ltd

Next verification falls due on or before: 2027-08-21

Signature of the verifying officer

HARISH RAMANKUTTY

Name of Principal Officer

Note:

1. The above-mentioned weights and measures are verified and stamped:

- ☐ i. In the premises of GATC
☒ ii. In-situ at the place of User

2. In the case of rejected weights, measures, weighing and measuring instruments, the principal officer shall give a separate certificate of rejection mentioning the reasons of such rejection against each item.



Scan QR to verify



Photo 1



Photo 2



Photo 3



Photo 4



Photo 5