

VISVESVARAYA TECHNOLOGICAL UNIVERSITY

Jnana Sangama, Belgaum -590014, Karnataka State, India



A PROJECT REPORT ON

“G-care: An AI Healthcare Assistant”

BACHELOR OF ENGINEERING

IN

DEPARTMENT

OF

COMPUTER SCIENCE ENGINEERING [AI&ML]

Submitted by: -

Yeshrutha S

1EW23CI063

Abhay Vadagi

1EW23CI002



EAST WEST INSTITUTE OF TECHNOLOGY

Off Magadi Main Road, Bengaluru, Karnataka — 560 091

AY 2026-27

CERTIFICATE

This is to certify that the project report entitled G-care: An AI Healthcare Assistant submitted by Yeshrutha S and Abhay Vadagi bearing the USN 1EW23CI063 and 1EW23CI002 to the Visvesvaraya Technological University, Belagavi, in partial fulfillment for the award of the degree of B. Tech in CSE[AI&ML] is a *bona fide* record of project work carried out by her under my supervision. The contents of this report, in full or in parts, have not been submitted to any other Institution or University for the award of any degree.

Signature:

Name of the Guide: Asst. Prof. Sreehari M

Department: CSE[AI&ML]

Signature of HOD with seal

DECLARATION

I/We declare that this project report titled G-care: An AI Healthcare Assistant submitted in partial fulfillment of the degree of B. Tech in CSE[AI&ML] is a record of original work carried out by me under the supervision of Sreehari M, Assitant Professor, CSE[AI&ML], and has not formed the basis for the award of any other degree, in this or any other Institution or University. In keeping with the ethical practice in reporting scientific information, due acknowledgements have been made wherever the findings of others have been cited.

Signature of the student(s)

Name

Signature

1. Yeshrutha S

2. Abhay Vadagi



EAST WEST INSTITUTE OF TECHNOLOGY

Off Magadi Main Road, Bengaluru, Karnataka — 560 091

Department of CSE[AI&ML]

CERTIFICATE

Certified that the project work entitled G-care: An AI Healthcare Assistant was carried out by Mr./Ms. Yeshrutha S and Abhay Vadagi, USN 1EW23CI063 and 1EW23CI063, a bonafide student of Asst. Prof. Sreehari M in partial fulfilment for the award of Bachelor of Engineering / Bachelor of Technology in CSE[AI&ML] of the Visveswaraiah Technological University, Belgaum during the year 2026-27. It is certified that all corrections/suggestions indicated for the Internal Assessment have been incorporated in the Report deposited in the departmental library. The project report has been approved as it satisfies the academic requirements in respect of Project work prescribed for the said Degree.

Name & Signature of the Guide

Name Signature of the HOD

Signature of the Principal

External Viva

Sl. No. Name of the examiners

Signature with date

1

2

ABSTRACT

With the rising global geriatric population and the prevalence of nuclear family structures, ensuring the continuous health monitoring, safety, and independence of senior citizens has emerged as a critical societal and technological challenge. Elderly individuals living independently frequently face life-threatening risks due to unmonitored chronic conditions, missed medication schedules, undetected physical falls, and delayed emergency responses. Existing remote healthcare systems often suffer from high operational costs, steep learning curves, fragmented communication, and interfaces unsuitable for seniors.

To resolve these limitations, this project presents **G-Care (Guardian Care)**, an integrated, full-stack, and intelligent remote elderly care platform that bridges the communication gap between elderly patients, family guardians, and healthcare practitioners. The system incorporates a tri-tier, Role-Based Access Control (RBAC) architecture featuring a **Senior Wearable Interface / Simulator**, a **Caretaker & Guardian Dashboard**, and a **Specialized Clinical Doctor Portal**. Real-time physiological telemetry—including heart rate, peripheral oxygen saturation (SpO₂), blood pressure, body temperature, stress, and hydration—is monitored continuously and processed through real-time anomaly detection heuristics to immediately identify critical vitals deviations, physical falls, and geofence breaches.

The platform provides an automated medication and routine scheduling engine with interactive adherence check-offs, instant SOS emergency dispatch protocols, and senior-centric accessibility features including a multilingual voice assistant supporting regional Indian languages (English, Hindi, Kannada, and Tamil). Developed using React 18, TypeScript, Tailwind CSS, Zustand, and Node.js/Express with PostgreSQL and JWT authentication, G-Care provides a scalable, secure, and cost-effective solution for proactive geriatric telecare. Experimental evaluations demonstrate responsive vitals synchronization, low alert dispatch latency, and significant reduction in caregiver dependency.

Keywords: *Elderly Care Monitoring, Telehealth, Internet of Medical Things (IoMT), Vitals Anomaly Detection, Role-Based Access Control, Medication Adherence, Fall Detection, Multilingual Voice Assistant.*

ACKNOWLEDGEMENT

With Great pride we would like to convey our gratitude and appreciation to our alma-mater **“East West Institute of Technology”** for giving us the required platform for the fulfillment of the Project on “G-care: An AI Healthcare Assistant” as per the V.T.U requirements, for the major Project.

We express our sincere thanks to **Dr. Nagaraj G Cholli**, Principal of **East West Institute of Technology**, for providing us with excellent infrastructure to complete the Project.

We express wholehearted gratitude to **Dr. Venkatesha G** who is the respectable **HOD of the Department of CSE[AI&ML] and CSE[IoT, CyS and BCT]**. We wish to acknowledge her help in making my task easy by providing us with his valuable help and encouragement.

It is our pleasure to thank my guide **Assistant Prof. Sreehari M** and the project co-ordinator **Prof. Shubha C**, EWIT for their guidance, encouragement and valuable suggestion from the beginning of the Project till the completion without which would not have been accomplished. We are greatly indebted to them.

And last but not the least, we would be very pleased to express our heartfelt thanks to the teaching and non-teaching staff of the Department CSE[AI&ML], EWIT for their motivation and support. We also thank all those who extended their support and co-operation while bringing out this internship.

YESHRUTHA S	1EW23CI063
ABHAY VADAGI	1EW23CI002

LIST OF FIGURES

Figure No	Name	Page No
1.1	High-Level Conceptual Architecture and Dataflow of the G-Care Ecosystem	5
3.1	Multi-Tier Layered System Architecture of G-Care	11
3.2	DFD Level 0 — Context Level Diagram	14
3.3	DFD Level 1 — Detailed Functional Data Flow Diagram	15
5.1	End-to-End Operational Methodology Flowchart of G-Care	23
5.2	Shared I2C Bus Interfacing on Prototype Breadboard	25
5.3	Entity-Relationship Diagram of G-Care Platform	26
5.4	Use Case Diagram for the Tri-Tier G-Care System	27
7.1	Multi-Tier Testing Pipeline for G-Care Platform	38

LIST OF TABLES

Table No	Name	Page No
2.1	Comparison Matrix of State-of-the-Art Solutions vs. Proposed G-Care Platform	8
4.1	Edge Hardware and Prototyping Components	16
4.2	Software Technology Stack	18
6.1	Breadboard Pin Interfacing and Wiring Connections	28
7.1	Comprehensive System Test Cases and Empirical Verification Results	39

CONTENTS

Chapter No.	Chapter Name	Page No.
CHAPTER 1	INTRODUCTION	1
	1.1	2
	1.2	3
CHAPTER 2	LITERATURE REVIEW	6
	2.1	6
	2.2	8
	2.3	9
CHAPTER 3	SYSTEM ARCHITECTURE	11
	3.1	11
	3.2	12
	3.3	14
CHAPTER 4	SYSTEM REQUIREMENTS	16
	4.1	16
	4.2	18
	4.3	20
CHAPTER 5	METHODOLOGY	22
	5.1	22
	5.2	24
CHAPTER 6	IMPLEMENTATION	28
	6.1	28
	6.2	32
	6.3	36
CHAPTER 7	TESTING	38
	7.1	38
	7.2	39
CHAPTER 8	CONCLUSION AND FUTURE SCOPE	42
	BIBILIOGRAPY	45
	APPENDIX I	46

Chapter 1

INTRODUCTION

With advancements in modern medicine and improved living standards, the global geriatric population is expanding at an unprecedented rate. According to the United Nations World Population Prospects, the proportion of individuals aged 60 and older is projected to double by 2050, reaching over 2.1 billion globally. In developing nations such as India, this demographic shift is intensified by rapid urbanization and the migration of young professionals for employment, accelerating the transition from traditional joint family structures to nuclear households. Consequently, an increasing number of senior citizens live alone or remain unassisted for extended periods during the day.

Aging is inherently accompanied by progressive physiological decline and an increased incidence of chronic degenerative illnesses, including cardiovascular disorders, hypertension, Type-2 diabetes, chronic obstructive pulmonary disease (COPD), and cognitive impairments. Managing these conditions requires continuous physiological observation, strict adherence to complex polypharmacy schedules, and rapid intervention during acute medical crises. However, traditional elderly care remains predominantly reactive and manual—relying on periodic physical visits, memory-based medication intake, and informal telephone check-ins. When elderly individuals live independently or their family guardians are geographically distant, subtle physiological deteriorations often go unnoticed until they escalate into acute medical emergencies, such as hypertensive crises, hypoxemia, or traumatic slip-and-fall injuries.

To address these compounding societal and technological challenges, G-Care (Guardian Remote Elder Care Dashboard & Clinic Portal) is developed as an integrated, full-stack, and proactive remote geriatric monitoring ecosystem. G-Care bridges the critical communication and visibility gap between elderly individuals, their family guardians/caregivers, and consulting healthcare professionals by unifying edge-level hardware sensing with cloud-assisted clinical workflows.

The system combines a physical Internet of Medical Things (IoMT) hardware prototype built on a breadboard with a secure multi-portal web telecare platform:

1. **Physical Hardware Sensing Node:** An edge embedded system assembled on a breadboard utilizing an ESP32-S3 microcontroller, an optical MAX30102 Photoplethysmography (PPG) biosensor (for real-time Heart Rate and SpO_2), a MAX30205 clinical-grade digital thermometer (for body temperature), an MPU6050 6-axis inertial motion tracking sensor (for fall and movement detection), and a local 0.96-inch OLED screen. The OLED display features a structured rectangular dashboard interface that continuously renders dynamic numeric vitals in real time, providing immediate feedback directly at the physical edge.

2. **Multi-Tier Web Platform:** A responsive, type-safe web ecosystem comprising a Guardian/Caretaker Telemetry Dashboard, a Specialized Clinical Doctor Portal, an interactive Wearable Watch Simulator (for software verification and development), and a Smart TV Ambient Display Mode.
3. **Automated Anomaly Detection & Closed-Loop Care:** Automated rule-based heuristics that continuously evaluate live physiological streams, detect critical threshold violations and falls, dispatch instantaneous audio-visual emergency alarms, track medication compliance with interactive check-offs, and deliver hands-free voice assistance in Indian regional languages (English, Hindi, Kannada, and Tamil).

1.1 Motivation

The conception and development of the G-Care platform is motivated by critical clinical, socioeconomic, and technological challenges:

1. **Vulnerability of Unmonitored Independent Seniors and the "Golden Hour":** A substantial portion of senior citizens live alone while managing chronic comorbidities. Acute medical events—such as myocardial infarctions, sudden arrhythmias, hypertensive surges, or physical falls—frequently incapacitate seniors, rendering them unable to operate a telephone or call for assistance. Clinical outcomes deteriorate drastically when help is delayed; the critical "golden hour" for trauma and cardiac intervention is often lost. There is an urgent need for edge devices that detect abnormal physiological drops or physical impacts autonomously and dispatch immediate alerts.
2. **Need for Physical Edge Hardware with Direct Local Feedback:** While purely software-based applications are common, elderly individuals require a tangible, physical edge device that monitors their vitals continuously without requiring them to navigate complex smartphone applications. Building a physical hardware breadboard prototype with a dedicated OLED display allows vitals (Heart Rate, SpO₂, Body Temperature) to be displayed in a clean, rectangular format in real time, giving seniors immediate reassurance regarding their health status.
3. **High Prevalence of Medication Non-Adherence:** Geriatric patients are frequently prescribed complex, multi-drug regimens with varying dosages, dietary requirements (before/after meals), and administration times. Normal age-related cognitive decline and memory lapses lead to missed doses, incorrect timing, or accidental double dosing. Medication non-adherence is a primary driver of avoidable hospital readmissions and therapeutic failures.
4. **Fragmented Communication Between Guardians and Doctors:** Traditional healthcare operates in disconnected silos. Family guardians often lack objective, continuous visibility into their loved one's daily health status, while consulting

physicians only observe isolated, single-point measurements during brief hospital visits. Doctors lack the longitudinal telemetry and verified medication compliance histories necessary to calibrate prescriptions accurately.

5. **Digital Literacy and Linguistic Barriers in Healthcare Interfaces:** The majority of contemporary health applications feature intricate touch navigation, dense text, small font sizes, and English-only menus. These design choices alienate elderly users, especially in multicultural regions. A viable elder-care system must incorporate high-contrast visuals, simplified typography, and natural language voice interactions in regional vernaculars.
6. **Economic Inaccessibility of Commercial Telecare Solutions:** Commercial medical-grade telecare devices and private assisted-living facilities are cost-prohibitive for the majority of middle- and lower-income families. G-Care demonstrates that an integrated, proactive telecare infrastructure can be constructed using cost-effective, readily available embedded sensors and open web standards without compromising clinical utility.

1.2 Objectives

The primary aim of this project is to conceptualize, design, and implement an end-to-end remote geriatric health monitoring and clinical management platform combining physical edge sensing with a multi-portal web dashboard.

The specific engineering and functional objectives are:

- **Physical Edge Hardware Prototyping:** To assemble and wire a functional physical hardware prototype on a solderless breadboard using an ESP32-S3 microcontroller, an optical MAX30102 PPG biosensor (measuring heart rate in BPM and oxygen saturation in $\text{SpO}_2\%$), a MAX30205 precision digital temperature sensor, and an MPU6050 6-axis accelerometer/gyroscope for motion and fall detection.
- **Embedded OLED Rectangular Dashboard Interface:** To design and program a dedicated firmware display driver on a 0.96-inch monochrome I²C OLED screen, structuring the layout into a high-contrast rectangular border with partitioned boxes displaying real-time, dynamic numeric streams for Heart Rate, SpO_2 , and Temperature, alongside system status alerts.
- **Real-Time Telemetry Tracking and Web Visualization:** To engineer an interactive web telemetry engine capable of continuously ingesting, charting, and tracking multi-dimensional physiological metrics (Heart Rate, SpO_2 , Blood Pressure, Body Temperature, Stress, and Hydration) using responsive Recharts time-series graphs and status indicators.

- **Automated Anomaly Detection & Emergency SOS Protocols:** To implement rule-based anomaly detection algorithms that evaluate continuous telemetry streams against clinical thresholds (categorizing states into *Normal*, *Warning*, and *Critical*), detect physical fall signatures from IMU acceleration spikes, and trigger immediate procedural audio alarms (Web Audio API), global visual alert banners, and emergency contact dispatches.
- **Closed-Loop Medication and Routine Scheduling:** To develop an interactive medication management module supporting custom frequencies (*once daily*, *twice daily*, *PRN*), time-triggered alerts, and an interactive checklist allowing caregivers and seniors to record administered doses for clinical compliance logging.
- **Specialized Clinical Doctor Portal:** To build an authenticated clinical workspace that enables physicians to evaluate longitudinal patient health trends, review medication compliance, upload and categorize diagnostic medical files (Complete Blood Counts, ECGs, scans, prescriptions), and author persistent clinical treatment notes.
- **Multilingual Senior-Centric Voice Assistance:** To integrate browser-native Web Speech synthesis and recognition supporting four languages (English, Hindi, Kannada, and Tamil) to speak medication reminders aloud and accept verbal confirmations, ensuring accessibility for seniors with visual or physical impairments.
- **Secure Role-Based Access Control (RBAC) & Dual Persistence:** To implement a secure RESTful API backend using Node.js, Express, Zod schema validation, and cryptographically signed JSON Web Tokens (JWT) with immediate token revocation upon logout, backed by a hybrid PostgreSQL relational database and a zero-configuration local JSON database fallback (data/db.json).

Architectural Description:

1. **Physical Hardware & Edge Layer:** The breadboard prototype acquires continuous biological signals through the MAX30102 pulse oximeter and MAX30205 temperature sensor, alongside spatial kinematic vectors from the MPU6050 IMU. The ESP32-S3 processes these signals locally, rendering real-time values inside a clean rectangular layout on the 0.96-inch OLED display while transmitting telemetry packets to the backend. The software watch simulator runs in parallel for development and testing.
2. **Backend Application & Anomaly Processing Tier:** Operates on an Express/Node.js pipeline. Incoming telemetry is validated through strict Zod contracts and evaluated by the anomaly detection engine. Validated metrics, medication regimens, and clinical files are persisted securely in the PostgreSQL / JSON database.
3. **Multi-Portal Presentation Tier:** Renders role-specific Single-Page Application (SPA) interfaces for family guardians (live tracking, medication compliance, geofencing, ambient TV mode), attending physicians (diagnostic documents, longitudinal trends,

clinical prescriptions), and emergency responders (immediate audible sirens and critical alert routing).




Figure 1.1: High-Level Conceptual Architecture and Dataflow of the G-Care Ecosystem

Chapter 2

LITERATURE REVIEW

The continuous expansion of the Internet of Medical Things (IoMT), wearable biosensors, and ubiquitous cloud computing has triggered a paradigm shift from traditional, episodic hospital care to decentralized, remote geriatric health monitoring. To construct a robust, clinically relevant, and user-centric platform such as **G-Care**, a comprehensive survey of contemporary literature was conducted across four foundational domains:

1. IoT-driven remote patient monitoring architectures.
2. Wearable biosensors for real-time physiological parameter estimation.
3. Inertial sensor-based fall detection algorithms.
4. Smart medication adherence systems and conversational voice interfaces for seniors.

2.1 Survey of Existing Literature and Methodologies

The literature is systematically reviewed and categorized according to the technical and functional dimensions of remote elderly healthcare.

2.1.1 IoT-Based Remote Patient Monitoring (RPM) Systems

- **Islam et al. [1]** presented a comprehensive survey on the application of the Internet of Things (IoT) in healthcare. The authors highlighted how edge-connected sensor nodes, wireless body area networks (WBANs), and centralized cloud repositories enable continuous tracking of vital parameters such as heart rate, peripheral oxygenation (SpO_2), and ambient temperature. However, the survey revealed that most existing RPM systems suffer from fragmented architectures—they lack unified multi-tier interfaces that integrate the elder, their family caregivers, and clinical doctors simultaneously. Furthermore, data security, end-to-end access control, and low-latency anomaly dispatch mechanisms were identified as major recurring bottlenecks.
- **Pantelopoulos and Bourbakis [2]** conducted an extensive review of wearable sensor-based systems designed for health prognosis. They categorized non-invasive physiological sensors and analyzed data fusion algorithms used to establish patient baselines. While the study demonstrated the diagnostic efficacy of wearable telemetry, it underscored a significant gap: the majority of systems operate strictly in data-collection mode, offering minimal automated decision support, negligible medication compliance tracking, and complex interfaces that hinder real-world adoption by older adults.
- **Majumder, Mondal, and Deen [3]** analyzed wearable sensors for home-based telecare, focusing on wireless communication protocols (BLE, Wi-Fi, ZigBee) and energy harvesting. The authors noted that while hardware miniaturization has advanced, existing commercial and academic prototypes fail to provide unified

emergency handling or automated alert verification, frequently leading to alert fatigue among caregivers.

2.1.2 Photoplethysmography (PPG) and Physiological Sensing

- **Allen [4]** investigated the physiological principles and clinical utility of Photoplethysmography (PPG) for estimating heart rate, arterial pulse wave velocity, and SpO₂. The research confirmed that optical biosensors (such as dual-wavelength red/infrared LED reflective sensors like the MAX30102) provide accurate, non-invasive cardiac telemetry suitable for low-power edge devices. Nevertheless, the study highlighted that raw PPG signals are highly susceptible to baseline wander and motion artifacts. It emphasized the necessity for software-level heuristics and threshold filters to validate physiological anomalies before triggering medical alerts.

2.1.3 Inertial Fall Detection and Emergency Alert Architectures

- **Igual, Medrano, and Plaza [5]** evaluated modern challenges in automated fall detection systems, categorizing methodologies into wearable sensor systems, ambient sensors, and computer vision techniques. They concluded that wearable tri-axial accelerometers and gyroscopes represent the most viable, privacy-preserving modality for geriatric monitoring. However, high rates of false positives during Activities of Daily Living (ADL)—such as sitting down abruptly or ascending stairs—remain a persistent challenge.
- **Bourke, O'Brien, and Lyons [6]** and **de Miguel et al. [7]** evaluated threshold-based algorithms applied to tri-axial accelerometer data (such as the MPU6050 sensor). Their experimental results proved that combining Upper Fall Thresholds (UFT) with secondary post-impact posture verification substantially reduces false alarms. The authors concluded that fall detection systems must be coupled with multi-channel dispatch frameworks (audible on-device alarms, cloud banners, and instant SOS caregiver notifications) to guarantee immediate medical response during the "golden hour."

2.1.4 Medication Adherence Systems and Senior Voice Interaction

- **Goldstein et al. [8]**, **Choudhry et al. [9]**, and **Dai et al. [10]** investigated the efficacy of electronic reminder devices and interactive prompts in improving medication compliance among elderly patients with chronic cardiovascular diseases. Randomized clinical trials proved that passive alarms (simple buzzers) achieve limited long-term adherence, whereas interactive reminders—requiring the user or caregiver to acknowledge dose administration—significantly improve therapeutic compliance and reduce preventable hospitalizations.
- **Thakkar et al. [11]** substantiated that bidirectional communication between patients and monitoring frameworks drastically enhances clinical outcomes in chronic disease management.
- **Nallam et al. [12]**, **Lifset et al. [13]**, and **Morrow, Lane, and Rogers [14]** explored the benefits and barriers of conversational AI and voice assistants among older adults.

Their findings demonstrated that senior citizens often abandon health applications due to small font sizes, complex touch menus, and cognitive overload. Conversational agents and voice-based reminders significantly lowered the barrier to digital adoption. However, a critical limitation in existing solutions is the lack of regional language localization (such as Indian vernacular languages), which excludes rural and non-English-speaking geriatric demographics.

2.2 Comparative Analysis of Existing Healthcare Monitoring Systems

A comparative synthesis of existing market solutions, research prototypes, and the proposed G-Care platform is summarized in Table 2.1.

Table 2.1: Comparison Matrix of State-of-the-Art Solutions vs. Proposed G-Care Platform

Feature / Capability	Commercial Smartwatches (<i>Apple Watch / Fitbit</i>)	Clinical RPM Hardware (e.g., <i>Philips Lifeline</i>)	Traditional Hospital Portals	Academic IoT Prototypes	Proposed G-Care Platform
Multi-Parametric Vitals (<i>HR, SpO₂, BP, Temp</i>)	Partial (<i>HR & SpO₂ only</i>)	Minimal (<i>Single-parameter</i>)	Retrospective (Manual entry)	Partial (<i>HR/Temp</i>)	Full Suite (<i>HR, SpO₂, BP, Temp, Stress, Hydration</i>)
Real-Time Anomaly Detection	Proprietary on-device alerts	Call-center dispatch only	Absent (Static records)	Basic single-threshold	Heuristic Multi-Level (<i>Normal, Warning, Critical</i>)
Fall Detection & GPS Tracking	High accuracy, closed ecosystem	Wearable pendant only	Absent	Accelerometer-only (high false positives)	6-Axis IMU Simulation + Geofencing + SOS
Tri-Tier RBAC Architecture	Disconnected (Single-user focus)	Caregiver-only	Doctor-only	Typically single-tier	Unified Tri-Tier (<i>Doctor, Caretaker, Guardian</i>)

Feature / Capability	Commercial Smartwatches (<i>Apple Watch / Fitbit</i>)	Clinical RPM Hardware (<i>e.g., Philips Lifeline</i>)	Traditional Hospital Portals	Academic IoT Prototypes	Proposed G-Care Platform
Interactive Medication Planner	Basic reminder notification	Absent	E-Prescription only (No adherence log)	Standalone alarm	Closed-Loop Scheduler (<i>Frequencies, logs, check-offs</i>)
Diagnostic Document Repository	Absent	Absent	Proprietary EMR/EHR	Absent	Built-in Doctor Workspace (<i>CBC, ECG, Scans, Notes</i>)
Senior Multilingual Voice Support	English / Major global languages	Call center (Manual voice)	Text-based only	Rare	Multilingual Voice Assistant (<i>EN, HI, KN, TA</i>)
Living Room / Ambient Display	Absent	Absent	Absent	Absent	Dedicated Smart TV Feed Simulator
Deployment Cost & Accessibility	High hardware cost (\$300–\$800)	Expensive monthly subscription	Enterprise licensing fees	Low, but lacks software maturity	Cost-Effective, Open-Web Architecture

2.3 Research Gaps and Problem Formulation

A critical analysis of the literature reveals several fundamental deficiencies in contemporary geriatric care systems:

1. **Siloed Architectures and Lack of Multi-Stakeholder Collaboration:** Most existing commercial and research platforms are designed either exclusively for the patient (wearable trackers) or exclusively for the clinician (hospital Electronic Health Records). The primary family caregiver/guardian—who is legally and emotionally responsible for daily care—is routinely omitted from the real-time feedback loop. There is a lack of unified platforms providing tailored, role-specific views for guardians, doctors, and seniors.
2. **Decoupled Vitals Tracking and Medication Adherence:** Current monitoring applications track vital signs in complete isolation from prescription compliance. When

an elder experiences an acute spike in blood pressure or pulse rate, physicians cannot ascertain whether the event is an organic clinical deterioration or simply the consequence of a missed antihypertensive dosage. Correlating vitals telemetry directly with verified medication check-offs represents an unaddressed necessity.

3. **Inadequate Emergency Dispatch Protocols and False Alarms:** Stand-alone wearable sensors frequently trigger false emergency alerts during normal physical activities, leading to alert fatigue. Conversely, when an actual critical event occurs, systems often rely solely on push notifications that may go unnoticed. A multi-channel escalation framework—integrating persistent on-screen alerts, audio alarms, visual emergency banners, and direct contact routing—is absent in low-cost platforms.
4. **Linguistic and Digital Divide in Geriatric User Interfaces:** Although voice technology has matured, existing health assistants are predominantly restricted to English and optimized for technologically literate demographics. In diverse societies like India, the absence of regional language voice support (such as Hindi, Kannada, and Tamil) severely restricts the accessibility and independence of elderly users.
5. **High Economic Barrier and Proprietary Vendor Lock-in:** Proprietary clinical monitoring solutions entail steep upfront costs and recurring subscription charges, placing them out of reach for middle- and lower-income families. Furthermore, proprietary data silos prevent cross-platform data export and collaborative clinical review.

Problem Formulation

Based on the identified gaps, there is an urgent need for an **integrated, affordable, and culturally accessible telecare ecosystem**. The objective of this project is to bridge these gaps by developing **G-Care**: an end-to-end web-based platform that unifies simulated edge IoT telemetry, rule-based vitals anomaly detection, closed-loop medication scheduling, a specialized doctor clinical portal, and senior-centric multilingual voice assistance into a cohesive, secure, and cost-effective solution.

Chapter 3

SYSTEM ARCHITECTURE

The architecture of **G-Care (Guardian Remote Elder Care Dashboard & Clinic Portal)** is engineered as a secure, distributed, and multi-tier Internet of Medical Things (IoMT) system. It seamlessly integrates a physical edge-sensing breadboard prototype, an embedded real-time OLED telemetry display, server-side data processing and anomaly detection, persistent clinical storage, and role-differentiated client presentation.

The primary design principle is **separation of concerns**, ensuring that continuous high-frequency vitals telemetry streaming, asynchronous event detection (falls and geofence breaches), scheduled medication reminders, and heavy clinical document management operate reliably without inducing mutual latency or single points of failure.

3.1 Architectural Framework and Layered Organization

The G-Care platform is organized into five cohesive layers: the Edge Sensing and Telemetry Simulation Layer, the Communication and Network Layer, the Backend Application and Logic Tier, the Data Persistence Tier, and the Presentation Layer (Role-Based Portals).




Figure 3.1: Multi-Tier Layered System Architecture of G-Care

3.2 Subsystem and Module Decomposition

The G-Care platform is modularized into seven interacting functional subsystems:

3.2.1 Physical Breadboard IoT Sensing Subsystem

- Constructed on a solderless breadboard controlled by an **ESP32-S3 microcontroller**.
- Houses three dedicated biosensors and an integrated display:
 - **MAX30102 Biosensor:** High-sensitivity red and infrared photoplethysmography sensor acquiring continuous pulse rates (40–180 BPM) and peripheral blood oxygen saturation (SpO_2 , 70–100%).
 - **MAX30205 Clinical Thermometer:** Digital human body temperature sensor providing clinical accuracy ($\pm 0.1^\circ\text{C}$).
 - **MPU6050 6-Axis Motion Tracking Sensor:** Tri-axial accelerometer and gyroscope sampling at 50Hz to detect sudden impact vectors and orientation changes during physical falls.
 - **0.96-inch OLED Display:** Structured with a continuous rectangular border and partitioned data boxes rendering real-time dynamic numeric streams for heart rate, SpO_2 , and temperature directly to the user.

3.2.2 Real-Time Anomaly Detection Engine

- Implements multi-threshold and derivative analysis heuristics on live telemetry:
 - **Heart Rate:** Flagged as *Warning* if $< 55 \text{ BPM}$ or $> 100 \text{ BPM}$; *Critical* if $< 48 \text{ BPM}$ or $> 120 \text{ BPM}$.
 - SpO_2 : Flagged as *Warning* if $< 93\%$; *Critical* (severe hypoxemia) if $< 90\%$.
 - **Blood Pressure:** Flagged if systolic $\geq 140 \text{ mmHg}$ or diastolic $\geq 90 \text{ mmHg}$; *Critical* if systolic $\geq 160 \text{ mmHg}$ or diastolic $\geq 100 \text{ mmHg}$.
 - **Fall Events:** Confirmed when acceleration magnitude $|\vec{A}|$ exhibits a free-fall trough ($< 0.5g$) followed by an impact spike ($\geq 3.0g$) and a static post-fall rest period.
- Automatically synthesizes procedural siren alerts via the Web Audio API and broadcasts high-priority visual warning banners to all connected guardian dashboards.

3.2.3 Medication and Routine Scheduling Subsystem

- Provides a structured scheduling engine supporting complex polypharmacy regimens:
 - Configures medicine names, dosages, instructions (*before/after meals*), and frequency patterns (*once daily, twice daily, PRN/as needed*).
 - Continuously evaluates client clock times against scheduled doses.

- Interactive compliance verification: guardians or elders mark doses as "Taken" or "Missed," creating historical adherence records for clinical analysis.

3.2.4 Caretaker and Guardian Telemetry Dashboard

- Central command center for family members:
 - **Live Vitals Grid:** Real-time visual cards displaying current metrics with color-coded status badges.
 - **Longitudinal Trends:** Historical metric charting powered by Recharts (line and area charts spanning hours, days, or weeks).
 - **Living Room / TV Mode:** High-contrast, large-format interface optimized for smart TV displays, enabling continuous ambient room monitoring.
 - **Geofence Map:** Interactive spatial tracking powered by React Leaflet showing the elder's real-time position within designated safe zones.

3.2.5 Clinical Doctor Portal & Diagnostic Workspace

- A specialized workspace for registered medical practitioners:
 - **Diagnostic Record Repository:** Allows uploading, categorizing, and previewing diagnostic files (ECGs, Complete Blood Counts, MRI/CT scans, prescriptions) with thumbnail badges and metadata.
 - **Clinical Recommendations Log:** A rich-text clinical note editor to write, update, and persist clinical directives, dosage alterations, and follow-up consultation dates.
 - **Care Team Directory:** Collaborative roster listing primary care physicians, cardiologists, and emergency hospital clinics linked to the patient profile.

3.2.6 Senior-Centric Accessibility & Multilingual Voice Subsystem

- Incorporates the **Web Speech API** (SpeechSynthesis and SpeechRecognition) coupled with an i18n localization store.
- Supports four regional languages: **English, Hindi, Kannada, and Tamil.**
- Speaks out incoming medicine reminders, daily routine prompts, and emergency warnings aloud.
- Accepts basic voice commands to confirm medication intake or query current vitals, addressing physical and cognitive limitations of senior users.

3.2.7 Security, Authentication, and RBAC Subsystem

- Enforces Role-Based Access Control (doctor, caretaker, guardian) using cryptographically signed JSON Web Tokens (JWT).

- Utilizes a stateful token revocation blacklist in memory/database to invalidate tokens immediately upon logout.
- Validates all incoming API payloads using strict schema definitions (Zod), rejecting malformed packets or unauthorized role elevation attempts.

3.3 Data Flow and System Interaction Modeling

3.3.1 Level 0: Context Data Flow Diagram

The Level 0 Context DFD establishes the operational boundaries of the G-Care platform, illustrating the interactions between external entities (Breadboard Hardware Node, Senior, Guardian/Caretaker, Doctor, and Emergency Responders) and the centralized G-Care System.




Figure 3.2: DFD Level 0 — Context Level Diagram

3.3.2 Level 1: Functional Data Flow Diagram

The Level 1 DFD decomposes the system into core internal processes, tracing data transformation across telemetry ingestion, anomaly analysis, medication scheduling, clinical management, and persistent storage.




Figure 3.3: DFD Level 1 — Detailed Functional Data Flow Diagram

Chapter 4

SYSTEM REQUIREMENTS

Defining the system requirements establishes the technical and operational boundaries necessary to build, deploy, and execute the **G-Care (Guardian Remote Elder Care Dashboard & Clinic Portal)** platform. These requirements are divided into:

1. **Hardware Requirements** (Physical breadboard prototyping components, edge sensors, display module, host machine, and client devices).
2. **Software Requirements** (Embedded firmware tooling, web application frameworks, state engines, databases, and APIs).
3. **Functional and Non-Functional Requirements** (System behavioral features, latency metrics, accessibility standards, and data security contracts).

4.1 Hardware Requirements

The hardware requirements encompass the physical embedded edge prototype assembled on the breadboard, the development environment, and client viewing terminals.

4.1.1 Physical Breadboard IoT Prototype Specifications

Table 4.1: Edge Hardware and Prototyping Components

Component	Model / Specification	Functional Role in G-Care
Prototyping Platform	Half-Size Solderless Breadboard (400 tie-points) + Male-to-Male / Male-to-Female Jumper Wires	Base physical prototyping surface interconnecting sensor breakout boards, display, and power distribution rails.
Main Microcontroller	ESP32-S3-WROOM-1 (Dual-core 32-bit Xtensa LX7 @ 240 MHz, 512 KB SRAM, Wi-Fi & BLE 5.0)	High-speed edge controller orchestrating sensor polling over I^2C , running local fall filtering, driving OLED display, and sending telemetry.
Optical PPG Biosensor	MAX30102 Integrated Pulse Oximeter & Heart-Rate Sensor	Emits dual-wavelength red (660nm) and infrared (880nm) light to measure pulse rate (40–180 BPM) and blood oxygen saturation (SpO_2 , 70–100%).
Clinical Temperature Sensor	MAX30205 Human Body Temperature Sensor	Provides high clinical accuracy ($\pm 0.1^\circ C$ over $37^\circ C$ to $39^\circ C$) via I^2C digital output to detect hypothermia or fever.

Component	Model / Specification	Functional Role in G-Care
Inertial Motion Unit (IMU)	MPU6050 6-Axis Motion Tracking (Tri-axial Accelerometer + Tri-axial Gyroscope)	Samples acceleration and angular velocity at 50Hz to detect rapid impact vectors ($> 3.0g$) and post-fall rest states.
Local Visual Display	0.96-inch OLED Display (128×64 pixels, I^2C , SSD1306 driver)	Renders a structured rectangular dashboard interface displaying real-time dynamic numeric values for HR, SpO_2 , and Temperature.
Pull-Up Resistors	4.7k Ω Metal Film Resistors	Pull-up resistors on SDA (GPIO 21) and SCL (GPIO 22) lines to maintain I^2C bus signal integrity across multiple sensor nodes.
Power Supply	5V DC via Micro-USB / Type-C or 3.7V 800mAh Li-Po battery with 3.3V LDO regulator	Delivers clean, regulated 3.3V DC power to the microcontroller and all sensor modules on the breadboard.

4.1.2 Development, Server, and Client Workstation Requirements

- **Processor:** Intel Core i5 / AMD Ryzen 5 (quad-core, $\geq 2.4GHz$) or Apple Silicon M-series.
- **System Memory (RAM):** Minimum 8GB (16 GB recommended for concurrent client, server, and local database execution).
- **Storage Space:** Minimum 20GB available Solid-State Drive (SSD) capacity.
- **Network Interface:** Standard 802.11ac/ax Wi-Fi or Gigabit Ethernet connection ($\geq 10Mbps$).
- **Client Viewing Terminals:**
 - Desktop / Laptop display with minimum screen resolution of 1366×768 pixels (1080p Full HD recommended).
 - Tablet or Smartphone display (iOS / Android) with modern HTML5-compliant mobile web browser.
 - Smart TV display (1920×1080 Full HD or 4K) for living room ambient monitor mode.

4.2 Software Requirements

The platform is constructed on modern web standards and embedded C++ libraries, utilizing type-safe languages across all tiers.

Table 4.2: Software Technology Stack

Domain	Technology / Tool	Version / Release	Role in G-Care System
Operating System	Windows 10/11, Ubuntu Linux 22.04 LTS, or macOS	64-bit Architecture	Base development and server hosting operating system.
Embedded Firmware IDE	Arduino IDE / PlatformIO	v2.3+	Writing, compiling, and flashing C++ firmware to the ESP32-S3 microcontroller.
Display & Sensor Libraries	Adafruit_SSD1306, Adafruit_GFX, Wire.h	Latest Stable	Controlling the I^2C bus and drawing the rectangular UI layout and dynamic numbers on the OLED screen.
Frontend Framework	React.js	v18.3+	Component-driven user interface architecture with declarative UI rendering.
Language & Typings	TypeScript	v5.5+	Enforces compile-time type safety across UI models, API responses, and store schemas.
Build & Bundler Tool	Vite	v8.0+	Fast build engine with Hot Module Replacement (HMR) and optimized tree-shaken production bundling.
CSS & Design System	Tailwind CSS + shadcn/ui	v3.4+ / Radix Primitives	Utility-first responsive styling framework with accessible, high-contrast UI primitives.
State Management	Zustand	v4.5+	Lightweight, reactive global store managing real-time vitals, medication states, and session tokens.
Data Visualization	Recharts	v2.12+	Composable charting library rendering responsive time-series graphs for heart rate, SpO_2 , and blood pressure.

Domain	Technology / Tool	Version / Release	Role in G-Care System
Spatial Mapping	React Leaflet + Leaflet.js	v4.2+ / v1.9+	Interactive mapping interface rendering elder GPS coordinates, safe-zone markers, and geofence radii.
Voice & Speech APIs	Web Speech API + Web Audio API	W3C Standard	Implements browser-native voice synthesis (SpeechSynthesis), voice recognition, and procedural alarm generation.
Backend Runtime	Node.js	v18.x / v20.x LTS	Asynchronous, event-driven JavaScript server runtime.
Web Server Framework	Express.js	v4.19+	Lightweight HTTP server handling RESTful routing, middleware pipelines, and API endpoints.
Payload Validation	Zod	v3.23+	Schema declaration and static type inference validating incoming API payloads against strict data contracts.
Authentication & Tokens	JSON Web Tokens (JWT) + bcryptjs	jsonwebtoken v9.0+	Stateless signed authentication with password hashing and stateful token revocation on logout.
Database Systems	PostgreSQL (Production) / db.json (Fallback)	PostgreSQL 14+ / Node FS	Relational data persistence for users, clinical notes, and reports, with local JSON file fallback for offline execution.
Testing Frameworks	Vitest, Playwright	Latest Stable	Automated unit and integration testing of stores, APIs, and anomaly triggers.

4.3 Functional and Non-Functional Requirements

4.3.1 Functional Requirements (FR)

- **FR1: Physical Sensor Acquisition and Local Rectangular Display**
 - The physical hardware prototype must continuously sample Heart Rate and SpO₂ (MAX30102), Body Temperature (MAX30205), and Motion/Fall Vectors (MPU6050).
 - The 0.96-inch OLED display must render a high-contrast rectangular border framing dedicated partitioned boxes that update dynamic numeric values continuously in real time.
- **FR2: Role-Based Authentication and Session Control**
 - The system must authenticate users under three distinct roles: doctor, caretaker, and guardian.
 - The system must issue cryptographically signed JWT tokens and maintain a token-revocation store that invalidates tokens immediately upon user sign-out.
- **FR3: Multi-Parametric Telemetry Ingestion and Web Monitoring**
 - The web platform must ingest, process, and display vital metrics in real time: Heart Rate (BPM), Blood Oxygen Saturation (SpO₂), Blood Pressure, Body Temperature (°C), Stress, and Hydration.
- **FR4: Real-Time Anomaly Detection and Emergency Dispatch**
 - The system must continuously evaluate live vitals against clinical thresholds and classify states as *Normal*, *Warning*, or *Critical*.
 - The system must detect physical fall events and geofence boundary deviations.
 - Upon detection of a Critical anomaly or manual SOS activation, the system must trigger visual alert banners, audible alarm frequencies, and emergency notifications.
- **FR5: Medication Management and Adherence Logging**
 - The system must allow guardians to configure medication regimens with customizable names, dosages, instructions, and frequency selectors (*once daily*, *twice daily*, *PRN*).
 - The system must trigger time-matched alerts and provide an interactive checklist enabling users to record doses as "Taken" or "Missed" for compliance tracking.
- **FR6: Clinical Workspace and Diagnostic Document Repository**
 - The system must enable physicians to inspect longitudinal vitals graphs, review medication compliance, and upload/categorize diagnostic files (CBC, ECG, radiology scans, prescriptions).

- The system must provide a clinical note editor for physicians to author and persist treatment directives and follow-up schedules.
- **FR7: Multilingual Senior Voice Assistance & Smart TV Mode**
 - The system must provide a voice assistant capable of speaking reminder notifications and responding to basic voice queries in four languages: **English, Hindi, Kannada, and Tamil**.
 - The system must provide a living room ambient display mode optimized for smart TVs.

4.3.2 Non-Functional Requirements (NFR)

- **NFR1: Performance and Low Latency**
 - The OLED display must refresh at $\geq 2\text{Hz}$ to render continuous dynamic vitals smoothly.
 - The client web dashboard must reflect live telemetry updates within 500ms of transmission.
 - Anomaly detection heuristics must evaluate incoming telemetry packets in under 50ms.
- **NFR2: Reliability, Availability, and Fault Tolerance**
 - The backend API server must maintain 99.9% operational uptime.
 - In the event of PostgreSQL database connectivity interruption, the system must automatically fall back to local file-backed JSON storage (data/db.json) to prevent data loss.
- **NFR3: Usability and Senior-Centric Accessibility**
 - The user interface must adhere to **WCAG 2.1 Level AA** standards: minimum body text size of 16–18px, high-contrast color palettes ($\geq 4.5:1$ contrast ratio), and large touch targets ($\geq 48 \times 48\text{px}$).
- **NFR4: Security, Data Integrity, and Confidentiality**
 - All communications between client terminals and backend endpoints must be transmitted over encrypted HTTPS/TLS. Passwords must be salted and hashed using bcrypt.
 - All incoming payloads must undergo strict schema validation via Zod to prevent injection and tampering.
- **NFR5: Portability and Cross-Platform Compatibility**
 - The responsive web frontend must render consistently across modern desktop and mobile browsers (Google Chrome, Mozilla Firefox, Microsoft Edge, Safari).

Chapter 5

METHODOLOGY

This chapter details the operational methodology and formal system design of **G-Care (Guardian Remote Elder Care Dashboard & Clinic Portal)**.

Section **5.1** outlines the end-to-end algorithmic and procedural methodology governing sensor data acquisition, dual-threshold anomaly detection, fall kinematics, and closed-loop medication workflows. Section **5.2** presents the concrete system design artifacts, including the physical breadboard I^2C circuit design, the rectangular OLED screen layout, the Entity-Relationship (ER) database schema, and Use Case behavioral models.

5.1 Proposed Methodology and Operational Flow

The G-Care methodology follows a reactive, event-driven pipeline organized into five sequential stages:

1. Physical Breadboard Sensor Telemetry Acquisition & Filtering.
2. Dual-Threshold Anomaly Detection and Fall Kinematics Analysis.
3. Closed-Loop Medication Adherence State Machine.
4. Emergency Alert Escalation & Broadcast Dispatch.
5. Multilingual Natural Speech Interaction Pipeline.

5.1.1 Physical Breadboard Telemetry Acquisition and Signal Processing

Telemetry is continuously sampled from the physical breadboard hardware at 50Hz for inertial motion and 10Hz for optical biosignals:

- **Optical PPG Processing:** Raw infrared and red AC/DC signal ratios from the MAX30102 are processed using a moving average window ($N = 16$ samples) to eliminate baseline wander and high-frequency motion artifacts.
- **Temperature Calibration:** The MAX30205 delta-sigma ADC outputs digital temperature registers, filtered to 0.1°C resolution.
- **Telemetry Data Frame:** Formatted into a standardized JSON payload: $T = \{\text{elder_id}, \text{timestamp}, \text{HR}, \text{SpO}_2, \text{BP}_{\text{sys}}, \text{BP}_{\text{dia}}, \text{Temp}, \text{Stress}, \text{Hydration}, \vec{A}, \vec{\omega}\}$




Figure 5.1: End-to-End Operational Methodology Flowchart of G-Care

5.1.2 Dual-Threshold Anomaly and Kinematics Heuristics

The heuristic engine evaluates vital parameters against clinical safety intervals and classifies severity into **Normal**, **Warning**, or **Critical**:

1. **Heart Rate Evaluation (HR in BPM):**

$$\text{Severity}(\text{HR}) = \begin{cases} \text{Critical (Severe Tachycardia),} & \text{HR} > 120 \\ \text{Warning (Elevated Pulse),} & 100 < \text{HR} \leq 120 \\ \text{Normal,} & 55 \leq \text{HR} \leq 100 \\ \text{Warning (Bradycardia),} & 48 \leq \text{HR} < 55 \\ \text{Critical (Severe Bradycardia),} & \text{HR} < 48 \end{cases}$$

2. **Peripheral Oxygen Saturation (SpO₂ in %):**

$$\text{Severity}(\text{SpO}_2) = \begin{cases} \text{Normal,} & 94\% \leq \text{SpO}_2 \leq 100\% \\ \text{Warning (Suboptimal Saturation),} & 90\% \leq \text{SpO}_2 < 94\% \\ \text{Critical (Acute Hypoxemia),} & \text{SpO}_2 < 90\% \end{cases}$$

3. **Inertial Fall Detection Kinematics:** Acceleration vector magnitude $|\vec{A}|$ is continuously computed from tri-axial accelerometer vectors: $|\vec{A}| = \sqrt{A_x^2 + A_y^2 + A_z^2}$. A fall event is confirmed if a severe impact spike ($|\vec{A}| \geq 3.0g$) is preceded by free-fall weightlessness ($|\vec{A}| \leq 0.5g$) and followed by a stationary post-fall posture period ($\Delta\theta \geq 60^\circ$ for > 5 seconds).
4. **Alert De-bouncing Logic:** To prevent alert fatigue, duplicate warning alerts for the same metric are suppressed during a sliding cooldown window ($T_{\text{cooldown}} = 60$ seconds), unless an event escalates from *Warning* to *Critical*.

5.2 Detailed System Design and Schema Modeling

5.2.1 Physical Breadboard I²C Bus and Hardware Interfacing Design

The physical hardware prototype is implemented on a half-size solderless breadboard. All peripheral modules share a common two-wire I²C serial bus:




Figure 5.2: Shared I^2C Bus Interfacing on Prototype Breadboard

I^2C Bus Addressing and Multiplexing

All four peripheral modules communicate over a shared two-wire I^2C bus operating at 400kHz Fast Mode:

- **OLED Display (SSD1306):** Slave address 0x3C
- **MAX30102 Pulse Oximeter:** Slave address 0x57
- **MAX30205 Clinical Thermometer:** Slave address 0x48
- **MPU6050 Inertial Measurement Unit:** Slave address 0x68

5.2.2 OLED Screen Rectangular UI Layout Design

The 0.96-inch monochrome OLED display (128×64 pixels) is designed with a structured rectangular dashboard layout to ensure high legibility for senior users:

- **Outer Rectangular Boundary:** A continuous border drawn from coordinate (0, 0) to (127, 63) framing the display.
- **Top Header (0 to 14 px):** System title (G-CARE MONITOR) with dividing horizontal separator line.
- **Partitioned Vitals Matrix (15 to 50 px):**
 - **Heart Rate Row:** Dynamic numeric pulse stream updated in real time (HR : 76 bpm).
 - **SpO₂ Row:** Real-time oxygen saturation percentage (SpO₂: 98 %).

- **Temperature Row:** Continuous clinical temperature reading (TEMP: 36.8 C).
- **Lower Status Footer (51 to 63 px):** Separated footer indicating system health: STATUS: NORMAL or flashing ! FALL DETECTED !.

5.2.3 Entity-Relationship (ER) Schema Modeling

The data layer is modeled relationally to guarantee referential integrity across users, elderly patient profiles, longitudinal vitals telemetry, medication schedules, and clinical reports.




Figure 5.3: Entity-Relationship Diagram of G-Care Platform

5.2.4 Use Case Modeling

The behavioral interactions between the system's human and device actors are formalized in the Use Case model:




Figure 5.4: Use Case Diagram for the Tri-Tier G-Care System

Chapter 6

IMPLEMENTATION

This chapter details the technical implementation of **G-Care (Guardian Remote Elder Care Dashboard & Clinic Portal)**.

Section **6.1** covers the physical breadboard hardware wiring, OLED display firmware, client web presentation, Zustand state stores, and clinical workspace modules. Section **6.2** formalizes the system's operational logic through rigorous algorithms and pseudocode. Section **6.3** describes the security controls, API contract validation using Zod, and the hybrid database persistence mechanism.

6.1 Implementation of Core Functional Modules

6.1.1 Physical Breadboard Hardware Prototype and Sensor Integration

A functional physical hardware prototype was implemented on a half-size solderless breadboard. This prototype validates edge sensor signal acquisition, on-device signal processing, and direct local visual telemetry feedback.

Table 6.1: Breadboard Pin Interfacing and Wiring Connections

Sensor / Module	Module Pin	ESP32-S3 Microcontroller Pin	Description
Common Power Rail	VCC / VIN	3V3 Pin (3.3V DC)	Regulated positive voltage rail for all sensors.
Common Ground Rail	GND	GND Pin	Common ground reference across breadboard.
0.96" OLED Display	VCC, GND	3.3V Rail, GND Rail	Power lines for SSD1306 driver.
	SDA	GPIO 21	I ² C Serial Data for display buffer.
	SCL	GPIO 22	I ² C Serial Clock line (400kHz).
MAX30102 (HR & SpO ₂)	VIN, GND	3.3V Rail, GND Rail	Optical biosensor power supply.
	SDA	GPIO 21	Shared I ² C Serial Data line.
	SCL	GPIO 22	Shared I ² C Serial Clock line.
	INT	GPIO 19	Active-low hardware interrupt for sample ready.

Sensor / Module	Module Pin	ESP32-S3 Microcontroller Pin	Description
MAX30205 (Temp)	VDD, GND	3.3V Rail, GND Rail	Precision body temperature sensor power.
	SDA, SCL	GPIO 21, GPIO 22	Shared I^2C bus.
MPU6050 (Fall IMU)	VCC, GND	3.3V Rail, GND Rail	6-axis accelerometer/gyroscope power.
	SDA, SCL	GPIO 21, GPIO 22	Shared I^2C bus.
	INT	GPIO 18	Motion interrupt pin to signal sudden shock/fall.

6.1.2 Embedded OLED Rectangular Display Firmware Implementation

The firmware running on the ESP32-S3 controls the screen rendering loop using the Adafruit_SSD1306 and Adafruit_GFX libraries. The code establishes a continuous polling routine that reads raw sensor values, calculates moving averages to filter physiological noise, and continuously refreshes the dynamic numbers inside the rectangular UI template:

```
#include <Wire.h>

#include <Adafruit_GFX.h>

#include <Adafruit_SSD1306.h>

#define SCREEN_WIDTH 128
#define SCREEN_HEIGHT 64
#define OLED_RESET -1

Adafruit_SSD1306 display(SCREEN_WIDTH, SCREEN_HEIGHT, &Wire,
OLED_RESET);

void setup() {
  Wire.begin(21, 22); // Initialize I2C bus: SDA=21, SCL=22
  display.begin(SSD1306_SWITCHCAPVCC, 0x3C);
  display.clearDisplay();
}

void loop() {
  // 1. Read real-time telemetry from physical sensors
```

```
int heartRate = readMAX30102HeartRate(); // Dynamic real-time pulse (e.g. 72-78 bpm)
int spo2Level = readMAX30102SpO2();      // Dynamic SpO2 level (e.g. 97-99 %)
float bodyTemp = readMAX30205Temperature(); // Precision temp (e.g. 36.7 C)
bool fallDetected = checkMPU6050Fall();

// 2. Render Rectangular Dashboard UI
display.clearDisplay();

// Outer frame boundary
display.drawRect(0, 0, 128, 64, SSD1306_WHITE);

// Header section
display.setTextSize(1);
display.setTextColor(SSD1306_WHITE);
display.setCursor(18, 4);
display.print("G-CARE MONITOR");
display.drawLine(0, 14, 127, 14, SSD1306_WHITE);

// Vitals Grid Boxes
// Box 1: Heart Rate
display.setCursor(6, 20);
display.print("HR : ");
display.print(heartRate);
display.print(" bpm");

// Box 2: SpO2
display.setCursor(6, 32);
display.print("SpO2: ");
display.print(spo2Level);
display.print(" %");

// Box 3: Temperature
display.setCursor(6, 44);
display.print("TEMP: ");
```

```
display.print(bodyTemp, 1);
display.print(" C");
// Footer Alert Status
display.drawLine(0, 53, 127, 53, SSD1306_WHITE);
display.setCursor(6, 55);
if (fallDetected) {
  display.print("! FALL DETECTED !");
} else {
  display.print("STATUS: NORMAL");
}
// Push buffer to OLED screen
display.display();
delay(500); // 2 Hz continuous dynamic screen refresh
}
```

6.1.3 Client Web Presentation and State Engine

- **Framework & UI:** Built with **React 18**, **TypeScript**, and **Vite 8** using **Tailwind CSS** and **shadcn/ui**.
- **Global State (Zustand):** Partitioned into `authStore.ts` (JWT session), `useAppStore` (live vitals grid, patient metrics, alert banners), and `useGuardianStore` (medication compliance checklist, routine logs).
- **Charting:** Dynamic time-series graphs rendered with **Recharts** (`ResponsiveContainer`, `LineChart`, `AreaChart`), updating dynamically upon incoming telemetry ticks.

6.1.4 Clinical Doctor Workspace (`DoctorPortal.tsx`)

- Implements medical record uploads categorized into Complete Blood Counts (CBC), Electrocardiograms (ECG), Radiology Scans (CT/MRI), and Prescriptions.
- Includes a clinical note editor for authoring persistent diagnostic notes and follow-up schedules.

6.1.5 Multilingual Voice Assistant & Web Audio Siren

- Implemented via **Web Speech API** (`SpeechSynthesisUtterance`), supporting **English, Hindi, Kannada, and Tamil** voice prompts.
- Procedural emergency siren generated using **Web Audio API** (`OscillatorNode` at alternating 880Hz/440Hz frequencies), eliminating static audio asset dependencies.

6.2 Algorithms and Pseudocode

Algorithm 6.1: Physiological Vitals Anomaly Detection and Cooldown De-Bouncing

Algorithm: DetectAndDispatchVitalsAnomalies

Input:

elder: Elder profile {id, name}

vitals: Telemetry frame {heart_rate, spo2, systolic_bp, diastolic_bp, breathing_rate, stress}

lastAlertTimestampMap: In-memory hash map tracking {elderId:metric -> (timestamp, severity)}

ALERT_COOLDOWN_MS = 60000 ms

Output:

dispatchedAnomalies: List of active anomalies routed to UI, Audio, and DB

1: Initialize anomalies $\leftarrow$ Empty List

2: Initialize dispatchedAnomalies $\leftarrow$ Empty List

3: currentTime $\leftarrow$ GetCurrentTimestampMs()

// 1. Evaluate Heart Rate (BPM)

4: if vitals.heart_rate > 120 then

5: anomalies.append({metric: "heart_rate", severity: "critical", title: "Severe Tachycardia", val: vitals.heart_rate})

6: else if vitals.heart_rate > 100 then

7: anomalies.append({metric: "heart_rate", severity: "warning", title: "Elevated Heart Rate", val: vitals.heart_rate})

8: else if vitals.heart_rate < 48 then

9: anomalies.append({metric: "heart_rate", severity: "critical", title: "Severe Bradycardia", val: vitals.heart_rate})

10: else if vitals.heart_rate < 55 then

11: anomalies.append({metric: "heart_rate", severity: "warning", title: "Bradycardia", val: vitals.heart_rate})

12: end if

// 2. Evaluate Oxygen Saturation (SpO2 %)

13: if vitals.spo2 < 90 then

14: anomalies.append({metric: "spo2", severity: "critical", title: "Acute Hypoxemia", val: vitals.spo2})

```
15: else if vitals.spo2 < 93 then
16:   anomalies.append({metric: "spo2", severity: "warning", title: "Suboptimal SpO2", val:
vitals.spo2})
17: end if

// 3. Evaluate Blood Pressure (mmHg)
18: if vitals.systolic_bp ≥ 160 or vitals.diastolic_bp ≥ 100 then
19:   anomalies.append({metric: "systolic_bp", severity: "critical", title: "Hypertensive
Urgency", val: vitals.systolic_bp})
20: else if vitals.systolic_bp ≥ 140 or vitals.diastolic_bp ≥ 90 then
21:   anomalies.append({metric: "systolic_bp", severity: "warning", title: "Elevated BP", val:
vitals.systolic_bp})
22: else if vitals.systolic_bp < 90 or vitals.diastolic_bp < 60 then
23:   anomalies.append({metric: "systolic_bp", severity: "warning", title: "Hypotension", val:
vitals.systolic_bp})
24: end if

// 4. Cooldown De-Bouncing & Dispatch Pipeline
25: for each item in anomalies do
26:   key ← elder.id + ":" + item.metric
27:   prev ← lastAlertTimestampMap.get(key)
28:   isCooldownExpired ← (prev == null) or (currentTime - prev.timestamp >
ALERT_COOLDOWN_MS)
29:   isSeverityEscalated ← (prev != null and prev.severity == "warning" and item.severity
== "critical")
30:   if isCooldownExpired or isSeverityEscalated then
31:     lastAlertTimestampMap.set(key, {timestamp: currentTime, severity: item.severity})
32:     dispatchedAnomalies.append(item)

33:   DispatchToAppStore(item)      // Notify Doctor & Caretaker Dashboards
34:   DispatchToGuardianStore(item)  // Notify Family Portal
35:   TriggerWebAudioAlarm(item.severity) // Trigger Procedural Siren
36:   AsyncPostAlertToBackend(item)  // Persist to Server Database
37: end if
```

38: end for

39: return dispatchedAnomalies

Algorithm 6.2: Tri-Axial IMU Fall Detection and Emergency SOS Escalation

Algorithm: InertialFallDetectionAndSOSEscalation

Input:

accel: Real-time 3D acceleration vectors (A_x, A_y, A_z) in units of g

gyro: Angular velocity (G_x, G_y, G_z) in deg/s

COUNTDOWN_DURATION = 15 seconds

Output:

Dispatches verified fall alert and initiates emergency protocols

1: $A_{mag} \leftarrow \sqrt{A_x^2 + A_y^2 + A_z^2}$

2: Initialize state $\leftarrow$ "MONITORING"

// Phase 1: Free-fall weightlessness detection

3: if $A_{mag} < 0.5$ then

4: $t_{\text{freefall}} \leftarrow \text{GetCurrentTimestamp}()$

5: while ($\text{GetCurrentTimestamp}() - t_{\text{freefall}} < 500 \text{ ms}$) do

6: $A_{mag} \leftarrow \sqrt{A_x^2 + A_y^2 + A_z^2}$

// Phase 2: High-g impact collision

7: if $A_{mag} \geq 3.0$ then

8: state $\leftarrow$ "IMPACT_DETECTED"

9: break

10: end if

11: end while

12: end if

13: if state == "IMPACT_DETECTED" then

14: Sleep(2000 ms) // Allow body to settle

// Phase 3: Post-impact immobility and posture verification

15: $A_{mag_post} \leftarrow \sqrt{A_x^2 + A_y^2 + A_z^2}$

16: if ($|A_{mag_post} - 1.0| < 0.2$) and ($G_{\text{magnitude}} < 15 \text{ deg/s}$) then

```

17:     state ← "FALL_CONFIRMED"
18:   end if
19: end if
20: if state == "FALL_CONFIRMED" then
21:   TriggerLocalHapticVibration()
22:   isCancelled ← StartCountdownTimer(COUNTDOWN_DURATION)
23:   if isCancelled == false then
24:     alertPacket ← {type: "FALL_EMERGENCY", timestamp: Now()}
25:     BroadcastSOSAlert(alertPacket)
26:     RenderGlobalCriticalBanner("CRITICAL: Fall Detected! Immediate Assistance
Required")
27:   end if
28: end if

```

Algorithm 6.3: Closed-Loop Medication Adherence Verification

Algorithm: MedicationAdherenceWorkflow

Input:

activeMedications: List of prescribed medicines {id, name, dose, frequency, times, instructions}

currentClock: Current system time "HH:mm"

Output:

Updates adherence state and triggers timely reminders

```

1: for each med in activeMedications do
2:   for each scheduledTime in med.times do
3:     if currentClock == scheduledTime and med.status == "PENDING" then
4:       med.status ← "DUE"
5:       spokenText ← "Time to take " + med.brand_name + ", " + med.dose_amount + " "
+ med.dose_unit + ". " + med.instructions
6:       SynthesizeRegionalSpeech(spokenText, userLanguagePreference)
7:       ShowDashboardReminderCard(med)
8:       StartAdherenceTimer(timeout = 45 minutes)

```

```

9:     end if
10:  end for
11:  if med.status == "DUE" then
12:    if UserClickedTaken() or VoiceRecognized("Taken") then
13:      med.status ← "TAKEN"
14:      med.takenAt ← GetCurrentTimestamp()
15:      PersistAdherenceLog(med.id, status="TAKEN", timestamp=med.takenAt)
16:      SynthesizeRegionalSpeech("Medication recorded. Thank you.",
userLanguagePreference)
17:    else if AdherenceTimerExpired() then
18:      med.status ← "MISSED"
19:      PersistAdherenceLog(med.id, status="MISSED", timestamp=Now())
20:      EscalateAlertToGuardian("Missed Dose Alert: " + med.brand_name + " was not
acknowledged.")
21:    end if
22:  end if
23: end for

```

6.3 Security, API Contracts, and Database Connectivity

6.3.1 Role-Based Access Control and Stateful Token Revocation

- Cryptographic JWT signed with HMAC-SHA256 (jwt.sign) with {userId, email, role}.
- Role gating middleware: requireRole(['doctor', 'caretaker', 'guardian']).
- Stateful revocation: When a user calls /api/logout, their token is inserted into a blacklist store (revokedTokens), blocking replay attacks.

6.3.2 Strict Schema Validation using Zod

All incoming payloads are strictly validated using **Zod** schemas (server/handlers.js):

```

const vitalsSchema = z.object({
  elderId: z.string().min(1),
  heart_rate: z.coerce.number().int().min(0).max(300),
  systolic_bp: z.coerce.number().int().min(0).max(300),
  diastolic_bp: z.coerce.number().int().min(0).max(200),

```

```
spo2: z.coerce.number().int().min(0).max(100),  
stress: z.coerce.number().int().min(0).max(100),  
hydration: z.coerce.number().int().min(0).max(100),  
breathing_rate: z.coerce.number().int().min(0).max(100),  
skin_temp: z.coerce.number().min(0).max(50),  
source: z.enum(['manual', 'simulator', 'device']).default('manual'),  
});
```

6.3.3 Dual-Engine Data Persistence Architecture (server/db.js)

- **Production Database (PostgreSQL):** Uses pg.Pool for concurrent connection pooling and ACID compliance.
- **Offline / Development Database (data/db.json):** Automatic fallback when DATABASE_URL is absent, persisting JSON records via asynchronous serialized write queues (writeQueue).

Chapter 7

TESTING

System testing validates that the physical breadboard hardware prototype, embedded OLED display, firmware routines, backend APIs, and multi-portal web dashboards operate reliably under both normal conditions and medical anomaly scenarios.

Section 7.1 outlines the testing methodologies and verification strategies. Section 7.2 presents the comprehensive test case matrix and empirical test results.

7.1 Testing Methodologies and Verification Strategies

A multi-phase verification strategy was conducted across four testing tiers:




Figure 7.1: Multi-Tier Testing Pipeline for G-Care Platform

7.1.1 Hardware Prototype & Circuit Testing

- **I²C Bus Peripheral Discovery:** An embedded I²C scanner sketch was executed on the ESP32-S3 to verify that all modules communicate on the shared bus without address collision:
 - OLED Display detected at 0x3C (ACK received).
 - MAX30102 Biosensor detected at 0x57 (ACK received).
 - MAX30205 Clinical Thermometer detected at 0x48 (ACK received).
 - MPU6050 IMU detected at 0x68 (ACK received).
- **OLED Screen Refresh & Display Stability:** Tested continuous dynamic number updates inside the rectangular border. The display sustained a stable 2Hz refresh rate without memory leaks, screen tearing, or I2C bus lockup.
- **Physical Sensor Calibration:** MAX30102 PPG values were validated against a commercial clinical fingertip pulse oximeter. Temperature readings from the MAX30205 were compared against a calibrated digital clinical thermometer.

7.1.2 Software Unit & State Testing (Vitest)

Unit tests were executed using **Vitest** (vitest.config.ts):

- `anomalyAlertRouting.test.ts`: Validates that pure evaluation functions correctly classify normal, warning, and critical vitals.
- `vitalsSimulator.test.ts`: Verifies synthetic telemetry generation and boundary values.
- `dataPersistenceAndSync.test.ts`: Tests that state transitions in Zustand correctly synchronize with the backend REST endpoints.

7.1.3 Integration, Security, and RBAC Testing

- **Payload Validation:** Tested malformed JSON payloads (e.g., negative heart rates or $\text{SpO}_2 > 100\%$). Verified that Zod middleware returns HTTP 400 Bad Request with precise error paths.
- **Role Authorization:** Verified that a user logged in as guardian is rejected with HTTP 403 Forbidden when attempting to access doctor-only endpoints (`/api/reports/upload`).
- **Session Revocation:** Tested that after logging out, submitting an old Bearer token returns HTTP 401 Unauthorized because the token is actively registered in `revokedTokens`.

7.2 Test Cases and Verification Results

The formal verification results are compiled in **Table 7.1**.

Table 7.1: Comprehensive System Test Cases and Empirical Verification Results

Test Case ID	Test Description	Input / Test Action	Expected Result	Actual Result	Status
TC01	I²C Hardware Peripheral Bus Discovery	Run I ² C address scanner firmware on ESP32-S3 breadboard setup.	All 4 devices (0x3C, 0x48, 0x57, 0x68) respond with ACK.	Addresses 0x3C, 0x48, 0x57, and 0x68 verified successfully.	PASS
TC02	OLED Rectangular Display Real-Time Rendering	Execute OLED display firmware loop on breadboard.	Outer rectangular frame drawn; dynamic numbers for HR, SpO ₂ , and Temp update continuously at 2Hz.	Display renders stable rectangular frame; numbers update smoothly without flicker.	PASS
TC03	MAX30102 PPG Telemetry Acquisition	Place finger on optical sensor on breadboard.	Sensor outputs fluctuating IR/Red values; software calculates valid HR (60–100 BPM) and SpO ₂ (95–99%).	Pulse rate read as 74BPM, SpO ₂ read as 98%.	PASS
TC04	MPU6050 Fall Detection Shock Trigger	Rapidly invert and tap breadboard to simulate fall impact (\$	$\vec{A}$	$\geq 3.0g$).	Firmware detects impact peak followed by rest; status changes to ! FALL DETECTED !.
TC05	Critical Tachycardia Anomaly & Audio Siren	Inject Heart Rate = 128BPM(> 120BPM critical limit).	Anomaly engine flags Critical; Web Audio API procedural siren sounds; global alert banner renders.	Audible dual-tone siren plays; red alert banner displays on Caretaker Dashboard.	PASS
TC06	Alert Cooldown De-Bouncing	Send consecutive abnormal vitals within	Second identical warning is suppressed by 60-second cooldown window.	Duplicate alarm suppressed; console logs cooldown active.	PASS

Test Case ID	Test Description	Input / Test Action	Expected Result	Actual Result	Status
		10 seconds of initial alert.			
TC07	Closed-Loop Medication Adherence Flow	System clock reaches scheduled medication time (08:00).	Voice prompt speaks reminder in chosen language; dashboard card shows "DUE"; user clicks "Taken".	Speech synthesis announces dose; status transitions from "DUE" to "TAKEN" in DB.	PASS
TC08	Multilingual Voice Assistant Synthesis	Set language preference to Kannada (kn) and trigger medication prompt.	Web Speech API synthesizes prompt using Kannada phonetics/voice pack.	Spoken reminder delivered clearly in Kannada.	PASS
TC09	Doctor Clinical Document Ingestion	Doctor uploads ECG PDF report with diagnostic comments.	File accepted, validated, categorized as "ECG", and associated with patient record.	Report appears in patient history with visual badge and download link.	PASS
TC10	JWT Authentication and Logout Revocation	Authenticate user, obtain JWT token, call /api/logout, then call /api/me with old token.	Token is stored in revocation blacklist; subsequent request returns HTTP 401.	HTTP 401 Unauthorized returned; token invalidated successfully.	PASS

Chapter 8

CONCLUSION AND FUTURE SCOPE

The rapid growth of the global geriatric population, paired with the proliferation of nuclear families, has created an urgent need for intelligent, non-intrusive, and accessible remote healthcare solutions. Senior citizens living independently face heightened vulnerabilities from unmonitored chronic degenerative conditions, complex medication regimens, and fatal delays during acute medical crises such as cardiac events or slip-and-fall trauma.

To resolve these challenges, this project successfully conceptualized, designed, implemented, and verified **G-Care**: an integrated, full-stack, and proactive remote elderly monitoring ecosystem. G-Care bridges the operational and communication gap between elderly patients, family guardians, and attending healthcare practitioners by unifying physical embedded edge sensing with a modern web telecare infrastructure.

Key Accomplishments of the Project:

1. Physical Breadboard Hardware & Local Telemetry Display:

- A functional, low-cost embedded IoT prototype was successfully engineered on a solderless breadboard powered by an ESP32-S3 microcontroller.
- High-accuracy physiological sensing was achieved over a shared $400\text{kHz I}^2\text{C}$ bus integrating an optical MAX30102 PPG biosensor (Heart Rate and SpO_2), a MAX30205 clinical digital thermometer, and an MPU6050 6-axis inertial motion tracking sensor.
- An embedded 0.96-inch OLED display was programmed with a structured rectangular dashboard interface, rendering smooth, dynamic numeric telemetry streams at 2Hz to provide immediate, localized visual feedback to the senior.

2. Automated Real-Time Anomaly Detection & Emergency SOS:

- An automated heuristic decision engine was deployed to evaluate continuous physiological streams against validated clinical thresholds.
- Multi-stage fall kinematics algorithms successfully distinguished high- g collision impacts ($|\vec{A}| \geq 3.0g$) and subsequent immobility from normal activities of daily living.
- Emergency protocols—including procedural dual-tone sirens (Web Audio API), global visual alert banners, and cooldown de-bouncing ($T_{\text{cooldown}} = 60\text{s}$)—guaranteed rapid alert escalation without causing alarm fatigue.

3. Closed-Loop Medication Management:

- Replaced conventional, unreliable manual reminders with an interactive scheduling engine supporting custom polypharmacy frequencies (*once daily*, *twice daily*, *PRN*).

- Bidirectional adherence verification enabled seniors and guardians to check off administered doses, generating historical compliance records essential for clinical diagnostics.

4. Specialized Tri-Tier Clinical Workspace:

- Built distinct, role-segregated portals enforcing Role-Based Access Control (RBAC) across doctors, caretakers, and guardians using cryptographically signed JWTs with stateful token revocation.
- Provided medical practitioners with a centralized clinical portal to review longitudinal vitals trends (Recharts), upload and inspect diagnostic records (ECG, CBC blood panels, radiology scans, prescriptions), and record persistent clinical notes.

5. Senior Accessibility & Regional Inclusion:

- Integrated browser-native Web Speech synthesis and recognition supporting four languages (English, Hindi, Kannada, and Tamil), eliminating digital literacy barriers for Indian senior citizens.
- Delivered an ambient Smart TV living room display mode optimized for passive, long-range household monitoring.

In conclusion, G-Care demonstrates that a comprehensive, responsive, and reliable telecare infrastructure can be realized at an estimated hardware prototype cost of approximately ₹6,000—a fraction of the cost of proprietary commercial smartwatches or assisted-living facilities. Empirical testing confirmed stable sensor bus discovery, low alert dispatch latency (< 500ms), robust tokenized API security, and high user acceptance, establishing G-Care as a viable, proactive solution for modern geriatric telecare.

Future Scope

While the current implementation delivers a complete, validated hardware-software telecare ecosystem, several avenues exist for commercialization, algorithmic refinement, and clinical expansion:

1. Hardware Miniaturization and Ergonomic Wearable Casing

- Custom Multi-Layer PCB: Transition the current solderless breadboard prototype onto a miniaturized, multi-layer surface-mount device (SMD) Printed Circuit Board (PCB).
- Ergonomic 3D-Printed Enclosure: Design an injection-molded, water-resistant (IP67/IP68 rated) smartwatch enclosure and medical-grade hypoallergenic silicone wristband, integrating a rechargeable Lithium-Polymer battery and a magnetic charging dock for daily wearable usage.

2. Edge AI and Predictive Machine Learning

- **Arrhythmia and Anomaly Classification:** Deploy lightweight embedded machine learning models (such as TinyML or TensorFlow Lite for Microcontrollers) directly

onto the ESP32-S3 to analyze continuous PPG pulse waveforms for early atrial fibrillation (AFib) and irregular beat detection.

- **Predictive Health Analytics:** Incorporate Long Short-Term Memory (LSTM) recurrent neural networks on the server tier to correlate historical vitals with medication compliance, enabling proactive forecasting of hypertensive surges or respiratory decline hours before clinical onset.

3. Cellular IoT (NB-IoT / LTE-M) and GPS Outdoor Tracking

- **Untethered Connectivity:** Equip the edge wearable with an integrated SIM7080G / SIM7600 LTE-M / NB-IoT cellular module. This will allow seniors to remain connected to the G-Care monitoring network when walking outdoors or traveling outside the range of domestic home Wi-Fi networks.
- **Active Satellite Geofencing:** Integrate continuous satellite tracking via the NEO-6M GPS receiver to automatically alert family members if an elderly individual experiencing dementia or Alzheimer's wanders beyond designated safe-zone boundaries.

4. Multi-Channel Telecommunications Gateway Integration

- **Automated IVR and SMS Dispatch:** Integrate telecommunications APIs (such as Twilio, Exotel, or AWS SNS) to broadcast immediate automated voice calls, SMS text alerts, and WhatsApp notifications to family guardians and registered emergency ambulance services upon confirmed fall or SOS events.

5. Hospital EHR and Interoperability Standards (HL7 / FHIR)

- **Clinical System Integration:** Standardize diagnostic reports, medication logs, and vital telemetry outputs into the HL7 (Health Level Seven) FHIR (Fast Healthcare Interoperability Resources) data standard. This will allow seamless bidirectional synchronization with major hospital Electronic Health Record (EHR) and Electronic Medical Record (EMR) databases.

6. Native Cross-Platform Mobile Applications


- **Dedicated Mobile Apps:** Develop native mobile applications using React Native or Flutter for Android and iOS. This will provide family guardians and doctors with native OS push notifications, widget glance views, biometric login (FaceID / Fingerprint), and background telemetry synchronization.

BIBLIOGRAPHY


- [1] S. M. R. Islam, D. Kwak, M. H. Kabir, M. Hossain, and K.-S. Kwak, "The Internet of Things for Health Care: A Comprehensive Survey," *IEEE Access*, vol. 3, pp. 678–708, 2015, doi: [10.1109/ACCESS.2015.2437951](https://doi.org/10.1109/ACCESS.2015.2437951).
- [2] S. Majumder, T. Mondal, and M. J. Deen, "Wearable Sensors for Remote Health Monitoring," *Sensors*, vol. 17, no. 1, Art. no. 130, 2017, doi: [10.3390/s17010130](https://doi.org/10.3390/s17010130).
- [3] J. Allen, "Photoplethysmography and its application in clinical physiological measurement," *Physiological Measurement*, vol. 28, no. 3, pp. R1–R39, 2007, doi: [10.1088/0967-3334/28/3/R01](https://doi.org/10.1088/0967-3334/28/3/R01).
- [4] R. Igual, C. Medrano, and I. Plaza, "Challenges, Issues and Trends in Fall Detection Systems," *BioMedical Engineering OnLine*, vol. 12, Art. no. 66, 2013, doi: [10.1186/1475-925X-12-66](https://doi.org/10.1186/1475-925X-12-66).
- [5] N. K. Choudhry *et al.*, "Effect of Reminder Devices on Medication Adherence: The REMIND Randomized Clinical Trial," *JAMA Internal Medicine*, vol. 177, no. 5, pp. 624–631, 2017, doi: [10.1001/jamainternmed.2016.9627](https://doi.org/10.1001/jamainternmed.2016.9627).
- [6] P. Nallam, S. Bhandari, J. Sanders, and A. Martin-Hammond, "Exploring the Perceived Benefits and Barriers of Intelligent Voice Assistants for Older Adults," *Gerontology and Geriatric Medicine*, vol. 6, pp. 1–13, 2020, doi: [10.1177/2333721420985975](https://doi.org/10.1177/2333721420985975).
- [7] Espressif Systems, "ESP32-S3 Wireless Microcontroller & MAX30102 Biosensor Hardware Integration Datasheets," Maxim Integrated / Espressif Co., Ltd., 2023.

APPENDIX A

SNAPSHOTS



G Care Telecare Platform



G-Care Multilingual Watch Simulator Interface


G Care

Every second matters.

- 10 vitals tracked
- 4 languages
- < 3s alerts
- Guardian Portal

Caretaker
Doctor
Guardian

Welcome back

Email address

Password

☐ Remember me [Forgot password?](#)

Log In

New to GuardianWatch? [Create an account](#)

Multi-Role User Authentication and Login Interface


G Care

Y
Yeshrutha S
Caretaker

Dashboard
Elders
Medications
Alarms
Alerts
Reports
Doctor Portal
Settings

Dashboard
Demo Mode
Log out

AI risk index
24
Weighted from vitals, battery, alerts, and baseline status.

Live devices
3/3
Connected watches streaming recent health signals.

Unresolved alerts
99
Active care events that still need attention.

Stable profiles
2
0 watch needs charging soon.

Highest priority
Watch 57
Venkatesh Rao
Check breathing comfort and keep oxygen trend under review.

SpO2
92.6%

BP
120/77

Stress
61/100

Review vitals
Call family
Locate

AI care queue
Live scoring
Ranked by current risk and monitoring reliability.

VR Venkatesh Rao
Age 71

HR 81 SPO2 92.6% 91%

Check breathing comfort and keep oxygen trend under...

Risk: 57

U Usha
Age 77

HR 72 SPO2 97% 78%

Vitals are inside personal range. Keep routine monitoring active.

Risk: 8

LD Lakshmi Devi
Age 82


HR 79 SPO2 97.1% 54%

Vitals are inside personal range. Keep routine monitorin...


Risk: 8

Elders
Log Vitals
Add Elder


Caretaker Dashboard for Elder Health Monitoring




Caretaker Elders Health Monitoring Dashboard



Caretaker Medication Schedule and Management Interface



Doctor Portal Dashboard for Patient Monitoring and Clinical Alerts



Clinical Alert Management and Urgent Appointment Scheduling

**VITAL ANOMALY ALERT**

WARNING

Usha: Vital Alert: Usha's biometric stress index spiked to 90/100. Dr. Ramesh Kumar and caretaker have been notified.

Logged at 2:23:50 PM · Clinical Doctor & Caretaker Notified

Acknowledge & Confirm Safe

View Alerts (12)

Feed

Logs

Alerts

Reminders

SOS

Smart TV

Watch Connected — Live

Motion Monitor

Sitting

Accelerometer + Gyroscope

Shiver Monitor

Normal

Micro-tremor analysis

Heart Rate

71 bpm

Blood Pressure

128/83 mmHg SpO₂**97.8** %

Stress

18 /100

Hydration

64 %

Temperature

36.6 °C

Guardian Portal for Real-Time Elder Health Monitoring