



TCT 2025, October 27, 2025



Valve-in-Valve TAVR with Bioprosthetic Valve Fracture and Snaring Technique in Extremely Challenging Anatomy

Subtitle: VIV-TAVR in Extremely Challenging Anatomy

Ju Han Kim, MD, PhD, and Seok Oh, MD, PhD



TCT[®]

TRANSCATHETER
CARDIOVASCULAR
THERAPEUTICS[®]


Disclosure of Relevant Financial Relationships

We, Ju Han Kim, and Seok Oh DO NOT have any financial relationships to disclose.

Case – An 88-year-old Korean woman

- Referred to valve-in-valve transcatheter aortic valve replacement (VIV-TAVR)
- Degeneration of a prior 21mm Carpentier-Edwards PERIMOUNT Magna Ease bioprosthesis
- P/H : Hypertension, previous history of AVR
- Echo :
 - S/P AVR with prosthetic valve dysfunction
 - LVEF 57.3%, severe AS, moderate AR

Pre-TAVR Echocardiogram




- **EOA 0.75 cm², mean PG 54.0 mmHg, and peak AoV velocity 4.63m/s**
- **Indicating significant patient-prosthesis mismatch d/t the prior small surgical bioprosthesis.**

Pre-TAVR CT Evaluation

1. Small annulus

Min. Ø: 18.5 mm
Max. Ø: 19.6 mm
Avg. Ø: 19.0 mm
Area derived Ø: 19.0 mm
Perimeter derived Ø: 19.0 mm
Area: 282.4 mm²
Perimeter: 59.7 mm




2. Extreme horizontal angulation



- The true inner diameter of the surgical bioprosthesis was 19.0mm based on CT scan (small annulus – 19mm)
- Also, CT scan showed the extreme horizontal angulation of the ascending aorta (extreme horizontal angulation – 97 degree)

Previously Implanted Surgical Prostheses

Size: 21



Stent ID Height True ID ⓘ
20 15 19

Fracturable ⓘ

True Balloon Size: 22mm

After fracture

THV size needed may be larger

21mm Carpentier-Edwards PERIMOUNT Magna Ease bioprosthesis

THV Selector: Current



ACURATE
USE WITH CAUTION

Allegra
23

Evolut

23

Myval


20/21.5

NAVITOR

23

S3





20/23





- While supra-annular valves are generally preferred for small annuli to achieve lower residual gradients, the patients severely angulated aorta and horizontal orientation raised significant concern about device deliverability and alignment.
- Navitor valve (Abbott), with its flexible delivery system and low profile, was considered advantageous in this extremely tortuous anatomy.



TAVR in Progress




Cerebral embolic protection device
(SENTINEL™, Boston Scientific)



Bioprosthetic valve fracture


- To offset the hemodynamic limitation of intra-annular valves (Navitor, Abbott), optimized the orifice area, then achieve acceptable hemodynamics post-VIV-TAVR, we performed bioprosthetic valve fracture.
- Nonetheless, advancement of the Navitor valve failed due to interference from the angulated aorta and the prior valve frame.

TAVR in Progress




- To overcome this point (failed advancement of the Navitor valve), an ipsilateral snare technique was employed.
- A 25-mm Amplatz Goose Neck snare (Medtronic) was pre-mounted on the FlexNav delivery system and introduced alongside the device into the ascending aorta.
- Gentle external traction enabled coaxial alignment and advancement of the prosthesis across the degenerative valve.
- The Navitor (Abbott) valve was successfully deployed without complications.

Post-TAVR Echocardiogram




EOA 1.63 cm², mean PG 11.7 mmHg, and peak AoV velocity 2.40 m/s

Pre-TAVR peak-to-peak PG



Post-TAVR peak-to-peak PG



Why This Case?

- A technically complex VIV-TAVR in an elderly patient with **a small annulus and extreme aortic angulation**
- **Valve selection: self-expanding supra-annular valve vs. infra-annular valve**
 - Supra-annular valves are generally accepted for small annuli to achieve lower residual gradients.
 - The patient's severely angulated aorta and horizontal orientation raised significant concerns about device deliverability and alignment.
 - **The Navitor valve, with its flexible delivery system and low profile, was considered advantageous in this extremely tortuous anatomy.**
- **Valve fracture timing: pre-TAVR vs. post-TAVR fracture**
 - Current consensus generally favors post-deployment fracture to reduce the risk of acute aortic regurgitation
 - Our team opted for **pre-dilatation to ensure full expansion of the small, rigid surgical valve frame, thus improving deliverability in a severely angulated aorta.**
- **Snare-assisted device system delivery**
 - To overcome this point (failed advancement of the Navitor valve), **an ipsilateral snare technique was employed.**

- **Like other cases using snaring or BVF for difficult VIV-TAVR, this case demonstrates that combined BVF and snare technique may serve as a feasible bailout strategy when standard delivery fails in anatomically challenging cases.**