HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 1 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Std 1 MD 1.2	**3.430**	**12516.7299**	**1.09**	**5393.6**
Std 2 MD 1.2	**3.414**	**12506.3845**	**1.13**	**5445.5**
Std 3 MD 1.2	**3.426**	**12484.7359**	**1.11**	**5477.8**
Std 4 MD 1.2	**3.408**	**12509.8093**	**1.12**	**5474.1**
Std 5 MD 1.2	**3.409**	**12517.4184**	**1.12**	**5438.6**
Jardiance F0006H26A MD 1.2 5' 1	3.417	8582.8984	1.14	5481.6
Jardiance F0006H26A MD 1.2 5' 2	3.404	8806.9938	1.13	5419.7
Jardiance F0006H26A MD 1.2 5' 3	3.429	8716.4087	1.12	5427.4
Jardiance F0006H26A MD 1.2 5' 4	3.417	8646.9581	1.14	5363.9
Jardiance F0006H26A MD 1.2 5' 5	3.435	8542.6461	1.11	5354.6
Jardiance F0006H26A MD 1.2 5' 6	3.434	8466.5169	1.10	5490.9
Jardiance F0006H26A MD 1.2 5' 7	3.428	8448.8602	1.09	5465.2
Jardiance F0006H26A MD 1.2 5' 8	3.420	8754.4788	1.11	5476.9
Jardiance F0006H26A MD 1.2 5' 9	3.404	8583.6119	1.12	5425.7
Jardiance F0006H26A MD 1.2 5' 10	3.421	8659.0587	1.13	5393.3
Jardiance F0006H26A MD 1.2 5' 11	3.410	8422.9295	1.12	5472.0
Jardiance F0006H26A MD 1.2 5' 12	3.407	8542.7026	1.15	5433.9
Jardiance F0006H26A MD 1.2 10' 1	3.411	10420.5117	1.09	5352.3
Jardiance F0006H26A MD 1.2 10' 2	3.421	10677.4679	1.13	5370.2
Jardiance F0006H26A MD 1.2 10' 3	3.435	10717.6655	1.11	5420.5
Jardiance F0006H26A MD 1.2 10' 4	3.407	10600.0108	1.14	5414.1
Jardiance F0006H26A MD 1.2 10' 5	3.403	10541.7119	1.12	5398.6
Jardiance F0006H26A MD 1.2 10' 6	3.440	10579.8359	1.14	5386.1
Jardiance F0006H26A MD 1.2 10' 7	3.421	10606.6396	1.15	5398.9
Jardiance F0006H26A MD 1.2 10' 8	3.425	10731.3393	1.09	5383.8
Jardiance F0006H26A MD 1.2 10' 9	3.426	10427.8225	1.08	5422.1
Jardiance F0006H26A MD 1.2 10' 10	3.424	10501.4555	1.13	5422.6
Jardiance F0006H26A MD 1.2 10' 11	3.427	10827.8821	1.13	5499.4
Jardiance F0006H26A MD 1.2 10' 12	3.417	10687.5297	1.14	5361.4
Jardiance F0006H26A MD 1.2 15' 1	3.407	11522.2893	1.10	5369.9
Jardiance F0006H26A MD 1.2 15' 2	3.428	11674.1861	1.11	5362.9
Jardiance F0006H26A MD 1.2 15' 3	3.402	11470.0549	1.09	5503.9
Jardiance F0006H26A MD 1.2 15' 4	3.433	11653.3193	1.13	5393.1
Jardiance F0006H26A MD 1.2 15' 5	3.402	11607.1448	1.14	5379.0
Jardiance F0006H26A MD 1.2 15' 6	3.423	11557.8713	1.10	5403.3
Jardiance F0006H26A MD 1.2 15' 7	3.427	11612.4683	1.10	5501.2
Jardiance F0006H26A MD 1.2 15' 8	3.433	11532.2444	1.10	5505.2
Jardiance F0006H26A MD 1.2 15' 9	3.416	11631.8810	1.12	5416.1

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 2 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 1.2 15' 10	3.408	11541.6292	1.13	5382.6
Jardiance F0006H26A MD 1.2 15' 11	3.422	11653.8961	1.09	5505.7
Jardiance F0006H26A MD 1.2 15' 12	3.424	11390.5103	1.12	5364.9
Jardiance F0006H26A MD 1.2 20' 1	3.440	12073.9244	1.14	5429.7
Jardiance F0006H26A MD 1.2 20' 2	3.429	11938.9705	1.15	5389.0
Jardiance F0006H26A MD 1.2 20' 3	3.422	12013.1733	1.14	5488.2
Jardiance F0006H26A MD 1.2 20' 4	3.411	12052.8310	1.12	5481.3
Jardiance F0006H26A MD 1.2 20' 5	3.411	11939.5846	1.11	5373.8
Jardiance F0006H26A MD 1.2 20' 6	3.422	11933.3246	1.11	5401.5
Jardiance F0006H26A MD 1.2 20' 7	3.408	12033.4685	1.13	5357.0
Jardiance F0006H26A MD 1.2 20' 8	3.430	12085.2899	1.13	5481.3
Jardiance F0006H26A MD 1.2 20' 9	3.439	11880.6085	1.12	5489.2
Jardiance F0006H26A MD 1.2 20' 10	3.435	11954.4155	1.13	5415.6
Jardiance F0006H26A MD 1.2 20' 11	3.426	11873.4140	1.10	5460.8
Jardiance F0006H26A MD 1.2 20' 12	3.426	12151.1239	1.11	5400.7
BS Mid 1 MD 1.2	**3.427**	**12510.3761**	**1.12**	**5374.7**
BS Mid 2 MD 1.2	**3.407**	**12505.2464**	**1.13**	**5499.7**
Jardiance F0006H26A MD 1.2 30' 1	3.430	12187.3107	1.14	5421.0
Jardiance F0006H26A MD 1.2 30' 2	3.428	12302.9892	1.10	5410.0
Jardiance F0006H26A MD 1.2 30' 3	3.406	12274.8082	1.09	5362.6
Jardiance F0006H26A MD 1.2 30' 4	3.435	12259.9162	1.09	5423.8
Jardiance F0006H26A MD 1.2 30' 5	3.410	12136.6453	1.14	5458.4
Jardiance F0006H26A MD 1.2 30' 6	3.406	12218.7553	1.12	5417.9
Jardiance F0006H26A MD 1.2 30' 7	3.410	12295.2689	1.11	5388.0
Jardiance F0006H26A MD 1.2 30' 8	3.423	12324.1464	1.15	5455.3
Jardiance F0006H26A MD 1.2 30' 9	3.411	12297.0473	1.08	5447.0
Jardiance F0006H26A MD 1.2 30' 10	3.420	12333.2789	1.12	5406.4
Jardiance F0006H26A MD 1.2 30' 11	3.425	12309.0712	1.13	5484.1
Jardiance F0006H26A MD 1.2 30' 12	3.425	12363.4115	1.11	5442.7
Jardiance F0006H26A MD 1.2 45' 1	3.429	12298.4022	1.09	5371.1
Jardiance F0006H26A MD 1.2 45' 2	3.422	12538.4553	1.14	5391.1
Jardiance F0006H26A MD 1.2 45' 3	3.433	12428.2643	1.10	5368.4
Jardiance F0006H26A MD 1.2 45' 4	3.439	12356.3997	1.13	5506.8
Jardiance F0006H26A MD 1.2 45' 5	3.439	12482.7283	1.08	5435.9
Jardiance F0006H26A MD 1.2 45' 6	3.418	12462.7652	1.14	5500.4
Jardiance F0006H26A MD 1.2 45' 7	3.404	12321.5466	1.12	5496.4
Jardiance F0006H26A MD 1.2 45' 8	3.409	12329.5768	1.10	5415.3
Jardiance F0006H26A MD 1.2 45' 9	3.406	12505.3669	1.09	5492.0

Confidential & Proprietary - Generated for UDT Studio

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 3 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 1.2 45' 10	3.405	12412.6105	1.15	5417.1
Jardiance F0006H26A MD 1.2 45' 11	3.430	12302.5317	1.11	5415.3
Jardiance F0006H26A MD 1.2 45' 12	3.438	12279.6387	1.15	5486.9
Jardiance F0006H26A MD 1.2 60' 1	3.404	12562.9786	1.11	5413.7
Jardiance F0006H26A MD 1.2 60' 2	3.408	12421.5831	1.13	5452.7
Jardiance F0006H26A MD 1.2 60' 3	3.424	12386.9862	1.09	5440.3
Jardiance F0006H26A MD 1.2 60' 4	3.403	12543.1258	1.14	5402.6
Jardiance F0006H26A MD 1.2 60' 5	3.406	12586.0267	1.09	5370.8
Jardiance F0006H26A MD 1.2 60' 6	3.410	12412.4506	1.08	5476.4
Jardiance F0006H26A MD 1.2 60' 7	3.410	12426.2271	1.13	5434.2
Jardiance F0006H26A MD 1.2 60' 8	3.430	12508.3595	1.09	5430.6
Jardiance F0006H26A MD 1.2 60' 9	3.438	12411.0120	1.12	5423.3
Jardiance F0006H26A MD 1.2 60' 10	3.439	12386.2040	1.13	5428.4
Jardiance F0006H26A MD 1.2 60' 11	3.436	12375.6611	1.08	5394.0
Jardiance F0006H26A MD 1.2 60' 12	3.425	12518.1872	1.12	5422.6
BS End 1 MD 1.2	**3.423**	**12482.3639**	**1.12**	**5460.9**
BS End 2 MD 1.2	**3.425**	**12511.8344**	**1.12**	**5378.0**
Std 1 MD 4.5	**3.419**	**12510.9688**	**1.11**	**5431.6**
Std 2 MD 4.5	**3.407**	**12484.1569**	**1.10**	**5457.8**
Std 3 MD 4.5	**3.411**	**12503.8398**	**1.10**	**5490.6**
Std 4 MD 4.5	**3.427**	**12512.9880**	**1.13**	**5408.3**
Std 5 MD 4.5	**3.416**	**12502.2194**	**1.10**	**5455.3**
Jardiance F0006H26A MD 4.5 5' 1	3.411	4192.2849	1.11	5448.5
Jardiance F0006H26A MD 4.5 5' 2	3.410	4394.3402	1.12	5400.0
Jardiance F0006H26A MD 4.5 5' 3	3.431	4378.1237	1.10	5353.7
Jardiance F0006H26A MD 4.5 5' 4	3.438	4484.7593	1.10	5358.1
Jardiance F0006H26A MD 4.5 5' 5	3.420	4527.3845	1.12	5485.5
Jardiance F0006H26A MD 4.5 5' 6	3.417	4554.5294	1.11	5414.1
Jardiance F0006H26A MD 4.5 5' 7	3.423	4283.9083	1.12	5472.3
Jardiance F0006H26A MD 4.5 5' 8	3.411	4555.3912	1.15	5413.0
Jardiance F0006H26A MD 4.5 5' 9	3.432	4444.1117	1.09	5431.5
Jardiance F0006H26A MD 4.5 5' 10	3.420	4142.8413	1.11	5423.1
Jardiance F0006H26A MD 4.5 5' 11	3.424	4340.0352	1.09	5508.5
Jardiance F0006H26A MD 4.5 5' 12	3.430	4283.7123	1.12	5463.4
Jardiance F0006H26A MD 4.5 10' 1	3.412	6965.0507	1.15	5443.4
Jardiance F0006H26A MD 4.5 10' 2	3.432	7018.4103	1.11	5504.7
Jardiance F0006H26A MD 4.5 10' 3	3.431	7020.0014	1.12	5377.3
Jardiance F0006H26A MD 4.5 10' 4	3.425	7038.2187	1.13	5355.1

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 4 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 4.5 10' 5	3.438	6633.2974	1.09	5460.5
Jardiance F0006H26A MD 4.5 10' 6	3.416	7131.3390	1.13	5360.0
Jardiance F0006H26A MD 4.5 10' 7	3.406	6768.6473	1.12	5400.2
Jardiance F0006H26A MD 4.5 10' 8	3.407	6976.2431	1.13	5465.3
Jardiance F0006H26A MD 4.5 10' 9	3.426	6989.4108	1.10	5439.3
Jardiance F0006H26A MD 4.5 10' 10	3.409	6978.5479	1.09	5462.8
Jardiance F0006H26A MD 4.5 10' 11	3.413	7005.2845	1.10	5369.5
Jardiance F0006H26A MD 4.5 10' 12	3.427	7029.7597	1.15	5478.7
Jardiance F0006H26A MD 4.5 15' 1	3.412	8735.2153	1.11	5391.0
Jardiance F0006H26A MD 4.5 15' 2	3.419	8683.2556	1.13	5372.0
Jardiance F0006H26A MD 4.5 15' 3	3.425	8716.7587	1.12	5452.1
Jardiance F0006H26A MD 4.5 15' 4	3.402	8766.3463	1.13	5350.8
Jardiance F0006H26A MD 4.5 15' 5	3.422	8868.5077	1.14	5426.4
Jardiance F0006H26A MD 4.5 15' 6	3.408	8848.8431	1.14	5364.3
Jardiance F0006H26A MD 4.5 15' 7	3.423	8829.4252	1.14	5378.0
Jardiance F0006H26A MD 4.5 15' 8	3.435	8772.5605	1.11	5429.3
Jardiance F0006H26A MD 4.5 15' 9	3.437	9006.4516	1.08	5440.9
Jardiance F0006H26A MD 4.5 15' 10	3.404	8855.7116	1.12	5429.0
Jardiance F0006H26A MD 4.5 15' 11	3.438	8838.1075	1.09	5376.0
Jardiance F0006H26A MD 4.5 15' 12	3.418	8605.8278	1.14	5440.8
Jardiance F0006H26A MD 4.5 20' 1	3.430	9932.1055	1.14	5432.4
Jardiance F0006H26A MD 4.5 20' 2	3.403	9847.7392	1.10	5389.7
Jardiance F0006H26A MD 4.5 20' 3	3.410	9919.8264	1.10	5465.7
Jardiance F0006H26A MD 4.5 20' 4	3.400	9999.4589	1.09	5399.1
Jardiance F0006H26A MD 4.5 20' 5	3.416	10036.5873	1.10	5368.8
Jardiance F0006H26A MD 4.5 20' 6	3.437	9872.1556	1.09	5374.5
Jardiance F0006H26A MD 4.5 20' 7	3.430	9913.1505	1.11	5436.0
Jardiance F0006H26A MD 4.5 20' 8	3.427	9750.4986	1.14	5363.1
Jardiance F0006H26A MD 4.5 20' 9	3.418	9837.4620	1.08	5437.6
Jardiance F0006H26A MD 4.5 20' 10	3.420	9813.4829	1.12	5368.6
Jardiance F0006H26A MD 4.5 20' 11	3.408	9803.6748	1.14	5422.5
Jardiance F0006H26A MD 4.5 20' 12	3.421	9753.9536	1.14	5465.7
BS Mid 1 MD 4.5	**3.417**	**12512.6581**	**1.12**	**5445.2**
BS Mid 2 MD 4.5	**3.406**	**12493.7812**	**1.08**	**5499.2**
Jardiance F0006H26A MD 4.5 30' 1	3.437	10924.1507	1.11	5473.0
Jardiance F0006H26A MD 4.5 30' 2	3.431	11091.1644	1.15	5411.2
Jardiance F0006H26A MD 4.5 30' 3	3.435	11013.5851	1.13	5436.3
Jardiance F0006H26A MD 4.5 30' 4	3.426	10820.1786	1.12	5466.5

Confidential & Proprietary - Generated for UDT Studio

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 5 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 4.5 30' 5	3.409	10688.0030	1.13	5437.1
Jardiance F0006H26A MD 4.5 30' 6	3.421	10950.6029	1.10	5500.0
Jardiance F0006H26A MD 4.5 30' 7	3.429	10915.6579	1.13	5491.4
Jardiance F0006H26A MD 4.5 30' 8	3.401	10750.2519	1.12	5462.5
Jardiance F0006H26A MD 4.5 30' 9	3.402	10965.2318	1.10	5488.2
Jardiance F0006H26A MD 4.5 30' 10	3.402	10895.2634	1.13	5401.0
Jardiance F0006H26A MD 4.5 30' 11	3.424	10929.8522	1.09	5455.6
Jardiance F0006H26A MD 4.5 30' 12	3.409	10791.3203	1.14	5495.9
Jardiance F0006H26A MD 4.5 45' 1	3.425	11651.3978	1.15	5457.4
Jardiance F0006H26A MD 4.5 45' 2	3.431	11584.7210	1.11	5470.7
Jardiance F0006H26A MD 4.5 45' 3	3.419	11712.4506	1.14	5496.4
Jardiance F0006H26A MD 4.5 45' 4	3.422	11772.9606	1.10	5382.5
Jardiance F0006H26A MD 4.5 45' 5	3.426	11814.6680	1.13	5380.5
Jardiance F0006H26A MD 4.5 45' 6	3.411	11730.1745	1.09	5471.9
Jardiance F0006H26A MD 4.5 45' 7	3.429	11678.1832	1.10	5503.1
Jardiance F0006H26A MD 4.5 45' 8	3.401	11840.6990	1.15	5426.6
Jardiance F0006H26A MD 4.5 45' 9	3.410	11639.4004	1.12	5474.1
Jardiance F0006H26A MD 4.5 45' 10	3.415	11631.1205	1.12	5429.6
Jardiance F0006H26A MD 4.5 45' 11	3.404	11781.0510	1.14	5409.3
Jardiance F0006H26A MD 4.5 45' 12	3.417	11608.9978	1.10	5499.6
Jardiance F0006H26A MD 4.5 60' 1	3.423	12002.8676	1.11	5483.3
Jardiance F0006H26A MD 4.5 60' 2	3.420	12159.7805	1.14	5504.6
Jardiance F0006H26A MD 4.5 60' 3	3.426	12038.0420	1.09	5465.3
Jardiance F0006H26A MD 4.5 60' 4	3.420	12210.1298	1.08	5355.4
Jardiance F0006H26A MD 4.5 60' 5	3.434	12244.5176	1.11	5484.8
Jardiance F0006H26A MD 4.5 60' 6	3.409	12053.6784	1.11	5448.9
Jardiance F0006H26A MD 4.5 60' 7	3.427	12119.3663	1.08	5438.3
Jardiance F0006H26A MD 4.5 60' 8	3.400	12073.1927	1.08	5480.1
Jardiance F0006H26A MD 4.5 60' 9	3.427	12235.3349	1.09	5373.8
Jardiance F0006H26A MD 4.5 60' 10	3.408	12152.4864	1.11	5409.7
Jardiance F0006H26A MD 4.5 60' 11	3.416	12193.6950	1.13	5485.7
Jardiance F0006H26A MD 4.5 60' 12	3.426	12219.7822	1.11	5438.4
BS End 1 MD 4.5	**3.407**	**12478.4158**	**1.09**	**5415.0**
BS End 2 MD 4.5	**3.418**	**12512.5126**	**1.11**	**5461.5**
Std 1 MD 6.8	**3.408**	**12502.6608**	**1.12**	**5422.0**
Std 2 MD 6.8	**3.420**	**12483.5854**	**1.12**	**5424.2**
Std 3 MD 6.8	**3.419**	**12495.7873**	**1.09**	**5520.4**
Std 4 MD 6.8	**3.431**	**12501.9031**	**1.13**	**5493.7**

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 6 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Std 5 MD 6.8	**3.418**	**12493.5849**	**1.10**	**5490.8**
Jardiance F0006H26A MD 6.8 5' 1	3.411	5591.6294	1.14	5378.7
Jardiance F0006H26A MD 6.8 5' 2	3.432	5995.0118	1.11	5395.2
Jardiance F0006H26A MD 6.8 5' 3	3.434	5696.7022	1.09	5440.2
Jardiance F0006H26A MD 6.8 5' 4	3.412	5780.5477	1.11	5473.6
Jardiance F0006H26A MD 6.8 5' 5	3.408	5927.6074	1.08	5415.6
Jardiance F0006H26A MD 6.8 5' 6	3.409	5800.1366	1.15	5350.2
Jardiance F0006H26A MD 6.8 5' 7	3.426	5899.6240	1.12	5388.1
Jardiance F0006H26A MD 6.8 5' 8	3.420	5748.5579	1.09	5497.4
Jardiance F0006H26A MD 6.8 5' 9	3.436	5802.3143	1.15	5363.6
Jardiance F0006H26A MD 6.8 5' 10	3.404	5707.2898	1.12	5419.6
Jardiance F0006H26A MD 6.8 5' 11	3.439	5605.7295	1.14	5462.1
Jardiance F0006H26A MD 6.8 5' 12	3.429	5625.6469	1.09	5474.8
Jardiance F0006H26A MD 6.8 10' 1	3.423	8071.9720	1.11	5435.7
Jardiance F0006H26A MD 6.8 10' 2	3.439	8126.5873	1.09	5375.2
Jardiance F0006H26A MD 6.8 10' 3	3.427	8426.9003	1.14	5402.4
Jardiance F0006H26A MD 6.8 10' 4	3.404	8161.0390	1.13	5391.6
Jardiance F0006H26A MD 6.8 10' 5	3.440	8109.2632	1.13	5390.0
Jardiance F0006H26A MD 6.8 10' 6	3.418	8245.2131	1.14	5507.2
Jardiance F0006H26A MD 6.8 10' 7	3.405	8342.3359	1.13	5443.2
Jardiance F0006H26A MD 6.8 10' 8	3.410	8212.5708	1.14	5420.9
Jardiance F0006H26A MD 6.8 10' 9	3.405	8500.6437	1.08	5423.8
Jardiance F0006H26A MD 6.8 10' 10	3.407	8400.7919	1.08	5459.3
Jardiance F0006H26A MD 6.8 10' 11	3.407	8091.2952	1.14	5355.5
Jardiance F0006H26A MD 6.8 10' 12	3.425	8138.3156	1.12	5359.8
Jardiance F0006H26A MD 6.8 15' 1	3.411	9752.6087	1.12	5450.9
Jardiance F0006H26A MD 6.8 15' 2	3.432	10094.9245	1.11	5437.1
Jardiance F0006H26A MD 6.8 15' 3	3.419	10027.1533	1.09	5401.0
Jardiance F0006H26A MD 6.8 15' 4	3.409	10063.9631	1.15	5434.7
Jardiance F0006H26A MD 6.8 15' 5	3.431	10133.0070	1.13	5392.6
Jardiance F0006H26A MD 6.8 15' 6	3.433	9963.0525	1.10	5485.4
Jardiance F0006H26A MD 6.8 15' 7	3.432	9936.6522	1.09	5444.8
Jardiance F0006H26A MD 6.8 15' 8	3.432	10029.0054	1.13	5501.6
Jardiance F0006H26A MD 6.8 15' 9	3.423	9881.3667	1.14	5428.8
Jardiance F0006H26A MD 6.8 15' 10	3.412	9818.5329	1.10	5434.2
Jardiance F0006H26A MD 6.8 15' 11	3.400	9964.5581	1.13	5497.0
Jardiance F0006H26A MD 6.8 15' 12	3.434	9766.2836	1.14	5473.9
Jardiance F0006H26A MD 6.8 20' 1	3.434	10969.5483	1.15	5413.3

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 7 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 6.8 20' 2	3.404	10589.3676	1.15	5476.7
Jardiance F0006H26A MD 6.8 20' 3	3.438	10727.3670	1.13	5498.6
Jardiance F0006H26A MD 6.8 20' 4	3.431	10679.4693	1.10	5460.2
Jardiance F0006H26A MD 6.8 20' 5	3.405	10944.3073	1.11	5509.1
Jardiance F0006H26A MD 6.8 20' 6	3.439	10647.9431	1.15	5385.3
Jardiance F0006H26A MD 6.8 20' 7	3.423	10798.5275	1.10	5507.2
Jardiance F0006H26A MD 6.8 20' 8	3.407	10769.0466	1.12	5491.9
Jardiance F0006H26A MD 6.8 20' 9	3.418	10603.9828	1.10	5367.3
Jardiance F0006H26A MD 6.8 20' 10	3.400	10788.0003	1.14	5428.9
Jardiance F0006H26A MD 6.8 20' 11	3.415	10663.5490	1.14	5454.2
Jardiance F0006H26A MD 6.8 20' 12	3.427	10552.5972	1.10	5401.0
BS Mid 1 MD 6.8	**3.434**	**12482.9680**	**1.09**	**5418.8**
BS Mid 2 MD 6.8	**3.427**	**12501.8819**	**1.12**	**5382.1**
Jardiance F0006H26A MD 6.8 30' 1	3.424	11542.3541	1.15	5376.4
Jardiance F0006H26A MD 6.8 30' 2	3.436	11361.9283	1.09	5461.6
Jardiance F0006H26A MD 6.8 30' 3	3.429	11538.1763	1.10	5357.9
Jardiance F0006H26A MD 6.8 30' 4	3.423	11547.0866	1.11	5397.1
Jardiance F0006H26A MD 6.8 30' 5	3.426	11485.7650	1.09	5396.7
Jardiance F0006H26A MD 6.8 30' 6	3.403	11578.1826	1.11	5455.8
Jardiance F0006H26A MD 6.8 30' 7	3.404	11494.7615	1.11	5421.3
Jardiance F0006H26A MD 6.8 30' 8	3.411	11464.2418	1.08	5470.4
Jardiance F0006H26A MD 6.8 30' 9	3.427	11399.6312	1.15	5444.4
Jardiance F0006H26A MD 6.8 30' 10	3.422	11673.7092	1.12	5488.7
Jardiance F0006H26A MD 6.8 30' 11	3.419	11609.5813	1.14	5464.3
Jardiance F0006H26A MD 6.8 30' 12	3.420	11693.0243	1.12	5369.2
Jardiance F0006H26A MD 6.8 45' 1	3.430	12054.4537	1.14	5370.2
Jardiance F0006H26A MD 6.8 45' 2	3.417	12156.1243	1.12	5483.5
Jardiance F0006H26A MD 6.8 45' 3	3.410	12198.6191	1.11	5413.6
Jardiance F0006H26A MD 6.8 45' 4	3.413	11978.6587	1.14	5449.7
Jardiance F0006H26A MD 6.8 45' 5	3.421	11984.5524	1.12	5471.3
Jardiance F0006H26A MD 6.8 45' 6	3.438	12137.2650	1.10	5435.7
Jardiance F0006H26A MD 6.8 45' 7	3.414	11986.6107	1.12	5356.0
Jardiance F0006H26A MD 6.8 45' 8	3.411	11968.3959	1.08	5492.3
Jardiance F0006H26A MD 6.8 45' 9	3.413	12024.3480	1.14	5353.2
Jardiance F0006H26A MD 6.8 45' 10	3.433	11940.5773	1.13	5365.7
Jardiance F0006H26A MD 6.8 45' 11	3.435	12027.0558	1.14	5393.1
Jardiance F0006H26A MD 6.8 45' 12	3.417	12000.4684	1.14	5468.6
Jardiance F0006H26A MD 6.8 60' 1	3.430	12507.3825	1.09	5399.8

Confidential & Proprietary - Generated for UDT Studio

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 8 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD 6.8 60' 2	3.420	12339.2044	1.08	5380.8
Jardiance F0006H26A MD 6.8 60' 3	3.433	12426.6239	1.14	5462.7
Jardiance F0006H26A MD 6.8 60' 4	3.411	12460.5797	1.09	5354.9
Jardiance F0006H26A MD 6.8 60' 5	3.415	12369.3233	1.12	5406.0
Jardiance F0006H26A MD 6.8 60' 6	3.407	12510.4086	1.10	5411.9
Jardiance F0006H26A MD 6.8 60' 7	3.407	12461.8926	1.13	5458.4
Jardiance F0006H26A MD 6.8 60' 8	3.411	12300.7925	1.13	5351.4
Jardiance F0006H26A MD 6.8 60' 9	3.436	12409.7395	1.10	5478.8
Jardiance F0006H26A MD 6.8 60' 10	3.412	12507.3129	1.11	5405.8
Jardiance F0006H26A MD 6.8 60' 11	3.410	12398.5533	1.11	5366.4
Jardiance F0006H26A MD 6.8 60' 12	3.420	12454.7262	1.13	5447.4
BS End 1 MD 6.8	**3.414**	**12482.7566**	**1.08**	**5478.5**
BS End 2 MD 6.8	**3.424**	**12515.9056**	**1.09**	**5442.4**
Std 1 MD Air	**3.428**	**12497.7563**	**1.10**	**5482.5**
Std 2 MD Air	**3.419**	**12501.2048**	**1.09**	**5444.7**
Std 3 MD Air	**3.429**	**12501.6461**	**1.13**	**5410.2**
Std 4 MD Air	**3.425**	**12514.9101**	**1.12**	**5492.1**
Std 5 MD Air	**3.416**	**12516.1807**	**1.11**	**5467.7**
Jardiance F0006H26A MD Air 5' 1	3.405	3576.3691	1.15	5398.8
Jardiance F0006H26A MD Air 5' 2	3.430	3747.3879	1.13	5432.0
Jardiance F0006H26A MD Air 5' 3	3.406	3720.0968	1.09	5477.9
Jardiance F0006H26A MD Air 5' 4	3.434	3969.8844	1.11	5437.9
Jardiance F0006H26A MD Air 5' 5	3.412	3413.8024	1.11	5402.9
Jardiance F0006H26A MD Air 5' 6	3.405	3587.7443	1.14	5400.8
Jardiance F0006H26A MD Air 5' 7	3.418	3652.0561	1.12	5408.3
Jardiance F0006H26A MD Air 5' 8	3.413	3472.6894	1.10	5468.4
Jardiance F0006H26A MD Air 5' 9	3.417	3659.3869	1.14	5424.1
Jardiance F0006H26A MD Air 5' 10	3.414	3638.7916	1.12	5484.7
Jardiance F0006H26A MD Air 5' 11	3.434	3586.8793	1.14	5362.2
Jardiance F0006H26A MD Air 5' 12	3.403	3531.7168	1.12	5414.8
Jardiance F0006H26A MD Air 10' 1	3.424	6364.5175	1.15	5425.5
Jardiance F0006H26A MD Air 10' 2	3.431	6065.6066	1.13	5396.6
Jardiance F0006H26A MD Air 10' 3	3.419	6176.7821	1.13	5485.7
Jardiance F0006H26A MD Air 10' 4	3.432	6107.1914	1.11	5378.7
Jardiance F0006H26A MD Air 10' 5	3.438	6478.2859	1.11	5407.4
Jardiance F0006H26A MD Air 10' 6	3.409	6019.8775	1.08	5383.1
Jardiance F0006H26A MD Air 10' 7	3.436	6162.4972	1.08	5506.9
Jardiance F0006H26A MD Air 10' 8	3.413	6299.1109	1.14	5392.1

Confidential & Proprietary - Generated for UDT Studio

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M

Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) |

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD Air 10' 9	3.422	6184.0968	1.13	5483.7
Jardiance F0006H26A MD Air 10' 10	3.421	6360.3083	1.10	5488.8
Jardiance F0006H26A MD Air 10' 11	3.419	6259.0304	1.09	5376.5
Jardiance F0006H26A MD Air 10' 12	3.403	6126.1890	1.11	5464.8
Jardiance F0006H26A MD Air 15' 1	3.405	8104.9013	1.11	5426.6
Jardiance F0006H26A MD Air 15' 2	3.408	8071.5591	1.08	5441.8
Jardiance F0006H26A MD Air 15' 3	3.406	7930.7268	1.08	5391.3
Jardiance F0006H26A MD Air 15' 4	3.407	8239.1387	1.08	5465.1
Jardiance F0006H26A MD Air 15' 5	3.426	7840.6682	1.15	5488.8
Jardiance F0006H26A MD Air 15' 6	3.407	8208.2621	1.14	5379.1
Jardiance F0006H26A MD Air 15' 7	3.426	8071.2064	1.11	5444.5
Jardiance F0006H26A MD Air 15' 8	3.409	7872.4228	1.14	5430.7
Jardiance F0006H26A MD Air 15' 9	3.417	7805.2564	1.12	5356.2
Jardiance F0006H26A MD Air 15' 10	3.431	7919.5346	1.09	5446.6
Jardiance F0006H26A MD Air 15' 11	3.433	8289.1440	1.12	5409.3
Jardiance F0006H26A MD Air 15' 12	3.402	7979.0713	1.14	5411.0
Jardiance F0006H26A MD Air 20' 1	3.424	9488.8967	1.09	5371.4
Jardiance F0006H26A MD Air 20' 2	3.419	9360.3968	1.14	5415.4
Jardiance F0006H26A MD Air 20' 3	3.408	9486.1177	1.09	5351.2
Jardiance F0006H26A MD Air 20' 4	3.426	9295.6613	1.09	5454.3
Jardiance F0006H26A MD Air 20' 5	3.410	9386.2980	1.10	5458.9
Jardiance F0006H26A MD Air 20' 6	3.430	9392.3631	1.11	5438.8
Jardiance F0006H26A MD Air 20' 7	3.410	9202.2944	1.13	5486.8
Jardiance F0006H26A MD Air 20' 8	3.426	9372.7967	1.09	5384.3
Jardiance F0006H26A MD Air 20' 9	3.416	9288.9092	1.13	5375.7
Jardiance F0006H26A MD Air 20' 10	3.439	9000.1053	1.14	5419.7
Jardiance F0006H26A MD Air 20' 11	3.403	9057.6786	1.11	5507.5
Jardiance F0006H26A MD Air 20' 12	3.425	9401.0780	1.15	5481.2
BS Mid 1 MD Air	**3.424**	**12497.5589**	**1.09**	**5455.4**
BS Mid 2 MD Air	**3.429**	**12491.4403**	**1.08**	**5447.9**
Jardiance F0006H26A MD Air 30' 1	3.421	10284.4630	1.10	5350.0
Jardiance F0006H26A MD Air 30' 2	3.406	10675.9459	1.11	5443.8
Jardiance F0006H26A MD Air 30' 3	3.421	10380.8792	1.10	5482.8
Jardiance F0006H26A MD Air 30' 4	3.437	10568.8464	1.12	5381.1
Jardiance F0006H26A MD Air 30' 5	3.437	10410.0368	1.09	5361.6
Jardiance F0006H26A MD Air 30' 6	3.412	10688.8746	1.14	5480.1
Jardiance F0006H26A MD Air 30' 7	3.423	10664.8354	1.13	5471.2
Jardiance F0006H26A MD Air 30' 8	3.438	10448.7920	1.08	5368.6

Confidential & Proprietary - Generated for UDT Studio

HPLC Sequence Summary Report - Dissolution Testing

Sequence: SEQ_REF_ALL_4MEDIA | Instrument: HPLC-02 | Operator: Analyst Lab | Method: DISS_EMPA.M
Medium: Multi-Media (pH 1.2, 4.5, 6.8, Air) | Batch: Jardiance (F0006H26A) | Page 10 of 10

Sample Name	RetTime	Area	USP Tail	Plates USP
Jardiance F0006H26A MD Air 30' 9	3.403	10661.9097	1.10	5400.5
Jardiance F0006H26A MD Air 30' 10	3.438	10329.6299	1.10	5364.9
Jardiance F0006H26A MD Air 30' 11	3.427	10429.4696	1.11	5406.6
Jardiance F0006H26A MD Air 30' 12	3.436	10556.2366	1.14	5429.8
Jardiance F0006H26A MD Air 45' 1	3.408	11455.7462	1.09	5390.0
Jardiance F0006H26A MD Air 45' 2	3.425	11140.7769	1.13	5456.1
Jardiance F0006H26A MD Air 45' 3	3.425	11375.1974	1.11	5449.3
Jardiance F0006H26A MD Air 45' 4	3.420	11544.1533	1.11	5507.5
Jardiance F0006H26A MD Air 45' 5	3.437	11190.3898	1.08	5477.9
Jardiance F0006H26A MD Air 45' 6	3.419	11320.1223	1.13	5378.0
Jardiance F0006H26A MD Air 45' 7	3.439	11357.7551	1.08	5383.7
Jardiance F0006H26A MD Air 45' 8	3.423	11289.8446	1.12	5387.9
Jardiance F0006H26A MD Air 45' 9	3.419	11307.2151	1.09	5462.1
Jardiance F0006H26A MD Air 45' 10	3.415	11336.3856	1.09	5442.1
Jardiance F0006H26A MD Air 45' 11	3.409	11147.0289	1.11	5410.9
Jardiance F0006H26A MD Air 45' 12	3.406	11348.8440	1.09	5429.9
Jardiance F0006H26A MD Air 60' 1	3.440	11910.7160	1.13	5398.0
Jardiance F0006H26A MD Air 60' 2	3.422	11809.7754	1.08	5480.2
Jardiance F0006H26A MD Air 60' 3	3.419	11827.4690	1.11	5439.7
Jardiance F0006H26A MD Air 60' 4	3.437	11620.5247	1.14	5504.1
Jardiance F0006H26A MD Air 60' 5	3.408	11858.9925	1.13	5406.8
Jardiance F0006H26A MD Air 60' 6	3.427	11880.1940	1.11	5395.8
Jardiance F0006H26A MD Air 60' 7	3.425	11649.8375	1.10	5451.0
Jardiance F0006H26A MD Air 60' 8	3.423	11782.4212	1.10	5400.7
Jardiance F0006H26A MD Air 60' 9	3.412	11783.6666	1.11	5488.0
Jardiance F0006H26A MD Air 60' 10	3.415	11948.0311	1.09	5353.7
Jardiance F0006H26A MD Air 60' 11	3.429	11969.1359	1.09	5376.7
Jardiance F0006H26A MD Air 60' 12	3.404	11799.5739	1.08	5413.5
BS End 1 MD Air	**3.428**	**12481.6631**	**1.09**	**5370.2**
BS End 2 MD Air	**3.425**	**12516.9607**	**1.10**	**5382.2**