

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

\\USER\UserProtocols\Yuhui\mtVASO\_anat\anat\_sVAPER\_dfa12.4\_dur180

TA: 9:18 PAT: 3 Voxel size: 0.8x0.8x0.9 mm Rel. SNR: 1.00 USER: Chai\VAPER\_rfdur2s

## Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

## Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R48.6 A11.7 H19.9
Orientation	S > T16.3 > C6.4
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	5 %
Slice oversampling	7.1 %
Slices per slab	28
FoV read	138.0 mm
FoV phase	100.0 %
Slice thickness	0.90 mm
TR	3321.2 ms
TE	28 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

## Contrast

Perfusion mode	Picore Q2TIPS
TI2	1100 ms
TI1	50 ms
TI1s	50 ms
Flip angle	25.0 deg
Fat suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	168
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	1100.0 ms
Flow limit	100 cm/s

## Resolution

Base resolution	168
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	7/8
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA

Accel. factor PE	3
Ref. lines PE	48
Accel. factor 3D	1
Ref. lines 3D	24
Reference scan mode	Separate
Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

## Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel R
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

## System

V32	Off
A32	On
Positioning mode	FIX
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	275.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R48.6 A11.7 H19.9
Orientation	S > T16.3 > C6.4
Rotation	0.00 deg
F >> H	138 mm
A >> P	138 mm
R >> L	26 mm

## Physio

1st Signal/Mode	None
-----------------	------

## BOLD

Motion correction	Off
Spatial filter	Off

## Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1102 Hz/Px
Free echo spacing	Off
Echo spacing	1.03 ms
EPI factor	168
RF pulse type	Normal

# SIEMENS MAGNETOM Investigational\_Device\_7T syngo MR B17

Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Read Diff Amp	0.0 mT/m
Phase Diff Amp	0.0 mT/m
Slice Diff Amp	0.0 mT/m
Dante puls # in 1st par	230
Dante puls # in 2nd par	36
Pulses FA in DANTE	12.4 degree
TAU in DANTE	1100 us
Vari readFA	0
Blank bef/aft DANTE-RF	50 us
Grad # bef DANTE	0
DANTE-RF dur	180 us
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	2.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2200 us
RF BWTP	25.0
EFFECTIVE TR	101 ms
PatPartitions	30
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	168
FlashRef BW	100 Hz/px
FlashRef TE	10000 us
FlashRef FA	5 deg
use CAIPI	Off