OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:		**Test Date:**		**Date Initiated:**	
Date of Incident:		**Test Name:**		**Sample ID:**	
Name of Analyst who Performed the Test:		**Sample / Active Name**		**Lot #:**	
		Dosage Form			
Description of Incident:		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS			
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis			
Personnel Discrepancy		**Comments**	
Analyst interviewed?	☐Yes ☐ No ☐ N/A		
Correct samples analyzed?	☐Yes ☐ No ☐ N/A		
Correct SOP / Test Method used/followed?	☐Yes ☐ No ☐ N/A		
Correct sample collection/technique?	☐Yes ☐ No ☐ N/A		
Are the Test Result Records Attached & Calculations Verified and Correct?	☐Yes ☐ No ☐ N/A		
Is the analyst qualified to perform the test?	☐Yes ☐ No ☐ N/A		
Is formula worksheet provided and verified?	☐Yes ☐ No ☐ N/A		
Does the processing method/method set conform to the required specification?	☐Yes ☐ No ☐ N/A		
Was the sample stored appropriately?	☐Yes ☐ No ☐ N/A		
Material Discrepancy		**Lot Number**	**Expiry Date**
Is the integrity of the sample(s) questionable?	☐Yes ☐ No ☐N/A		
Was the sample transported appropriately?	☐Yes ☐ No ☐N/A		
Are the other test results from this run under investigation?	☐Yes ☐ No ☐N/A		

OOS Number: OOS-

Appropriate/correct glassware used?	☐Yes ☐No ☐N/A		
Are reagents stored in proper condition and not expired?	☐Yes ☐No ☐N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐Yes ☐No ☐N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐Yes ☐No ☐N/A		
Equipment Discrepancy		**Equipment ID**	**Calibration Due Date**
Was the instrument set up correctly?	☐Yes ☐No ☐N/A		
Is the clean room facility certified?	☐Yes ☐No ☐N/A		
System Performance Check (if applicable)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 2% (PDA)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐Yes ☐No ☐N/A		
Do the environmental monitoring results conform to required standards?	☐Yes ☐No ☐N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐Yes ☐No ☐N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐Yes ☐No ☐N/A		

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?	☐ Yes ☐ No ☐ N/A		

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

- ☐ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: _____

Section D – Phase I Summary

Phase I Summary:

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?
Note: Root cause must be conclusive, verifiable, and well-documented.
☐ Yes
☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☐ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (*Print*)**:**	Signature/Date:
Lab Manager/Designee (*Print*)**:**	Signature/Date:

Section E – Quality Disposition

Can investigation be closed?
☐ Yes
☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (*Print*)**:**	Signature/Date:

Table 1: Reading analyst, Days of incubation, and Microbial ID for Sample under Investigation

Reading Analyst	Day of Incubation	Sample ID	Related Microbial ID	Media with growth	Microbial ID
CCD	Day 7	ETX-261002-0288	ETX-261012-0654	1 x 300mL FTM	*Bacillus paralicheniformis* (Gram-positive rod)

Environmental Monitoring (EM) Sampling Site	Frequency	Date	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM Bracketing 05Oct2026								
Personal (Left Touch and Right Touch)	Daily	02Oct2026	ES	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	05Oct2026	ES	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	06Oct2026	ES	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM Bracketing E001316 for 05Oct2026								
Surface Sampling of ISO 5 BSC E001316 (4 Locations)	Daily	02Oct2026	ALA	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001316 (4 Locations)	Daily	05Oct2026	ES	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001316 (4 Locations)	Daily	06Oct2026	ES	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001316	Daily	02Oct2026	ALA	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001316	Daily	05Oct2026	ES	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001316	Daily	06Oct2026	ES	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air and Surface Sampling of Cleanroom Suite 114 with Processing BSC For 05Oct2026								
Environmental Monitoring (EM) Sampling Site	**Frequency**	**Date**	**Analyst**	**Observation**	**EM Plate ETX Number**	**Microbial ID**		**Notes**
Active Air Sampling of Cleanrooms	Weekly	05Oct2026	SMO	No Growth	Not Applicable	Not Applicable		None
Surface Sampling of Cleanrooms	Weekly	05Oct2026	SMO	No Growth	Not Applicable	Not Applicable		None

Table 2: Environmental Monitoring from Celsis Processing Performed on 05Oct2026

Environmental Monitoring (EM) Sampling Site	Frequency	Date	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM Bracketing 12Oct2026								
Personal (Left Touch and Right Touch)	Daily	09Oct2026	CCD	Date Before Aliquoting	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	12Oct2026	CCD	Date of Aliquoting	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	13Oct2026	CCD	Date After Aliquoting	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM for 12Oct2026								
Surface Sampling of ISO 5 BSC E001798 (4 Locations)	Daily	09Oct2026	CCD	Date Before Aliquoting	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001798 (4 Locations)	Daily	12Oct2026	CCD	Date of Aliquoting	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001798 (4 Locations)	Daily	13Oct2026	CCD	Date After Aliquoting	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001798	Daily	09Oct2026	CCD	Date Before Aliquoting	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001798	Daily	12Oct2026	CCD	Date of Aliquoting	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001798	Daily	13Oct2026	CCD	Date After Aliquoting	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air and Surface Sampling of Cleanroom Suite 114 with Processing BSC For 12Oct2026								
Environmental Monitoring (EM) Sampling Site	**Frequency**	**Date**	**Analyst**	**Observation**	**EM Plate ETX Number**	**Microbial ID**		**Notes**
Active Air Sampling of Cleanrooms	Weekly	12Oct2026	SMO	6 CFUs (ISO 8 114)	ETX-261021-0578	*Moraxella osloensis* *Sporisorium graminicola* *Staphylococcus haemolyticus* *Staphylococcus hominis* *Kocuria indica* *Ustilago maydis* *Kocuria rhizophila*		None
Surface Sampling of Cleanrooms	Weekly	12Oct2026	SMO	No Growth	Not Applicable	Not Applicable		None

Table 3: Environmental Monitoring from Celsis Aliquoting Performed on 12Oct2026