

Table 1: Information for ETX-260805-0189 under investigation

| Processing Analyst | Reading Analyst    | Sample ID                       | Events | Confirmed Microbial Events | Morphology Description |
|--------------------|--------------------|---------------------------------|--------|----------------------------|------------------------|
| Sonal Uprety       | Varsha Subramanian | <a href="#">ETX-260805-0189</a> | 187    | 4                          | Rod-shaped Morphology  |

Table 2: Environmental Monitoring from Testing Performed on 07Aug26

| Environmental Monitoring (EM) Sampling Site                                                        | Frequency | Date (DDMMM YY) | Analyst (Initials) | Day /Week(s)            | Observation*                                                                              | Environmental Monitoring Plate ETX ID                 | Microbial ID                                                                                                                                                                                                                                   | Notes                              |
|----------------------------------------------------------------------------------------------------|-----------|-----------------|--------------------|-------------------------|-------------------------------------------------------------------------------------------|-------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------|
| <b>Personnel EM Bracketing</b>                                                                     |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Personal (Left Touch and Right Touch)                                                              | Daily     | 07Aug26         | SU                 | Date of Testing Scan    | 1 CFU on right touch                                                                      | ETX-260817-0447                                       | Micrococcus luteus                                                                                                                                                                                                                             | Plate was desiccated on 5 day read |
| Personal (Left Touch and Right Touch)                                                              | Daily     | 07Aug26         | GA                 | Date of Testing ScanC/O | No growth                                                                                 | N/A                                                   | N/A                                                                                                                                                                                                                                            | None                               |
| <b>Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC) E001313 and E001937</b> |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Surface Sampling of ISO 5 E001313 (4 locations)                                                    | Daily     | 07Aug26         | SU                 | Date of TestingScan     | No growth                                                                                 | N/A                                                   | N/A                                                                                                                                                                                                                                            | None                               |
| Surface Sampling of ISO 5 E001937 (4 locations)                                                    | Daily     | 07Aug26         | GA                 | Date of Testing ScanC/O | No growth                                                                                 | N/A                                                   | N/A                                                                                                                                                                                                                                            | None                               |
| Settling Sampling of ISO 5 E001313 (2 locations)                                                   | Daily     | 07Aug26         | SU                 | Date of TestingScan     | 2 CFUs on Sett1                                                                           | ETX-260817-0507                                       | Sporisorium graminicola<br>Ustilago maydis                                                                                                                                                                                                     | Plate was desiccated on 5 day read |
| Settling Sampling of ISO 5 E001937 (2 locations)                                                   | Daily     | 07Aug26         | GA                 | Date of Testing ScanC/O | No growth                                                                                 | N/A                                                   | N/A                                                                                                                                                                                                                                            | None                               |
| <b>Weekly Active Air Sampling Bracketing - Cleanroom 115 - CR1737</b>                              |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Active Air Sampling of Cleanrooms                                                                  | Weekly    | 06Aug26         | SMO                | Week of testing Scan    | 1 CFU on ISO 8 115                                                                        | ETX-260817-0370                                       | Gram (+) short rods                                                                                                                                                                                                                            | None                               |
| <b>Weekly Active Air Sampling Bracketing - Cleanroom 144 - CR1978 (L-Suite)</b>                    |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Active Air Sampling of Cleanrooms                                                                  | Weekly    | 07Aug26         | SMO                | Week of testing ScanC/O | 6 CFUs on 143 SecI air ISO 8 32 CFUs on 143 SecII air ISO 8, and 12 CFUs on 142 air ISO 8 | ETX-260817-0416<br>ETX-260817-0422<br>ETX-260817-0429 | Staphylococcus epidermidis<br>Staphylococcus capitis<br>Micrococcus luteus<br>Kocuria palustris<br><br>Corynebacterium segmentosum<br>Staphylococcus hominis<br>Staphylococcus aureus<br>Corynebacterium aurimucosum<br>Kytococcus sedentarius | None                               |
| <b>Surface Sampling of Anteroom and Cleanroom Bracketing - Cleanroom 115 - CR1737</b>              |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Surface Sampling of Cleanrooms                                                                     | Weekly    | 06Aug26         | SMO                | Week of testing Scan    | 2 CFU on floor in ISO 7 115A                                                              | ETX-260817-0366                                       | Bacillus megaterium                                                                                                                                                                                                                            | None                               |
| <b>Surface Sampling of Anteroom and Cleanroom Bracketing - Cleanroom 144 - CR1978 (L-Suite)</b>    |           |                 |                    |                         |                                                                                           |                                                       |                                                                                                                                                                                                                                                |                                    |
| Surface Sampling of Cleanrooms                                                                     | Weekly    | 07Aug26         | SMO                | Week of testing ScanC/O | No growth                                                                                 | N/A                                                   | N/A                                                                                                                                                                                                                                            | None                               |