

Table 1: Information for ETX-260805-0189 under investigation

Processing Analyst	Reading Analyst	Sample ID	Events	Confirmed Microbial Events	Morphology Description
Sonal Uprety	Varsha Subramanian	ETX-260805-0189	187	4	Rod-shaped Morphology

Table 2: Environmental Monitoring from Testing Performed on 07Aug26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMM YYYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing								
Personal (Left Touch and Right Touch)	Daily	07AUG 2026	SU	Date of Testing Scan	1 CFU on right touch	ETX-260817-0447	Micrococcus luteus	None
Personal (Left Touch and Right Touch)	Daily	07AUG 2026	GA	Date of Testing ScanC/O	No growth	N/A	N/A	None
Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC) 1313 and 1937								
Surface Sampling of ISO 5 (4 locations)	Daily	07AUG 2026	SU	Date of TestingScan	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	07AUG 2026	GA	Date of Testing ScanC/O	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	07AUG 2026	SU	Date of TestingScan	2 CFUs on Sett1	ETX-260817-0507		None
Settling Sampling of ISO 5 (2 locations)	Daily	07AUG 2026	GA	Date of Testing ScanC/O	No growth	N/A	N/A	None
Weekly Active Air Sampling Bracketing 115								
Active Air Sampling of Cleanrooms	Weekly	06AUG 2026	SMO	Week of testing Scan	1 CFU on ISO 8 115	ETX-260817-0370	Gram (+) short rods	None
Weekly Active Air Sampling Bracketing L-suite								
Active Air Sampling of Cleanrooms	Weekly	07AUG 2026	SMO	Week of testing ScanC/O	6 CFUs on 143 SecI air ISO 8 32 CFUs on 143 SecII air ISO 8, and 12 CFUs on 142 air ISO 8	ETX-260817-0416 ETX-260817-0422 ETX-260817-0429	Staphylococcus epidermidis Staphylococcus capitis Micrococcus luteus Kocuria palustris Corynebacterium segmentosum Staphylococcus hominis Staphylococcus aureus Corynebacterium aurimucosum Kytococcus sedentarius	None
Surface Sampling of Anteroom and Cleanroom Bracketing 115								
Surface Sampling of Cleanrooms	Weekly	06AUG 2026	SMO	Week of testing Scan	2 CFU on floor in ISO 7 115A	ETX-260817-0366	Bacillus megaterium	None
Weekly Active Air Sampling Bracketing L-suite								
Surface Sampling of Cleanrooms	Weekly	07AUG 2026	SMO	Week of testing ScanC/O	No growth	N/A	N/A	None