

	CORP-FORM-5 v9.1	Category: Form Corporate Name: Nonconformance Report Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.007.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC			Page 1 of 2

NCR#: N26436

IDENTIFICATION	
Initiator: Simin Mohammad	Date: 21-Aug-2026
Department: Microbiology	Sample Name/Material Description:
Equipment ID #: Cleanroom Suite 114 (Anteroom & Buffer Room) ETX-260915-0653, ETX-260916-0044, ETX-260916-0374	Semaglutide/Cyanocobalamin; Tirzepatide 22mg/mL-Pyridoxine HCl; SEMAGLUTIDE/B12
Sample ID #	Sample/Material Lot #: 0216202602, 22666, LG342010613
NONCONFORMITY DESCRIPTION (Who, What, When, Where, Why, & How)	
Nonconformance Description: It was identified that the Environmental Monitoring (EM) Analyst, SMO, did not perform the required routine weekly surface and active air sampling for Cleanroom Suite 114, as specified in MICRO-SOP-2, Environmental Monitoring of the Clean Room Facility, during the period from 02-Aug-2026 through 15-Aug-2026. As a result, weekly EM data for Cleanroom Suite 114 were unavailable for two consecutive weeks. However, the majority of sterility test batches processed concurrently in Suite 114 met all acceptance criteria, and no adverse or widespread positive trend was observed. In addition, environmental monitoring performed within the ISO 5 Biological Safety Cabinet (BSC), continuous 24/7 Lighthouse monitoring data for HVAC performance and room differential pressure, and all method negative controls remained within established specifications throughout the period in question. Based on the lack of a widespread contamination trend, the satisfactory performance of concurrent sterility testing activities, and the acceptable results of all other relevant environmental and process controls, the missed weekly background environmental monitoring is not considered to have contributed to or be related to these isolated sterility test failures.	
CAUSE(S) ANALYSIS	
Description of cause(s): Analyst oversight Category (select all that apply): ☐ Equipment/System ☐ Process/Method ☒ Personnel ☐ External Phenomena ☒ Other N/A 21Aug26 SMO	
INTERIM CORRECTIVE ACTION(S)	
Description of corrective action(s): The microbiology lab supervisor instructed the analysts to be more vigilant during the environmental monitoring process to ensure that the surface sampling of the anteroom and buffer room is performed as required. On 21-Aug-2026, Microbiology Laboratory Supervisor Robin Seymour instructed the microbiology team to have actively update supervisor and lead analysts every week going forward to ensure rigorous tracking, accountability, and verification of all cleanroom sampling points. Analyst SMO will need to complete Read & Understood retraining on MICRO-SOP-2.	
QUALITY EVALUATION	
Was action effective in addressing the root cause of the nonconformity? ☒ Yes ☐ No ☐ N/A CAR required: ☐ YES ☐ NO If yes, CAR #: _____ If yes, explain how effectiveness was determined: Effectiveness confirmed via immediate reinstatement of weekly EM with acceptable results and launch of Weekly EM Tracking Sheet. Additional Comments: Risk Assessment / Comments: Impact evaluated for 3 correlated Celsius sterility OOSs in Room 114B (OOS-261824, OOS-261861, OOS-261877). Over 470+ concurrent samples tested negative (>99.3% pass rate). Continuous 24/7 Lighthouse HVAC logs, session-specific ISO 5 BSC EM, and negative controls confirmed operational control. See attached Risk Assessment Addendum.	
ACKNOWLEDGEMENT	
Prepared By (Print): Simin Mohammad	_____ Signature Date
Reviewed By (Print): Qiyue Chen	_____ Signature Date



CORP-FORM-5
v9.1

Category: Form Corporate
Name: Nonconformance Report Form

Effective
10-JUL-2026

Location
Houston (Corporate)

ETX ID
3.100.007.F01

Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC

Page 2 of 2

NCR#: N26436

CLOSING THE NONCONFORMANCE (*Closure by Quality*)

Assessment Impact:

Frequency: 2

<2 = 1 Between 2 to 10 = 2 >10 = 3

Severity: 2

Typo Corrections = 1

Changes in processes and/or systems with no retroactive implications = 2

Changes in processes and/or systems with retroactive implication = 3

Magnitude: 2

<2 Process/Material/Equipment Affected = 1

>2 to <10 Process/Material/Equipment Affected = 2

>10 Process/Material/Equipment = 3

1 to 2: Low
3 to 17: Major
18 to 27: Critical

3	18	27
2	17	18
1	2	3

Assessment Impact Grading: **8**

Impact Level

☐ Low

☒ Major

☐ Critical

Quality Reviewed By (*Print*):

Signature

Date

Quality Approved By (*Print*):

Signature

Date

Samples Processed in Suite 114 Between 03 Aug 2026 to 21 Aug 2026							
Submission_Number	Client_Facility	Sample_Name	Lot_Number	Test_Name	Estimated_Completion_Date	Status	Analyst
ETX-260806-0771	ounders, LLC dba Northwest	one cypionate (MCT) MDV 5m	08-04-2026@3	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260807-0309	Boothwyn Pharmacy LLC 4548	ONE CYPIONATE 200 MG/ML (	08032026@5	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260728-0370	Boothwyn Pharmacy LLC 4548	ypionate (MCT) (4mL) 200mg	07272026@15	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260804-0285	Eagle Pharma Outsourcing LL	ethylprednisolone Acetate 40r	2507MA40	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260702-0567	acific Compounding Pharmac	Sermorelin Acetate 2mg/ml	07022026@14	Celsis Sterility Test	08/21/2026	Completed	HSmith
ETX-260807-0493	evive Rx Pharmacy - PO Requir	ine Cypionate 200 mg/ml Injec	16510120	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260807-0342	evive Rx Pharmacy - PO Requir	ine Cypionate 200 mg/ml Injec	16510114	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260730-0696	Las Colinas Pharmacy	TRADIOL CYPIONATE 6.25mg/	30JUL2026-001	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260807-0567	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227454	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260807-0568	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227455	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0542	Boudreaux's New Drug Store	G/GLYCINE 5MG/ML SOLUTIO	227384	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260804-0140	VLS Pharmacy Inc	e Acetate 50mcg/0.1mL Inj. S	08042026@5	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260807-0261	Mills Pharmacy & Apothecary	200mg/mL Sterile Solution for	1630530	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260804-0789	Potters House Apothecary	0 mg/ml injection-Nicotinamid	03Aug2026-059	Celsis Sterility Test	08/21/2026	Completed	GSurber
ETX-260811-0608	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227601	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0709	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227605	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260805-0764	Boudreaux's New Drug Store	G/GLYCINE 5MG/ML SOLUTIO	227326	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0682	Hillstone Pharmacy	CYPIONATE 200MG/ML (GRAP	06AUG2026-024	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0689	Hillstone Pharmacy	CYPIONATE 200MG/ML (GRAP	06AUG2026-016	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0468	Hillstone Pharmacy	CYPIONATE 200MG/ML (GRAP	06AUG2026-007	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260804-0229	Las Colinas Pharmacy	NE CYP/PROP 200MG/ML ***M	03AUG2026-011	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260807-0227	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227448	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0707	Boudreaux's New Drug Store	G/GLYCINE 5MG/ML SOLUTIO	227383	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260811-0752	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227607	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260807-0496	evive Rx Pharmacy - PO Requir	ine Cypionate 200 mg/ml Injec	16510118	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260807-0501	evive Rx Pharmacy - PO Requir	ine Cypionate 200 mg/ml Injec	16510115	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260810-0523	evive Rx Pharmacy - PO Requir	mL / Ipamorelin 0.5 mg/mL **	16523457	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260702-0137	Allenhill Pharmacy & Medical	ALPROSTADIL 500MCG/ML	S622070	SP <71> / EP 2.6.1 Sterility Tes	08/21/2026	Completed	asaravia
ETX-260803-0529	Carie Boyd Pharmaceuticals	CYPIONATE 200MG/ with MIG	07162026@3	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260810-0438	evive Rx Pharmacy - PO Requir	ine Cypionate 200 mg/ml Injec	16523507	Celsis Sterility Test	08/18/2026	Completed	asaravia

ETX-260810-0554	evive Rx Pharmacy - PO Requ	ne Cypionate 200 mg/ml Injec	16523512	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260803-0531	Carie Boyd Pharmaceuticals	CYPIONATE 200MG/ with MIG	07172026@3	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260806-0153	Tailor Made Compounding	erone Cypionate (GSO) 100mg	5938398	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260810-0085	Tailor Made Compounding	erone Cypionate (GSO) 50mg	5978877	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260807-0115	Tailor Made Compounding	erone Cypionate (GSO) 200mg	5938414	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260810-0130	Tailor Made Compounding	erone Cypionate (GSO) 200mg	5938410	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0149	Tailor Made Compounding	erone Cypionate (GSO) 100mg	5978878	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260806-0147	Tailor Made Compounding	erone Cypionate (GSO) 50mg	5938402	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260807-0325	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227452	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0690	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227392	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260807-0495	Hillstone Pharmacy	DE/L-CARNITINE 3.5MG/50M	07AUG2026-010	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260806-0459	Hillstone Pharmacy	L-CARNITINE 2.5MG/50MG/M	05AUG2026-035	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260813-0220	evive Rx Pharmacy - PO Requ	mL / Ipamorelin 0.5 mg/mL **	16554773	Celsis Sterility Test	08/21/2026	Completed	GSurber
ETX-260807-0305	Connect Pharmaceuticals	Testosterone 100mg/ml	TESG080626	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260806-0673	Hillstone Pharmacy	CYPIONATE 200MG/ML (GRAP	06AUG2026-009	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260807-0231	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227449	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260807-0271	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227450	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0415	Boudreaux's New Drug Store	G/GLYCINE 5MG/ML SOLUTIO	227382	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260729-0412	Strive Pharmacy - 15523	TB-500 5 MG/ML - 6ml	LG364287558	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260810-0022	Southend Pharmacy	Sermorelin 3mg/mL 5 mL	21303	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260622-0620	Southend Pharmacy	MOTs-C 2mg/ml	19472	SP <71> / EP 2.6.1 Sterility Tes	08/21/2026	Completed	asaravia
ETX-260806-0456	Hillstone Pharmacy	L-CARNITINE 2.5MG/50MG/M	05AUG2026-034	Celsis Sterility Test	08/17/2026	On Hold	asaravia
ETX-260806-0678	Hillstone Pharmacy	CYPIONATE 200MG/ML (GRAP	06aug2026-008	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260730-0299	Strive Pharmacy - 15523	aglutide/Glycine 9/5/ mg/ml (0	LG364288126	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260729-0434	Strive Pharmacy - 15523	aglutide/Glycine 5/5/ mg/ml (	LG364287567	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260729-0417	Strive Pharmacy - 15523	aglutide/Glycine 5/5/ mg/ml (0	LG364287562	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260811-0599	GoGoMeds Select	pionate 50 mg/mL Oil Injunctio	18698	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0700	GoGoMeds Select	pionate 50 mg/mL Oil Injunctio	18699	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260810-0742	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227537	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260806-0306	Boudreaux's New Drug Store	G/GLYCINE 5MG/ML SOLUTIO	227380	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260807-0406	evive Rx Pharmacy - PO Requ	ne Enanthate 200 mg/ml Injec	16510105	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260728-0333	Strive Pharmacy - 15523	TB-500 5 MG/ML - 3ml	LG364287114	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260727-0181	Strive Pharmacy - 15523	ine/B12 (Cyanocobalamin) 9/	LG364286114	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia

ETX-260810-0534	evive Rx Pharmacy - PO Requ	ne Cypionate 200 mg/ml Injec	16523510	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260811-0589	GoGoMeds Select	pionate 25 mg/mL Oil Injec	18695	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0591	GoGoMeds Select	pionate 25 mg/mL Oil Injec	18696	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260708-0662	so Home Care Pharmacy 710	Ferric Carboxymaltose	2620708	SP <71> / EP 2.6.1 Sterility Tes	08/21/2026	Completed	asaravia
ETX-260806-0786	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227391	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260716-0266	PGRrx LLC	tric Acid 62.5mM (12.01mg/m	PGR-15072026-01	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0450	evive Rx Pharmacy - PO Requ	ne Cypionate 200 mg/ml Injec	16523514	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260722-0295	Strive Pharmacy - 15523	epatide/Glycine 20mg/5mg (6	LG364285210	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260807-0039	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	EDQGUIPPHTR	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0043	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	JGFGKKCHPUGE	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0417	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227457	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0414	Boudreaux's New Drug Store	IG/GLYCINE 5MG/ML SOLUTIO	227381	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0785	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227393	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260807-0032	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	HGJFJRDSRUPC	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260810-0125	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	ECCEUDIIRPL	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260728-0408	Strive Pharmacy - 15523	aglutide/Glycine 5/5/ mg/ml (	LG364287121	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260709-0407	Strive Pharmacy - 15523	cine/B12 (Cyanocobalamin) 9	LG364280643	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260810-0113	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	QKISHUHTTPDR	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0145	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	TQJRGQHSQEGT	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260807-0050	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	IHKIGJGTFRPD	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0034	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	UFEIJTJHLGH	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0037	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	LLQEHIRHGFLC	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260730-0230	Strive Pharmacy - 15523	TB-500 5 MG/ML - 3ml	LG364288225	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260730-0280	Strive Pharmacy - 15523	aglutide/Glycine 9/5/ mg/ml (0	LG364288157	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia
ETX-260807-0040	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	PHJGDESFDG	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260811-0114	re Pharmacy-Quality (PO Requ	lutide 5mg/2ML Solution for In	FQTJSKCJGCUI	Celsis Sterility Test	08/20/2026	Completed	asaravia
ETX-260811-0572	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227602	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0732	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJEC	227606	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260609-0486	's Regional Medical (PO # REQ	Sulfate 3gm in 50ml D5W (no	SJR060926MLNHD	SP <71> / EP 2.6.1 Sterility Tes	08/21/2026	Completed	asaravia
ETX-260811-0129	re Pharmacy-Quality (PO Requ	lutide 5mg/2ML Solution for In	IDIEQQQSTIIP	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260807-0044	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	UPTFLCJJRGSE	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260810-0129	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	QQTCLRTPITPI	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260722-0285	Strive Pharmacy - 15523	epatide/Glycine 20mg/5mg (6	LG364285207	SP <71> / EP 2.6.1 Sterility Tes	08/20/2026	Completed	asaravia

ETX-260811-0382	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227600	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260807-0572	Boudreaux's New Drug Store	COBALAMIN 10MG/ GLYCINE 5	227458	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260728-0410	Strive Pharmacy - 15523	emaglutide/Glycine 5/5/ mg/ml (	LG364287125	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	asaravia
ETX-260810-0616	Puramint, LLC	rom/Moxi/Pred 1936113 Cels	1936113	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260803-0773	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227197	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260811-0630	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227604	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260810-0158	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	TRIQPPRRTSRC	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0212	re Pharmacy-Quality (PO Requ	emaglutide Injection 1.6mg/0.6	HKJERGTEKQJL	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260807-0327	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227453	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260811-0380	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227599	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260810-0741	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227536	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0230	re Pharmacy-Quality (PO Requ	emaglutide Injection 2.5mg/1r	FCIELQUIGUKC	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260806-0073	re Pharmacy-Quality (PO Requ	lutide 5mg/2ml solution for in	JCJPDPUURPGC	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260811-0119	re Pharmacy-Quality (PO Requ	lutide 5mg/2ML Solution for Ir	UCEUQRJSCGHD	Celsis Sterility Test	08/20/2026	Completed	asaravia
ETX-260810-0743	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227538	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260807-0398	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227456	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0349	Boudreaux's New Drug Store	TE/ ESTRADIOL VALERATE 50	227394	Celsis Sterility Test	08/17/2026	Completed	asaravia
ETX-260811-0676	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227603	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260807-0235	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227451	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260811-0120	re Pharmacy-Quality (PO Requ	lutide 5mg/2ML Solution for Ir	CRUSSLTRTSSS	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260806-0788	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227390	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0416	Boudreaux's New Drug Store	MG/ GLYCINE 5MG/ML INJE	227534	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0599	Puramint, LLC	rom/Moxi/Pred 1936110 Cels	1936110	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260807-0199	bear's Compounding Pharmac	G/ML SUBCUTANEOUS INJECT	23751	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260804-0122	Prince Sterilization Service	Fold Hydrophilic CenterFIT / C	PSS-032420	SP <71> / EP 2.6.1 Sterility Test	08/21/2026	Completed	qchen
ETX-260803-0224	Prince Sterilization Service	Amber TS Vials SKU # 154 - [3	PSS-032300	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	qchen
ETX-260810-0114	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	CUELIURRCSSG	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0196	re Pharmacy-Quality (PO Requ	lutide 6mg/2.4ml solution for i	JRFDCPLPSCFG	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0123	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	QDURGDCSQULD	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0124	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	SLDUKFUEQGUD	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0134	re Pharmacy-Quality (PO Requ	emaglutide Injection 5mg/2m	JESUQIPSSIJE	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260803-0226	Prince Sterilization Service	Amber TS Vials SKU # 154 - [3	PSS-032301	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	qchen
ETX-260731-0048	Prince Sterilization Service	Amber TS Vials SKU # 154 - [3	PSS-32298	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	qchen

ETX-260731-0047	Prince Sterilization Service	Amber Vials SKU# 168 - Upright	PSS-032416	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	qchen
ETX-260807-0053	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	QIERIGLLKEUG	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260810-0138	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	PSTFIQCPCRPP	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0146	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	ELESUKKHQHKH	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260803-0216	Prince Sterilization Service	Amber TS Vials SKU # 154 - [	PSS-032299	SP <71> / EP 2.6.1 Sterility Test	08/20/2026	Completed	qchen
ETX-260807-0033	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	JPTCRSSHJLGC	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0047	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	DIQUTESJTEHH	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0030	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	DQFIFCJPGSEU	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0036	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	IGICHGGKSGSD	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0038	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	SKPPLRKEPUKF	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0045	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	FJDKSLUJDGIT	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0031	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	JIQIDKJLSFIR	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260807-0046	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	TFLGUPGSIDCU	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260810-0205	re Pharmacy-Quality (PO Request)	lutide 6mg/2.4mL solution for in	RQGJJPKPTSIQ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0127	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	PEGFQJLPCPCT	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0209	re Pharmacy-Quality (PO Request)	maglutide Injection 1.6mg/0.64	EKLPKCCGITDQ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0223	re Pharmacy-Quality (PO Request)	Semaglutide Injection 2.5mg/1mL	EGQHDSJSPPP	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0240	re Pharmacy-Quality (PO Request)	Semaglutide Injection 2.5mg/1mL	SFCRHHUCTQSS	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260807-0041	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	HCUFJSSGEFKF	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260810-0112	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	DDFHUJLLIHU	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0160	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	QDULRGDCJTCF	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0233	re Pharmacy-Quality (PO Request)	Semaglutide Injection 2.5mg/1mL	RURREJRIEJHK	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260807-0029	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	REHLEFHGPHID	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260805-0221	re Pharmacy-Quality (PO Request)	lutide 5mg/2mL solution for in	ITPFJETDEIUH	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0184	re Pharmacy-Quality (PO Request)	lutide 6mg/2.4mL solution for in	CSPTEHDQTTTJ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0116	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	LKDGLGTHIDK	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260810-0201	re Pharmacy-Quality (PO Request)	lutide 6mg/2.4mL solution for in	KFHKIJFGLHC	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0109	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	JFICFTCQHCHK	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0156	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	UFERJFQQDFRQ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0249	re Pharmacy-Quality (PO Request)	Semaglutide 5mg/2mL	KFLTDHLQURID	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0186	re Pharmacy-Quality (PO Request)	lutide 6mg/2.4mL solution for in	LKURUQTGSGGL	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0120	re Pharmacy-Quality (PO Request)	Semaglutide Injection 5mg/2mL	KIJRSTEIGSGQ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260811-0130	re Pharmacy-Quality (PO Request)	lutide 5mg/2ML Solution for In	ECITKRLRTJLF	Celsis Sterility Test	08/20/2026	Completed	asaravia

ETX-260805-0146	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	SKFGDUSDESQD	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0103	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	LIKKUSSTGKTP	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260805-0140	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	HJUIPGGPHFTE	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260805-0160	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	RPHRGLFGTRRS	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0100	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	IRFSRTGJTISI	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0141	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	TPLLUIECIDFQ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0082	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	HHCHSHGJDQTQ	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0094	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	UUSIGKPLQTGC	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0110	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	JULHEPPQTHKD	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0167	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	IQFDTKTPGSQD	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0150	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	DJFRCLSDQDPU	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260810-0151	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	FUFFLFHDHSCKT	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260811-0113	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	JSKITCFSDLDR	Celsis Sterility Test	08/20/2026	Completed	asaravia
ETX-260810-0102	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	LPUIEHJKUGPU	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0118	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	QRRDRFRJDLHIH	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0153	re Pharmacy-Quality (PO Req	emaglutide Injection 5mg/2m	PUSGRULEJSQP	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260811-0116	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	EQKGDGLETCSH	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0124	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	DRDFPEQDEHLI	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260805-0190	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	HJIJRDRHEEE	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260806-0094	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	KTGUGKURKUEI	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0096	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	QEGHIJFUGFD	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260806-0103	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	KKIREUCUERKH	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0097	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	FKSUFSEDTSS	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0124	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	FGEIGJJKIPPS	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260811-0127	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	DSHPHEHHUQRRF	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260811-0121	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	LUGSEDESIJKE	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260806-0102	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	TIQQCJRDEDPC	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260806-0131	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	FDDPQHJIDICK	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0146	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	CFLUTQQTKGTS	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260806-0148	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	GGIUGDJLUJJC	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260811-0115	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	FJPQUCCGLSSU	Celsis Sterility Test	08/20/2026	Completed	asaravia
ETX-260811-0131	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	KTRSEPELUHU	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260806-0088	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	ETJQKHDDFLDH	Celsis Sterility Test	08/18/2026	Completed	asaravia

ETX-260806-0100	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	UUISGLFEDPGU	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260806-0140	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	QHQDSGGUELGT	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260811-0123	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	JSTGSQCDQCHS	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260810-0097	re Pharmacy-Quality (PO Req	semaglutide Injection 5mg/2m	KJGCEJCJSDKU	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260806-0129	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	TDPLFJQGPDP	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260806-0164	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	FKJGCESUKIL	Celsis Sterility Test	08/18/2026	Completed	asaravia
ETX-260805-0154	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	PTJJLLQFKDI	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0117	re Pharmacy-Quality (PO Req	semaglutide Injection 5mg/2m	RKTKKQQIFCGG	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260806-0178	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	KFCTCFTLSQUH	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260805-0150	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	LHFDKFKTCSUD	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0111	re Pharmacy-Quality (PO Req	semaglutide Injection 5mg/2m	RPKLUDTFIKDE	Celsis Sterility Test	08/19/2026	Completed	asaravia
ETX-260810-0193	re Pharmacy-Quality (PO Req	lutide 6mg/2.4ml solution for i	STJJURLKTGTJ	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260806-0170	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	DRHTFDDFUUJG	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260811-0134	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	ULLFPTCHETLE	Celsis Sterility Test	08/20/2026	Completed	asaravia
ETX-260811-0133	re Pharmacy-Quality (PO Req	lutide 5mg/2ML Solution for In	SRJEKEQLJEEC	Celsis Sterility Test	08/20/2026	Completed	GSurber
ETX-260805-0224	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	IFQHURDSKTFQ	Celsis Sterility Test	08/17/2026	Completed	GSurber
ETX-260810-0098	re Pharmacy-Quality (PO Req	semaglutide Injection 5mg/2m	RSJELTEISSQK	Celsis Sterility Test	08/19/2026	Completed	GSurber
ETX-260806-0134	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	PKQICKHDKFDG	Celsis Sterility Test	08/18/2026	Completed	GSurber
ETX-260806-0156	re Pharmacy-Quality (PO Req	lutide 5mg/2ml solution for in	GPTISJUSKLUD	Celsis Sterility Test	08/18/2026	Completed	GSurber