**OOS Number: OOS-**

**Phase I**

| **Section A – Test Details** | | | | | |
|---|---|---|---|---|---|
| **Initiator:** | | **Test Date:** | | **Date Initiated:** | |
| **Date of Incident:** | | **Test Name:** | | **Sample ID:** | |
| **Name of Analyst who Performed the Test:** | | **Sample / Active Name** | | **Lot #:** | |
| | | **Dosage Form** | | | |
| **Description of Incident:** | | ☐ Client Care notified of OOS<br>☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS | | | |
| | SOP / Test Method # | Effective Date: | SOP / Test Method Rev: | Limits/Specification: | |
| **Manager notified** | Manager/designee name: | Date: | Quality Assurance Department notified: | Yes ☐ No ☐ | |

| **Section B – Investigation and Root Cause Analysis** | | | |
|---|---|---|---|
| **Personnel Discrepancy** | | **Comments** | |
| Analyst interviewed? | ☐Yes ☐ No ☐ N/A | | |
| Correct samples analyzed? | ☐Yes ☐ No ☐ N/A | | |
| Correct SOP / Test Method used/followed? | ☐Yes ☐ No ☐ N/A | | |
| Correct sample collection/technique? | ☐Yes ☐ No ☐ N/A | | |
| Are the Test Result Records Attached & Calculations Verified and Correct? | ☐Yes ☐ No ☐ N/A | | |
| Is the analyst qualified to perform the test? | ☐Yes ☐ No ☐ N/A | | |
| Is formula worksheet provided and verified? | ☐Yes ☐ No ☐ N/A | | |
| Does the processing method/method set conform to the required specification? | ☐Yes ☐ No ☐ N/A | | |
| Was the sample stored appropriately? | ☐Yes ☐ No ☐ N/A | | |
| **Material Discrepancy** | | **Lot Number** | **Expiry Date** |
| Is the integrity of the sample(s) questionable? | ☐Yes ☐ No ☐N/A | | |
| Was the sample transported appropriately? | ☐Yes ☐ No ☐N/A | | |
| Are the other test results from this run under investigation? | ☐Yes ☐ No ☐N/A | | |

**OOS Number: OOS-**

| | | | |
|---|---|---|---|
| Appropriate/correct glassware used? | ☐Yes ☐ No ☐N/A | | |
| Are reagents stored in proper condition and not expired? | ☐Yes ☐ No ☐N/A | | |
| Standard Used | | | |
| Solvent Used | | | |
| Is the standard the source of the root cause? | ☐Yes ☐ No ☐N/A | | |
| Has the analyst been instructed to keep the original test solution for possible retesting? | ☐Yes ☐ No ☐N/A | | |
| **Equipment Discrepancy** | | **Equipment ID** | **Calibration Due Date** |
| Was the instrument set up correctly? | ☐Yes ☐ No ☐ N/A | | |
| Is the clean room facility certified? | ☐Yes ☐ No ☐ N/A | | |
| System Performance Check (if applicable)? | ☐Yes ☐ No ☐ N/A | | |
| System Suitability RSD < 2% (PDA)? | ☐Yes ☐ No ☐ N/A | | |
| System Suitability RSD < 10% (QDA, CAD)? | ☐Yes ☐ No ☐ N/A | | |
| Do the environmental monitoring results conform to required standards? | ☐Yes ☐ No ☐ N/A | | |
| Do the controls (i.e., positive, negative, etc.) conform to the required standards? | ☐Yes ☐ No ☐ N/A | | |
| Is equipment/instrument functional and calibrated/qualified/certified/PM? | ☐Yes ☐ No ☐ N/A | | |

**OOS Number: OOS-**

| | | | |
|---|---|---|---|
| Are chromatogram/spectra/raw data correct? | ☐ Yes ☐ No ☐ N/A | | |

**Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)**

- ☐ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: _____

**Section D – Phase I Summary**

**Phase I Summary:**

**OOS Number: OOS-**

**Phase I Summary (continued):**

**OOS Number: OOS-**

**Phase I Summary (continued):**

**OOS Number: OOS-**

**Has root cause been identified as a possible laboratory error?**
*Note: Root cause must be conclusive, verifiable, and well-documented.*

☐ Yes

☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☐ No, the OOS result is valid.

**Reviewer comment(s):**

| | |
|---|---|
| **Prepared by** (*Print*)**:** | Signature/Date: |
| **Lab Manager/Designee** (*Print*)**:** | Signature/Date: |

**Section E – Quality Disposition**

**Can investigation be closed?**

☐ Yes

☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

| | |
|---|---|
| **Approved by QA Representative** (*Print*)**:** | Signature/Date: |

**Table 1: Reading analyst, Days of incubation, and Microbial ID for Sample under Investigation**

| Reading Analyst | Day of Incubation | Sample ID | Related Microbial ID | Media with growth | Microbial ID |
|---|---|---|---|---|---|
| CCD | Day 7 | ETX-261002-0288 | ETX-261012-0654 | 1 x 300mL FTM | *Bacillus paralicheniformis* (Gram-positive rod) |

## Table 2: Environmental Monitoring from Celsis Processing Performed on 05Oct2026

| Environmental Monitoring (EM) Sampling Site | Frequency | Date | Analyst | Day /Week(s) | Observation | EM Plate ETX Number | Microbial ID | Notes |
|---|---|---|---|---|---|---|---|---|
| **Personnel EM Bracketing 05Oct2026** | | | | | | | | |
| Personal (Left Touch and Right Touch) | Daily | 02Oct2026 | ES | Date Before Testing | No Growth | Not Applicable | Not Applicable | None |
| Personal (Left Touch and Right Touch) | Daily | 05Oct2026 | ES | Date of Testing | No Growth | Not Applicable | Not Applicable | None |
| Personal (Left Touch and Right Touch) | Daily | 06Oct2026 | ES | Date After Testing | No Growth | Not Applicable | Not Applicable | None |
| **Biological Safety Cabinet EM Bracketing E001316 for 05Oct2026** | | | | | | | | |
| Surface Sampling of ISO 5 BSC E001314 (4 Locations) | Daily | 02Oct2026 | ALA | Date Before Testing | No Growth | Not Applicable | Not Applicable | None |
| Surface Sampling of ISO 5 BSC E001314 (4 Locations) | Daily | 05Oct2026 | ES | Date After Testing | No Growth | Not Applicable | Not Applicable | None |
| Surface Sampling of ISO 5 BSC E001314 (4 Locations) | Daily | 06Oct2026 | ES | Date After Testing | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001314 | Daily | 02Oct2026 | ALA | Date Before Testing | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001314 | Daily | 05Oct2026 | ES | Date of Testing | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001314 | Daily | 06Oct2026 | ES | Date After Testing | No Growth | Not Applicable | Not Applicable | None |
| **Weekly Active Air and Surface Sampling of Cleanroom Suite 114 with Processing BSC For 05Oct2026** | | | | | | | | |
| **Environmental Monitoring (EM) Sampling Site** | **Frequency** | **Date** | **Analyst** | **Observation** | **EM Plate ETX Number** | **Microbial ID** | | **Notes** |
| Active Air Sampling of Cleanrooms | Weekly | 05Oct2026 | SMO | No Growth | Not Applicable | Not Applicable | | None |
| Surface Sampling of Cleanrooms | Weekly | 05Oct2026 | SMO | No Growth | Not Applicable | Not Applicable | | None |

## Table 3: Environmental Monitoring from Celsis Aliquoting Performed on 12Oct2026

| Environmental Monitoring (EM) Sampling Site | Frequency | Date | Analyst | Day /Week(s) | Observation | EM Plate ETX Number | Microbial ID | Notes |
|---|---|---|---|---|---|---|---|---|
| **Personnel EM Bracketing 12Oct2026** | | | | | | | | |
| Personal (Left Touch and Right Touch) | Daily | 09Oct2026 | CCD | Date Before Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Personal (Left Touch and Right Touch) | Daily | 12Oct2026 | CCD | Date of Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Personal (Left Touch and Right Touch) | Daily | 13Oct2026 | CCD | Date After Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| **Biological Safety Cabinet EM for 12Oct2026** | | | | | | | | |
| Surface Sampling of ISO 5 BSC E001798 (4 Locations) | Daily | 09Oct2026 | ALA | Date Before Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Surface Sampling of ISO 5 BSC E001798 (4 Locations) | Daily | 12Oct2026 | CCD | Date of Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Surface Sampling of ISO 5 BSC E001798 (4 Locations) | Daily | 13Oct2026 | CCD | Date After Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001798 | Daily | 09Oct2026 | ALA | Date Before Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001798 | Daily | 12Oct2026 | CCD | Date of Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| Settling Sampling of ISO 5 BSC E001798 | Daily | 13Oct2026 | CCD | Date After Aliquoting | No Growth | Not Applicable | Not Applicable | None |
| **Weekly Active Air and Surface Sampling of Cleanroom Suite 114 with Processing BSC For 12Oct2026** | | | | | | | | |
| **Environmental Monitoring (EM) Sampling Site** | **Frequency** | **Date** | **Analyst** | **Observation** | **EM Plate ETX Number** | **Microbial ID** | | **Notes** |
| Active Air Sampling of Cleanrooms | Weekly | 12Oct2026 | SMO | 6 CFUs (ISO 8 114) | ETX-261021-0578 | *Moraxella osloensis*<br>*Sporisorium graminicola*<br>*Staphylococcus haemolyticus*<br>*Staphylococcus hominis*<br>*Kocuria indica*<br>*Ustilago maydis*<br>*Kocuria rhizophila* | | None |
| Surface Sampling of Cleanrooms | Weekly | 12Oct2026 | SMO | No Growth | Not Applicable | Not Applicable | | None |