

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 1 of 7

OOS Number: OOS- 261186

Phase I

Section A – Test Details					
Initiator:	Simin Mohammad	Test Date:	07-May-2026	Date Initiated:	18-May-2026
Date of Incident:	18-May-2026	Test Name:	Environmental Monitoring	Sample ID:	ETX-260518-0254
Name of Analyst who Performed the Test:	Simin Mohammad (Weekly Active Air Sampling Plate Setup)	Sample / Active Name	EM SMO 116A Air 07MAY2026	Lot #:	EM SMO 116A Air 07MAY2026
	Sophia Santamaria (Weekly Active Air Sampling Plate Readers)	Dosage Form	Plate		
Description of Incident:	The CFU count for environmental monitoring plate exceeded the action level.		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
Manager notified	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
	MICRO-SOP-2	05-AUG-2026	15	Action level: >10	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	
	Kathan Parikh	18-May-2026			
Section B – Investigation and Root Cause Analysis					
Personnel Discrepancy			Comments		
Analyst interviewed?	☒ Yes ☐ No ☐ N/A		Yes, analyst Simin Mohammad, and Sophia Santamaria were comprehensively interviewed.		
Correct samples analyzed?	☐ Yes ☐ No ☒ N/A		N/A		
Correct SOP / Test Method used/followed?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Correct sample collection/technique?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Are the Test Result Records Attached & Calculations Verified and Correct?	☒ Yes ☐ No ☐ N/A		Yes, information is available in EagleTrax under ETX-260518-0254.		
Is the analyst qualified to perform the test?	☒ Yes ☐ No ☐ N/A		Yes, the analysts are trained and qualified by quality to perform the test		
Is formula worksheet provided and verified?	☐ Yes ☐ No ☒ N/A		N/A		
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☒ N/A		N/A		
Was the sample stored appropriately?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Material Discrepancy			Lot Number	Expiry Date	
Is the integrity of the sample(s) questionable?	☐ Yes ☐ No ☒ N/A				
Was the sample transported appropriately?	☐ Yes ☐ No ☒ N/A				
Are the other test results from this run under investigation?	☐ Yes ☐ No ☒ N/A				

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 2 of 7

OOS Number: OOS- 261186

Appropriate/correct glassware used?	☐ Yes ☐ No ☒ N/A		
Are reagents stored in proper condition and not expired?	☒ Yes ☐ No ☐ N/A	TSA Plate: 1011543730	TSA Plate: 29-SEP-2026
Standard Used	N/A	N/A	N/A
Solvent Used	N/A	N/A	N/A
Is the standard the source of the root cause?	☐ Yes ☐ No ☒ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☒ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☒ Yes ☐ No ☐ N/A	Please see below	Please see below
Is the clean room facility certified?	☒ Yes ☐ No ☐ N/A	CR116 (E001738)	Dec-2026
System Performance Check (if applicable)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the environmental monitoring results conform to required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☒ Yes ☐ No ☐ N/A	Incubator E001034 (Sensor E001501)	Aug-2026 Feb-2027
		Incubator E001031 (Sensor E001505)	Aug-2026 Feb-2027

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 3 of 7

OOS Number: OOS- 261186

Are chromatogram/spectra/raw data correct?				☐ Yes	☐ No	☒ N/A	N/A	N/A
Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)								
☒ N/A		☐ Re-vialing of original working sample solution						
☐ Re-injection of sample vial		☐ Re-dilution of a working sample solution from the original stock						
☐ Re-injection of original standard vial		☐ Re-dilution of a working standard solution from the original stock standard						
☐ Re-vialing of original working standard solution		☐ Other: N/A SMO 18-Aug-2026						
Section D – Phase I Summary								
Phase I Summary: The analyst involved in the Active air Sampling plate setup, Simin Mohammad, and the analysts involved in reading the plate Sophia Sanatamaria, were interviewed comprehensively. Their answers are recorded throughout this document. The EM plates were stored in compliance with the supplier's recommendations, and their integrity was visually inspected prior to use. Furthermore, the plates were confirmed to be within their valid expiration dates. All the supplies were thoroughly disinfected according to MICRO-SOP-9 Cleaning and Disinfecting Procedure for Microbiology. The functionality of both incubators were verified through a review of data obtained from our comprehensive in-house continuous monitoring system. Active air Sampling was performed by analyst Simin Mohammad during weekly environmental monitoring processing, in ISO 8 116, and ISO 7 116A, and ISO 7 116B on 07 May 2026 as per MICRO-SOP-2 - Environmental Monitoring of the Cleanroom Facility. The plates were initially incubated at a temperature of 30–35°C in incubator E001031 for a minimum duration of 48 hours, commencing on 07 May 2026. Following completion of minimum of 48 hours of incubation on 11 May 2026, the plates were further incubated for minimum of 5 days, with the incubation concluding on 18 May 2026. Please see Table 1 for detailed information on the observations during respective incubations. Based on the observations in Table 1, since the CFU count exceeded the action level for the one of the Active air Sampling plates (116A), the plate was submitted for Microbial Identification under ETX-260518-0254. The colonies were identified as Kocuria palustris (Gram (+) cocci) To observe if the organisms identified, were transient in nature or recurring, weekly monitoring environmental monitoring plates for the clean room, were bracketed to include the week before testing, the week of testing and week after testing as detailed in Table 2 (please see attached).								

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 4 of 7

OOS Number: OOS- 261186

Phase I Summary (continued):

Environmental Monitoring Summary: Weekly environmental sampling for the previous week showed no growth for the active air and weekly surface sample. Week of testing showed 748 CFUs in 116A ISO 7 and 1 CFU in 116B ISO 7 for active air identified as *Kocuria palustris* (Gram (+) cocci). Lastly, the following week of testing showed 6 CFUs in ISO 8 116 for active air, 1 CFU on cart in ISO 8 116, and 2 CFUs on floor in 116A ISO 7 identified as Gram (+) cocci, Gram-variable coccobacilli, Gram (+) cocci, *Sporosarcina psychrophila*, and *Bacillus pumilus*.

During the interview with the analyst, they indicated that no obvious abnormalities or deviations in the testing procedure were observed. All the samples were thoroughly disinfected prior to testing. Moreover, The CR116 was thoroughly cleaned and prepared before initiating the testing as per MICRO-SOP-2 (Environmental Monitoring of the Clean Room Facility), and MICRO-SOP-9 (Cleaning and Disinfecting Procedure for Microbiology).

Monthly cleaning and disinfection of the outermost ISO 8 Anteroom (Suite 116), the middle ISO 7 Buffer room (Suite 116A), the innermost ISO 7 clean room (Suite 116B), and its containing ISO 5 Biosafety Cabinets were performed on 31-May-2026 as per MICRO-SOP--9 (Cleaning and Disinfecting Procedure for Microbiology) by analyst Tamiru Kotisso and Cuong Du on 31-May-2026 It was documented that all H₂O₂ indicators passed. This confirms the efficient monthly cleaning of all three suites 116, 116A, and 116B. Additionally, cleaning and disinfecting was performed both prior to and after the testing process as per MICRO-SOP-9.

It is also important to note that no samples processed in suite 115 for the week of testing (03-May-2026 to 09-May-2026) failed testing.

Based on the findings outlined in the preceding sections, the Out-Of-Specification (OOS) result observed for the Environmental Monitoring (EM) Active Air Sampling plate may be attributed to a potential analyst error. It is to be noted that the growth observed on the weekly active air and weekly surface sampling plates for the day of testing do not follow a trend, indicating that the contamination was transient in nature and that routine daily disinfection procedures were effective in eliminating the contamination. Furthermore, no trend was observed in the clean rooms previous weekly EM data, therefore, no preventive and corrective actions are deemed necessary at this time.

OOS Number: OOS- 261186

Phase I Summary (continued):

N/A SMO 18-AUG-2026

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 7 of 7

VERSION HISTORY

Laboratory OOS Investigation Form (v11.1)

Approved: 10-JUL-2026 | Effective: 10-JUL-2026 | Retired: N/A
Minor up version to fillable PDF

Laboratory OOS Investigation Form (v11.0)

Approved: 30-SEP-2025 | Effective: 30-SEP-2025 | Retired: 10-JUL-2026
Refer to original approved version on ETX

ELECTRONIC SIGNATURES

Approvals

Preeti Purwaha (DPP) Director of Quality Assurance ppurwaha@eagleanalytical.com <i>Comment: Author review completed.</i>	Document Author I am the author of this document. Signed 05:05:09 PM UTC 09-JUL-2026
Minh-Anh (Alison) Eng (AE) Quality Site Manager meng@eagleanalytical.com <i>Comment: approved. AE10JUL2026</i>	Technical Manager Approval I have reviewed and approve this Minor Revision. Signed 05:32:28 PM UTC 10-JUL-2026
Jacqueline Esqueda (JE) VP of Quality and Regulatory Comp jesqueda@eagleanalytical.com <i>Comment: Approved</i>	Quality Approval I have reviewed and approve this Minor Revision. Signed 05:41:18 PM UTC 10-JUL-2026

CURRENT AUTHORS

Preeti Purwaha (DPP)	Director of Quality Assurance
--------------------------------------	-------------------------------

CUSTOM FIELDS

Location	Houston (Corporate)
ETX ID	3.100.019.F01

LINKS/ATTACHMENTS

Document CORP-SOP-13: Out-of-Specification (OOS) - Test Failure Investigation (10.0) [effective]