OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:		**Test Date:**		**Date Initiated:**	
Date of Incident:		**Test Name:**		**Sample ID:**	
Name of Analyst who Performed the Test:		**Sample / Active Name**		**Lot #:**	
		Dosage Form			
Description of Incident:		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS			
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis			
Personnel Discrepancy		**Comments**	
Analyst interviewed?	☐Yes ☐ No ☐ N/A		
Correct samples analyzed?	☐Yes ☐ No ☐ N/A		
Correct SOP / Test Method used/followed?	☐Yes ☐ No ☐ N/A		
Correct sample collection/technique?	☐Yes ☐ No ☐ N/A		
Are the Test Result Records Attached & Calculations Verified and Correct?	☐Yes ☐ No ☐ N/A		
Is the analyst qualified to perform the test?	☐Yes ☐ No ☐ N/A		
Is formula worksheet provided and verified?	☐Yes ☐ No ☐ N/A		
Does the processing method/method set conform to the required specification?	☐Yes ☐ No ☐ N/A		
Was the sample stored appropriately?	☐Yes ☐ No ☐ N/A		
Material Discrepancy		**Lot Number**	**Expiry Date**
Is the integrity of the sample(s) questionable?	☐Yes ☐ No ☐N/A		
Was the sample transported appropriately?	☐Yes ☐ No ☐N/A		
Are the other test results from this run under investigation?	☐Yes ☐ No ☐N/A		

OOS Number: OOS-

Appropriate/correct glassware used?	☐Yes ☐No ☐N/A		
Are reagents stored in proper condition and not expired?	☐Yes ☐No ☐N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐Yes ☐No ☐N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐Yes ☐No ☐N/A		
Equipment Discrepancy		**Equipment ID**	**Calibration Due Date**
Was the instrument set up correctly?	☐Yes ☐No ☐N/A		
Is the clean room facility certified?	☐Yes ☐No ☐N/A		
System Performance Check (if applicable)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 2% (PDA)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐Yes ☐No ☐N/A		
Do the environmental monitoring results conform to required standards?	☐Yes ☐No ☐N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐Yes ☐No ☐N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐Yes ☐No ☐N/A		

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?	☐ Yes ☐ No ☐ N/A		

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

- ☐ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: _____

Section D – Phase I Summary

Phase I Summary:

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?
Note: Root cause must be conclusive, verifiable, and well-documented.
☐ Yes
☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☐ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (*Print*)**:**	Signature/Date:
Lab Manager/Designee (*Print*)**:**	Signature/Date:

Section E – Quality Disposition

Can investigation be closed?
☐ Yes
☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (*Print*)**:**	Signature/Date:

Table 1: Information for ETX-260624-0251 (BUD-T90) under investigation

Processing Analyst	Reading Analyst	Sample ID	Related Microbial ID	Media with microbial growth	Microbial ID
Guanchen Li	Andrew Carrillo	ETX-260624-0251 (BUD-T90)	ETX-261012-0367	TSB	*Bacillus megaterium*

Table 2: Environmental Monitoring from Processing Performed on 23Sep26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMMYY)	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM								
Personnel (Left Touch and Right Touch)	Daily	22Sep26	GL	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Personnel (Left Touch and Right Touch)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Personnel (Left Touch and Right Touch)	Daily	24Sep26	GL	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM for E001314								
Surface Sampling of ISO 5 (4 locations)	Daily	22Sep26	AOD	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 (4 locations)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 (4 locations)	Daily	24Sep26	AOD	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	22Sep26	AOD	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	24Sep26	AOD	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air Sampling of Cleanroom Suite 115 (E001737) with Processing BSC								
Active Air Sampling of Cleanrooms	Weekly	23Sep26	SMO	Week of Testing	6 CFUs (ISO 8 115)	ETX-261001-0625	*Micrococcus luteus* *Kocuria indica* *Kocuria rhizophila* Pending FunITS	None
Weekly Surface Sampling of Cleanroom Suite 115 (E001737) with Processing BSC								
Surface Sampling of Cleanrooms	Weekly	23Sep26	SMO	Week of Testing	No Growth	Not Applicable	Not Applicable	None