

	CORP-FORM-5 v9.1	Category: Form Corporate Name: Nonconformance Report Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.007.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC			Page 1 of 2

NCR#: N26487

IDENTIFICATION	
Initiator: Qiyue Chen	Date: 18-Sep-2026
Department: Microbiology	Sample Name/Material Description: Semaglutide/Cyanocobalamin; Tirzepatide 22mg/mL-Pyridoxine HCl; SEMAGLUTIDE/B12
Equipment ID #: E001230	
Sample ID #: ETX-260915-0653, ETX-260916-0044, ETX-260916-0374	Sample/Material Lot #: 0216202602, 22666, LG342010613
NONCONFORMITY DESCRIPTION (Who, What, When, Where, Why, & How)	
Nonconformance Description: On 17-Sep-2026, during routine Scan RDI sterility testing operations, an unexpected software crash occurred on the Scan RDI instrument (Equipment ID: E001230) while running session '17 Sep 2026 - 3'. Third-shift analyst JOC immediately notified Project Microbiologist QYC regarding the incident and software failure. JOC attempted to re-open the application and recover the testing session data; however, the instrument repeatedly displayed error messages stating, 'Unable to Create New Session Unknown Error' and 'Unable to Load Session Unknown Error', preventing access to the session. The following day, on 18-Sep-2026, QYC attempted to re-open the application and recover the testing session data again; the session '17 Sep 2026 - 3' could be opened, but no data was present within it. Work Order WO-260470 was promptly issued to Engineering/IT by RS to formally investigate the software failure and initiate technical remediation. On 21-Sep-2026, QYC proactively reached out to Eagle IT Coordinator KT (who was previously unaware of the incident) and attempted to restore and retrieve the session data. Following technical review, KT confirmed that the session file was corrupted, exporting as only 7 KB, and the corresponding raw acquisition dataset within the D:\ directory was completely empty and unrecoverable. Three ScanRDI test sample submissions, ETX-260915-0653, ETX-260916-0044 and ETX-260916-0374 were actively processed in this corrupted session and impacted by data loss. No valid test results could be generated or verified for these submissions from the corrupted run.	
CAUSE(S) ANALYSIS	
Description of cause(s): Scan RDI software crash and session file corruption on instrument E001230, resulting in unrecoverable raw acquisition data in the D:\ directory. Category (select all that apply): ☒ Equipment/System ☐ Process/Method ☐ Personnel ☐ External Phenomena ☐ Other	
INTERIM CORRECTIVE ACTION(S)	
Description of corrective action(s): - Work Order WO-260470 was issued to Engineering and IT to investigate and resolve the software crash, clear corrupted temporary files, and evaluate database integrity on instrument E001230. - Instrument E001230 was taken out of service for live sample testing pending completion and sign-off of WO-260470. - Client Care was instructed to request additional sample vials for retesting from ANG Labs (ETX-260915-0653) and Southend Pharmacy (ETX-260916-0044). Additional sample for ETX-260916-0374 (Optimal Balance Pharmacy) was received and is in-process with the third shift for retesting on an alternate qualified Scan RDI unit.	
QUALITY EVALUATION	
Was action effective in addressing the root cause of the nonconformity? ☒ Yes ☐ No ☐ N/A CAR required: ☐ YES ☐ NO If yes, CAR #: _____ If yes, explain how effectiveness was determined: Impact is isolated to the 3 sample submissions processed in the corrupted session (ETX-260915-0653, ETX-260916-0044, ETX-260916-0374). No impact to other testing sessions or alternate operational Scan RDI instruments (E002017, E002225). Additional vials requested / retesting initiated. Instrument E001230 remains under observation and repair under WO-260470. Additional Comments:	
ACKNOWLEDGEMENT	
Prepared By (Print): Qiyue Chen	SIGN _____ Signature _____ Date
Reviewed By (Print): Robin Seymour	SIGN _____ Signature _____ Date

	CORP-FORM-5 v9.1	Category: Form Corporate Name: Nonconformance Report Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.007.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC			Page 2 of 2

NCR#: N26487

CLOSING THE NONCONFORMANCE <i>(Closure by Quality)</i>										
Assessment Impact: Frequency: ²_____ Severity: ²_____ Magnitude: ²_____ <2 = 1 Between 2 to 10 = 2 >10 = 3 Typo Corrections = 1 Changes in processes and/or systems with no retroactive implications = 2 Changes in processes and/or systems with retroactive implication =3 <2 Process/Material/Equipment Affected = 1 >2 to <10 Process/Material/Equipment Affected = 2 >10 Process/Material/Equipment = 3 1 to 2: Low 3 to 17: Major 18 to 27: Critical <table border="1"> <tr> <td>3</td> <td>18</td> <td>27</td> </tr> <tr> <td>2</td> <td>17</td> <td>18</td> </tr> <tr> <td>1</td> <td>2</td> <td>3</td> </tr> </table>		3	18	27	2	17	18	1	2	3
3	18	27								
2	17	18								
1	2	3								