OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:		**Test Date:**		**Date Initiated:**	
Date of Incident:		**Test Name:**	Celsis Sterility Test	**Sample ID:**	
Name of Analyst who Performed the Test:		**Sample / Active Name**		**Lot #:**	
		Dosage Form			
Description of Incident:		☑ Client Care notified of OOS ☑ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS			
[redacted]	SOP / Test Method # 2.600.059	Effective Date: 03 Jan 2025	SOP / Test Method Rev: 01	Limits/Specification: <1 event	
Manager notified	Manager/designee name: Kathan Parikh	Date:	Quality Assurance Department notified:	Yes ☑ No ☐	

Section B – Investigation and Root Cause Analysis		
Personnel Discrepancy		**Comments**
Analyst interviewed?	☑ Yes ☐ No ☐ N/A	
Correct samples analyzed?	☑ Yes ☐ No ☐ N/A	
Correct SOP / Test Method used/followed?	☑ Yes ☐ No ☐ N/A	Yes, as per SOP 2.600.059
Correct sample collection/technique?	☑ Yes ☐ No ☐ N/A	Yes, as per SOP 2.600.059
Are the Test Result Records Attached & Calculations Verified and Correct?	☑ Yes ☐ No ☐ N/A	
Is the analyst qualified to perform the test?	☑ Yes ☐ No ☐ N/A	Yes, all analysts are trained and qualified by the quality department to perform the test.
Is formula worksheet provided and verified?	☐ Yes ☐ No ☑ N/A	Not Applicable
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☑ N/A	Not Applicable
Was the sample stored appropriately?	☑ Yes ☐ No ☐ N/A	

Material Discrepancy		Lot Number	Expiry Date
Is the integrity of the sample(s) questionable?	☐ Yes ☑ No ☐ N/A	[redacted]	
Was the sample transported appropriately?	☑ Yes ☐ No ☐ N/A		
Are the other test results from this run under investigation?	☐ Yes ☑ No ☐ N/A		

OOS Number: OOS-

Appropriate/correct glassware used?	☐Yes ☐ No ☑N/A		
Are reagents stored in proper condition and not expired?	☑Yes ☐ No ☐N/A	Celsis Process Consumables: Information is available in EagleTrax Sample Consumables section Celsis Aliquoting Consumables: See attached Celsis data packet Environmental Plates: TSA and Surface Plate: see attached environmental logs Monthly Cleaning: H2O2 strips and IPA: See attached monthly cleaning logs	Celsis Process Consumables: Information is available in EagleTrax Sample Consumables section Celsis Aliquoting Consumables: See attached Celsis data packet Environmental Plates: TSA and Surface Plate: see attached environmental logs Monthly Cleaning: H2O2 strips and IPA: See attached monthly cleaning logs
Standard Used	Celsis ATP Positive Control		
Solvent Used	Not Applicable	Not Applicable	Not Applicable
Is the standard the source of the root cause?	☐Yes ☐ No ☑N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐Yes ☐ No ☑N/A		
Equipment Discrepancy		**Equipment ID**	**Calibration Due Date**
Was the instrument set up correctly?	☑Yes ☐ No ☐ N/A		Mar 2026
Is the clean room facility certified?	☑Yes ☐ No ☐ N/A		Jun 2026 Dec 2026 Jun 2026
System Performance Check (if applicable)?	☑Yes ☐ No ☐ N/A		Mar 2026
System Suitability RSD < 2% (PDA)?	☐Yes ☐ No ☑ N/A	Not Applicable	Not Applicable
System Suitability RSD < 10% (QDA, CAD)?	☐Yes ☐ No ☑ N/A	Not Applicable	Not Applicable
Do the environmental monitoring results conform to required standards?	☑Yes ☐ No ☐ N/A	See Phase I Summary	See Phase I Summary
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☑Yes ☐ No ☐ N/A	See Phase I Summary	See Phase I Summary
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☑Yes ☐ No ☐ N/A	Incubator E001034 (Sensor E001501) Incubator E001031 (Sensor E001505) Incubator E001356 (Sensor E001450) Incubator E001357 (Sensor E001449)	Jun 2026 Feb 2027 Jun 2026 Feb 2027 Jun 2026 Feb 2027 Jun 2026 Feb 2027

OOS Number: OOS-

[redacted]

Are chromatogram/spectra/raw data correct?	☐ Yes ☐ No ☑ N/A	Not Applicable	Not Applicable

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

- ☑ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: N/A QYC

Section D – Phase I Summary

Phase I Summary:

OOS Number: OOS-

Phase I Summary (continued):

Phase I Summary (continued):

N/A QYC

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?
Note: Root cause must be conclusive, verifiable, and well-documented.

☐ Yes

☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☑ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (*Print*): Qiyue Chen	Signature/Date:
Lab Manager/Designee (*Print*):	Signature/Date:

Section E – Quality Disposition

Can investigation be closed?

☐ Yes

☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (*Print*):	Signature/Date: