Table 1: Information for ETX-260929-0931 under investigation

Processing Analyst	Reading Analyst	Sample ID	Events	Confirmed Microbial Events	Morphology Description
Ellen Bennett	Varsha Subramanian	ETX-260929-0931	15	4	***Cocci shaped morphology***

Table 2: Environmental Monitoring from Testing Performed on 01Oct26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMMYY)	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM								
Personal (Left Touch and Right Touch)	Daily	01Oct26	ELB	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	01Oct26	KSM	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM for E001938 and E001937								
Surface Sampling of ISO 5 BSC E001938 (4 Locations)	Daily	01Oct26	ELB	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001937 (4 Locations)	Daily	01Oct26	KSM	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001938	Daily	01Oct26	ELB	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001937	Daily	01Oct26	KSM	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Active Air Sampling of Cleanrooms	Weekly	01Oct26	SMO	Week of Testing	1 CFU (ISO 7 144 Sec II)	ETX-261012-0430	*Staphylococcus epidermidis* (Gram (+) cocci)	None
					5 CFUs (ISO 8 142)	ETX-261012-0433	*Staphylococcus epidermidis* *Pseudomonas parafulva* *Bacillus safensis* *Corynebacterium xerosis* *Corynebacterium ureicelerivorans*	
Weekly Surface Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Surface Sampling of Cleanrooms	Weekly	01Oct26	SMO	Week of Testing	No Growth	Not Applicable	Not Applicable	None