OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:		**Test Date:**		**Date Initiated:**	
Date of Incident:		**Test Name:**		**Sample ID:**	
Name of Analyst who Performed the Test:		**Sample / Active Name**		**Lot #:**	
		Dosage Form			
Description of Incident:		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS			
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis			
Personnel Discrepancy		**Comments**	
Analyst interviewed?	☐Yes ☐ No ☐ N/A		
Correct samples analyzed?	☐Yes ☐ No ☐ N/A		
Correct SOP / Test Method used/followed?	☐Yes ☐ No ☐ N/A		
Correct sample collection/technique?	☐Yes ☐ No ☐ N/A		
Are the Test Result Records Attached & Calculations Verified and Correct?	☐Yes ☐ No ☐ N/A		
Is the analyst qualified to perform the test?	☐Yes ☐ No ☐ N/A		
Is formula worksheet provided and verified?	☐Yes ☐ No ☐ N/A		
Does the processing method/method set conform to the required specification?	☐Yes ☐ No ☐ N/A		
Was the sample stored appropriately?	☐Yes ☐ No ☐ N/A		
Material Discrepancy		**Lot Number**	**Expiry Date**
Is the integrity of the sample(s) questionable?	☐Yes ☐ No ☐N/A		
Was the sample transported appropriately?	☐Yes ☐ No ☐N/A		
Are the other test results from this run under investigation?	☐Yes ☐ No ☐N/A		

OOS Number: OOS-

Appropriate/correct glassware used?	☐Yes ☐No ☐N/A		
Are reagents stored in proper condition and not expired?	☐Yes ☐No ☐N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐Yes ☐No ☐N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐Yes ☐No ☐N/A		
Equipment Discrepancy		**Equipment ID**	**Calibration Due Date**
Was the instrument set up correctly?	☐Yes ☐No ☐N/A		
Is the clean room facility certified?	☐Yes ☐No ☐N/A		
System Performance Check (if applicable)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 2% (PDA)?	☐Yes ☐No ☐N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐Yes ☐No ☐N/A		
Do the environmental monitoring results conform to required standards?	☐Yes ☐No ☐N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐Yes ☐No ☐N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐Yes ☐No ☐N/A		

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?	☐ Yes ☐ No ☐ N/A		

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

- ☐ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: _____

Section D – Phase I Summary

Phase I Summary:

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?
Note: Root cause must be conclusive, verifiable, and well-documented.

☐ Yes
- ☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☐ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (*Print*)**:**	Signature/Date:
Lab Manager/Designee (*Print*)**:**	Signature/Date:

Section E – Quality Disposition

Can investigation be closed?

☐ Yes
- ☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (*Print*)**:**	Signature/Date:

Table 1: Read Dates and Incubation Observation

ETX Submission ID	Set-up Analyst and Date	Sampling Site and Location	Plate Reading Analyst (≥ 48H)	CFUs Observed after 48 Hour Incubation at 30–35°C (E001031)	Plate Reading Analyst (NLT 5 days)	CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034)	Microbial Identification
ETX-260914-0493 ETX-260914-0496	SMO 04SEP 2026	Active air sampling plate	MC	53 CFUs on 143 sec I ISO 8, and 44 CFUs on 143 sec II ISO 8	MC	53 CFUs on 143 sec I ISO 8, and 44 CFUs on 143 sec II ISO 8	*Staphylococcus capitis* *Micrococcus luteus* *Moraxella osloensis* *Corynebacterium segmentosum* *Kytococcus sedentarius* *Staphylococcus hominis* *Paracoccus yeei* *Brevibacterium* sp. CS2 *Corynebacterium ureicelerivorans* *Micrococcus* sp. *Corynebacterium sanguinis* *Staphylococcus haemolyticus* *Corynebacterium genitalium* *Cordyceps memorabilis* *Allophoma tropica* *Corynebacterium jeikeium* *Pseudomonas fulva* Lophiotrematacеae *Acrodontium salmoneum* *Quambalaria cyanescens* Sordariales *Streptococcus* sp. A12

Table 2: Environmental Monitoring for Analyst and Cleanroom

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMMYY)	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Weekly Active Air Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Active Air Sampling of Cleanrooms	Weekly	28AUG 2026	SMO	Week (Before Testing Date)	5 CFUs on 142 ISO 8	ETX-260908-0584	Gram (+) rods Gram (+) cocci	None
Active Air Sampling of Cleanrooms	Weekly	04SEP 2026	SMO	Week (Of Testing Date)	53 CFUs on 143 sec I ISO 8, 44 CFUs on 143 sec II ISO 8, and 17 CFUs on 142 ISO 8	ETX-260914-0493 ETX-260914-0496 ETX-260914-0498	*Staphylococcus capitis* *Micrococcus luteus* *Moraxella osloensis* *Corynebacterium segmentosum* *Kytococcus sedentarius* *Staphylococcus hominis* *Paracoccus yeei* *Brevibacterium* sp. CS2 *Corynebacterium ureicelerivorans* *Micrococcus* sp. *Corynebacterium sanguinis* *Staphylococcus haemolyticus* *Corynebacterium genitalium* *Cordyceps memorabilis* *Allophoma tropica* *Corynebacterium jeikeium* *Pseudomonas fulva* Lophiotrematacеae *Acrodontium salmoneum* *Quambalaria cyanescens* Sordariales *Streptococcus* sp. A12	None
Active Air Sampling of Cleanrooms	Weekly	09SEP 2026	SMO	Week (After Testing Date)	1 CFU on 144 sec II ISO 7, 16 CFUs on 143 sec I ISO 8, 17 CFUs on 143 sec II ISO 8, and 3 CFUs on 142 ISO 8	ETX-260916-0728 ETX-260916-0730 ETX-260916-0735 ETX-260916-0737	*Corynebacterium simulans* *Staphylococcus epidermidis* *Staphylococcus pasteuri* *Micrococcus luteus* *Staphylococcus hominis* *Rothia dentocariosa* *Staphylococcus pasteuri* *Corynebacterium imitans* *Staphylococcus* sp. M0911 *Moraxella osloensis* *Streptococcus thermophilus* *Aeromicrobium erythreum* *Roseomonas* sp. FDAARGOS_362 *Agrococcus carbonis* *Corynebacterium segmentosum* *Streptomyces xiamenensis*	None
Weekly Surface Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Surface Sampling of Cleanrooms	Weekly	28AUG 2026	SMO	Week (Before Testing Date)	1 CFU on table 2 in 143 ISO 8	ETX-260908-0580	Gram (+) rods	None
Surface Sampling of Cleanrooms	Weekly	04SEP 2026	SMO	Week (Of Testing Date)	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of Cleanrooms	Weekly	09SEP 2026	SMO	Week (After Testing Date)	No Growth	Not Applicable	Not Applicable	None