

Table 1: Information for ETX-260804-0335 under investigation

Processing Analyst	Reading Analyst	Sample ID	Related Microbial ID	Media with microbial growth	Microbial ID
Pending	Qiyue Chen	ETX-260804-0335	ETX-260904-0331	1 x 100mL TSB bottle	Pending

Table 2: Environmental Monitoring from Processing Performed on 25Aug26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing								
Personnel (Left Touch and Right Touch)	Daily	24Aug26	Pending	Date Before Testing	No growth	N/A	N/A	None
Personnel (Left Touch and Right Touch)	Daily	25Aug26	Pending	Date of Testing	No growth	N/A	N/A	None
Personnel (Left Touch and Right Touch)	Daily	26Aug26	Pending	Date After Testing	No growth	N/A	N/A	None
Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC) E001316								
Surface Sampling of ISO 5 (4 locations)	Daily	24Aug26	Pending	Date Before Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	25Aug26	Pending	Date of Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	26Aug26	Pending	Date After Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	24Aug26	Pending	Date Before Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	25Aug26	Pending	Date of Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	26Aug26	Pending	Date After Testing	No growth	N/A	N/A	None
Weekly Active Air Sampling of Cleanroom - CR114 (E001736)								
Active Air Sampling of Cleanrooms	Weekly	25Aug26	SMO	Week on Testing Date	4 CFU (ISO 8 114)	ETX-260901-0112	3 Gram (+) cocci, 1 Hyphae	None
Surface Sampling of Anteroom and Cleanroom - CR114 (E001736)								
Surface Sampling of Cleanrooms	Weekly	25Aug26	SMO	Week on Testing Date	No growth	N/A	N/A	None