



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED
JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 1 of 6

OOS Number: OOS-

Phase I

Section A – Test Details

Initiator:		Test Date:		Date Initiated:	
Date of Incident:		Test Name:		Sample ID:	
Name of Analyst who Performed the Test:		Sample / Active Name		Lot #:	
		Dosage Form			
Description of Incident:			☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis

Personnel Discrepancy		Comments
Analyst interviewed?	☐ Yes ☐ No ☐ N/A	
Correct samples analyzed?	☐ Yes ☐ No ☐ N/A	
Correct SOP / Test Method used/followed?	☐ Yes ☐ No ☐ N/A	
Correct sample collection/technique?	☐ Yes ☐ No ☐ N/A	
Are the Test Result Records Attached & Calculations Verified and Correct?	☐ Yes ☐ No ☐ N/A	
Is the analyst qualified to perform the test?	☐ Yes ☐ No ☐ N/A	
Is formula worksheet provided and verified?	☐ Yes ☐ No ☐ N/A	
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☐ N/A	
Was the sample stored appropriately?	☐ Yes ☐ No ☐ N/A	
Material Discrepancy		Lot Number
Is the integrity of the sample(s) questionable?	☐ Yes ☐ No ☐ N/A	
Was the sample transported appropriately?	☐ Yes ☐ No ☐ N/A	
Are the other test results from this run under investigation?	☐ Yes ☐ No ☐ N/A	



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED
JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 2 of 6

OOS Number: OOS-

Appropriate/correct glassware used?	☐ Yes ☐ No ☐ N/A		
Are reagents stored in proper condition and not expired?	☐ Yes ☐ No ☐ N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐ Yes ☐ No ☐ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☐ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☐ Yes ☐ No ☐ N/A		
Is the clean room facility certified?	☐ Yes ☐ No ☐ N/A		
System Performance Check (if applicable)?	☐ Yes ☐ No ☐ N/A		
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☐ N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☐ N/A		
Do the environmental monitoring results conform to required standards?	☐ Yes ☐ No ☐ N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐ Yes ☐ No ☐ N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐ Yes ☐ No ☐ N/A		

**QUALITY SYSTEM**FORM**DOCUMENT #: 3.100.019.F01****LEVEL: 3**

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:**LABORATORY OOS INVESTIGATION FORM**

Page 3 of 6

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?

☐ Yes☐ No☐ N/A**Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)**☐ N/A☐ Re-injection of sample vial☐ Re-injection of original standard vial☐ Re-vialing of original working standard solution☐ Re-vialing of original working sample solution☐ Re-dilution of a working sample solution from the original stock☐ Re-dilution of a working standard solution from the original stock standard☐ Other: _____**Section D – Phase I Summary****Phase I Summary:**



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 4 of 6

OOS Number: OOS-

Phase I Summary (continued):



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 5 of 6

OOS Number: OOS-

Phase I Summary (continued):



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 6 of 6

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?*Note: Root cause must be conclusive, verifiable, and well-documented.*☐ Yes☐ If retesting is **required**, complete Phase II of the OOS (initiate Form 3.100.019.F02).☐ No, the OOS result is valid.**Reviewer comment(s):****Prepared by (Print):**

Signature/Date:

Lab Manager/Designee (Print):

Signature/Date:

Section E – Quality Disposition**Can investigation be closed?**☐ Yes☐ Original result is valid.☐ No, initiate Form 3.100.019.F02.**Approved by QA Representative (Print):**

Signature/Date:

Table 1: Read Dates and Incubation Observation

Sampling Location	Read Date (30-35°C, NLT 48h)	Read By	CFU Count / Observation	Read Date (20-25°C, NLT 5d)	Read By	CFU Count / Observation	Microbial Identification
Surface Sampling (Changeover) Plate (BSC E001309)	06 Jun 2026	MC	No microbial growth was observed	11 Jun 2026	SMO	1 CFU on ScanC/O CGS E001309 S1 11MAY2026	colony-like artifact (White, shiny, hardened, smooth area embedded on agar)

Table 2: Environmental Monitoring Plates for Analyst and Cleanroom Bracketing

Sampling Type	Sampling Location	Date Before Testing (08 May 2026)		Date of Testing (11-May-2026)		Date After Testing (12 May 2026)	
		Analyst / Result	ETX / Identification	Analyst / Result	ETX / Identification	Analyst / Result	ETX / Identification
Personnel Monitoring	Glove Touch (Left / Right)	CGS No growth	N/A N/A	CGS No growth	N/A N/A	CGS No growth	N/A N/A
Surface Sampling	ISO 5 BSC E001309	CGS No growth	N/A N/A	CGS 1 CFU	ETX-260518-0273 colony-like artifact (White, shiny, hardened, smooth area embedded on agar)	CGS No growth	N/A N/A
Settling Sampling	ISO 5 BSC E001309	CGS No growth	N/A N/A	CGS No growth	N/A N/A	CGS No growth	N/A N/A
Weekly Cleanroom Active Air	Cleanroom CR117 (E001739)	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A
Weekly Cleanroom Surface	Anteroom & Buffer (CR117 (E001739))	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A