

Table 1: Information for ETX-260828-0527 under investigation

| Processing Analyst | Aliquoting Analyst | Sample ID                       | Related Microbial ID | Media with microbial growth | Microbial ID           |
|--------------------|--------------------|---------------------------------|----------------------|-----------------------------|------------------------|
| Gabrielle Surber   | Cuong Du           | <a href="#">ETX-260828-0527</a> | ETX-260908-0593      | 1 x 300mL FTM               | No growth was obtained |

Table 2: Environmental Monitoring from Processing Performed on 01Sep26

| Environmental Monitoring (EM) Sampling Site                                    | Frequency | Date (DDMMYY) | Analyst (Initials) | Day /Week(s)         | Observation*      | Environmental Monitoring Plate ETX ID | Microbial ID                          | Notes |
|--------------------------------------------------------------------------------|-----------|---------------|--------------------|----------------------|-------------------|---------------------------------------|---------------------------------------|-------|
| <b>Personnel EM Bracketing</b>                                                 |           |               |                    |                      |                   |                                       |                                       |       |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 31Aug26       | GS                 | Date Before Testing  | No growth         | N/A                                   | N/A                                   | None  |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 01Sep26       | GS                 | Date of Testing      | No growth         | N/A                                   | N/A                                   | None  |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 02Sep26       | GS                 | Date After Testing   | No growth         | N/A                                   | N/A                                   | None  |
| <b>Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC)</b> |           |               |                    |                      |                   |                                       |                                       |       |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 31Aug26       | GS                 | Date Before Testing  | No growth         | N/A                                   | N/A                                   | None  |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 01Sep26       | GS                 | Date of Testing      | No growth         | N/A                                   | N/A                                   | None  |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 02Sep26       | GS                 | Date After Testing   | No growth         | N/A                                   | N/A                                   | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 31Aug26       | GS                 | Date Before Testing  | No growth         | N/A                                   | N/A                                   | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 01Sep26       | GS                 | Date of Testing      | No growth         | N/A                                   | N/A                                   | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 02Sep26       | GS                 | Date After Testing   | No growth         | N/A                                   | N/A                                   | None  |
| <b>Weekly Active Air Sampling Bracketing</b>                                   |           |               |                    |                      |                   |                                       |                                       |       |
| Active Air Sampling of Cleanrooms                                              | Weekly    | 04Sep26       | ISS                | Week on Testing Date | 1 CFU (ISO 8 114) | ETX-260914-0487                       | Gram (+) coccobacilli & Gram (+) rods | None  |
| <b>Surface Sampling of Anteroom and Cleanroom Bracketing</b>                   |           |               |                    |                      |                   |                                       |                                       |       |
| Surface Sampling of Cleanrooms                                                 | Weekly    | 04Sep26       | ISS                | Week on Testing Date | No growth         | N/A                                   | N/A                                   | None  |

Table 3: Environmental Monitoring from Aliquoting Performed on 08Sep26

| Environmental Monitoring (EM) Sampling Site                                    | Frequency | Date (DDMMYY) | Analyst (Initials) | Day /Week(s)         | Observation*      | Environmental Monitoring Plate ETX ID | Microbial ID       | Notes |
|--------------------------------------------------------------------------------|-----------|---------------|--------------------|----------------------|-------------------|---------------------------------------|--------------------|-------|
| <b>Personnel EM Bracketing</b>                                                 |           |               |                    |                      |                   |                                       |                    |       |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 03Sep26       | CCD                | Date Before Testing  | No growth         | N/A                                   | N/A                | None  |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 08Sep26       | CCD                | Date of Testing      | No growth         | N/A                                   | N/A                | None  |
| Personnel (Left Touch and Right Touch)                                         | Daily     | 15Sep26       | CCD                | Date After Testing   | No growth         | N/A                                   | N/A                | None  |
| <b>Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC)</b> |           |               |                    |                      |                   |                                       |                    |       |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 07Sep26       | ALA                | Date Before Testing  | No growth         | N/A                                   | N/A                | None  |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 08Sep26       | CCD                | Date of Testing      | No growth         | N/A                                   | N/A                | None  |
| Surface Sampling of ISO 5 (4 locations)                                        | Daily     | 09Sep26       | ALA                | Date After Testing   | No growth         | N/A                                   | N/A                | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 07Sep26       | ALA                | Date Before Testing  | No growth         | N/A                                   | N/A                | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 08Sep26       | CCD                | Date of Testing      | No growth         | N/A                                   | N/A                | None  |
| Settling Sampling of ISO 5 (2 locations)                                       | Daily     | 09Sep26       | ALA                | Date After Testing   | No growth         | N/A                                   | N/A                | None  |
| <b>Weekly Active Air Sampling Bracketing</b>                                   |           |               |                    |                      |                   |                                       |                    |       |
| Active Air Sampling of Cleanrooms                                              | Weekly    | 10Sep26       | SMO                | Week on Testing Date | 1 CFU (ISO 8 114) | <a href="#">ETX-260921-0520</a>       | Micrococcus luteus | None  |
| <b>Surface Sampling of Anteroom and Cleanroom Bracketing</b>                   |           |               |                    |                      |                   |                                       |                    |       |
| Surface Sampling of Cleanrooms                                                 | Weekly    | 10Sep26       | SMO                | Week on Testing Date | No growth         | N/A                                   | N/A                | None  |