

Table 1: Information for ETX-260813-0778 under investigation

| Processing Analyst | Reading Analyst | Sample ID                       | Events | Confirmed Microbial Events | Morphology Description             |
|--------------------|-----------------|---------------------------------|--------|----------------------------|------------------------------------|
| Varsha Subramanian | Sonal Uprety    | <a href="#">ETX-260813-0778</a> | 31     | 4                          | <i>Short rod-shaped morphology</i> |

Table 2: Environmental Monitoring from Testing Performed on 25Aug26

| Environmental Monitoring (EM) Sampling Site                                           | Frequency | Date (DDMMYY) | Analyst | Day /Week(s)               | Observation       | EM Plate ETX Number             | Microbial ID                  | Notes |
|---------------------------------------------------------------------------------------|-----------|---------------|---------|----------------------------|-------------------|---------------------------------|-------------------------------|-------|
| <b>Personnel EM for 25Aug26</b>                                                       |           |               |         |                            |                   |                                 |                               |       |
| Personal (Left Touch and Right Touch)                                                 | Daily     | 25Aug26       | VV      | Date of Testing (Scan)     | No Growth         | Not Applicable                  | Not Applicable                | None  |
| Personal (Left Touch and Right Touch)                                                 | Daily     | 25Aug26       | VV      | Date of Testing (Scan C/O) | No Growth         | Not Applicable                  | Not Applicable                | None  |
| <b>Biological Safety Cabinet (BSC) EM for BSC E001319 for 25Aug26</b>                 |           |               |         |                            |                   |                                 |                               |       |
| Surface Sampling of ISO 5 BSC E001319 (4 Locations)                                   | Daily     | 25Aug26       | VV      | Date of Testing (Scan)     | No Growth         | Not Applicable                  | Not Applicable                | None  |
| Surface Sampling of ISO 5 BSC E001319 (4 Locations)                                   | Daily     | 25Aug26       | VV      | Date of Testing (Scan C/O) | No Growth         | Not Applicable                  | Not Applicable                | None  |
| Settling Sampling of ISO 5 BSC E001319                                                | Daily     | 25Aug26       | VV      | Date of Testing (Scan)     | No Growth         | Not Applicable                  | Not Applicable                | None  |
| Settling Sampling of ISO 5 BSC E001319                                                | Daily     | 25Aug26       | VV      | Date of Testing (Scan C/O) | No Growth         | Not Applicable                  | Not Applicable                | None  |
| <b>Weekly Active Air Sampling of CR 145 (E001979) with Processing BSC for 25Aug26</b> |           |               |         |                            |                   |                                 |                               |       |
| Active Air Sampling of Cleanrooms                                                     | Weekly    | 28Aug26       | SMO     | Week of Testing            | 5 CFU (ISO 8 142) | <a href="#">ETX-260908-0584</a> | Gram (+) rods, Gram (+) cocci | None  |
| <b>Weekly Surface Sampling of CR 145 (E001979) with Processing BSC for 25Aug26</b>    |           |               |         |                            |                   |                                 |                               |       |
| Surface Sampling of Cleanrooms                                                        | Weekly    | 28Aug26       | SMO     | Week of Testing            | 1 CFU (ISO 8 143) | <a href="#">ETX-260908-0580</a> | Gram (+) rods                 | None  |